

Cover Page

Order ID : Q1985

Project ID : NWIRP Bethpage 112G08005-WE13

Client : Tetra Tech NUS, Inc.

Lab Sample Number

Q1985-01
Q1985-02

Client Sample Number

RW8-BW-20250507
TB-20250507

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 5/13/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage 112G08005-WE13

Project Manager # Ernie Wu

Order ID # Q1985

Test Name: VOCMS Group1

A. Number of Samples and Date of Receipt:

2 Water samples were received on 05/08/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Mercury, Metals ICP-TAL, METALS-TAL, pH, SVOC-SIMGroup1, SVOC-TCL BNA - 20, TSS and VOCMS Group1. This data package contains results for VOCMS Group1.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI. The analysis of VOCMS Group1 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank Spike Duplicate met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

E. Additional Comments:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

The not QT review data is reported in the Miscellaneous.

The laboratory certifies that the all-electronic diskette deliverable exactly match the data Summary forms (i.e. Form Is)."



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092
NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1985

MATRIX: Water

METHOD: 8260D

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The Initial Calibration met the requirements. The Continuous Calibration met the requirements.			✓
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The Blank Spike met requirements for all samples. The Blank Spike Duplicate met requirements for all samples.			✓
9.	Internal Standard Area/Retention Time Shift Meet Criteria Comments:			✓
10.	Analysis Holding Time Met If not met, list number of days exceeded for each sample:			✓

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

ADDITIONAL COMMENTS:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is)."

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is >20% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1985

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 05/13/2025

LAB CHRONICLE

OrderID: Q1985	OrderDate: 5/8/2025 10:49:00 AM
Client: Tetra Tech NUS, Inc.	Project: NWIRP Bethpage 112G08005-WE13
Contact: Ernie Wu	Location: L41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1985-01	RW8-BW-20250507	Water	VOCMS Group1	8260-Low	05/07/25		05/08/25	05/08/25
Q1985-02	TB-20250507	Water	VOCMS Group1	8260-Low	05/07/25		05/08/25	05/08/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1985

Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units

Client ID:

0

Total Voc :

Total Concentration:



QC SUMMARY

Surrogate Summary

SDG No.: Q1985

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1985-01	RW8-BW-20250507	1,2-Dichloroethane-d4	50	53.6	107	81	118
		Dibromofluoromethane	50	51.1	102	80	119
		Toluene-d8	50	49.5	99	89	112
		4-Bromofluorobenzene	50	49.3	99	85	114
Q1985-02	TB-20250507	1,2-Dichloroethane-d4	50	51.4	103	81	118
		Dibromofluoromethane	50	49.4	99	80	119
		Toluene-d8	50	49.2	98	89	112
		4-Bromofluorobenzene	50	49.4	99	85	114
VX0508WBL01	VX0508WBL01	1,2-Dichloroethane-d4	50	52.5	105	81	118
		Dibromofluoromethane	50	49.3	99	80	119
		Toluene-d8	50	49.8	100	89	112
		4-Bromofluorobenzene	50	48.3	97	85	114
VX0508WBS01	VX0508WBS01	1,2-Dichloroethane-d4	50	49.8	100	81	118
		Dibromofluoromethane	50	50.2	100	80	119
		Toluene-d8	50	49.1	98	89	112
		4-Bromofluorobenzene	50	48.2	96	85	114
VX0508WBSD0	VX0508WBSD01	1,2-Dichloroethane-d4	50	50.9	102	81	118
		Dibromofluoromethane	50	50.0	100	80	119
		Toluene-d8	50	49.9	100	89	112
		4-Bromofluorobenzene	50	49.4	99	85	114

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1985

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VX046081.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0508WBS01	Chloromethane	20	18.6	ug/L	93			50	139	
	Vinyl chloride	20	18.5	ug/L	93			58	137	
	Bromomethane	20	17.9	ug/L	90			53	141	
	Chloroethane	20	19.3	ug/L	97			60	138	
	Trichlorofluoromethane	20	19.3	ug/L	97			65	141	
	1,1,2-Trichlorotrifluoroethane	20	18.5	ug/L	93			70	136	
	1,1-Dichloroethene	20	18.4	ug/L	92			71	131	
	Acetone	100	93.0	ug/L	93			39	160	
	Carbon disulfide	20	17.3	ug/L	86			64	133	
	Methyl tert-butyl Ether	20	19.3	ug/L	97			71	124	
	Methylene Chloride	20	17.5	ug/L	88			74	124	
	trans-1,2-Dichloroethene	20	18.8	ug/L	94			75	124	
	1,1-Dichloroethane	20	19.4	ug/L	97			77	125	
	2-Butanone	100	96.5	ug/L	97			56	143	
	Carbon Tetrachloride	20	18.9	ug/L	95			72	136	
	cis-1,2-Dichloroethene	20	19.5	ug/L	98			78	123	
	Chloroform	20	19.8	ug/L	99			79	124	
	1,1,1-Trichloroethane	20	19.2	ug/L	96			74	131	
	Methylcyclohexane	20	18.2	ug/L	91			72	132	
	Benzene	20	19.4	ug/L	97			79	120	
	1,2-Dichloroethane	20	18.9	ug/L	95			73	128	
	Trichloroethene	20	18.6	ug/L	93			79	123	
	1,2-Dichloropropane	20	19.5	ug/L	98			78	122	
	Bromodichloromethane	20	19.2	ug/L	96			79	125	
	4-Methyl-2-Pentanone	100	97.8	ug/L	98			67	130	
	Toluene	20	19.1	ug/L	96			80	121	
	t-1,3-Dichloropropene	20	17.8	ug/L	89			73	127	
	cis-1,3-Dichloropropene	20	18.6	ug/L	93			75	124	
	1,1,2-Trichloroethane	20	19.3	ug/L	97			80	119	
	2-Hexanone	100	98.2	ug/L	98			57	139	
	Dibromochloromethane	20	19.3	ug/L	97			74	126	
	Tetrachloroethene	20	18.9	ug/L	95			74	129	
	Chlorobenzene	20	18.8	ug/L	94			82	118	
	Ethyl Benzene	20	19.6	ug/L	98			79	121	
	m/p-Xylenes	40	39.2	ug/L	98			80	121	
	o-Xylene	20	20.1	ug/L	101			78	122	
	Styrene	20	19.9	ug/L	100			78	123	
	Bromoform	20	18.7	ug/L	94			66	130	
	Isopropylbenzene	20	20.2	ug/L	101			72	131	
	1,1,2,2-Tetrachloroethane	20	19.5	ug/L	98			71	121	
	1,3-Dichlorobenzene	20	19.5	ug/L	98			80	119	
	1,4-Dichlorobenzene	20	19.1	ug/L	96			79	118	
	1,2-Dichlorobenzene	20	20.1	ug/L	101			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1985

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VX046082.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0508WBSD01	Chloromethane	20	19.2	ug/L	96	3		50	139	20
	Vinyl chloride	20	18.0	ug/L	90	3		58	137	20
	Bromomethane	20	18.0	ug/L	90	0		53	141	20
	Chloroethane	20	19.0	ug/L	95	2		60	138	20
	Trichlorofluoromethane	20	19.2	ug/L	96	1		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	19.3	ug/L	97	4		70	136	20
	1,1-Dichloroethene	20	18.3	ug/L	92	0		71	131	20
	Acetone	100	98.2	ug/L	98	5		39	160	20
	Carbon disulfide	20	17.4	ug/L	87	1		64	133	20
	Methyl tert-butyl Ether	20	20.0	ug/L	100	3		71	124	20
	Methylene Chloride	20	18.4	ug/L	92	4		74	124	20
	trans-1,2-Dichloroethene	20	19.1	ug/L	96	2		75	124	20
	1,1-Dichloroethane	20	19.8	ug/L	99	2		77	125	20
	2-Butanone	100	100	ug/L	100	3		56	143	20
	Carbon Tetrachloride	20	18.9	ug/L	95	0		72	136	20
	cis-1,2-Dichloroethene	20	19.8	ug/L	99	1		78	123	20
	Chloroform	20	19.9	ug/L	100	1		79	124	20
	1,1,1-Trichloroethane	20	19.6	ug/L	98	2		74	131	20
	Methylcyclohexane	20	18.5	ug/L	93	2		72	132	20
	Benzene	20	19.3	ug/L	97	0		79	120	20
	1,2-Dichloroethane	20	20.2	ug/L	101	6		73	128	20
	Trichloroethene	20	19.3	ug/L	97	4		79	123	20
	1,2-Dichloropropane	20	20.3	ug/L	102	4		78	122	20
	Bromodichloromethane	20	19.8	ug/L	99	3		79	125	20
	4-Methyl-2-Pentanone	100	110	ug/L	110	12		67	130	20
	Toluene	20	19.5	ug/L	98	2		80	121	20
	t-1,3-Dichloropropene	20	19.1	ug/L	96	8		73	127	20
	cis-1,3-Dichloropropene	20	19.9	ug/L	100	7		75	124	20
	1,1,2-Trichloroethane	20	20.6	ug/L	103	6		80	119	20
	2-Hexanone	100	110	ug/L	110	12		57	139	20
	Dibromochloromethane	20	20.4	ug/L	102	5		74	126	20
	Tetrachloroethene	20	19.0	ug/L	95	0		74	129	20
	Chlorobenzene	20	18.8	ug/L	94	0		82	118	20
	Ethyl Benzene	20	19.5	ug/L	98	0		79	121	20
	m/p-Xylenes	40	39.4	ug/L	99	1		80	121	20
	o-Xylene	20	19.8	ug/L	99	2		78	122	20
	Styrene	20	20.0	ug/L	100	0		78	123	20
	Bromoform	20	19.4	ug/L	97	3		66	130	20
	Isopropylbenzene	20	20.1	ug/L	101	0		72	131	20
	1,1,2,2-Tetrachloroethane	20	20.4	ug/L	102	4		71	121	20
	1,3-Dichlorobenzene	20	19.7	ug/L	99	1		80	119	20
	1,4-Dichlorobenzene	20	19.5	ug/L	98	2		79	118	20
	1,2-Dichlorobenzene	20	20.3	ug/L	102	1		80	119	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0508WBL01

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: Q1985

SAS No.: Q1985 SDG NO.: Q1985

Lab File ID: VX046080.D

Lab Sample ID: VX0508WBL01

Date Analyzed: 05/08/2025

Time Analyzed: 10:56

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0508WBS01	VX0508WBS01	VX046081.D	05/08/2025
VX0508WBSD01	VX0508WBSD01	VX046082.D	05/08/2025
TB-20250507	Q1985-02	VX046084.D	05/08/2025
RW8-BW-20250507	Q1985-01	VX046085.D	05/08/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1985 SAS No.: Q1985 SDG NO.: Q1985
Lab File ID: VX046038.D BFB Injection Date: 05/05/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:37
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.1
75	30.0 - 60.0% of mass 95	56.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.5 (0.7) 1
174	50.0 - 100.0% of mass 95	68.8
175	5.0 - 9.0% of mass 174	5 (7.3) 1
176	95.0 - 101.0% of mass 174	66.7 (97) 1
177	5.0 - 9.0% of mass 176	4.6 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC020	VSTDIC020	VX046041.D	05/05/2025	11:35
VSTDIC050	VSTDIC050	VX046042.D	05/05/2025	11:58
VSTDIC100	VSTDIC100	VX046043.D	05/05/2025	12:21
VSTDIC150	VSTDIC150	VX046044.D	05/05/2025	12:45
VSTDIC005	VSTDIC005	VX046046.D	05/05/2025	16:04
VSTDIC001	VSTDIC001	VX046047.D	05/05/2025	16:27



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1985 SAS No.: Q1985 SDG NO.: Q1985
Lab File ID: VX046077.D BFB Injection Date: 05/08/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:36
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.3
75	30.0 - 60.0% of mass 95	56.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.8 (1.2) 1
174	50.0 - 100.0% of mass 95	67.8
175	5.0 - 9.0% of mass 174	5.1 (7.5) 1
176	95.0 - 101.0% of mass 174	64.8 (95.5) 1
177	5.0 - 9.0% of mass 176	4.5 (7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX046078.D	05/08/2025	10:05
VX0508WBL01	VX0508WBL01	VX046080.D	05/08/2025	10:56
VX0508WBS01	VX0508WBS01	VX046081.D	05/08/2025	11:20
VX0508WBSD01	VX0508WBSD01	VX046082.D	05/08/2025	11:45
TB-20250507	Q1985-02	VX046084.D	05/08/2025	12:32
RW8-BW-20250507	Q1985-01	VX046085.D	05/08/2025	12:55
VSTDCCC050EC	VSTDCCC050	VX046100.D	05/08/2025	18:46

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q1985 SAS No.: Q1985 SDG NO.: Q1985
 Lab File ID: VX046078.D Date Analyzed: 05/08/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:05
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	105825	5.54	181991	6.75	158314	10.05
UPPER LIMIT	211650	6.044	363982	7.251	316628	10.549
LOWER LIMIT	52912.5	5.044	90995.5	6.251	79157	9.549
EPA SAMPLE NO.						
RW8-BW-20250507	65710	5.54	131109	6.76	121918	10.05
TB-20250507	64053	5.55	126115	6.76	114932	10.05
VX0508WBL01	66087	5.54	132156	6.76	121254	10.05
VX0508WBS01	97732	5.54	172839	6.76	147787	10.05
VX0508WBSD01	89125	5.54	157859	6.76	138205	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q1985 SAS No.: Q1985 SDG NO.: Q1985
 Lab File ID: VX046078.D Date Analyzed: 05/08/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:05
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	73554	12.018				
UPPER LIMIT	147108	12.518				
LOWER LIMIT	36777	11.518				
EPA SAMPLE NO.						
RW8-BW-20250507	50270	12.02				
TB-20250507	49527	12.02				
VX0508WBL01	49493	12.02				
VX0508WBS01	68652	12.02				
VX0508WBSD01	63610	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	05/07/25		
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	05/08/25		
Client Sample ID:	RW8-BW-20250507			SDG No.:	Q1985		
Lab Sample ID:	Q1985-01			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046085.D	1		05/08/25 12:55	VX050825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/07/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/08/25	
Client Sample ID:	RW8-BW-20250507		SDG No.:	Q1985	
Lab Sample ID:	Q1985-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046085.D	1		05/08/25 12:55	VX050825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.6		81 - 118		107%	SPK: 50
1868-53-7	Dibromofluoromethane	51.1		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	49.4		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.3		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	65700	5.544				
540-36-3	1,4-Difluorobenzene	131000	6.757				
3114-55-4	Chlorobenzene-d5	122000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	50300	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	05/07/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	05/08/25	
Client Sample ID:	RW8-BW-20250507		SDG No.:	Q1985	
Lab Sample ID:	Q1985-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046085.D	1		05/08/25 12:55	VX050825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046085.D
 Acq On : 08 May 2025 12:55
 Operator : JC/MD
 Sample : Q1985-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-BW-20250507

Quant Time: May 09 01:56:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.544	168	65710	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	131109	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	121918	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50270	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65644	53.585	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.160%	
35) Dibromofluoromethane	5.379	113	48204	51.057	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.120%	
50) Toluene-d8	8.647	98	161572	49.445	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.880%	
62) 4-Bromofluorobenzene	11.079	95	61840	49.336	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 98.680%	
Target Compounds						Qvalue
29) Tetrahydrofuran	5.026	42	3978	8.902	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046085.D
Acq On : 08 May 2025 12:55
Operator : JC/MD
Sample : Q1985-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-BW-20250507

Quant Time: May 09 01:56:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration





#1
Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.544 min Scan# 71

Delta R.T. -0.000 min

Lab File: VX046085.D

Acq: 08 May 2025 12:55

Instrument :

MSVOA_X

ClientSampleId :





RW8-BW-20250507

Tgt Ion:168 Resp: 65710

Ion Ratio Lower Upper

168 100

99 75.2 54.9 82.3



#29

Tetrahydrofuran

Concen: 8.902 ug/l

RT: 5.026 min Scan# 646

Delta R.T. 0.024 min

Lab File: VX046085.D

Acq: 08 May 2025 12:55




Tgt Ion: 42 Resp: 3978


Ion Ratio Lower Upper

42 100

72 39.6 33.7 50.5

71 39.0 31.4 47.2





#33

1,2-Dichloroethane-d4

Concen: 53.585 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046085.D

Acq: 08 May 2025 12:55

Instrument :

MSVOA_X

ClientSampleId :





RW8-BW-20250507

Tgt Ion: 65 Resp: 65644

Ion Ratio Lower Upper

65 100

67 48.7 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046085.D

Acq: 08 May 2025 12:55




Tgt Ion: 114 Resp: 131109


Ion Ratio Lower Upper

114 100

63 23.9 0.0 49.2

88 17.8 0.0 33.6





#35

Dibromofluoromethane

Concen: 51.057 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX046085.D


Acq: 08 May 2025 12:55

Instrument :




MSVOA_X

ClientSampleId :

RW8-BW-20250507



Tgt Ion:	113	Resp:	48204
Ion	Ratio	Lower	Upper
113	100		
111	102.2	83.1	124.7
192	16.6	13.3	19.9



#50

Toluene-d8

Concen: 49.445 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. -0.000 min

Lab File: VX046085.D

Acq: 08 May 2025 12:55

Tgt Ion:	98	Resp:	161572
Ion	Ratio	Lower	Upper
98	100		
100	66.1	53.5	80.3









#62
4-Bromofluorobenzene
Concen: 49.336 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046085.D
Acq: 08 May 2025 12:55




Instrument :
MSVOA_X
ClientSampleId :
RW8-BW-20250507


Tgt Ion: 95 Resp: 61840
Ion Ratio Lower Upper
95 100
174 67.8 0.0 135.8
176 66.4 0.0 131.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046085.D
Acq: 08 May 2025 12:55

Tgt Ion: 117 Resp: 121918
Ion Ratio Lower Upper
117 100
82 62.8 50.6 76.0
119 32.1 25.8 38.6





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046085.D


Acq: 08 May 2025 12:55

Instrument :

MSVOA_X

ClientSampleId :

RW8-BW-20250507





Tgt Ion:152 Resp: 50270

Ion Ratio Lower Upper

152 100

115 65.8 46.9 140.7

150 161.0 0.0 351.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046085.D
Acq On : 08 May 2025 12:55
Operator : JC/MD
Sample : Q1985-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-BW-20250507

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M

Title : SW846 8260

Signal : TIC: VX046085.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.246	23	26	36	rBV3	3898	7271	1.59%	0.298%
2	1.563	69	78	79	rBV4	3278	6439	1.41%	0.264%
3	1.593	79	83	89	rVB3	5078	7968	1.74%	0.326%
4	5.026	637	646	658	rBV4	3700	12740	2.79%	0.522%
5	5.379	693	704	718	rBV2	54376	161580	35.34%	6.619%
6	5.544	722	731	743	rBV	76348	217690	47.61%	8.918%
7	5.952	789	798	814	rBV	64472	170910	37.38%	7.001%
8	6.757	920	930	946	rBV	136474	334367	73.12%	13.697%
9	8.647	1234	1240	1258	rBV	289690	457270	100.00%	18.732%
10	10.049	1465	1470	1491	rBV	304604	414854	90.72%	16.995%
11	11.079	1634	1639	1648	rBV	232943	304161	66.52%	12.460%
12	12.018	1788	1793	1802	rBV	273833	345847	75.63%	14.168%




Sum of corrected areas: 2441097

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046085.D
Acq On : 08 May 2025 12:55
Operator : JC/MD
Sample : Q1985-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-BW-20250507

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046085.D
Acq On : 08 May 2025 12:55
Operator : JC/MD
Sample : Q1985-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-BW-20250507

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046085.D
Acq On : 08 May 2025 12:55
Operator : JC/MD
Sample : Q1985-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-BW-20250507

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	05/07/25		
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	05/08/25		
Client Sample ID:	TB-20250507			SDG No.:	Q1985		
Lab Sample ID:	Q1985-02			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046084.D	1		05/08/25 12:32	VX050825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	05/07/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	05/08/25
Client Sample ID:	TB-20250507	SDG No.:	Q1985
Lab Sample ID:	Q1985-02	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046084.D	1		05/08/25 12:32	VX050825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.4		81 - 118		103%	SPK: 50
1868-53-7	Dibromofluoromethane	49.4		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	49.2		89 - 112		98%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.4		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	64100	5.55				
540-36-3	1,4-Difluorobenzene	126000	6.757				
3114-55-4	Chlorobenzene-d5	115000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	49500	12.018				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	05/07/25	
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	05/08/25	
Client Sample ID:	TB-20250507			SDG No.:	Q1985	
Lab Sample ID:	Q1985-02			Matrix:	Water	
Analytical Method:	SW8260			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046084.D	1		05/08/25 12:32	VX050825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046084.D
 Acq On : 08 May 2025 12:32
 Operator : JC/MD
 Sample : Q1985-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-20250507

Quant Time: May 09 01:56:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	64053	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	126115	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	114932	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49527	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61356	51.380	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.760%
35) Dibromofluoromethane	5.385	113	44864	49.401	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.800%
50) Toluene-d8	8.647	98	154594	49.183	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.360%
62) 4-Bromofluorobenzene	11.079	95	59587	49.421	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.840%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046084.D
Acq On : 08 May 2025 12:32
Operator : JC/MD
Sample : Q1985-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20250507

Quant Time: May 09 01:56:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration



#1
Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.550 min Scan# 731

Delta R.T. 0.006 min

Lab File: VX046084.D

Acq: 08 May 2025 12:32

Instrument :

MSVOA_X

ClientSampleId :





TB-20250507

Tgt Ion:168 Resp: 64053

Ion Ratio Lower Upper

168 100

99 68.1 54.9 82.3



#33

1,2-Dichloroethane-d4

Concen: 51.380 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046084.D




Acq: 08 May 2025 12:32


Tgt Ion: 65 Resp: 61356

Ion Ratio Lower Upper

65 100

67 48.6 0.0 99.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX046084.D
Acq: 08 May 2025 12:32




Instrument :
MSVOA_X
ClientSampleId :
TB-20250507


Tgt Ion:114 Resp: 126115
Ion Ratio Lower Upper
114 100
63 24.7 0.0 49.2
88 16.9 0.0 33.6



#35
Dibromofluoromethane
Concen: 49.401 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX046084.D
Acq: 08 May 2025 12:32

Tgt Ion:113 Resp: 44864
Ion Ratio Lower Upper
113 100
111 103.4 83.1 124.7
192 16.9 13.3 19.9





#50

Toluene-d8

Concen: 49.183 ug/l

RT: 8.647 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046084.D

Acq: 08 May 2025 12:32

Instrument :

MSVOA_X

ClientSampleId :





TB-20250507

Tgt Ion: 98 Resp: 154594

Ion Ratio Lower Upper

98 100

100 65.1 53.5 80.3



#62

4-Bromofluorobenzene

Concen: 49.421 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.000 min

Lab File: VX046084.D

Acq: 08 May 2025 12:32




Tgt Ion: 95 Resp: 59587


Ion Ratio Lower Upper

95 100

174 67.6 0.0 135.8

176 64.9 0.0 131.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.049 min Scan# 1470

Delta R.T. -0.000 min

Lab File: VX046084.D

Acq: 08 May 2025 12:32

Instrument :

MSVOA_X

ClientSampleId :

TB-20250507





Tgt Ion:117 Resp: 114932

Ion Ratio Lower Upper

117 100

82 61.7 50.6 76.0

119 33.0 25.8 38.6



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX046084.D

Acq: 08 May 2025 12:32




Tgt Ion:152 Resp: 49527

Ion Ratio Lower Upper

152 100

115 64.4 46.9 140.7

150 156.3 0.0 351.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046084.D
Acq On : 08 May 2025 12:32
Operator : JC/MD
Sample : Q1985-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20250507

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M

Title : SW846 8260

Signal : TIC: VX046084.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.252	22	27	33	rBV2	4136	7538	1.75%	0.327%
2	1.581	70	81	83	rBV4	4774	13524	3.13%	0.586%
3	5.385	692	705	722	rBV	50777	153432	35.55%	6.646%
4	5.544	722	731	747	rVV2	73913	210959	48.88%	9.138%
5	5.952	787	798	812	rBV	60412	159514	36.96%	6.910%
6	6.757	922	930	942	rBV2	128018	313498	72.63%	13.580%
7	8.647	1234	1240	1255	rBV	272703	431613	100.00%	18.697%
8	10.049	1465	1470	1486	rBV	275587	390212	90.41%	16.903%
9	11.079	1634	1639	1648	rBV	226453	290977	67.42%	12.605%
10	12.018	1788	1793	1808	rBV	263431	337231	78.13%	14.608%




Sum of corrected areas: 2308498

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046084.D
Acq On : 08 May 2025 12:32
Operator : JC/MD
Sample : Q1985-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20250507

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046084.D
Acq On : 08 May 2025 12:32
Operator : JC/MD
Sample : Q1985-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20250507

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046084.D
Acq On : 08 May 2025 12:32
Operator : JC/MD
Sample : Q1985-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20250507

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc



CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q1985 SAS No.: Q1985 SDG No.: Q1985
 Instrument ID: MSVOA_X Calibration Date(s): 05/05/2025 05/05/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:35 16:27
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:	RRF020 = VX046041.D		RRF050 = VX046042.D		RRF100 = VX046043.D			
	RRF150 = VX046044.D		RRF005 = VX046046.D		RRF001 = VX046047.D			
COMPOUND	RRF020	RRF050	RRF100	RRF150	RRF005	RRF001	RRF	% RSD
Chloromethane	0.727	0.775	0.787	0.791	0.679	0.694	0.742	6.6
Vinyl Chloride	0.660	0.710	0.727	0.755	0.619	0.673	0.691	7.2
Bromomethane	0.296	0.326	0.340	0.334	0.305		0.320	5.8
Chloroethane	0.354	0.378	0.329	0.317	0.368	0.467	0.369	14.4
Trichlorofluoromethane	1.035	1.068	0.983	0.985	0.990	1.064	1.021	3.9
1,1,2-Trichlorotrifluoroethane	0.628	0.641	0.629	0.648	0.610	0.633	0.632	2.1
1,1-Dichloroethene	0.565	0.601	0.607	0.625	0.567	0.594	0.593	3.9
Acetone	0.361	0.362	0.361	0.370	0.408	0.380	0.374	4.9
Carbon Disulfide	1.295	1.455	1.522	1.597	1.141	1.423	1.406	11.7
Methyl tert-butyl Ether	2.044	2.160	2.172	2.239	1.908	1.949	2.079	6.4
Methylene Chloride	0.689	0.684	0.691	0.691	0.689	0.853	0.716	9.4
trans-1,2-Dichloroethene	0.573	0.610	0.612	0.622	0.557	0.604	0.596	4.3
1,1-Dichloroethane	1.233	1.263	1.263	1.286	1.154	1.116	1.219	5.6
2-Butanone	0.540	0.555	0.558	0.569	0.539	0.495	0.543	4.8
Carbon Tetrachloride	0.528	0.558	0.552	0.577	0.505	0.541	0.544	4.6
cis-1,2-Dichloroethene	0.716	0.737	0.738	0.755	0.642	0.719	0.718	5.5
Chloroform	1.287	1.296	1.277	1.300	1.199	1.265	1.271	3
1,1,1-Trichloroethane	1.106	1.131	1.155	1.188	1.013	1.015	1.101	6.6
Methylcyclohexane	0.596	0.641	0.627	0.658	0.587	0.627	0.623	4.3
Benzene	1.426	1.474	1.441	1.477	1.337	1.348	1.417	4.3
1,2-Dichloroethane	0.632	0.627	0.611	0.625	0.594	0.579	0.612	3.5
Trichloroethene	0.344	0.355	0.345	0.362	0.315	0.324	0.341	5.3
1,2-Dichloropropane	0.356	0.371	0.368	0.378	0.324	0.317	0.352	7.4
Bromodichloromethane	0.557	0.577	0.573	0.594	0.498	0.485	0.547	8.2
4-Methyl-2-Pentanone	0.620	0.634	0.630	0.631	0.555	0.561	0.605	6
Toluene	0.884	0.898	0.885	0.904	0.838	0.803	0.869	4.5
t-1,3-Dichloropropene	0.468	0.528	0.555	0.591	0.406	0.371	0.487	17.9
cis-1,3-Dichloropropene	0.531	0.578	0.602	0.623	0.469	0.423	0.538	14.6
1,1,2-Trichloroethane	0.349	0.354	0.351	0.356	0.337	0.308	0.343	5.3
2-Hexanone	0.466	0.473	0.477	0.473	0.414	0.385	0.448	8.7

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q1985 SAS No.: Q1985 SDG No.: Q1985
 Instrument ID: MSVOA_X Calibration Date(s): 05/05/2025 05/05/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:35 16:27
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF020 = VX046041.D		RRF050 = VX046042.D		RRF100 = VX046043.D			
		RRF150 = VX046044.D		RRF005 = VX046046.D		RRF001 = VX046047.D			
COMPOUND	RRF020	RRF050	RRF100	RRF150	RRF005	RRF001	RRF	% RSD	
Dibromochloromethane	0.378	0.400	0.415	0.431	0.326	0.306	0.376	13.3	
Tetrachloroethene	0.390	0.375	0.345	0.344	0.323	0.347	0.354	6.8	
Chlorobenzene	1.093	1.098	1.085	1.114	1.046	1.131	1.094	2.7	
Ethyl Benzene	1.919	2.022	1.979	2.036	1.816	1.803	1.929	5.2	
m/p-Xylenes	0.706	0.740	0.721	0.740	0.678	0.648	0.706	5.2	
o-Xylene	0.688	0.727	0.706	0.726	0.639	0.642	0.688	5.7	
Styrene	1.135	1.219	1.214	1.230	1.012	0.951	1.127	10.6	
Bromoform	0.270	0.304	0.312	0.327	0.236	0.234	0.281	14.2	
Isopropylbenzene	3.843	4.130	3.876	4.156	3.562	3.789	3.893	5.7	
1,1,2,2-Tetrachloroethane	1.315	1.338	1.284	1.345	1.350	1.552	1.364	7	
1,3-Dichlorobenzene	1.633	1.701	1.656	1.730	1.558	1.619	1.649	3.7	
1,4-Dichlorobenzene	1.629	1.693	1.639	1.722	1.606	1.817	1.684	4.6	
1,2-Dichlorobenzene	1.613	1.696	1.634	1.702	1.577	1.710	1.655	3.3	
1,2-Dichloroethane-d4	0.953	0.910	0.930	0.932	0.935		0.932	1.6	
Dibromofluoromethane	0.359	0.355	0.364	0.368	0.354		0.360	1.7	
Toluene-d8	1.246	1.223	1.266	1.275	1.221		1.246	2	
4-Bromofluorobenzene	0.455	0.470	0.500	0.500	0.464		0.478	4.4	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
 Method File : 82X050525W.M
 Title : SW846 8260
 Last Update : Tue May 06 07:12:22 2025
 Response Via : Initial Calibration

Calibration Files

1 =VX046047.D 5 =VX046046.D 20 =VX046041.D 50 =VX046042.D 100 =VX046043.D 150 =VX046044.D

	Compound	1	5	20	50	100	150	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.658	0.639	0.697	0.864	0.859	0.875	0.765	14.62
3) P	Chloromethane	0.694	0.679	0.727	0.775	0.787	0.791	0.742	6.58
4) C	Vinyl Chloride	0.673	0.619	0.660	0.710	0.727	0.755	0.691	7.17#
5) T	Bromomethane		0.305	0.296	0.326	0.340	0.334	0.320	5.84
6) T	Chloroethane	0.467	0.368	0.354	0.378	0.329	0.317	0.369	14.44
7) T	Trichlorofluor...	1.064	0.990	1.035	1.068	0.983	0.985	1.021	3.92
8) T	Diethyl Ether	0.403	0.311	0.340	0.337	0.338	0.355	0.347	8.83
9) T	1,1,2-Trichlor...	0.633	0.610	0.628	0.641	0.629	0.648	0.632	2.10
10) T	Methyl Iodide		0.608	0.767	0.806	0.793	0.763	0.747	10.68
11) T	Tert butyl alc...		0.114	0.122	0.129	0.144	0.146	0.131	10.45
12) CM	1,1-Dichloroet...	0.594	0.567	0.565	0.601	0.607	0.625	0.593	3.94#
13) T	Acrolein		0.117	0.158	0.152	0.154	0.163	0.149	12.33
14) T	Allyl chloride	1.052	1.058	1.127	1.179	1.187	1.196	1.133	5.75
15) T	Acrylonitrile	0.363	0.345	0.378	0.388	0.381	0.390	0.374	4.52
16) T	Acetone	0.380	0.408	0.361	0.362	0.361	0.370	0.374	4.90
17) T	Carbon Disulfide	1.423	1.141	1.295	1.455	1.522	1.597	1.406	11.68
18) T	Methyl Acetate	1.006	0.816	0.814	0.848	0.845	0.875	0.867	8.27
19) T	Methyl tert-bu...	1.949	1.908	2.044	2.160	2.172	2.239	2.079	6.39
20) T	Methylene Chlo...	0.853	0.689	0.689	0.684	0.691	0.691	0.716	9.39
21) T	trans-1,2-Dich...	0.604	0.557	0.573	0.610	0.612	0.622	0.596	4.27
22) T	Diisopropyl ether	2.095	1.924	2.219	2.278	2.295	2.321	2.189	6.97
23) T	Vinyl Acetate	1.660	1.698	1.928	2.048	2.082	2.134	1.925	10.52
24) P	1,1-Dichloroet...	1.116	1.154	1.233	1.263	1.263	1.286	1.219	5.60
25) T	2-Butanone	0.495	0.539	0.540	0.555	0.558	0.569	0.543	4.77
26) T	2,2-Dichloropr...	0.965	0.850	0.910	0.957	1.003	1.039	0.954	7.03
27) T	cis-1,2-Dichlo...	0.719	0.642	0.716	0.737	0.738	0.755	0.718	5.55
28) T	Bromochloromet...	0.576	0.553	0.628	0.578	0.595	0.590	0.587	4.28
29) T	Tetrahydrofuran	0.318	0.318	0.340	0.350	0.351	0.362	0.340	5.36
30) C	Chloroform	1.265	1.199	1.287	1.296	1.277	1.300	1.271	2.95#
31) T	Cyclohexane		1.059	1.090	1.128	1.128	1.150	1.111	3.26
32) T	1,1,1-Trichlor...	1.015	1.013	1.106	1.131	1.155	1.188	1.101	6.60
33) S	1,2-Dichloroet...		0.935	0.953	0.910	0.930	0.932	0.932	1.65
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...		0.354	0.359	0.355	0.364	0.368	0.360	1.70
36) T	1,1-Dichloropr...	0.493	0.462	0.463	0.495	0.483	0.505	0.484	3.68
37) T	Ethyl Acetate	0.586	0.582	0.569	0.611	0.609	0.631	0.598	3.83
38) T	Carbon Tetrach...	0.541	0.505	0.528	0.558	0.552	0.577	0.544	4.61
39) T	Methylcyclohexane	0.627	0.587	0.596	0.641	0.627	0.658	0.623	4.34
40) TM	Benzene	1.348	1.337	1.426	1.474	1.441	1.477	1.417	4.30
41) T	Methacrylonitrile	0.233	0.288	0.318	0.346	0.343	0.348	0.313	14.50
42) TM	1,2-Dichloroet...	0.579	0.594	0.632	0.627	0.611	0.625	0.612	3.45
43) T	Isopropyl Acetate	0.764	0.826	0.905	0.963	0.982	1.030	0.912	11.05
44) TM	Trichloroethene	0.324	0.315	0.344	0.355	0.345	0.362	0.341	5.28
45) C	1,2-Dichloropr...	0.317	0.324	0.356	0.371	0.368	0.378	0.352	7.37#
46) T	Dibromomethane	0.263	0.262	0.285	0.287	0.280	0.289	0.278	4.35
47) T	Bromodichlorom...	0.485	0.498	0.557	0.577	0.573	0.594	0.547	8.22
48) T	Methyl methacr...	0.370	0.426	0.465	0.502	0.500	0.531	0.466	12.75
49) T	1,4-Dioxane	0.007	0.009	0.009	0.009	0.009	0.010	0.009	8.45
50) S	Toluene-d8		1.221	1.246	1.223	1.266	1.275	1.246	1.95
51) T	4-Methyl-2-Pen...	0.561	0.555	0.620	0.634	0.630	0.631	0.605	6.04
52) CM	Toluene	0.803	0.838	0.884	0.898	0.885	0.904	0.869	4.54#
53) T	t-1,3-Dichloro...	0.371	0.406	0.468	0.528	0.555	0.591	0.487	17.86
54) T	cis-1,3-Dichlo...	0.423	0.469	0.531	0.578	0.602	0.623	0.538	14.61
55) T	1,1,2-Trichlor...	0.308	0.337	0.349	0.354	0.351	0.356	0.343	5.30
56) T	Ethyl methacry...	0.377	0.508	0.540	0.595	0.617	0.639	0.546	17.58

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
 Method File : 82X050525W.M

57)	T	1,3-Dichloropr...	0.610	0.601	0.618	0.623	0.613	0.627	0.615	1.53
58)	T	2-Chloroethyl ...	0.230	0.247	0.270	0.307	0.303	0.313	0.278	12.41
59)	T	2-Hexanone	0.385	0.414	0.466	0.473	0.477	0.473	0.448	8.69
60)	T	Dibromochlorom...	0.306	0.326	0.378	0.400	0.415	0.431	0.376	13.28
61)	T	1,2-Dibromoethane	0.322	0.333	0.359	0.373	0.368	0.381	0.356	6.54
62)	S	4-Bromofluorob...		0.464	0.455	0.470	0.500	0.500	0.478	4.41
63)	I	Chlorobenzene-d5	-----ISTD-----							
64)	T	Tetrachloroethene	0.347	0.323	0.390	0.375	0.345	0.344	0.354	6.84
65)	PM	Chlorobenzene	1.131	1.046	1.093	1.098	1.085	1.114	1.094	2.65
66)	T	1,1,1,2-Tetrac...	0.369	0.341	0.365	0.390	0.382	0.395	0.374	5.22
67)	C	Ethyl Benzene	1.803	1.816	1.919	2.022	1.979	2.036	1.929	5.24#
68)	T	m/p-Xylenes	0.648	0.678	0.706	0.740	0.721	0.740	0.706	5.21
69)	T	o-Xylene	0.642	0.639	0.688	0.727	0.706	0.726	0.688	5.75
70)	T	Styrene	0.951	1.012	1.135	1.219	1.214	1.230	1.127	10.56
71)	P	Bromoform	0.234	0.236	0.270	0.304	0.312	0.327	0.281	14.17
72)	I	1,4-Dichlorobenzen...	-----ISTD-----							
73)	T	Isopropylbenzene	3.789	3.562	3.843	4.130	3.876	4.156	3.893	5.72
74)	T	N-amyl acetate	1.715	1.652	1.846	2.067	2.068	2.192	1.924	11.32
75)	P	1,1,2,2-Tetrac...	1.552	1.350	1.315	1.338	1.284	1.345	1.364	6.97
76)	T	1,2,3-Trichlor...	1.405	1.167	1.151	1.187	1.131	1.181	1.204	8.36
77)	T	Bromobenzene	0.926	0.862	0.896	0.928	0.883	0.928	0.904	3.08
78)	T	n-propylbenzene	4.272	4.186	4.394	4.854	4.583	4.868	4.526	6.45
79)	T	2-Chlorotoluene	3.184	2.748	2.832	2.994	2.805	2.953	2.919	5.45
80)	T	1,3,5-Trimethy...	3.036	3.053	3.275	3.487	3.255	3.405	3.252	5.60
81)	T	trans-1,4-Dich...		0.269	0.335	0.385	0.410	0.449	0.370	18.91
82)	T	4-Chlorotoluene	3.226	2.939	3.196	3.430	3.255	3.379	3.238	5.32
83)	T	tert-Butylbenzene	3.341	3.098	3.115	3.435	3.255	3.411	3.276	4.44
84)	T	1,2,4-Trimethy...	3.150	3.034	3.274	3.522	3.335	3.444	3.293	5.52
85)	T	sec-Butylbenzene	3.708	3.767	3.937	4.282	4.095	4.343	4.022	6.55
86)	T	p-Isopropyltol...	3.025	3.084	3.206	3.555	3.450	3.599	3.320	7.44
87)	T	1,3-Dichlorobe...	1.619	1.558	1.633	1.701	1.656	1.729	1.649	3.71
88)	T	1,4-Dichlorobe...	1.817	1.606	1.629	1.693	1.639	1.722	1.684	4.64
89)	T	n-Butylbenzene	2.443	2.650	2.748	3.147	3.139	3.346	2.912	12.00
90)	T	Hexachloroethane	0.511	0.523	0.551	0.622	0.622	0.680	0.585	11.44
91)	T	1,2-Dichlorobe...	1.710	1.577	1.613	1.696	1.634	1.702	1.655	3.34
92)	T	1,2-Dibromo-3-...	0.259	0.248	0.299	0.322	0.329	0.356	0.302	13.89
93)	T	1,2,4-Trichlor...	0.862	0.842	0.861	0.981	1.035	1.123	0.951	12.03
94)	T	Hexachlorobuta...	0.393	0.414	0.394	0.427	0.418	0.445	0.415	4.79
95)	T	Naphthalene	3.499	2.929	3.204	3.613	3.690	3.984	3.487	10.69
96)	T	1,2,3-Trichlor...	0.941	0.846	0.921	1.019	1.051	1.107	0.981	9.74

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046041.D
 Acq On : 05 May 2025 11:35
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:08:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	83671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	147096	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127829	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	31909	12.820	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	25.640%#		
35) Dibromofluoromethane	5.385	113	21112	12.804	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	25.600%#		
50) Toluene-d8	8.646	98	73333	13.200	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	26.400%#		
62) 4-Bromofluorobenzene	11.079	95	26760	13.236	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	26.480%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	23341	12.638	ug/l	96
3) Chloromethane	1.306	50	24316	13.371	ug/l	98
4) Vinyl Chloride	1.373	62	22096	13.868	ug/l	94
5) Bromomethane	1.593	94	9920	12.279	ug/l	93
6) Chloroethane	1.666	64	11832	14.414	ug/l	95
7) Trichlorofluoromethane	1.873	101	34636	14.271	ug/l	100
8) Diethyl Ether	2.129	74	11392	14.278	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.324	101	21027	14.499	ug/l	98
10) Methyl Iodide	2.446	142	25661	15.070	ug/l	99
11) Tert butyl alcohol	2.971	59	20370	68.662	ug/l	97
12) 1,1-Dichloroethene	2.312	96	18899	13.385	ug/l	98
13) Acrolein	2.233	56	26427	76.689	ug/l	95
14) Allyl chloride	2.660	41	37708	14.133	ug/l	99
15) Acrylonitrile	3.062	53	63290	71.312	ug/l	98
16) Acetone	2.379	43	60363	70.884	ug/l	99
17) Carbon Disulfide	2.501	76	43348	13.251	ug/l	100
18) Methyl Acetate	2.702	43	27234	13.397	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	68402	13.890	ug/l	100
20) Methylene Chloride	2.782	84	23050	13.362	ug/l	96
21) trans-1,2-Dichloroethene	3.086	96	19161	13.294	ug/l	96
22) Diisopropyl ether	3.763	45	74257	14.896	ug/l	95
23) Vinyl Acetate	3.720	43	322598	73.230	ug/l	100
24) 1,1-Dichloroethane	3.605	63	41257	14.082	ug/l	98
25) 2-Butanone	4.556	43	90331	73.489	ug/l	98
26) 2,2-Dichloropropane	4.470	77	30472	13.729	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	23970	13.791	ug/l	98
28) Bromochloromethane	4.897	49	21028	13.497	ug/l	99
29) Tetrahydrofuran	5.007	42	56819	71.175	ug/l	100
30) Chloroform	5.086	83	43065	14.144	ug/l	95
31) Cyclohexane	5.458	56	36470	14.769	ug/l	95
32) 1,1,1-Trichloroethane	5.379	97	37003	14.092	ug/l	98
36) 1,1-Dichloropropene	5.684	75	27242	14.231	ug/l	98
37) Ethyl Acetate	4.720	43	33458	13.931	ug/l	98
38) Carbon Tetrachloride	5.671	117	31091	14.137	ug/l	98
39) Methylcyclohexane	7.372	83	35077	14.623	ug/l	99
40) Benzene	6.031	78	83898	14.140	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046041.D
 Acq On : 05 May 2025 11:35
 Operator : JC/MD
 Sample : VSTDIC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:08:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.915	41	18732	14.038	ug/l	95
42) 1,2-Dichloroethane	6.080	62	37203	15.165	ug/l	100
43) Isopropyl Acetate	6.342	43	53275	14.512	ug/l	99
44) Trichloroethene	7.116	130	20239	14.429	ug/l	99
45) 1,2-Dichloropropane	7.427	63	20940	14.204	ug/l	98
46) Dibromomethane	7.580	93	16781	14.439	ug/l	99
47) Bromodichloromethane	7.817	83	32762	14.545	ug/l	94
48) Methyl methacrylate	7.689	41	27374	14.521	ug/l	100
49) 1,4-Dioxane	7.659	88	10585	282.442	ug/l	98
51) 4-Methyl-2-Pentanone	8.573	43	182455	75.775	ug/l	98
52) Toluene	8.713	92	52030	14.788	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	27549	14.566	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	31226	14.154	ug/l	98
55) 1,1,2-Trichloroethane	9.152	97	20536	14.442	ug/l	96
56) Ethyl methacrylate	9.116	69	31787	14.425	ug/l	98
57) 1,3-Dichloropropane	9.305	76	36341	14.417	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	79531	80.312	ug/l	100
59) 2-Hexanone	9.427	43	137054	74.997	ug/l	99
60) Dibromochloromethane	9.518	129	22267	14.387	ug/l	99
61) 1,2-Dibromoethane	9.610	107	21113	14.458	ug/l	96
64) Tetrachloroethene	9.268	164	19924	15.263	ug/l	97
65) Chlorobenzene	10.079	112	55863	14.202	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.158	131	18671	14.231	ug/l	98
67) Ethyl Benzene	10.189	91	98113	14.631	ug/l	100
68) m/p-Xylenes	10.299	106	72202	29.806	ug/l	99
69) o-Xylene	10.640	106	35178	14.408	ug/l	96
70) Styrene	10.652	104	58034	14.853	ug/l	98
71) Bromoform	10.799	173	13825	14.030	ug/l #	100
73) Isopropylbenzene	10.957	105	93013	14.096	ug/l	99
74) N-amyl acetate	10.841	43	44684	13.581	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	31830	13.560	ug/l	99
76) 1,2,3-Trichloropropane	11.237	75	27845m	11.039	ug/l	
77) Bromobenzene	11.195	156	21695	14.212	ug/l	96
78) n-propylbenzene	11.298	91	106337	14.376	ug/l	99
79) 2-Chlorotoluene	11.359	91	68534	13.748	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	79261	14.468	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	8104	13.371	ug/l	91
82) 4-Chlorotoluene	11.451	91	77338	14.116	ug/l	100
83) tert-Butylbenzene	11.713	119	75382	14.011	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	79232	14.480	ug/l	99
85) sec-Butylbenzene	11.890	105	95271	14.263	ug/l	100
86) p-Isopropyltoluene	12.006	119	77599	14.387	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	39520	13.898	ug/l	98
88) 1,4-Dichlorobenzene	12.036	146	39419	14.146	ug/l	99
89) n-Butylbenzene	12.329	91	66506	14.184	ug/l	99
90) Hexachloroethane	12.536	117	13332	13.287	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	39031	14.092	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.938	75	7232	13.535	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	20846	13.745	ug/l	99
94) Hexachlorobutadiene	13.725	225	9524	13.426	ug/l	97
95) Naphthalene	13.774	128	77542	14.040	ug/l	99
96) 1,2,3-Trichlorobenzene	13.956	180	22301	14.023	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046041.D
Acq On : 05 May 2025 11:35
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:08:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.543 min Scan# 71
Delta R.T. -0.001 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion:168 Resp: 8367
Ion Ratio Lower Upper
168 100
99 72.7 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 12.638 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 85 Resp: 23341
Ion Ratio Lower Upper
85 100
87 33.5 15.7 47.1









#3
Chloromethane
Concen: 13.371 ug/l
RT: 1.306 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 50 Resp: 24310
Ion Ratio Lower Upper
50 100
52 33.0 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 13.868 ug/l
RT: 1.373 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 62 Resp: 22096
Ion Ratio Lower Upper
62 100
64 34.9 25.2 37.8





#5
Bromomethane
Concen: 12.279 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion: 94 Resp: 9920
Ion Ratio Lower Upper
94 100
96 98.1 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 14.414 ug/l
RT: 1.666 min Scan# 95
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 64 Resp: 11832
Ion Ratio Lower Upper
64 100
66 32.9 24.3 36.5





#7
Trichlorofluoromethane
Concen: 14.271 ug/l
RT: 1.873 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion:101 Resp: 34630
Ion Ratio Lower Upper
101 100
103 63.5 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 14.278 ug/l
RT: 2.129 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 74 Resp: 11392
Ion Ratio Lower Upper
74 100
45 103.7 54.9 164.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 14.499 ug/l

RT: 2.324 min Scan# 202

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:	101	Resp:	2102
Ion	Ratio	Lower	Upper
101	100		
85	46.6	38.6	58.0
151	70.4	55.2	82.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#10

Methyl Iodide

Concen: 15.070 ug/l




RT: 2.446 min Scan# 223


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	142	Resp:	25661
Ion	Ratio	Lower	Upper
142	100		
127	53.0	41.7	62.5
141	14.7	11.5	17.3





#11

Tert butyl alcohol

Concen: 68.662 ug/l

RT: 2.971 min Scan# 309

Delta R.T. -0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 59 Resp: 20370

Ion Ratio Lower Upper

59 100





57 11.7 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 13.385 ug/l

RT: 2.312 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 96 Resp: 18899


Ion Ratio Lower Upper

96 100

61 179.8 146.2 219.2

98 61.8 51.0 76.6





#13

Acrolein

Concen: 76.689 ug/l

RT: 2.233 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 56 Resp: 2642

Ion Ratio Lower Upper

56 100





55 74.3 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 14.133 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 41 Resp: 37708


Ion Ratio Lower Upper

41 100

39 75.0 60.6 90.8

76 30.6 24.9 37.3





#15

Acrylonitrile

Concen: 71.312 ug/l

RT: 3.062 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion:	53	Resp:	63290
Ion	Ratio	Lower	Upper
53	100		
52	83.5	65.3	97.9
51	38.4	29.8	44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 70.884 ug/l




RT: 2.379 min Scan# 212


Delta R.T. -0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	43	Resp:	60363
Ion	Ratio	Lower	Upper
43	100		
58	26.0	21.2	31.8





#17

Carbon Disulfide

Concen: 13.251 ug/l

RT: 2.501 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 76 Resp: 4334

Ion Ratio Lower Upper

76 100





78 8.8 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 13.397 ug/l

RT: 2.702 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 43 Resp: 27234

Ion Ratio Lower Upper

43 100

74 20.7 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 13.890 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion: 73 Resp: 6840
Ion Ratio Lower Upper
73 100
57 22.2 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#20
Methylene Chloride
Concen: 13.362 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 84 Resp: 23050
Ion Ratio Lower Upper
84 100
49 146.0 113.9 170.9
51 43.8 33.5 50.3
86 62.2 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 13.294 ug/l
RT: 3.086 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion: 96 Resp: 1916
Ion Ratio Lower Upper
96 100
61 155.6 119.5 179.3
98 63.2 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 14.896 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 45 Resp: 74257
Ion Ratio Lower Upper
45 100
43 78.2 66.6 100.0
87 24.0 19.8 29.6
59 11.7 8.6 12.8





#23

Vinyl Acetate

Concen: 73.230 ug/l

RT: 3.720 min Scan# 41

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 43 Resp: 322598

Ion Ratio Lower Upper

43 100





86 9.3 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#24

1,1-Dichloroethane

Concen: 14.082 ug/l

RT: 3.605 min Scan# 413

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 63 Resp: 41257


Ion Ratio Lower Upper

63 100

98 7.9 3.6 10.8

100 5.0 2.1 6.3





#25

2-Butanone

Concen: 73.489 ug/l

RT: 4.556 min Scan# 50

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 9033

Ion Ratio Lower Upper

43 100





72 23.8 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 13.729 ug/l

RT: 4.470 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 77 Resp: 30472

Ion Ratio Lower Upper

77 100

97 20.9 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 13.791 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 96 Resp: 23970

Ion Ratio Lower Upper

96 100

61 158.3 0.0 322.8





98 63.5 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 13.497 ug/l

RT: 4.897 min Scan# 625

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 49 Resp: 21028


Ion Ratio Lower Upper

49 100

129 1.9 0.0 4.0

130 70.9 56.2 84.2





#29

Tetrahydrofuran

Concen: 71.175 ug/l

RT: 5.007 min Scan# 642

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:	42	Resp:	56819
Ion Ratio	Lower	Upper	
42	100		
72	42.3	33.7	50.5
71	39.2	31.4	47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 14.144 ug/l




RT: 5.086 min Scan# 656


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	83	Resp:	43065
Ion Ratio	Lower	Upper	
83	100		
85	65.7	49.3	73.9









#31
Cyclohexane
Concen: 14.769 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 56 Resp: 36470
Ion Ratio Lower Upper
56 100
69 28.8 24.4 36.6
84 78.5 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#32
1,1,1-Trichloroethane
Concen: 14.092 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 97 Resp: 37003
Ion Ratio Lower Upper
97 100
99 62.8 51.8 77.6
61 49.5 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 12.820 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 65 Resp: 31909

Ion Ratio Lower Upper

65 100





67 49.9 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:114 Resp: 147096


Ion Ratio Lower Upper

114 100

63 25.2 0.0 49.2


88 18.0 0.0 33.6





#35
Dibromofluoromethane
Concen: 12.804 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion: 113 Resp: 2111
Ion Ratio Lower Upper
113 100
111 103.0 83.1 124.7
192 17.0 13.3 19.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#36
1,1-Dichloropropene
Concen: 14.231 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 75 Resp: 27242
Ion Ratio Lower Upper
75 100
110 31.6 16.3 48.9
77 31.9 24.3 36.5





#37

Ethyl Acetate

Concen: 13.931 ug/l

RT: 4.720 min Scan# 595

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion:	43	Resp:	3345
Ion	Ratio	Lower	Upper
43	100		
61	12.6	10.3	15.5
70	10.9	7.9	11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 14.137 ug/l




RT: 5.671 min Scan# 752


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	117	Resp:	31091
Ion	Ratio	Lower	Upper
117	100		
119	95.6	75.2	112.8
121	32.1	24.2	36.4





#39

Methylcyclohexane

Concen: 14.623 ug/l

RT: 7.372 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion: 83	Resp: 3507
Ion Ratio	Lower Upper
83 100	
55 80.7	64.7 97.1
98 44.8	37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 14.140 ug/l




RT: 6.031 min Scan# 811


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35


Tgt Ion: 78	Resp: 83898
Ion Ratio	Lower Upper
78 100	
77 22.3	19.0 28.4





#41
Methacrylonitrile
Concen: 14.038 ug/l
RT: 4.915 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion: 41 Resp: 18732
Ion Ratio Lower Upper
41 100
39 62.6 47.2 70.8
67 68.3 50.7 76.1
52 32.5 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 15.165 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 62 Resp: 37203
Ion Ratio Lower Upper
62 100
98 7.6 0.0 15.2





#43

Isopropyl Acetate

Concen: 14.512 ug/l

RT: 6.342 min Scan# 80

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion: 43	Resp: 5327
Ion Ratio	Lower Upper
43	100
61	17.3 14.3 21.5
87	11.6 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 14.429 ug/l




RT: 7.116 min Scan# 989


Delta R.T. -0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:130	Resp: 20239
Ion Ratio	Lower Upper
130	100
95	102.8 0.0 204.2





#45

1,2-Dichloropropane

Concen: 14.204 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 63 Resp: 20940

Ion Ratio Lower Upper

63 100





65 29.1 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 14.439 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 93 Resp: 16781


Ion Ratio Lower Upper

93 100

95 81.5 65.6 98.4

174 83.1 68.2 102.2





#47

Bromodichloromethane

Concen: 14.545 ug/l

RT: 7.817 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:	83	Resp:	32762
Ion	Ratio	Lower	Upper
83	100		
85	68.4	50.5	75.7
127	8.3	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 14.521 ug/l




RT: 7.689 min Scan# 1083


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	41	Resp:	27374
Ion	Ratio	Lower	Upper
41	100		
69	73.5	58.5	87.7
39	64.9	51.7	77.5





#49

1,4-Dioxane

Concen: 282.442 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion:	88	Resp:	1058
Ion	Ratio	Lower	Upper
88	100		
43	45.5	33.4	50.2
58	73.6	58.6	88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 13.200 ug/l




RT: 8.646 min Scan# 1240


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35


Tgt Ion:	98	Resp:	73333
Ion	Ratio	Lower	Upper
98	100		
100	64.8	53.5	80.3





#51
4-Methyl-2-Pentanone
Concen: 75.775 ug/l
RT: 8.573 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion: 43 Resp: 18245
Ion Ratio Lower Upper
43 100
58 34.8 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 14.788 ug/l
RT: 8.713 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 92 Resp: 52030
Ion Ratio Lower Upper
92 100
91 172.2 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 14.566 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 75 Resp: 27549

Ion Ratio Lower Upper

75 100





77 31.4 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 14.154 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 75 Resp: 31226


Ion Ratio Lower Upper

75 100

77 30.8 25.4 38.0

39 64.0 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 14.442 ug/l

RT: 9.152 min Scan# 1323

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 97 Resp: 20530

Ion Ratio Lower Upper

97 100

83 87.0 69.0 103.6

85 58.7 44.3 66.5





99 66.8 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 14.425 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 69 Resp: 31787


Ion Ratio Lower Upper

69 100

41 78.2 60.8 91.2

39 49.3 39.0 58.6





#57

1,3-Dichloropropane

Concen: 14.417 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 76 Resp: 3634

Ion Ratio Lower Upper

76 100





78 31.8 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 80.312 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 63 Resp: 79531

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3





#59

2-Hexanone

Concen: 74.997 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 137054

Ion Ratio Lower Upper

43 100





58 48.7 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 14.387 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion:129 Resp: 22267

Ion Ratio Lower Upper

129 100

127 78.0 39.3 117.8









#61
1,2-Dibromoethane
Concen: 14.458 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion:107 Resp: 2111
Ion Ratio Lower Upper
107 100
109 95.3 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carbone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#62
4-Bromofluorobenzene
Concen: 13.236 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 95 Resp: 26760
Ion Ratio Lower Upper
95 100
174 67.4 0.0 135.8
176 67.0 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion:117 Resp: 127829
Ion Ratio Lower Upper
117 100
82 57.7 50.6 76.0
119 32.7 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 15.263 ug/l
RT: 9.268 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:164 Resp: 19924
Ion Ratio Lower Upper
164 100
166 127.2 105.0 157.6
129 107.6 83.5 125.3
131 97.7 76.5 114.7









#65
Chlorobenzene
Concen: 14.202 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion:112 Resp: 5586
Ion Ratio Lower Upper
112 100
114 32.0 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 14.231 ug/l
RT: 10.158 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:131 Resp: 18671
Ion Ratio Lower Upper
131 100
133 94.5 47.3 141.9
119 66.7 31.6 95.0









#67
Ethyl Benzene
Concen: 14.631 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 91 Resp: 9811
Ion Ratio Lower Upper
91 100
106 29.3 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#68
m/p-Xylenes
Concen: 29.806 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:106 Resp: 72202
Ion Ratio Lower Upper
106 100
91 216.1 171.2 256.8





#69

o-Xylene

Concen: 14.408 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion:106 Resp: 35178

Ion Ratio Lower Upper

106 100





91 231.7 112.7 338.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#70

Styrene

Concen: 14.853 ug/l

RT: 10.652 min Scan# 1569

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:104 Resp: 58034


Ion Ratio Lower Upper

104 100

78 58.7 45.7 68.5


103 55.6 43.7 65.5





#71
Bromoform
Concen: 14.030 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:173 Resp: 13825
Ion Ratio Lower Upper
173 100
175 46.4 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:152 Resp: 60503
Ion Ratio Lower Upper
152 100
115 76.6 46.9 140.7
150 163.6 0.0 351.0









#73
Isopropylbenzene
Concen: 14.096 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion:105 Resp: 9301
Ion Ratio Lower Upper
105 100
120 26.1 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 13.581 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 43 Resp: 44684
Ion Ratio Lower Upper
43 100
70 38.5 30.9 46.3
55 24.1 18.7 28.1
61 22.3 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 13.560 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 83 Resp: 31830
Ion Ratio Lower Upper
83 100
131 10.1 5.0 14.9
85 63.3 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#76
1,2,3-Trichloropropane
Concen: 11.039 ug/l m
RT: 11.237 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 75 Resp: 27845
Ion Ratio Lower Upper
75 100
77 43.0 20.5 61.5





#77
Bromobenzene
Concen: 14.212 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion	Ratio	Lower	Upper
156	100		
77	179.9	92.8	278.3
158	93.1	48.8	146.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#78
n-propylbenzene
Concen: 14.376 ug/l
RT: 11.298 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion	Ratio	Lower	Upper
91	100		
120	22.0	10.8	32.4









#79
2-Chlorotoluene
Concen: 13.748 ug/l
RT: 11.359 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 91 Resp: 68534
Ion Ratio Lower Upper
91 100
126 31.4 15.6 46.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#80
1,3,5-Trimethylbenzene
Concen: 14.468 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:105 Resp: 79261
Ion Ratio Lower Upper
105 100
120 45.9 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 13.371 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 75 Resp: 8104
Ion Ratio Lower Upper
75 100
53 138.3 100.5 150.7
89 44.6 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 14.116 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 91 Resp: 77338
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83
tert-Butylbenzene
Concen: 14.011 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:119 Resp: 75382
Ion Ratio Lower Upper
119 100
91 66.8 32.9 98.7
134 23.7 11.4 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#84
1,2,4-Trimethylbenzene
Concen: 14.480 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:105 Resp: 79232
Ion Ratio Lower Upper
105 100
120 42.0 21.2 63.6









#85
sec-Butylbenzene
Concen: 14.263 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion:105 Resp: 9527
Ion Ratio Lower Upper
105 100
134 19.3 9.7 29.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#86
p-Isopropyltoluene
Concen: 14.387 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:119 Resp: 77599
Ion Ratio Lower Upper
119 100
134 25.9 12.5 37.5
91 28.8 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 13.898 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion:146 Resp: 39520

Ion Ratio Lower Upper

146 100

111 42.7 22.1 66.3





148 62.9 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 14.146 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:146 Resp: 39419


Ion Ratio Lower Upper

146 100

111 42.8 21.3 63.9

148 64.2 31.9 95.5





#89

n-Butylbenzene

Concen: 14.184 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 91 Resp: 66500

Ion Ratio Lower Upper

91 100

92 52.9 26.9 80.7





134 23.1 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 13.287 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion:117 Resp: 13332

Ion Ratio Lower Upper

117 100


201 62.9 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 14.092 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion:146 Resp: 3903
Ion Ratio Lower Upper
146 100
111 46.6 22.4 67.3
148 64.2 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 13.535 ug/l
RT: 12.938 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 75 Resp: 7232
Ion Ratio Lower Upper
75 100
155 69.4 34.9 104.8
157 88.4 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 13.745 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:180 Resp: 20840
Ion Ratio Lower Upper
180 100
182 96.6 47.9 143.8
145 34.1 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#94
Hexachlorobutadiene
Concen: 13.426 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion:225 Resp: 9524
Ion Ratio Lower Upper
225 100
223 65.6 30.8 92.4
227 61.4 30.9 92.7





#95
Naphthalene
Concen: 14.040 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:128 Resp: 7754
Ion Ratio Lower Upper
128 100
127 12.9 10.4 15.6
129 11.1 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 14.023 ug/l
RT: 13.956 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:180 Resp: 22301
Ion Ratio Lower Upper
180 100
182 92.1 47.8 143.3
145 36.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046042.D
 Acq On : 05 May 2025 11:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	97076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	167947	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	146257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67976	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	88371	30.601	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	61.200%#		
35) Dibromofluoromethane	5.379	113	59550	31.633	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	63.260%#		
50) Toluene-d8	8.647	98	205479	32.395	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	64.780%#		
62) 4-Bromofluorobenzene	11.079	95	79012	34.229	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	68.460%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	83826	39.120	ug/l	100
3) Chloromethane	1.307	50	75238	35.658	ug/l	100
4) Vinyl Chloride	1.374	62	68901	37.273	ug/l	100
5) Bromomethane	1.593	94	31648	33.765	ug/l	100
6) Chloroethane	1.666	64	36722	38.558	ug/l	100
7) Trichlorofluoromethane	1.874	101	103693	36.824	ug/l	100
8) Diethyl Ether	2.136	74	32724	35.350	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	62251	36.997	ug/l	100
10) Methyl Iodide	2.447	142	78277	39.622	ug/l	100
11) Tert butyl alcohol	2.977	59	62557	181.745	ug/l	100
12) 1,1-Dichloroethene	2.313	96	58334	35.610	ug/l	100
13) Acrolein	2.233	56	73927	184.907	ug/l	100
14) Allyl chloride	2.654	41	114445	36.971	ug/l	100
15) Acrylonitrile	3.062	53	188304	182.874	ug/l	100
16) Acetone	2.386	43	175777	177.910	ug/l	100
17) Carbon Disulfide	2.502	76	141281	37.223	ug/l	100
18) Methyl Acetate	2.703	43	82347	34.915	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	209665	36.695	ug/l	100
20) Methylene Chloride	2.782	84	66412	33.182	ug/l	100
21) trans-1,2-Dichloroethene	3.087	96	59234	35.421	ug/l	100
22) Diisopropyl ether	3.757	45	221172	38.241	ug/l	100
23) Vinyl Acetate	3.721	43	994236	194.528	ug/l	100
24) 1,1-Dichloroethane	3.605	63	122601	36.067	ug/l	100
25) 2-Butanone	4.556	43	269224	188.783	ug/l	100
26) 2,2-Dichloropropane	4.465	77	92857	36.059	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	71518	35.466	ug/l	100
28) Bromochloromethane	4.898	49	56091	31.031	ug/l	100
29) Tetrahydrofuran	5.001	42	170124	183.679	ug/l	100
30) Chloroform	5.087	83	125850	35.626	ug/l	100
31) Cyclohexane	5.464	56	109459	38.205	ug/l	100
32) 1,1,1-Trichloroethane	5.373	97	109781	36.034	ug/l	100
36) 1,1-Dichloropropene	5.684	75	83215	38.073	ug/l	100
37) Ethyl Acetate	4.715	43	102537	37.394	ug/l	100
38) Carbon Tetrachloride	5.672	117	93712	37.320	ug/l	100
39) Methylcyclohexane	7.373	83	107651	39.307	ug/l	100
40) Benzene	6.031	78	247476	36.530	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046042.D
 Acq On : 05 May 2025 11:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	58082	38.123	ug/l	100
42) 1,2-Dichloroethane	6.080	62	105332	37.605	ug/l	100
43) Isopropyl Acetate	6.336	43	161787	38.598	ug/l	100
44) Trichloroethene	7.123	130	59623	37.230	ug/l	100
45) 1,2-Dichloropropane	7.428	63	62387	37.064	ug/l	100
46) Dibromomethane	7.574	93	48201	36.326	ug/l	100
47) Bromodichloromethane	7.818	83	96916	37.685	ug/l	100
48) Methyl methacrylate	7.690	41	84277	39.157	ug/l	100
49) 1,4-Dioxane	7.659	88	30529	713.477	ug/l	100
51) 4-Methyl-2-Pentanone	8.574	43	532388	193.653	ug/l	100
52) Toluene	8.714	92	150757	37.530	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	88655	41.056	ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	97031	38.521	ug/l	100
55) 1,1,2-Trichloroethane	9.147	97	59499	36.648	ug/l	100
56) Ethyl methacrylate	9.116	69	99947	39.726	ug/l	100
57) 1,3-Dichloropropane	9.305	76	104606	36.347	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	257678	227.902	ug/l	100
59) 2-Hexanone	9.427	43	396784	190.168	ug/l	100
60) Dibromochloromethane	9.519	129	67198	38.027	ug/l	100
61) 1,2-Dibromoethane	9.604	107	62633	37.567	ug/l	100
64) Tetrachloroethene	9.269	164	54853	36.726	ug/l	100
65) Chlorobenzene	10.079	112	160600	35.686	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	57066	38.017	ug/l	100
67) Ethyl Benzene	10.189	91	295712	38.543	ug/l	100
68) m/p-Xylenes	10.299	106	216578	78.142	ug/l	100
69) o-Xylene	10.640	106	106335	38.064	ug/l	100
70) Styrene	10.653	104	178313	39.888	ug/l	100
71) Bromoform	10.799	173	44486	39.457	ug/l #	100
73) Isopropylbenzene	10.957	105	280719	37.865	ug/l	100
74) N-amyl acetate	10.842	43	140509	38.012	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.207	83	90975	34.495	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	80678m	28.468	ug/l	
77) Bromobenzene	11.195	156	63053	36.763	ug/l	100
78) n-propylbenzene	11.299	91	329981	39.707	ug/l	100
79) 2-Chlorotoluene	11.360	91	203509	36.336	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	237066	38.516	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	26179	38.445	ug/l	100
82) 4-Chlorotoluene	11.451	91	233172	37.880	ug/l	100
83) tert-Butylbenzene	11.713	119	233468	38.622	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	239433	38.946	ug/l	100
85) sec-Butylbenzene	11.890	105	291074	38.786	ug/l	100
86) p-Isopropyltoluene	12.006	119	241658	39.879	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	115651	36.199	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	115097	36.763	ug/l	100
89) n-Butylbenzene	12.329	91	213899	40.603	ug/l	100
90) Hexachloroethane	12.536	117	42314	37.534	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	115286	37.047	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21909	36.496	ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	66655	39.117	ug/l	100
94) Hexachlorobutadiene	13.725	225	29007	36.395	ug/l	100
95) Naphthalene	13.774	128	245568	39.575	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	69262	38.765	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046042.D
Acq On : 05 May 2025 11:58
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:168 Resp: 97070
Ion Ratio Lower Upper
168 100
99 68.6 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 39.120 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 85 Resp: 83826
Ion Ratio Lower Upper
85 100
87 31.4 15.7 47.1









#3
Chloromethane
Concen: 35.658 ug/l
RT: 1.307 min Scan# 36
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 50 Resp: 75238
Ion Ratio Lower Upper
50 100
52 31.8 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 37.273 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 62 Resp: 68901
Ion Ratio Lower Upper
62 100
64 31.5 25.2 37.8





#5
Bromomethane
Concen: 33.765 ug/l
RT: 1.593 min Scan# 83
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 94 Resp: 31648
Ion Ratio Lower Upper
94 100
96 91.0 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 38.558 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 64 Resp: 36722
Ion Ratio Lower Upper
64 100
66 30.4 24.3 36.5





#7
Trichlorofluoromethane
Concen: 36.824 ug/l
RT: 1.874 min Scan# 129
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:101 Resp: 10369
Ion Ratio Lower Upper
101 100
103 63.7 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 35.350 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 74 Resp: 32724
Ion Ratio Lower Upper
74 100
45 109.9 54.9 164.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 36.997 ug/l

RT: 2.319 min Scan# 202

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion:101 Resp: 6225

Ion Ratio Lower Upper

101 100

85 48.3 38.6 58.0





151 69.0 55.2 82.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#10

Methyl Iodide

Concen: 39.622 ug/l

RT: 2.447 min Scan# 223

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:142 Resp: 78277


Ion Ratio Lower Upper

142 100

127 52.1 41.7 62.5

141 14.4 11.5 17.3





#11

Tert butyl alcohol

Concen: 181.745 ug/l

RT: 2.977 min Scan# 310

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 59 Resp: 6255

Ion Ratio Lower Upper

59 100





57 10.7 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 35.610 ug/l

RT: 2.313 min Scan# 201

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 96 Resp: 58334


Ion Ratio Lower Upper

96 100

61 182.7 146.2 219.2

98 63.8 51.0 76.6





#13

Acrolein

Concen: 184.907 ug/l

RT: 2.233 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 56 Resp: 7392

Ion Ratio Lower Upper

56 100





55 70.3 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 36.971 ug/l

RT: 2.654 min Scan# 257

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 41 Resp: 114445


Ion Ratio Lower Upper

41 100

39 75.7 60.6 90.8

76 31.1 24.9 37.3





#15

Acrylonitrile

Concen: 182.874 ug/l

RT: 3.062 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:	53	Resp:	188304
Ion	Ratio	Lower	Upper
53	100		
52	81.6	65.3	97.9
51	37.3	29.8	44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 177.910 ug/l




RT: 2.386 min Scan# 213


Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:	43	Resp:	175777
Ion	Ratio	Lower	Upper
43	100		
58	26.5	21.2	31.8





#17

Carbon Disulfide

Concen: 37.223 ug/l

RT: 2.502 min Scan# 21

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 14128

Ion Ratio Lower Upper

76 100





78 8.7 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 34.915 ug/l

RT: 2.703 min Scan# 265

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 43 Resp: 82347

Ion Ratio Lower Upper

43 100

74 20.9 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 36.695 ug/l

RT: 3.111 min Scan# 311

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 73 Resp: 20966

Ion Ratio Lower Upper

73 100





57 22.1 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#20

Methylene Chloride

Concen: 33.182 ug/l

RT: 2.782 min Scan# 278

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion: 84 Resp: 66412




Ion Ratio Lower Upper


84 100

49 142.4 113.9 170.9

51 41.9 33.5 50.3


86 67.3 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 35.421 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 96 Resp: 59234
Ion Ratio Lower Upper
96 100
61 149.4 119.5 179.3
98 62.5 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 38.241 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 45 Resp: 221172
Ion Ratio Lower Upper
45 100
43 83.3 66.6 100.0
87 24.7 19.8 29.6
59 10.7 8.6 12.8





#23

Vinyl Acetate

Concen: 194.528 ug/l

RT: 3.721 min Scan# 413

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 994230

Ion Ratio Lower Upper

43 100





86 9.4 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#24

1,1-Dichloroethane

Concen: 36.067 ug/l

RT: 3.605 min Scan# 413

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 63 Resp: 122601


Ion Ratio Lower Upper

63 100

98 7.2 3.6 10.8

100 4.2 2.1 6.3





#25

2-Butanone

Concen: 188.783 ug/l

RT: 4.556 min Scan# 569

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 269224

Ion Ratio Lower Upper

43 100





72 23.0 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 36.059 ug/l

RT: 4.465 min Scan# 554

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 77 Resp: 92857

Ion Ratio Lower Upper

77 100

97 21.0 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 35.466 ug/l

RT: 4.483 min Scan# 557

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 96 Resp: 71518

Ion Ratio Lower Upper

96 100

61 161.4 0.0 322.8





98 64.5 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 31.031 ug/l

RT: 4.898 min Scan# 625

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 49 Resp: 56091


Ion Ratio Lower Upper

49 100

129 2.0 0.0 4.0

130 70.2 56.2 84.2





#29

Tetrahydrofuran

Concen: 183.679 ug/l

RT: 5.001 min Scan# 642

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 42 Resp: 170124

Ion Ratio Lower Upper

42 100

72 42.1 33.7 50.5





71 39.3 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 35.626 ug/l

RT: 5.087 min Scan# 656

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 83 Resp: 125850

Ion Ratio Lower Upper

83 100


85 61.6 49.3 73.9





#31
Cyclohexane
Concen: 38.205 ug/l
RT: 5.464 min Scan# 718
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 56 Resp: 10945
Ion Ratio Lower Upper
56 100
69 30.5 24.4 36.6
84 83.6 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#32
1,1,1-Trichloroethane
Concen: 36.034 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 97 Resp: 109781
Ion Ratio Lower Upper
97 100
99 64.7 51.8 77.6
61 50.1 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 30.601 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 65 Resp: 8837

Ion Ratio Lower Upper

65 100





67 49.5 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:114 Resp: 167947


Ion Ratio Lower Upper

114 100

63 24.6 0.0 49.2

88 16.8 0.0 33.6





#35

Dibromofluoromethane

Concen: 31.633 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:	113	Resp:	59550
Ion	Ratio	Lower	Upper
113	100		
111	103.9	83.1	124.7
192	16.6	13.3	19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 38.073 ug/l




RT: 5.684 min Scan# 754


Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:	75	Resp:	83215
Ion	Ratio	Lower	Upper
75	100		
110	32.6	16.3	48.9
77	30.4	24.3	36.5





#37

Ethyl Acetate

Concen: 37.394 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 10253

Ion Ratio Lower Upper

43 100

61 12.9 10.3 15.5





70 9.9 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 37.320 ug/l

RT: 5.672 min Scan# 752

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:117 Resp: 93712


Ion Ratio Lower Upper

117 100

119 94.0 75.2 112.8

121 30.3 24.2 36.4





#39

Methylcyclohexane

Concen: 39.307 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 83 Resp: 10765

Ion Ratio Lower Upper

83 100

55 80.9 64.7 97.1





98 46.8 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 36.530 ug/l

RT: 6.031 min Scan# 811

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 78 Resp: 247476

Ion Ratio Lower Upper

78 100

77 23.7 19.0 28.4









#41
Methacrylonitrile
Concen: 38.123 ug/l
RT: 4.916 min Scan# 612
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 41 Resp: 5808
Ion Ratio Lower Upper
41 100
39 59.0 47.2 70.8
67 63.4 50.7 76.1
52 30.8 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 37.605 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 62 Resp: 105332
Ion Ratio Lower Upper
62 100
98 7.6 0.0 15.2





#43

Isopropyl Acetate

Concen: 38.598 ug/l

RT: 6.336 min Scan# 861

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 16178

Ion Ratio Lower Upper

43

100

61

17.9

87

11.9

14.3

21.5





14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 37.230 ug/l

RT: 7.123 min Scan# 990

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:130 Resp: 59623

Ion Ratio Lower Upper

130




100


95

102.1

0.0

204.2





#45

1,2-Dichloropropane

Concen: 37.064 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 63 Resp: 6238

Ion Ratio Lower Upper

63 100





65 30.1 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 36.326 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 93 Resp: 48201


Ion Ratio Lower Upper

93 100

95 82.0 65.6 98.4

174 85.2 68.2 102.2





#47

Bromodichloromethane

Concen: 37.685 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 83	Resp: 96910
Ion Ratio	Lower Upper
83	100
85	63.1 50.5 75.7
127	8.1 6.5 9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 39.157 ug/l




RT: 7.690 min Scan# 1083


Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion: 41	Resp: 84277
Ion Ratio	Lower Upper
41	100
69	73.1 58.5 87.7
39	64.6 51.7 77.5









#49
1,4-Dioxane
Concen: 713.477 ug/l
RT: 7.659 min Scan# 1078
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 88 Resp: 30529
Ion Ratio Lower Upper
88 100
43 41.8 33.4 50.2
58 73.3 58.6 88.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#50
Toluene-d8
Concen: 32.395 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 98 Resp: 205479
Ion Ratio Lower Upper
98 100
100 66.9 53.5 80.3





#51

4-Methyl-2-Pentanone

Concen: 193.653 ug/l

RT: 8.574 min Scan# 1228

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 53238

Ion Ratio Lower Upper

43 100





58 36.1 28.9 43.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#52

Toluene

Concen: 37.530 ug/l

RT: 8.714 min Scan# 1251

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 92 Resp: 150757

Ion Ratio Lower Upper

92 100

91 170.8 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 41.056 ug/l

RT: 8.976 min Scan# 1294

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion: 75 Resp: 8865

Ion Ratio Lower Upper

75 100

77 31.3 25.0 37.6

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 38.521 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 75 Resp: 97031


Ion Ratio Lower Upper

75 100

77 31.7 25.4 38.0

39 65.3 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 36.648 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 97 Resp: 5949

Ion Ratio Lower Upper

97 100

83 86.3 69.0 103.6

85 55.4 44.3 66.5





99 59.7 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 39.726 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 69 Resp: 99947


Ion Ratio Lower Upper

69 100

41 76.0 60.8 91.2

39 48.8 39.0 58.6





#57

1,3-Dichloropropane

Concen: 36.347 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 104600

Ion Ratio Lower Upper

76 100





78 32.9 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 227.902 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 63 Resp: 257678

Ion Ratio Lower Upper

63 100

106 26.9 21.5 32.3





#59

2-Hexanone

Concen: 190.168 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 396784

Ion Ratio Lower Upper

43 100





58 49.7 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 38.027 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:129 Resp: 67198

Ion Ratio Lower Upper

129 100

127 78.5 39.3 117.8









#61
1,2-Dibromoethane
Concen: 37.567 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:107 Resp: 6263
Ion Ratio Lower Upper
107 100
109 91.6 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#62
4-Bromofluorobenzene
Concen: 34.229 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 95 Resp: 79012
Ion Ratio Lower Upper
95 100
174 67.9 0.0 135.8
176 65.7 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:117 Resp: 14625
Ion Ratio Lower Upper
117 100
82 63.3 50.6 76.0
119 32.2 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 36.726 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion:164 Resp: 54853
Ion Ratio Lower Upper
164 100
166 131.3 105.0 157.6
129 104.4 83.5 125.3
131 95.6 76.5 114.7





#65
Chlorobenzene
Concen: 35.686 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:112 Resp: 160600
Ion Ratio Lower Upper
112 100
114 31.8 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 38.017 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:131 Resp: 57066
Ion Ratio Lower Upper
131 100
133 94.6 47.3 141.9
119 63.3 31.6 95.0





#67

Ethyl Benzene

Concen: 38.543 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 29571

Ion Ratio Lower Upper

91 100





106 29.3 23.4 35.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#68

m/p-Xylenes

Concen: 78.142 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:106 Resp: 216578

Ion Ratio Lower Upper

106 100

91 214.0 171.2 256.8









#69
o-Xylene
Concen: 38.064 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:106 Resp: 106333
Ion Ratio Lower Upper
106 100
91 225.4 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#70
Styrene
Concen: 39.888 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion:104 Resp: 178313
Ion Ratio Lower Upper
104 100
78 57.1 45.7 68.5
103 54.6 43.7 65.5





#71
Bromoform
Concen: 39.457 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:173 Resp: 44480
Ion Ratio Lower Upper
173 100
175 46.7 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:152 Resp: 67976
Ion Ratio Lower Upper
152 100
115 93.8 46.9 140.7
150 175.5 0.0 351.0









#73
Isopropylbenzene
Concen: 37.865 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:105 Resp: 280719
Ion Ratio Lower Upper
105 100
120 25.6 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 38.012 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 43 Resp: 140509
Ion Ratio Lower Upper
43 100
70 38.6 30.9 46.3
55 23.4 18.7 28.1
61 21.4 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 34.495 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 83 Resp: 90975
Ion Ratio Lower Upper
83 100
131 9.9 5.0 14.9
85 63.8 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#76
1,2,3-Trichloropropane
Concen: 28.468 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 75 Resp: 80678
Ion Ratio Lower Upper
75 100
77 41.0 20.5 61.5





#77
Bromobenzene
Concen: 36.763 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:156 Resp: 6305
Ion Ratio Lower Upper
156 100
77 185.5 92.8 278.3
158 97.5 48.8 146.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#78
n-propylbenzene
Concen: 39.707 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 91 Resp: 329981
Ion Ratio Lower Upper
91 100
120 21.6 10.8 32.4









#79
2-Chlorotoluene
Concen: 36.336 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 91 Resp: 203509
Ion Ratio Lower Upper
91 100
126 31.1 15.6 46.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#80
1,3,5-Trimethylbenzene
Concen: 38.516 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:105 Resp: 237066
Ion Ratio Lower Upper
105 100
120 46.1 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 38.445 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 75 Resp: 26179
Ion Ratio Lower Upper
75 100
53 125.6 100.5 150.7
89 46.3 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 37.880 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 91 Resp: 233172
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83
tert-Butylbenzene
Concen: 38.622 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:119 Resp: 233468
Ion Ratio Lower Upper
119 100
91 65.8 32.9 98.7
134 22.7 11.4 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#84
1,2,4-Trimethylbenzene
Concen: 38.946 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:105 Resp: 239433
Ion Ratio Lower Upper
105 100
120 42.4 21.2 63.6









#85
sec-Butylbenzene
Concen: 38.786 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:105 Resp: 291074
Ion Ratio Lower Upper
105 100
134 19.4 9.7 29.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#86
p-Isopropyltoluene
Concen: 39.879 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:119 Resp: 241658
Ion Ratio Lower Upper
119 100
134 25.0 12.5 37.5
91 27.6 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 36.199 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion:146 Resp: 11565

Ion Ratio Lower Upper

146 100

111 44.2 22.1 66.3





148 64.3 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 36.763 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:146 Resp: 115097


Ion Ratio Lower Upper

146 100

111 42.6 21.3 63.9

148 63.7 31.9 95.5





#89

n-Butylbenzene

Concen: 40.603 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 21389

Ion Ratio Lower Upper

91 100

92 53.8 26.9 80.7





134 23.5 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 37.534 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:117 Resp: 42314

Ion Ratio Lower Upper

117 100


201 63.1 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 37.047 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:146 Resp: 115280
Ion Ratio Lower Upper
146 100
111 44.9 22.4 67.3
148 63.2 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 36.496 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 75 Resp: 21909
Ion Ratio Lower Upper
75 100
155 69.9 34.9 104.8
157 87.6 43.8 131.4









#93
1,2,4-Trichlorobenzene
Concen: 39.117 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:180 Resp: 6665
Ion Ratio Lower Upper
180 100
182 95.9 47.9 143.8
145 34.1 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#94
Hexachlorobutadiene
Concen: 36.395 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion:225 Resp: 29007
Ion Ratio Lower Upper
225 100
223 61.6 30.8 92.4
227 61.8 30.9 92.7





#95
Naphthalene
Concen: 39.575 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:128 Resp: 245568
Ion Ratio Lower Upper
128 100
127 13.0 10.4 15.6
129 10.8 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 38.765 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:180 Resp: 69262
Ion Ratio Lower Upper
180 100
182 95.5 47.8 143.3
145 36.2 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046043.D
 Acq On : 05 May 2025 12:21
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:10:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	87028	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	152958	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	135214	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66369	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	161877	62.526	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 125.060%#			
35) Dibromofluoromethane	5.379	113	111455	65.006	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 130.020%#			
50) Toluene-d8	8.647	98	387230	67.031	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 134.060%#			
62) 4-Bromofluorobenzene	11.079	95	153085	72.818	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 145.640%#			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	149508	77.828	ug/l	99
3) Chloromethane	1.307	50	136937	72.392	ug/l	98
4) Vinyl Chloride	1.374	62	126609	76.398	ug/l	98
5) Bromomethane	1.593	94	59246	70.507	ug/l	99
6) Chloroethane	1.660	64	57332	67.148	ug/l	99
7) Trichlorofluoromethane	1.867	101	171042	67.754	ug/l	99
8) Diethyl Ether	2.136	74	58762	70.807	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	109520	72.604	ug/l	99
10) Methyl Iodide	2.440	142	138045	77.943	ug/l	100
11) Tert butyl alcohol	2.995	59	125094	405.394	ug/l	100
12) 1,1-Dichloroethene	2.306	96	105566	71.884	ug/l	96
13) Acrolein	2.239	56	134456	375.132	ug/l	99
14) Allyl chloride	2.660	41	206519	74.419	ug/l	98
15) Acrylonitrile	3.068	53	331142	358.723	ug/l	97
16) Acetone	2.386	43	314383	354.936	ug/l	98
17) Carbon Disulfide	2.501	76	264956	77.867	ug/l	100
18) Methyl Acetate	2.703	43	147014	69.531	ug/l	96
19) Methyl tert-butyl Ether	3.117	73	378113	73.818	ug/l	100
20) Methylene Chloride	2.782	84	120270	67.030	ug/l	96
21) trans-1,2-Dichloroethene	3.087	96	106606	71.109	ug/l	99
22) Diisopropyl ether	3.763	45	399494	77.048	ug/l	90
23) Vinyl Acetate	3.721	43	1811550	395.363	ug/l	100
24) 1,1-Dichloroethane	3.605	63	219759	72.114	ug/l	99
25) 2-Butanone	4.556	43	485448	379.704	ug/l	100
26) 2,2-Dichloropropane	4.464	77	174579	75.621	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	128454	71.056	ug/l	99
28) Bromochloromethane	4.891	49	103602	63.932	ug/l	100
29) Tetrahydrofuran	5.001	42	305757	368.234	ug/l	99
30) Chloroform	5.086	83	222253	70.180	ug/l	95
31) Cyclohexane	5.458	56	196302	76.428	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	201112	73.635	ug/l	98
36) 1,1-Dichloropropene	5.690	75	147873	74.285	ug/l	99
37) Ethyl Acetate	4.714	43	186215	74.565	ug/l	99
38) Carbon Tetrachloride	5.672	117	168946	73.875	ug/l	99
39) Methylcyclohexane	7.372	83	191941	76.952	ug/l	96
40) Benzene	6.031	78	440885	71.457	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046043.D
 Acq On : 05 May 2025 12:21
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By :John Carlone 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:10:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	104863	75.574	ug/l	98
42) 1,2-Dichloroethane	6.080	62	187036	73.318	ug/l	99
43) Isopropyl Acetate	6.342	43	300491	78.714	ug/l	100
44) Trichloroethene	7.123	130	105628	72.420	ug/l	98
45) 1,2-Dichloropropane	7.427	63	112554	73.420	ug/l	100
46) Dibromomethane	7.574	93	85728	70.938	ug/l	100
47) Bromodichloromethane	7.818	83	175301	74.845	ug/l	99
48) Methyl methacrylate	7.689	41	153093	78.101	ug/l	99
49) 1,4-Dioxane	7.659	88	57313	1470.687	ug/l	99
51) 4-Methyl-2-Pentanone	8.573	43	963319	384.740	ug/l	100
52) Toluene	8.714	92	270763	74.010	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	169863	86.372	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	184134	80.264	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	107264	72.543	ug/l	99
56) Ethyl methacrylate	9.116	69	188717	82.359	ug/l	99
57) 1,3-Dichloropropane	9.305	76	187649	71.590	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	463169	449.791	ug/l	100
59) 2-Hexanone	9.433	43	729602	383.945	ug/l	100
60) Dibromochloromethane	9.518	129	127012	78.918	ug/l	98
61) 1,2-Dibromoethane	9.610	107	112669	74.200	ug/l	98
64) Tetrachloroethene	9.268	164	93195	67.493	ug/l	96
65) Chlorobenzene	10.079	112	293284	70.490	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	103177	74.349	ug/l	99
67) Ethyl Benzene	10.189	91	535122	75.444	ug/l	99
68) m/p-Xylenes	10.299	106	389935	152.181	ug/l	99
69) o-Xylene	10.640	106	190833	73.890	ug/l	97
70) Styrene	10.652	104	328195	79.411	ug/l	100
71) Bromoform	10.799	173	84353	80.927	ug/l #	98
73) Isopropylbenzene	10.957	105	514528	71.083	ug/l	100
74) N-amyl acetate	10.841	43	274553	76.073	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	170481	66.207	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	150182m	54.277	ug/l	
77) Bromobenzene	11.195	156	117151	69.959	ug/l	99
78) n-propylbenzene	11.305	91	608332	74.974	ug/l	100
79) 2-Chlorotoluene	11.360	91	372392	68.100	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	432096	71.903	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	54398	81.820	ug/l	93
82) 4-Chlorotoluene	11.451	91	432053	71.890	ug/l	100
83) tert-Butylbenzene	11.713	119	432011	73.197	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	442728	73.758	ug/l	100
85) sec-Butylbenzene	11.890	105	543597	74.188	ug/l	100
86) p-Isopropyltoluene	12.006	119	457898	77.393	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	219753	70.449	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	217613	71.190	ug/l	99
89) n-Butylbenzene	12.329	91	416602	80.996	ug/l	100
90) Hexachloroethane	12.536	117	82528	74.978	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	216834	71.367	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	43673	74.512	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	137430	82.605	ug/l	98
94) Hexachlorobutadiene	13.725	225	55494	71.314	ug/l	98
95) Naphthalene	13.774	128	489778	80.842	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	139525	79.981	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046043.D
Acq On : 05 May 2025 12:21
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:10:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046043.D
Acq On : 05 May 2025 12:21
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By : John Carlone 05/06/2025
Supervised By : Mahesh Dadoda 05/06/2025


Quant Time: May 06 06:10:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100



Tgt Ion:168 Resp: 87028
Ion Ratio Lower Upper
168 100
99 72.1 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 77.828 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 85 Resp: 149508
Ion Ratio Lower Upper
85 100
87 32.0 15.7 47.1





#3
Chloromethane
Concen: 72.392 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 50 Resp: 13693
Ion Ratio Lower Upper
50 100
52 32.9 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 76.398 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 62 Resp: 126609
Ion Ratio Lower Upper
62 100
64 30.6 25.2 37.8





#5
Bromomethane
Concen: 70.507 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 94 Resp: 59240
Ion Ratio Lower Upper
94 100
96 91.7 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 67.148 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 64 Resp: 57332
Ion Ratio Lower Upper
64 100
66 30.8 24.3 36.5





#7
Trichlorofluoromethane
Concen: 67.754 ug/l
RT: 1.867 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100



Tgt Ion:101 Resp: 17104
Ion Ratio Lower Upper
101 100
103 64.8 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 70.807 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 74 Resp: 58762
Ion Ratio Lower Upper
74 100
45 110.1 54.9 164.8





#9
1,1,2-Trichlorotrifluoroethane
Concen: 72.604 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion:101 Resp: 109520
Ion Ratio Lower Upper
101 100
85 48.1 38.6 58.0
151 69.9 55.2 82.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#10
Methyl Iodide
Concen: 77.943 ug/l
RT: 2.440 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21



Tgt Ion:142 Resp: 138045
Ion Ratio Lower Upper
142 100
127 52.0 41.7 62.5
141 14.5 11.5 17.3





#11

Tert butyl alcohol

Concen: 405.394 ug/l

RT: 2.995 min Scan# 310

Delta R.T. 0.018 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 59 Resp: 125094

Ion Ratio Lower Upper

59 100





57 10.9 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 71.884 ug/l

RT: 2.306 min Scan# 200

Delta R.T. -0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 96 Resp: 105566


Ion Ratio Lower Upper

96 100

61 189.9 146.2 219.2

98 63.3 51.0 76.6





#13

Acrolein

Concen: 375.132 ug/l

RT: 2.239 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 56 Resp: 134450

Ion Ratio Lower Upper

56 100





55 71.3 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 74.419 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 41 Resp: 206519


Ion Ratio Lower Upper

41 100

39 73.8 60.6 90.8


76 31.4 24.9 37.3





#15
Acrylonitrile
Concen: 358.723 ug/l
RT: 3.068 min Scan# 3114
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 53 Resp: 33114
Ion Ratio Lower Upper
53 100
52 84.7 65.3 97.9
51 37.9 29.8 44.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#16
Acetone
Concen: 354.936 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 43 Resp: 314383
Ion Ratio Lower Upper
43 100
58 27.6 21.2 31.8





#17

Carbon Disulfide

Concen: 77.867 ug/l

RT: 2.501 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 76 Resp: 264950

Ion Ratio Lower Upper

76 100





78 8.7 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 69.531 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 43 Resp: 147014

Ion Ratio Lower Upper

43 100

74 22.6 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 73.818 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 73 Resp: 37811
Ion Ratio Lower Upper
73 100
57 22.1 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#20
Methylene Chloride
Concen: 67.030 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 84 Resp: 120270
Ion Ratio Lower Upper
84 100
49 136.4 113.9 170.9
51 41.1 33.5 50.3
86 64.0 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 71.109 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 96 Resp: 106600
Ion Ratio Lower Upper
96 100
61 149.7 119.5 179.3
98 64.4 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 77.048 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 45 Resp: 399494
Ion Ratio Lower Upper
45 100
43 71.2 66.6 100.0
87 25.7 19.8 29.6
59 11.0 8.6 12.8









#23
Vinyl Acetate
Concen: 395.363 ug/l
RT: 3.721 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 43 Resp: 1811550
Ion Ratio Lower Upper
43 100
86 9.5 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#24
1,1-Dichloroethane
Concen: 72.114 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 63 Resp: 219759
Ion Ratio Lower Upper
63 100
98 6.8 3.6 10.8
100 4.5 2.1 6.3





#25

2-Butanone

Concen: 379.704 ug/l

RT: 4.556 min Scan# 569

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 485448

Ion Ratio Lower Upper

43 100





72 23.1 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 75.621 ug/l

RT: 4.464 min Scan# 554

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 77 Resp: 174579

Ion Ratio Lower Upper

77 100

97 21.1 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 71.056 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 96 Resp: 128454

Ion Ratio Lower Upper

96 100

61 160.1 0.0 322.8





98 64.9 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 63.932 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 49 Resp: 103602


Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 70.1 56.2 84.2





#29

Tetrahydrofuran

Concen: 368.234 ug/l

RT: 5.001 min Scan# 642

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 42 Resp: 30575

Ion Ratio Lower Upper

42 100

72 42.1 33.7 50.5





71 38.7 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 70.180 ug/l

RT: 5.086 min Scan# 656

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 83 Resp: 222253

Ion Ratio Lower Upper

83 100


85 65.1 49.3 73.9





#31
Cyclohexane
Concen: 76.428 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 56 Resp: 196302
Ion Ratio Lower Upper
56 100
69 31.2 24.4 36.6
84 82.3 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#32
1,1,1-Trichloroethane
Concen: 73.635 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 97 Resp: 201112
Ion Ratio Lower Upper
97 100
99 63.8 51.8 77.6
61 48.7 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 62.526 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 65 Resp: 16187

Ion Ratio Lower Upper

65 100





67 49.5 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:114 Resp: 152958


Ion Ratio Lower Upper

114 100

63 24.3 0.0 49.2

88 17.1 0.0 33.6





#35

Dibromofluoromethane

Concen: 65.006 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion:113 Resp: 11145

Ion Ratio Lower Upper

113 100

111 101.7 83.1 124.7





192 16.6 13.3 19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 74.285 ug/l

RT: 5.690 min Scan# 755

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 75 Resp: 147873


Ion Ratio Lower Upper

75 100

110 33.5 16.3 48.9

77 31.1 24.3 36.5





#37

Ethyl Acetate

Concen: 74.565 ug/l

RT: 4.714 min Scan# 595

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 186215

Ion Ratio Lower Upper

43 100

61 12.4 10.3 15.5





70 9.9 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 73.875 ug/l

RT: 5.672 min Scan# 752

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:117 Resp: 168946


Ion Ratio Lower Upper

117 100

119 95.0 75.2 112.8

121 29.6 24.2 36.4





#39

Methylcyclohexane

Concen: 76.952 ug/l

RT: 7.372 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 83 Resp: 19194

Ion Ratio Lower Upper

83 100

55 84.2 64.7 97.1





98 49.0 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 71.457 ug/l

RT: 6.031 min Scan# 811

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 78 Resp: 440885

Ion Ratio Lower Upper

78 100


77 24.6 19.0 28.4





#41
Methacrylonitrile
Concen: 75.574 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 41 Resp: 10486
Ion Ratio Lower Upper
41 100
39 57.0 47.2 70.8
67 64.6 50.7 76.1
52 29.7 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 73.318 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 62 Resp: 187036
Ion Ratio Lower Upper
62 100
98 7.4 0.0 15.2





#43
Isopropyl Acetate
Concen: 78.714 ug/l
RT: 6.342 min Scan# 861
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 43 Resp: 30049
Ion Ratio Lower Upper
43 100
61 18.0 14.3 21.5
87 11.6 9.5 14.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#44
Trichloroethene
Concen: 72.420 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:130 Resp: 105628
Ion Ratio Lower Upper
130 100
95 103.8 0.0 204.2





#45

1,2-Dichloropropane

Concen: 73.420 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 63 Resp: 112554

Ion Ratio Lower Upper

63 100





65 30.0 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 70.938 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 93 Resp: 85728


Ion Ratio Lower Upper

93 100

95 82.4 65.6 98.4

174 85.0 68.2 102.2





#47

Bromodichloromethane

Concen: 74.845 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 83 Resp: 17530

Ion Ratio Lower Upper

83 100

85 63.6 50.5 75.7





127 7.8 6.5 9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 78.101 ug/l

RT: 7.689 min Scan# 1083

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 41 Resp: 153093


Ion Ratio Lower Upper

41 100

69 74.1 58.5 87.7

39 65.5 51.7 77.5





#49

1,4-Dioxane

Concen: 1470.687 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 88 Resp: 5731

Ion Ratio Lower Upper

88 100

43 40.1 33.4 50.2





58 73.3 58.6 88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 67.031 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 98 Resp: 387230

Ion Ratio Lower Upper

98 100

100 65.2 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 384.740 ug/l
RT: 8.573 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 43 Resp: 963319
Ion Ratio Lower Upper
43 100
58 36.1 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 74.010 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 92 Resp: 270763
Ion Ratio Lower Upper
92 100
91 171.2 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 86.372 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 75 Resp: 16986

Ion Ratio Lower Upper

75 100





77 31.5 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 80.264 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 75 Resp: 184134


Ion Ratio Lower Upper

75 100

77 31.9 25.4 38.0

39 59.6 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 72.543 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 97 Resp: 107264

Ion Ratio Lower Upper

97 100

83 86.8 69.0 103.6

85 56.4 44.3 66.5





99 61.8 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 82.359 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 69 Resp: 188717


Ion Ratio Lower Upper

69 100

41 75.7 60.8 91.2

39 47.7 39.0 58.6





#57

1,3-Dichloropropane

Concen: 71.590 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 76 Resp: 187649

Ion Ratio Lower Upper

76 100





78 32.1 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 449.791 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 63 Resp: 463169

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3





#59

2-Hexanone

Concen: 383.945 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 729602

Ion Ratio Lower Upper

43 100





58 49.9 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 78.918 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion:129 Resp: 127012

Ion Ratio Lower Upper

129 100

127 77.0 39.3 117.8





#61

1,2-Dibromoethane

Concen: 74.200 ug/l

RT: 9.610 min Scan# 11266

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion:107 Resp: 11266

Ion Ratio Lower Upper

107 100





109 93.9 73.3 109.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#62

4-Bromofluorobenzene

Concen: 72.818 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 95 Resp: 153085


Ion Ratio Lower Upper

95 100

174 67.1 0.0 135.8

176 65.3 0.0 131.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.049 min Scan# 1470

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion:117 Resp: 135214

Ion Ratio Lower Upper

117 100

82 61.7 50.6 76.0





119 30.6 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#64

Tetrachloroethene

Concen: 67.493 ug/l

RT: 9.268 min Scan# 1342

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Tgt Ion:164 Resp: 93195




Ion Ratio Lower Upper


164 100

166 123.4 105.0 157.6

129 101.4 83.5 125.3


131 98.0 76.5 114.7





#65
Chlorobenzene
Concen: 70.490 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:112 Resp: 293284
Ion Ratio Lower Upper
112 100
114 31.9 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 74.349 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion:131 Resp: 103177
Ion Ratio Lower Upper
131 100
133 95.3 47.3 141.9
119 64.5 31.6 95.0





#67
Ethyl Benzene
Concen: 75.444 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 91 Resp: 53512
Ion Ratio Lower Upper
91 100
106 29.6 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#68
m/p-Xylenes
Concen: 152.181 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:106 Resp: 389935
Ion Ratio Lower Upper
106 100
91 215.8 171.2 256.8









#69
o-Xylene
Concen: 73.890 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion:106 Resp: 190833
Ion Ratio Lower Upper
106 100
91 230.3 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#70
Styrene
Concen: 79.411 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion:104 Resp: 328195
Ion Ratio Lower Upper
104 100
78 57.4 45.7 68.5
103 54.8 43.7 65.5





#71
Bromoform
Concen: 80.927 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:173 Resp: 8435
Ion Ratio Lower Upper
173 100
175 48.3 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion:152 Resp: 66369
Ion Ratio Lower Upper
152 100
115 120.4 46.9 140.7
150 190.4 0.0 351.0





#73
Isopropylbenzene
Concen: 71.083 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100



Tgt Ion:105 Resp: 514528
Ion Ratio Lower Upper
105 100
120 25.4 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 76.073 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 43 Resp: 274553
Ion Ratio Lower Upper
43 100
70 38.2 30.9 46.3
55 22.5 18.7 28.1
61 22.2 17.1 25.7





#75

1,1,2,2-Tetrachloroethane

Concen: 66.207 ug/l

RT: 11.213 min Scan# 1661

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 83 Resp: 17048

Ion Ratio Lower Upper

83 100

131 9.7 5.0 14.9





85 64.9 31.9 95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#76

1,2,3-Trichloropropane

Concen: 54.277 ug/l m

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 75 Resp: 150182

Ion Ratio Lower Upper

75 100


77 42.2 20.5 61.5





#77
Bromobenzene
Concen: 69.959 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion:156 Resp: 11715
Ion Ratio Lower Upper
156 100
77 187.4 92.8 278.3
158 96.1 48.8 146.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#78
n-propylbenzene
Concen: 74.974 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21



Tgt Ion: 91 Resp: 608332
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4









#79
2-Chlorotoluene
Concen: 68.100 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 91 Resp: 372392
Ion Ratio Lower Upper
91 100
126 30.9 15.6 46.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#80
1,3,5-Trimethylbenzene
Concen: 71.903 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:105 Resp: 432096
Ion Ratio Lower Upper
105 100
120 46.3 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 81.820 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 75 Resp: 54398
Ion Ratio Lower Upper
75 100
53 115.8 100.5 150.7
89 47.0 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 71.890 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 91 Resp: 432053
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83
tert-Butylbenzene
Concen: 73.197 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:119 Resp: 43201
Ion Ratio Lower Upper
119 100
91 64.4 32.9 98.7
134 22.8 11.4 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#84
1,2,4-Trimethylbenzene
Concen: 73.758 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion:105 Resp: 442728
Ion Ratio Lower Upper
105 100
120 42.4 21.2 63.6





#85
sec-Butylbenzene
Concen: 74.188 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion:105 Resp: 54359
Ion Ratio Lower Upper
105 100
134 19.4 9.7 29.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#86
p-Isopropyltoluene
Concen: 77.393 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21



Tgt Ion:119 Resp: 457898
Ion Ratio Lower Upper
119 100
134 24.6 12.5 37.5
91 26.8 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 70.449 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion:146 Resp: 21975

Ion Ratio Lower Upper

146 100

111 44.1 22.1 66.3





148 63.5 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 71.190 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:146 Resp: 217613


Ion Ratio Lower Upper

146 100

111 43.4 21.3 63.9

148 63.2 31.9 95.5





#89

n-Butylbenzene

Concen: 80.996 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 91 Resp: 416602

Ion Ratio Lower Upper

91 100

92 54.0 26.9 80.7





134 24.0 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 74.978 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion:117 Resp: 82528

Ion Ratio Lower Upper

117 100


201 63.3 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 71.367 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:146 Resp: 216834
Ion Ratio Lower Upper
146 100
111 46.2 22.4 67.3
148 64.0 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carbone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 74.512 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 75 Resp: 43673
Ion Ratio Lower Upper
75 100
155 69.4 34.9 104.8
157 90.1 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 82.605 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion:180 Resp: 137430

Ion Ratio Lower Upper

180 100

182 93.7 47.9 143.8





145 34.6 17.1 51.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#94

Hexachlorobutadiene

Concen: 71.314 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:225 Resp: 55494


Ion Ratio Lower Upper

225 100

223 62.3 30.8 92.4


227 64.9 30.9 92.7





#95
Naphthalene
Concen: 80.842 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:128 Resp: 489778
Ion Ratio Lower Upper
128 100
127 12.9 10.4 15.6
129 10.9 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 79.981 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:180 Resp: 139525
Ion Ratio Lower Upper
180 100
182 93.7 47.8 143.3
145 35.3 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046044.D
 Acq On : 05 May 2025 12:45
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By :John Carlone 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:11:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	82963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	144975	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	128557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60345	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	231952	93.983	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 187.960%#			
35) Dibromofluoromethane	5.385	113	160168	98.562	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 197.120%#			
50) Toluene-d8	8.647	98	554390	101.252	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 202.500%#			
62) 4-Bromofluorobenzene	11.079	95	217536	109.174	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 218.340%#			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	217793	118.929	ug/l	98
3) Chloromethane	1.307	50	196843	109.161	ug/l	99
4) Vinyl Chloride	1.374	62	187924	118.953	ug/l	100
5) Bromomethane	1.593	94	83008	103.625	ug/l	95
6) Chloroethane	1.660	64	78806	96.822	ug/l	98
7) Trichlorofluoromethane	1.868	101	245108	101.851	ug/l	97
8) Diethyl Ether	2.136	74	88441	111.791	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.313	101	161404	112.243	ug/l	99
10) Methyl Iodide	2.441	142	189843	112.441	ug/l	100
11) Tert butyl alcohol	3.002	59	181238	616.119	ug/l	100
12) 1,1-Dichloroethene	2.307	96	155451	111.039	ug/l	99
13) Acrolein	2.239	56	203257	594.872	ug/l	98
14) Allyl chloride	2.654	41	297776	112.561	ug/l	98
15) Acrylonitrile	3.069	53	484854	550.973	ug/l	98
16) Acetone	2.392	43	460820	545.754	ug/l	98
17) Carbon Disulfide	2.502	76	397353	122.499	ug/l	100
18) Methyl Acetate	2.709	43	217885	108.099	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	557210	114.112	ug/l	100
20) Methylene Chloride	2.782	84	172035	100.579	ug/l	96
21) trans-1,2-Dichloroethene	3.081	96	154694	108.241	ug/l	96
22) Diisopropyl ether	3.764	45	577610	116.858	ug/l	90
23) Vinyl Acetate	3.721	43	2655751	608.005	ug/l	100
24) 1,1-Dichloroethane	3.605	63	320083	110.182	ug/l	99
25) 2-Butanone	4.562	43	708463	581.292	ug/l	98
26) 2,2-Dichloropropane	4.471	77	258698	117.548	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	187835	108.994	ug/l	99
28) Bromochloromethane	4.898	49	146923	95.108	ug/l	100
29) Tetrahydrofuran	5.007	42	450456	569.081	ug/l	99
30) Chloroform	5.087	83	323659	107.208	ug/l	95
31) Cyclohexane	5.458	56	286312	116.934	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	295702	113.572	ug/l	98
36) 1,1-Dichloropropene	5.684	75	219731	116.462	ug/l	99
37) Ethyl Acetate	4.715	43	274312	115.889	ug/l	99
38) Carbon Tetrachloride	5.666	117	250916	115.760	ug/l	99
39) Methylcyclohexane	7.373	83	286311	121.107	ug/l	97
40) Benzene	6.031	78	642178	109.813	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046044.D
 Acq On : 05 May 2025 12:45
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:11:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	151424	115.139	ug/l	98
42) 1,2-Dichloroethane	6.080	62	271992	112.492	ug/l	99
43) Isopropyl Acetate	6.342	43	447871	123.780	ug/l	100
44) Trichloroethene	7.117	130	157641	114.032	ug/l	96
45) 1,2-Dichloropropane	7.428	63	164505	113.217	ug/l	99
46) Dibromomethane	7.574	93	125846	109.869	ug/l	100
47) Bromodichloromethane	7.818	83	258326	116.365	ug/l	98
48) Methyl methacrylate	7.690	41	231040	124.356	ug/l	99
49) 1,4-Dioxane	7.665	88	82800	2241.695	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	1371275	577.831	ug/l	100
52) Toluene	8.714	92	393323	113.430	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	257233	138.001	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	271054	124.658	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	154932	110.551	ug/l	97
56) Ethyl methacrylate	9.116	69	277921	127.968	ug/l	99
57) 1,3-Dichloropropane	9.305	76	272814	109.813	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	680608	697.344	ug/l	100
59) 2-Hexanone	9.433	43	1028597	571.094	ug/l	99
60) Dibromochloromethane	9.519	129	187497	122.915	ug/l	98
61) 1,2-Dibromoethane	9.604	107	165634	115.087	ug/l	97
64) Tetrachloroethene	9.269	164	132488	100.918	ug/l	96
65) Chlorobenzene	10.080	112	429656	108.615	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	152173	115.333	ug/l	99
67) Ethyl Benzene	10.189	91	785121	116.421	ug/l	99
68) m/p-Xylenes	10.299	106	570913	234.349	ug/l	100
69) o-Xylene	10.640	106	279906	113.990	ug/l	99
70) Styrene	10.653	104	474254	120.694	ug/l	99
71) Bromoform	10.799	173	125986	127.128	ug/l #	99
73) Isopropylbenzene	10.957	105	752337	114.312	ug/l	99
74) N-amyl acetate	10.842	43	396877	120.944	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	243532	104.018	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	213744m	84.960	ug/l	
77) Bromobenzene	11.195	156	167934	110.297	ug/l	99
78) n-propylbenzene	11.305	91	881218	119.448	ug/l	100
79) 2-Chlorotoluene	11.360	91	534560	107.514	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	616477	112.825	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	81372	134.609	ug/l	93
82) 4-Chlorotoluene	11.451	91	611736	111.948	ug/l	100
83) tert-Butylbenzene	11.713	119	617473	115.065	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	623456	114.236	ug/l	99
85) sec-Butylbenzene	11.890	105	786293	118.023	ug/l	99
86) p-Isopropyltoluene	12.006	119	651531	121.113	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	313100	110.394	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	311752	112.167	ug/l	100
89) n-Butylbenzene	12.329	91	605783	129.533	ug/l	100
90) Hexachloroethane	12.536	117	123188	123.091	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	308039	111.506	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	64438	120.915	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	203255	134.366	ug/l	98
94) Hexachlorobutadiene	13.719	225	80558	113.857	ug/l	99
95) Naphthalene	13.774	128	721236	130.929	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	200379	126.331	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046044.D
Acq On : 05 May 2025 12:45
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:11:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion:168 Resp: 8296
Ion Ratio Lower Upper
168 100
99 67.6 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 118.929 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 85 Resp: 217793
Ion Ratio Lower Upper
85 100
87 32.5 15.7 47.1









#3
Chloromethane
Concen: 109.161 ug/l
RT: 1.307 min Scan# 30
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion: 50 Resp: 19684
Ion Ratio Lower Upper
50 100
52 32.5 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 118.953 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion: 62 Resp: 187924
Ion Ratio Lower Upper
62 100
64 31.7 25.2 37.8





#5
Bromomethane
Concen: 103.625 ug/l
RT: 1.593 min Scan# 81
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion: 94 Resp: 83008
Ion Ratio Lower Upper
94 100
96 95.8 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 96.822 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion: 64 Resp: 78806
Ion Ratio Lower Upper
64 100
66 31.3 24.3 36.5





#7
Trichlorofluoromethane
Concen: 101.851 ug/l
RT: 1.868 min Scan# 128
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:101 Resp: 245108
Ion Ratio Lower Upper
101 100
103 66.4 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 111.791 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 74 Resp: 88441
Ion Ratio Lower Upper
74 100
45 104.7 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 112.243 ug/l
RT: 2.313 min Scan# 202
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion:101 Resp: 161404
Ion Ratio Lower Upper
101 100
85 46.8 38.6 58.0
151 69.8 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#10
Methyl Iodide
Concen: 112.441 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:142 Resp: 189843
Ion Ratio Lower Upper
142 100
127 52.4 41.7 62.5
141 14.4 11.5 17.3





#11

Tert butyl alcohol

Concen: 616.119 ug/l

RT: 3.002 min Scan# 310

Delta R.T. 0.024 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 59 Resp: 181238

Ion Ratio Lower Upper

59 100





57 10.8 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 111.039 ug/l

RT: 2.307 min Scan# 200

Delta R.T. -0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 96 Resp: 155451


Ion Ratio Lower Upper

96 100

61 184.5 146.2 219.2

98 64.3 51.0 76.6





#13

Acrolein

Concen: 594.872 ug/l

RT: 2.239 min Scan# 188

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Tgt Ion: 56 Resp: 20325

Ion Ratio Lower Upper

56 100





55 72.1 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 112.561 ug/l

RT: 2.654 min Scan# 257

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 41 Resp: 297776


Ion Ratio Lower Upper

41 100

39 73.7 60.6 90.8

76 31.8 24.9 37.3





#15

Acrylonitrile

Concen: 550.973 ug/l

RT: 3.069 min Scan# 31

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Tgt Ion: 53 Resp: 484854

Ion Ratio Lower Upper

53 100

52 84.2 65.3 97.9





51 37.5 29.8 44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 545.754 ug/l

RT: 2.392 min Scan# 214

Delta R.T. 0.006 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 43 Resp: 460820

Ion Ratio Lower Upper

43 100

58 27.3 21.2 31.8





#17

Carbon Disulfide

Concen: 122.499 ug/l

RT: 2.502 min Scan# 21

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 76 Resp: 39735

Ion Ratio Lower Upper

76 100





78 8.8 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 108.099 ug/l

RT: 2.709 min Scan# 266

Delta R.T. 0.006 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 43 Resp: 217885

Ion Ratio Lower Upper

43 100

74 21.9 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 114.112 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 73 Resp: 557210
Ion Ratio Lower Upper
73 100
57 22.0 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#20
Methylene Chloride
Concen: 100.579 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion: 84 Resp: 172035
Ion Ratio Lower Upper
84 100
49 136.0 113.9 170.9
51 41.5 33.5 50.3
86 63.7 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 108.241 ug/l
RT: 3.081 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 96 Resp: 154694
Ion Ratio Lower Upper
96 100
61 154.6 119.5 179.3
98 64.3 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 116.858 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion: 45 Resp: 577610
Ion Ratio Lower Upper
45 100
43 70.3 66.6 100.0
87 24.7 19.8 29.6
59 10.6 8.6 12.8





#23
Vinyl Acetate
Concen: 608.005 ug/l
RT: 3.721 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150




Tgt Ion: 43 Resp: 265575
Ion Ratio Lower Upper
43 100
86 9.6 7.5 11.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#24
1,1-Dichloroethane
Concen: 110.182 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45



Tgt Ion: 63 Resp: 320083
Ion Ratio Lower Upper
63 100
98 7.0 3.6 10.8
100 4.7 2.1 6.3





#25

2-Butanone

Concen: 581.292 ug/l

RT: 4.562 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 70846

Ion Ratio Lower Upper

43 100





72 24.0 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 117.548 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 77 Resp: 258698

Ion Ratio Lower Upper

77 100

97 21.1 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 108.994 ug/l

RT: 4.483 min Scan# 557

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 96 Resp: 187835

Ion Ratio Lower Upper

96 100

61 159.8 0.0 322.8





98 64.8 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 95.108 ug/l

RT: 4.898 min Scan# 625

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 49 Resp: 146923


Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 70.5 56.2 84.2





#29

Tetrahydrofuran

Concen: 569.081 ug/l

RT: 5.007 min Scan# 642

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 42 Resp: 450450

Ion Ratio Lower Upper

42 100

72 41.6 33.7 50.5





71 38.7 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 107.208 ug/l

RT: 5.087 min Scan# 656

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 83 Resp: 323659

Ion Ratio Lower Upper

83 100


85 65.1 49.3 73.9





#31
Cyclohexane
Concen: 116.934 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion: 56 Resp: 28631
Ion Ratio Lower Upper
56 100
69 31.2 24.4 36.6
84 82.8 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#32
1,1,1-Trichloroethane
Concen: 113.572 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 97 Resp: 295702
Ion Ratio Lower Upper
97 100
99 64.3 51.8 77.6
61 48.0 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 93.983 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 65 Resp: 231952

Ion Ratio Lower Upper

65 100





67 49.4 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:114 Resp: 144975


Ion Ratio Lower Upper

114 100

63 23.9 0.0 49.2

88 18.3 0.0 33.6





#35

Dibromofluoromethane

Concen: 98.562 ug/l

RT: 5.385 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:113 Resp: 160168

Ion Ratio Lower Upper

113 100

111 103.0 83.1 124.7





192 16.9 13.3 19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 116.462 ug/l

RT: 5.684 min Scan# 754

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 75 Resp: 219731


Ion Ratio Lower Upper

75 100

110 32.9 16.3 48.9

77 30.8 24.3 36.5





#37

Ethyl Acetate

Concen: 115.889 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 274312

Ion Ratio Lower Upper

43 100

61 12.5 10.3 15.5





70 10.2 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 115.760 ug/l

RT: 5.666 min Scan# 751

Delta R.T. -0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:117 Resp: 250916


Ion Ratio Lower Upper

117 100

119 94.3 75.2 112.8

121 29.7 24.2 36.4





#39

Methylcyclohexane

Concen: 121.107 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 83 Resp: 28631

Ion Ratio Lower Upper

83 100

55 83.8 64.7 97.1





98 47.8 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 109.813 ug/l

RT: 6.031 min Scan# 811

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 78 Resp: 642178

Ion Ratio Lower Upper

78 100


77 23.4 19.0 28.4





#41
Methacrylonitrile
Concen: 115.139 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 41 Resp: 151424
Ion Ratio Lower Upper
41 100
39 60.0 47.2 70.8
67 65.9 50.7 76.1
52 30.9 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 112.492 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 62 Resp: 271992
Ion Ratio Lower Upper
62 100
98 7.4 0.0 15.2





#43

Isopropyl Acetate

Concen: 123.780 ug/l

RT: 6.342 min Scan# 861

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Tgt Ion: 43 Resp: 44787

Ion Ratio Lower Upper

43 100

61 18.1 14.3 21.5





87 11.7 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 114.032 ug/l

RT: 7.117 min Scan# 989

Delta R.T. -0.006 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:130 Resp: 157641

Ion Ratio Lower Upper

130 100

95 105.9 0.0 204.2





#45

1,2-Dichloropropane

Concen: 113.217 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 63 Resp: 16450

Ion Ratio Lower Upper

63 100





65 30.8 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 109.869 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 93 Resp: 125846


Ion Ratio Lower Upper

93 100

95 82.4 65.6 98.4


174 85.0 68.2 102.2





#47
Bromodichloromethane
Concen: 116.365 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 83 Resp: 258320
Ion Ratio Lower Upper
83 100
85 64.8 50.5 75.7
127 8.0 6.5 9.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#48
Methyl methacrylate
Concen: 124.356 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 41 Resp: 231040
Ion Ratio Lower Upper
41 100
69 73.0 58.5 87.7
39 63.8 51.7 77.5





#49

1,4-Dioxane

Concen: 2241.695 ug/l

RT: 7.665 min Scan# 1078

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 88 Resp: 82800

Ion Ratio Lower Upper

88 100

43 40.0 33.4 50.2





58 74.8 58.6 88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 101.252 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 98 Resp: 554390

Ion Ratio Lower Upper

98 100

100 65.4 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 577.831 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 43 Resp: 137127
Ion Ratio Lower Upper
43 100
58 36.2 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 113.430 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 92 Resp: 393323
Ion Ratio Lower Upper
92 100
91 171.0 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 138.001 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 75 Resp: 25723

Ion Ratio Lower Upper

75 100





77 31.7 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 124.658 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 75 Resp: 271054


Ion Ratio Lower Upper

75 100

77 31.4 25.4 38.0


39 60.0 52.2 78.4





#55
1,1,2-Trichloroethane
Concen: 110.551 ug/l
RT: 9.147 min Scan# 1322
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion: 97 Resp: 154932
Ion Ratio Lower Upper
97 100
83 88.5 69.0 103.6
85 56.0 44.3 66.5
99 64.5 47.8 71.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#56
Ethyl methacrylate
Concen: 127.968 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 69 Resp: 277921
Ion Ratio Lower Upper
69 100
41 75.4 60.8 91.2
39 47.2 39.0 58.6





#57

1,3-Dichloropropane

Concen: 109.813 ug/l

RT: 9.305 min Scan# 1174

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 76 Resp: 272814

Ion Ratio Lower Upper

76 100





78 32.3 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 697.344 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.006 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 63 Resp: 680608

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3





#59

2-Hexanone

Concen: 571.094 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 102859

Ion Ratio Lower Upper

43 100





58 50.2 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 122.915 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:129 Resp: 187497

Ion Ratio Lower Upper

129 100


127 77.1 39.3 117.8





#61
1,2-Dibromoethane
Concen: 115.087 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:107 Resp: 165634
Ion Ratio Lower Upper
107 100
109 94.6 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#62
4-Bromofluorobenzene
Concen: 109.174 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion: 95 Resp: 217536
Ion Ratio Lower Upper
95 100
174 67.5 0.0 135.8
176 64.4 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:117 Resp: 12855
Ion Ratio Lower Upper
117 100
82 61.4 50.6 76.0
119 31.9 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 100.918 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:164 Resp: 132488
Ion Ratio Lower Upper
164 100
166 123.3 105.0 157.6
129 101.6 83.5 125.3
131 97.3 76.5 114.7









#65
Chlorobenzene
Concen: 108.615 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion:112 Resp: 429650
Ion Ratio Lower Upper
112 100
114 31.5 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 115.333 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:131 Resp: 152173
Ion Ratio Lower Upper
131 100
133 95.7 47.3 141.9
119 64.0 31.6 95.0









#67
Ethyl Benzene
Concen: 116.421 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion: 91 Resp: 78512
Ion Ratio Lower Upper
91 100
106 29.9 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#68
m/p-Xylenes
Concen: 234.349 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:106 Resp: 570913
Ion Ratio Lower Upper
106 100
91 214.2 171.2 256.8









#69
o-Xylene
Concen: 113.990 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion:106 Resp: 279900
Ion Ratio Lower Upper
106 100
91 226.9 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#70
Styrene
Concen: 120.694 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion:104 Resp: 474254
Ion Ratio Lower Upper
104 100
78 57.2 45.7 68.5
103 55.5 43.7 65.5





#71
Bromoform
Concen: 127.128 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:173 Resp: 125980
Ion Ratio Lower Upper
173 100
175 47.3 23.4 70.0
254 0.1 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:152 Resp: 60345
Ion Ratio Lower Upper
152 100
115 150.7 46.9 140.7#
150 211.6 0.0 351.0









#73
Isopropylbenzene
Concen: 114.312 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion:105 Resp: 75233
Ion Ratio Lower Upper
105 100
120 25.3 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 120.944 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 43 Resp: 396877
Ion Ratio Lower Upper
43 100
70 38.5 30.9 46.3
55 22.5 18.7 28.1
61 22.5 17.1 25.7





#75

1,1,2,2-Tetrachloroethane

Concen: 104.018 ug/l

RT: 11.213 min Scan# 1661

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 83 Resp: 24353

Ion Ratio Lower Upper

83 100

131 9.7 5.0 14.9





85 64.4 31.9 95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#76

1,2,3-Trichloropropane

Concen: 84.960 ug/l m

RT: 11.238 min Scan# 1665

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 75 Resp: 213744

Ion Ratio Lower Upper

75 100


77 42.4 20.5 61.5





#77
Bromobenzene
Concen: 110.297 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion:156 Resp: 167934
Ion Ratio Lower Upper
156 100
77 187.1 92.8 278.3
158 97.3 48.8 146.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#78
n-propylbenzene
Concen: 119.448 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 91 Resp: 881218
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4





#79

2-Chlorotoluene

Concen: 107.514 ug/l

RT: 11.360 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 91 Resp: 534560

Ion Ratio Lower Upper

91 100





126 31.0 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#80

1,3,5-Trimethylbenzene

Concen: 112.825 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:105 Resp: 616477

Ion Ratio Lower Upper

105 100

120 46.0 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 134.609 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 75 Resp: 81372
Ion Ratio Lower Upper
75 100
53 114.9 100.5 150.7
89 46.5 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 111.948 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 91 Resp: 611736
Ion Ratio Lower Upper
91 100
126 26.7 13.3 39.8





#83

tert-Butylbenzene

Concen: 115.065 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:119 Resp: 61747

Ion Ratio Lower Upper

119 100

91 65.6 32.9 98.7





134 22.6 11.4 34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#84

1,2,4-Trimethylbenzene

Concen: 114.236 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:105 Resp: 623456

Ion Ratio Lower Upper

105 100

120 43.1 21.2 63.6





#85

sec-Butylbenzene

Concen: 118.023 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:105 Resp: 78629

Ion Ratio Lower Upper

105 100





134 19.1 9.7 29.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#86

p-Isopropyltoluene

Concen: 121.113 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:119 Resp: 651531


Ion Ratio Lower Upper

119 100

134 25.4 12.5 37.5

91 27.0 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 110.394 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:146 Resp: 313100

Ion Ratio Lower Upper

146 100

111 44.5 22.1 66.3





148 64.0 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 112.167 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:146 Resp: 311752


Ion Ratio Lower Upper

146 100

111 43.0 21.3 63.9

148 63.9 31.9 95.5





#89

n-Butylbenzene

Concen: 129.533 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 91 Resp: 60578

Ion Ratio Lower Upper

91 100

92 53.6 26.9 80.7





134 24.0 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 123.091 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:117 Resp: 123188

Ion Ratio Lower Upper

117 100


201 63.2 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 111.506 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion:146 Resp: 308039
Ion Ratio Lower Upper
146 100
111 46.3 22.4 67.3
148 63.6 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 120.915 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 75 Resp: 64438
Ion Ratio Lower Upper
75 100
155 69.6 34.9 104.8
157 88.4 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 134.366 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:180 Resp: 20325

Ion Ratio Lower Upper

180 100

182 94.1 47.9 143.8





145 34.5 17.1 51.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#94

Hexachlorobutadiene

Concen: 113.857 ug/l

RT: 13.719 min Scan# 2072

Delta R.T. -0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:225 Resp: 80558


Ion Ratio Lower Upper

225 100

223 62.5 30.8 92.4

227 63.1 30.9 92.7





#95

Naphthalene

Concen: 130.929 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:128 Resp: 721236

Ion Ratio Lower Upper

128 100

127 12.9 10.4 15.6





129 11.0 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#96

1,2,3-Trichlorobenzene

Concen: 126.331 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:180 Resp: 200379

Ion Ratio Lower Upper

180 100

182 96.5 47.8 143.3

145 36.7 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046046.D
 Acq On : 05 May 2025 16:04
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:12:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	96964	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	168484	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	147167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67939	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	9067	3.143	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	6.280%#		
35) Dibromofluoromethane	5.373	113	5969	3.161	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	6.320%#		
50) Toluene-d8	8.647	98	20566	3.232	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	6.460%#		
62) 4-Bromofluorobenzene	11.079	95	7822	3.378	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	6.760%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	6195	2.894	ug/l	95
3) Chloromethane	1.307	50	6587	3.125	ug/l	99
4) Vinyl Chloride	1.374	62	5999	3.249	ug/l	96
5) Bromomethane	1.593	94	2962	3.164	ug/l	88
6) Chloroethane	1.666	64	3567	3.750	ug/l	94
7) Trichlorofluoromethane	1.874	101	9599	3.413	ug/l	88
8) Diethyl Ether	2.130	74	3020	3.266	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.325	101	5911	3.517	ug/l	99
10) Methyl Iodide	2.441	142	5900	2.990	ug/l	99
11) Tert butyl alcohol	2.965	59	5539	16.111	ug/l	98
12) 1,1-Dichloroethene	2.306	96	5496	3.359	ug/l	97
13) Acrolein	2.233	56	5673	14.206	ug/l	95
14) Allyl chloride	2.654	41	10262	3.319	ug/l	98
15) Acrylonitrile	3.062	53	16746	16.282	ug/l	97
16) Acetone	2.380	43	19773	20.036	ug/l	98
17) Carbon Disulfide	2.501	76	11068	2.919	ug/l	96
18) Methyl Acetate	2.703	43	7909	3.357	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	18497	3.241	ug/l	97
20) Methylene Chloride	2.782	84	6680	3.341	ug/l	97
21) trans-1,2-Dichloroethene	3.081	96	5402	3.234	ug/l	100
22) Diisopropyl ether	3.751	45	18657	3.230	ug/l #	65
23) Vinyl Acetate	3.715	43	82328	16.127	ug/l	99
24) 1,1-Dichloroethane	3.605	63	11193	3.297	ug/l	99
25) 2-Butanone	4.562	43	26125	18.340	ug/l	98
26) 2,2-Dichloropropane	4.465	77	8245	3.205	ug/l	96
27) cis-1,2-Dichloroethene	4.483	96	6222	3.089	ug/l	95
28) Bromochloromethane	4.891	49	5360	2.969	ug/l #	100
29) Tetrahydrofuran	5.001	42	15431	16.680	ug/l	98
30) Chloroform	5.080	83	11624	3.294	ug/l	89
31) Cyclohexane	5.458	56	10269	3.588	ug/l	96
32) 1,1,1-Trichloroethane	5.367	97	9827	3.229	ug/l	97
36) 1,1-Dichloropropene	5.678	75	7787	3.551	ug/l	98
37) Ethyl Acetate	4.721	43	9806m	3.565	ug/l	
38) Carbon Tetrachloride	5.659	117	8505	3.376	ug/l	97
39) Methylcyclohexane	7.373	83	9882	3.597	ug/l	93
40) Benzene	6.031	78	22528	3.315	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046046.D
 Acq On : 05 May 2025 16:04
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:12:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	4851	3.174	ug/l #	65
42) 1,2-Dichloroethane	6.080	62	10011	3.563	ug/l	100
43) Isopropyl Acetate	6.348	43	13914	3.309	ug/l	96
44) Trichloroethene	7.123	130	5313	3.307	ug/l	94
45) 1,2-Dichloropropane	7.427	63	5451	3.228	ug/l #	89
46) Dibromomethane	7.580	93	4417	3.318	ug/l	98
47) Bromodichloromethane	7.818	83	8394	3.254	ug/l	98
48) Methyl methacrylate	7.696	41	7171	3.321	ug/l	97
49) 1,4-Dioxane	7.659	88	2907	67.721	ug/l	96
51) 4-Methyl-2-Pentanone	8.567	43	46794	16.967	ug/l	98
52) Toluene	8.714	92	14126	3.505	ug/l	97
53) t-1,3-Dichloropropene	8.976	75	6834	3.155	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	7908	3.129	ug/l	91
55) 1,1,2-Trichloroethane	9.153	97	5685	3.490	ug/l	98
56) Ethyl methacrylate	9.116	69	8555	3.390	ug/l	98
57) 1,3-Dichloropropane	9.305	76	10127	3.508	ug/l	94
58) 2-Chloroethyl Vinyl ether	8.238	63	20809	18.346	ug/l	99
59) 2-Hexanone	9.427	43	34853	16.651	ug/l	97
60) Dibromochloromethane	9.518	129	5500	3.102	ug/l	98
61) 1,2-Dibromoethane	9.604	107	5614	3.356	ug/l	99
64) Tetrachloroethene	9.269	164	4752	3.162	ug/l	96
65) Chlorobenzene	10.079	112	15391	3.399	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	5024	3.326	ug/l	99
67) Ethyl Benzene	10.195	91	26732	3.463	ug/l	99
68) m/p-Xylenes	10.299	106	19957	7.156	ug/l	98
69) o-Xylene	10.640	106	9398	3.343	ug/l	95
70) Styrene	10.652	104	14900	3.312	ug/l	99
71) Bromoform	10.799	173	3470	3.059	ug/l #	99
73) Isopropylbenzene	10.957	105	24201	3.266	ug/l	98
74) N-amyl acetate	10.841	43	11224	3.038	ug/l	96
75) 1,1,2,2-Tetrachloroethane	11.207	83	9170	3.479	ug/l	97
76) 1,2,3-Trichloropropane	11.238	75	7928m	2.799	ug/l	
77) Bromobenzene	11.195	156	5857	3.417	ug/l	96
78) n-propylbenzene	11.299	91	28436	3.424	ug/l	100
79) 2-Chlorotoluene	11.360	91	18672	3.336	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	20742	3.372	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	1828	2.686	ug/l	99
82) 4-Chlorotoluene	11.451	91	19966	3.245	ug/l	98
83) tert-Butylbenzene	11.713	119	21049	3.484	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	20613	3.355	ug/l	97
85) sec-Butylbenzene	11.890	105	25590	3.412	ug/l	99
86) p-Isopropyltoluene	12.006	119	20955	3.460	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	10585	3.315	ug/l	97
88) 1,4-Dichlorobenzene	12.036	146	10908	3.486	ug/l	90
89) n-Butylbenzene	12.329	91	18004	3.419	ug/l	94
90) Hexachloroethane	12.536	117	3553	3.153	ug/l	94
91) 1,2-Dichlorobenzene	12.335	146	10712	3.444	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	1687	2.812	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	5720	3.359	ug/l	96
94) Hexachlorobutadiene	13.719	225	2816	3.535	ug/l	96
95) Naphthalene	13.774	128	19902	3.209	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	5745	3.217	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046046.D
Acq On : 05 May 2025 16:04
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:12:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046046.D
Acq On    : 05 May 2025  16:04
Operator  : JC/MD
Sample    : VSTDIC005
Misc      : 5.0mL/MSVOA_X/WATER
ALS Vial  : 10   Sample Multiplier: 1
```


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:12:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.538 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion:168 Resp: 9696
Ion Ratio Lower Upper
168 100
99 74.0 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 2.894 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 85 Resp: 6195
Ion Ratio Lower Upper
85 100
87 28.7 15.7 47.1









#3
Chloromethane
Concen: 3.125 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion: 50 Resp: 658
Ion Ratio Lower Upper
50 100
52 32.2 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 3.249 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion: 62 Resp: 5999
Ion Ratio Lower Upper
62 100
64 29.1 25.2 37.8





#5
Bromomethane
Concen: 3.164 ug/l
RT: 1.593 min Scan# 83
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion: 94 Resp: 2962
Ion Ratio Lower Upper
94 100
96 102.2 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 3.750 ug/l
RT: 1.666 min Scan# 95
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion: 64 Resp: 3567
Ion Ratio Lower Upper
64 100
66 33.7 24.3 36.5





#7
Trichlorofluoromethane
Concen: 3.413 ug/l
RT: 1.874 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion:101 Resp: 959
Ion Ratio Lower Upper
101 100
103 54.0 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 3.266 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 74 Resp: 3020
Ion Ratio Lower Upper
74 100
45 115.1 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 3.517 ug/l
RT: 2.325 min Scan# 202
Delta R.T. 0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion:101 Resp: 591
Ion Ratio Lower Upper
101 100
85 46.3 38.6 58.0
151 68.8 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#10
Methyl Iodide
Concen: 2.990 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:142 Resp: 5900
Ion Ratio Lower Upper
142 100
127 52.8 41.7 62.5
141 14.3 11.5 17.3





#11

Tert butyl alcohol

Concen: 16.111 ug/l

RT: 2.965 min Scan# 30

Delta R.T. -0.012 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 59 Resp: 553

Ion Ratio Lower Upper

59 100





57 10.0 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 3.359 ug/l

RT: 2.306 min Scan# 200

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 96 Resp: 5496


Ion Ratio Lower Upper

96 100

61 178.7 146.2 219.2

98 59.5 51.0 76.6





#13

Acrolein

Concen: 14.206 ug/l

RT: 2.233 min Scan# 188

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 56 Resp: 567

Ion Ratio Lower Upper

56 100





55 74.5 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carbone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 3.319 ug/l

RT: 2.654 min Scan# 257

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 41 Resp: 10262


Ion Ratio Lower Upper

41 100

39 74.1 60.6 90.8

76 29.1 24.9 37.3





#15

Acrylonitrile

Concen: 16.282 ug/l

RT: 3.062 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	53	Resp:	16740
Ion	Ratio	Lower	Upper
53	100		
52	84.3	65.3	97.9
51	38.5	29.8	44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 20.036 ug/l




RT: 2.380 min Scan# 212


Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	43	Resp:	19773
Ion	Ratio	Lower	Upper
43	100		
58	27.7	21.2	31.8





#17

Carbon Disulfide

Concen: 2.919 ug/l

RT: 2.501 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04

Tgt Ion: 76 Resp: 11068

Ion Ratio Lower Upper

76 100

78 10.2 7.0 10.4



Instrument :

MSVOA_X

ClientSampleId :


VSTDIC005

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 3.357 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 43 Resp: 7909

Ion Ratio Lower Upper

43 100

74 20.9 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 3.241 ug/l

RT: 3.111 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion: 73 Resp: 1849

Ion Ratio Lower Upper

73 100

57 20.5 17.7 26.5

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#20

Methylene Chloride

Concen: 3.341 ug/l

RT: 2.782 min Scan# 278

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion: 84 Resp: 6680




Ion Ratio Lower Upper


84 100

49 137.0 113.9 170.9

51 43.8 33.5 50.3


86 67.6 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 3.234 ug/l
RT: 3.081 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion: 96 Resp: 540
Ion Ratio Lower Upper
96 100
61 149.9 119.5 179.3
98 62.9 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 3.230 ug/l
RT: 3.751 min Scan# 437
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 45 Resp: 18657
Ion Ratio Lower Upper
45 100
43 128.0 66.6 100.0#
87 24.7 19.8 29.6
59 10.7 8.6 12.8









#23
Vinyl Acetate
Concen: 16.127 ug/l
RT: 3.715 min Scan# 41
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion: 43 Resp: 82328
Ion Ratio Lower Upper
43 100
86 9.1 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#24
1,1-Dichloroethane
Concen: 3.297 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 63 Resp: 11193
Ion Ratio Lower Upper
63 100
98 6.9 3.6 10.8
100 4.7 2.1 6.3





#25

2-Butanone

Concen: 18.340 ug/l

RT: 4.562 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 43 Resp: 2612

Ion Ratio Lower Upper

43 100





72 21.8 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 3.205 ug/l

RT: 4.465 min Scan# 554

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 77 Resp: 8245

Ion Ratio Lower Upper

77 100

97 23.1 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 3.089 ug/l

RT: 4.483 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 96 Resp: 622

Ion Ratio Lower Upper

96 100

61 168.5 0.0 322.8





98 67.6 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 2.969 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 49 Resp: 5360


Ion Ratio Lower Upper

49 100

129 0.0 0.0 4.0

130 70.5 56.2 84.2





#29

Tetrahydrofuran

Concen: 16.680 ug/l

RT: 5.001 min Scan# 642

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 42 Resp: 1543

Ion Ratio Lower Upper

42 100

72 42.2 33.7 50.5





71 37.4 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 3.294 ug/l

RT: 5.080 min Scan# 655

Delta R.T. -0.006 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 83 Resp: 11624

Ion Ratio Lower Upper

83 100

85 69.9 49.3 73.9





#31

Cyclohexane

Concen: 3.588 ug/l

RT: 5.458 min Scan# 718

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 56 Resp: 10269

Ion Ratio Lower Upper

56 100

69 31.6 24.4 36.6





84 79.1 66.9 100.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#32

1,1,1-Trichloroethane

Concen: 3.229 ug/l

RT: 5.367 min Scan# 702

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 97 Resp: 9827


Ion Ratio Lower Upper

97 100

99 62.6 51.8 77.6

61 51.8 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 3.143 ug/l

RT: 5.946 min Scan# 798

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 65 Resp: 906

Ion Ratio Lower Upper

65 100





67 50.1 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.751 min Scan# 929

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion:114 Resp: 168484


Ion Ratio Lower Upper

114 100

63 24.5 0.0 49.2

88 17.7 0.0 33.6





#35

Dibromofluoromethane

Concen: 3.161 ug/l

RT: 5.373 min Scan# 704

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	113	Resp:	596
Ion	Ratio	Lower	Upper
113	100		
111	104.4	83.1	124.7
192	17.5	13.3	19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 3.551 ug/l




RT: 5.678 min Scan# 753


Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	75	Resp:	7787
Ion	Ratio	Lower	Upper
75	100		
110	30.6	16.3	48.9
77	30.5	24.3	36.5





#37

Ethyl Acetate

Concen: 3.565 ug/l m

RT: 4.721 min Scan# 595

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 43 Resp: 9800

Ion Ratio Lower Upper

43 100

61 11.0 10.3 15.5





70 8.4 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 3.376 ug/l

RT: 5.659 min Scan# 750

Delta R.T. -0.012 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion:117 Resp: 8505


Ion Ratio Lower Upper

117 100

119 90.6 75.2 112.8

121 29.5 24.2 36.4





#39

Methylcyclohexane

Concen: 3.597 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC005

Tgt Ion: 83	Resp: 988
Ion Ratio	Lower Upper
83	100
55	73.3 64.7 97.1
98	50.3 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 3.315 ug/l




RT: 6.031 min Scan# 811


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04


Tgt Ion: 78	Resp: 22528
Ion Ratio	Lower Upper
78	100
77	23.9 19.0 28.4





#41
Methacrylonitrile
Concen: 3.174 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion: 41 Resp: 485
Ion Ratio Lower Upper
41 100
39 59.7 47.2 70.8
67 0.0 50.7 76.1
52 28.8 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 3.563 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 62 Resp: 10011
Ion Ratio Lower Upper
62 100
98 7.8 0.0 15.2





#43

Isopropyl Acetate

Concen: 3.309 ug/l

RT: 6.348 min Scan# 863

Delta R.T. 0.012 min

Lab File: VX046046.D


Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005






Tgt Ion: 43	Resp: 13914
Ion Ratio	Lower Upper
43	100
61	16.2 14.3 21.5
87	10.6 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 3.307 ug/l




RT: 7.123 min Scan# 990


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:130	Resp: 5313
Ion Ratio	Lower Upper
130	100
95	108.1 0.0 204.2





#45

1,2-Dichloropropane

Concen: 3.228 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 63 Resp: 545

Ion Ratio Lower Upper

63 100





65 36.2 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 3.318 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 93 Resp: 4417


Ion Ratio Lower Upper

93 100

95 80.1 65.6 98.4

174 87.6 68.2 102.2





#47

Bromodichloromethane

Concen: 3.254 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	83	Resp:	8394
Ion	Ratio	Lower	Upper
83	100		
85	61.9	50.5	75.7
127	9.0	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 3.321 ug/l




RT: 7.696 min Scan# 1084


Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	41	Resp:	7171
Ion	Ratio	Lower	Upper
41	100		
69	68.4	58.5	87.7
39	65.1	51.7	77.5





#49

1,4-Dioxane

Concen: 67.721 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	88	Resp:	290
Ion	Ratio	Lower	Upper
88	100		
43	45.3	33.4	50.2
58	71.0	58.6	88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 3.232 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	98	Resp:	20566
Ion	Ratio	Lower	Upper
98	100		
100	66.1	53.5	80.3









#51
4-Methyl-2-Pentanone
Concen: 16.967 ug/l
RT: 8.567 min Scan# 1227
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 43 Resp: 46794
Ion Ratio Lower Upper
43 100
58 34.8 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 3.505 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 92 Resp: 14126
Ion Ratio Lower Upper
92 100
91 167.2 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 3.155 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 75 Resp: 6834

Ion Ratio Lower Upper

75 100





77 31.1 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 3.129 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 75 Resp: 7908


Ion Ratio Lower Upper

75 100

77 31.9 25.4 38.0

39 55.3 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 3.490 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 97 Resp: 568

Ion Ratio Lower Upper

97 100

83 88.0 69.0 103.6

85 54.4 44.3 66.5





99 62.2 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 3.390 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 69 Resp: 8555


Ion Ratio Lower Upper

69 100

41 73.7 60.8 91.2

39 49.3 39.0 58.6





#57

1,3-Dichloropropane

Concen: 3.508 ug/l

RT: 9.305 min Scan# 1134

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 76 Resp: 1012

Ion Ratio Lower Upper

76 100





78 29.5 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 18.346 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 63 Resp: 20809

Ion Ratio Lower Upper

63 100

106 26.4 21.5 32.3





#59

2-Hexanone

Concen: 16.651 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 43 Resp: 3485

Ion Ratio Lower Upper

43 100





58 47.5 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 3.102 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion:129 Resp: 5500

Ion Ratio Lower Upper

129 100

127 80.4 39.3 117.8









#61
1,2-Dibromoethane
Concen: 3.356 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion:107 Resp: 5614
Ion Ratio Lower Upper
107 100
109 91.0 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carbone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#62
4-Bromofluorobenzene
Concen: 3.378 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion: 95 Resp: 7822
Ion Ratio Lower Upper
95 100
174 65.0 0.0 135.8
176 66.6 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion:117 Resp: 14716
Ion Ratio Lower Upper
117 100
82 61.1 50.6 76.0
119 32.1 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 3.162 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion:164 Resp: 4752
Ion Ratio Lower Upper
164 100
166 127.8 105.0 157.6
129 105.8 83.5 125.3
131 104.1 76.5 114.7





#65
Chlorobenzene
Concen: 3.399 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005



Tgt Ion:112 Resp: 1539
Ion Ratio Lower Upper
112 100
114 32.6 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 3.326 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:131 Resp: 5024
Ion Ratio Lower Upper
131 100
133 95.0 47.3 141.9
119 64.8 31.6 95.0





#67

Ethyl Benzene

Concen: 3.463 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 91 Resp: 2673

Ion Ratio Lower Upper

91 100





106 29.9 23.4 35.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#68

m/p-Xylenes

Concen: 7.156 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion:106 Resp: 19957

Ion Ratio Lower Upper

106 100

91 210.1 171.2 256.8









#69
o-Xylene
Concen: 3.343 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion:106 Resp: 9398
Ion Ratio Lower Upper
106 100
91 233.6 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#70
Styrene
Concen: 3.312 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion:104 Resp: 14900
Ion Ratio Lower Upper
104 100
78 57.9 45.7 68.5
103 55.6 43.7 65.5





#71
Bromoform
Concen: 3.059 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005



Tgt Ion:173 Resp: 3470
Ion Ratio Lower Upper
173 100
175 45.8 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:152 Resp: 67939
Ion Ratio Lower Upper
152 100
115 66.8 46.9 140.7
150 157.5 0.0 351.0









#73
Isopropylbenzene
Concen: 3.266 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 105 Resp: 2420
Ion Ratio Lower Upper
105 100
120 26.4 12.8 38.4




Manual Integrations
APPROVED


Reviewed By : John Carlone 05/06/2025
Supervised By : Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 3.038 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 43 Resp: 11224
Ion Ratio Lower Upper
43 100
70 35.4 30.9 46.3
55 25.0 18.7 28.1
61 19.6 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 3.479 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 83 Resp: 9170
Ion Ratio Lower Upper
83 100
131 8.3 5.0 14.9
85 61.7 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#76
1,2,3-Trichloropropane
Concen: 2.799 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion: 75 Resp: 7928
Ion Ratio Lower Upper
75 100
77 40.6 20.5 61.5





#77
Bromobenzene
Concen: 3.417 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion:156 Resp: 585
Ion Ratio Lower Upper
156 100
77 177.0 92.8 278.3
158 97.7 48.8 146.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#78
n-propylbenzene
Concen: 3.424 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 91 Resp: 28436
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4









#79
2-Chlorotoluene
Concen: 3.336 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion: 91 Resp: 18672
Ion Ratio Lower Upper
91 100
126 29.7 15.6 46.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#80
1,3,5-Trimethylbenzene
Concen: 3.372 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion:105 Resp: 20742
Ion Ratio Lower Upper
105 100
120 45.9 23.1 69.2





#81
trans-1,4-Dichloro-2-butene
Concen: 2.686 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion: 75 Resp: 1828
Ion Ratio Lower Upper
75 100
53 125.8 100.5 150.7
89 47.9 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 3.245 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 91 Resp: 19966
Ion Ratio Lower Upper
91 100
126 27.3 13.3 39.8





#83

tert-Butylbenzene

Concen: 3.484 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion:119 Resp: 21049

Ion Ratio Lower Upper

119 100

91 67.0 32.9 98.7





134 23.8 11.4 34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#84

1,2,4-Trimethylbenzene

Concen: 3.355 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion:105 Resp: 20613

Ion Ratio Lower Upper

105 100


120 44.5 21.2 63.6





#85
sec-Butylbenzene
Concen: 3.412 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005



Tgt Ion:105 Resp: 25590
Ion Ratio Lower Upper
105 100
134 19.1 9.7 29.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#86
p-Isopropyltoluene
Concen: 3.460 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion:119 Resp: 20955
Ion Ratio Lower Upper
119 100
134 25.6 12.5 37.5
91 28.0 13.8 41.4





#87
1,3-Dichlorobenzene
Concen: 3.315 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005




Tgt Ion:146 Resp: 1058
Ion Ratio Lower Upper
146 100
111 46.2 22.1 66.3
148 62.4 32.1 96.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#88
1,4-Dichlorobenzene
Concen: 3.486 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04



Tgt Ion:146 Resp: 10908
Ion Ratio Lower Upper
146 100
111 52.1 21.3 63.9
148 69.4 31.9 95.5





#89

n-Butylbenzene

Concen: 3.419 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 91 Resp: 18004

Ion Ratio Lower Upper

91 100

92 49.5 26.9 80.7





134 21.2 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 3.153 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion:117 Resp: 3553

Ion Ratio Lower Upper

117 100


201 58.5 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 3.444 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005




Tgt Ion: 146 Resp: 10712
Ion Ratio Lower Upper
146 100
111 44.7 22.4 67.3
148 64.2 31.6 94.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 2.812 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Tgt Ion: 75 Resp: 1687
Ion Ratio Lower Upper
75 100
155 71.7 34.9 104.8
157 87.0 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 3.359 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005



Tgt Ion:180 Resp: 5720
Ion Ratio Lower Upper
180 100
182 91.6 47.9 143.8
145 35.7 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#94
Hexachlorobutadiene
Concen: 3.535 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:225 Resp: 2816
Ion Ratio Lower Upper
225 100
223 65.1 30.8 92.4
227 58.8 30.9 92.7





#95

Naphthalene

Concen: 3.209 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:128	Resp:	1990
Ion	Ratio	Lower Upper
128	100	
127	13.0	10.4 15.6
129	11.5	8.6 13.0

Manual Integrations

APPROVED

Reviewed By :John Carbone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#96

1,2,3-Trichlorobenzene

Concen: 3.217 ug/l




RT: 13.957 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:180	Resp:	5745
Ion	Ratio	Lower Upper
180	100	
182	99.1	47.8 143.3
145	34.8	18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046047.D
 Acq On : 05 May 2025 16:27
 Operator : JC/MD
 Sample : VSTDIC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:13:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	92040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	167022	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	140108	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59123	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount	50.000	Range	74 - 125	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount	50.000	Range	86 - 113	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount	50.000	Range	77 - 121	Recovery	=	0.000%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	1211	0.596	ug/l	98
3) Chloromethane	1.307	50	1278	0.639	ug/l	97
4) Vinyl Chloride	1.374	62	1239	0.707	ug/l #	82
6) Chloroethane	1.678	64	859	0.951	ug/l	91
7) Trichlorofluoromethane	1.880	101	1959	0.734	ug/l	95
8) Diethyl Ether	2.130	74	742	0.845	ug/l	87
9) 1,1,2-Trichlorotrifluo...	2.319	101	1166	0.731	ug/l	96
12) 1,1-Dichloroethene	2.313	96	1093	0.704	ug/l #	87
14) Allyl chloride	2.654	41	1936	0.660	ug/l	89
15) Acrylonitrile	3.075	53	3343	3.424	ug/l	98
16) Acetone	2.380	43	3500	3.736	ug/l	97
17) Carbon Disulfide	2.508	76	2619	0.728	ug/l	98
18) Methyl Acetate	2.709	43	1852	0.828	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	3588	0.662	ug/l	96
20) Methylene Chloride	2.782	84	1571	0.828	ug/l #	90
21) trans-1,2-Dichloroethene	3.081	96	1111	0.701	ug/l #	71
22) Diisopropyl ether	3.757	45	3857	0.703	ug/l #	26
23) Vinyl Acetate	3.727	43	15283	3.154	ug/l	96
24) 1,1-Dichloroethane	3.605	63	2054	0.637	ug/l	97
25) 2-Butanone	4.593	43	4558	3.371	ug/l	92
26) 2,2-Dichloropropane	4.458	77	1777	0.728	ug/l	84
27) cis-1,2-Dichloroethene	4.483	96	1324	0.693	ug/l	91
28) Bromochloromethane	4.910	49	1061m	0.619	ug/l	
29) Tetrahydrofuran	5.019	42	2931	3.338	ug/l	99
30) Chloroform	5.093	83	2328	0.695	ug/l	98
32) 1,1,1-Trichloroethane	5.379	97	1869	0.647	ug/l #	51
36) 1,1-Dichloropropene	5.684	75	1648	0.758	ug/l #	83
37) Ethyl Acetate	4.733	43	1956m	0.717	ug/l	
38) Carbon Tetrachloride	5.678	117	1807	0.724	ug/l #	78
39) Methylcyclohexane	7.373	83	2096	0.770	ug/l	97
40) Benzene	6.044	78	4502	0.668	ug/l #	89
41) Methacrylonitrile	4.946	41	778	0.513	ug/l #	56
42) 1,2-Dichloroethane	6.092	62	1934	0.694	ug/l	81
43) Isopropyl Acetate	6.354	43	2553	0.612	ug/l #	81
44) Trichloroethene	7.135	130	1083	0.680	ug/l	85
45) 1,2-Dichloropropane	7.440	63	1059	0.633	ug/l	89

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046047.D
 Acq On : 05 May 2025 16:27
 Operator : JC/MD
 Sample : VSTDIC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:13:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	7.586	93	880	0.667	ug/l	97
47) Bromodichloromethane	7.818	83	1620	0.633	ug/l #	91
48) Methyl methacrylate	7.714	41	1235m	0.577	ug/l	
49) 1,4-Dioxane	7.677	88	498	11.703	ug/l #	66
51) 4-Methyl-2-Pentanone	8.580	43	9378	3.430	ug/l	97
52) Toluene	8.720	92	2684	0.672	ug/l	96
53) t-1,3-Dichloropropene	8.988	75	1238	0.576	ug/l #	79
54) cis-1,3-Dichloropropene	8.379	75	1414	0.564	ug/l #	87
55) 1,1,2-Trichloroethane	9.153	97	1029	0.637	ug/l	92
56) Ethyl methacrylate	9.128	69	1260	0.504	ug/l #	76
57) 1,3-Dichloropropane	9.311	76	2036	0.711	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.244	63	3846	3.420	ug/l	94
59) 2-Hexanone	9.439	43	6424	3.096	ug/l	88
60) Dibromochloromethane	9.525	129	1023	0.582	ug/l	93
61) 1,2-Dibromoethane	9.610	107	1077	0.650	ug/l	98
64) Tetrachloroethene	9.275	164	972	0.679	ug/l	95
65) Chlorobenzene	10.079	112	3170	0.735	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	1035	0.720	ug/l #	67
67) Ethyl Benzene	10.195	91	5052	0.687	ug/l	94
68) m/p-Xylenes	10.299	106	3630	1.367	ug/l	97
69) o-Xylene	10.640	106	1799	0.672	ug/l	94
70) Styrene	10.659	104	2664	0.622	ug/l	98
71) Bromoform	10.799	173	657	0.608	ug/l #	100
73) Isopropylbenzene	10.963	105	4480	0.695	ug/l	96
74) N-amyl acetate	10.854	43	2028	0.631	ug/l #	91
75) 1,1,2,2-Tetrachloroethane	11.213	83	1835	0.800	ug/l	94
76) 1,2,3-Trichloropropane	11.244	75	1661m	0.674	ug/l	
77) Bromobenzene	11.201	156	1095	0.734	ug/l	96
78) n-propylbenzene	11.305	91	5052	0.699	ug/l	97
79) 2-Chlorotoluene	11.360	91	3765	0.773	ug/l	94
80) 1,3,5-Trimethylbenzene	11.451	105	3590	0.671	ug/l	96
82) 4-Chlorotoluene	11.457	91	3815	0.713	ug/l	97
83) tert-Butylbenzene	11.713	119	3951	0.751	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	3725	0.697	ug/l	97
85) sec-Butylbenzene	11.890	105	4385	0.672	ug/l	100
86) p-Isopropyltoluene	12.006	119	3577	0.679	ug/l	94
87) 1,3-Dichlorobenzene	11.969	146	1914	0.689	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	2149m	0.789	ug/l	
89) n-Butylbenzene	12.335	91	2889	0.631	ug/l	92
90) Hexachloroethane	12.536	117	604	0.616	ug/l	99
91) 1,2-Dichlorobenzene	12.341	146	2022	0.747	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	12.945	75	306	0.586	ug/l	83
93) 1,2,4-Trichlorobenzene	13.591	180	1019	0.688	ug/l	94
94) Hexachlorobutadiene	13.725	225	465	0.671	ug/l	89
95) Naphthalene	13.780	128	4138	0.767	ug/l	97
96) 1,2,3-Trichlorobenzene	13.963	180	1113	0.716	ug/l	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001

Tgt Ion:168 Resp: 92040
Ion Ratio Lower Upper
168 100
99 71.5 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 0.596 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 85 Resp: 1211
Ion Ratio Lower Upper
85 100
87 32.4 15.7 47.1









#3
Chloromethane
Concen: 0.639 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 50 Resp: 1278
Ion Ratio Lower Upper
50 100
52 30.3 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 0.707 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 62 Resp: 1239
Ion Ratio Lower Upper
62 100
64 41.7 25.2 37.8#









#6
Chloroethane
Concen: 0.951 ug/l
RT: 1.678 min Scan# 91
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 64 Resp: 859
Ion Ratio Lower Upper
64 100
66 35.2 24.3 36.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#7
Trichlorofluoromethane
Concen: 0.734 ug/l
RT: 1.880 min Scan# 130
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:101 Resp: 1959
Ion Ratio Lower Upper
101 100
103 67.8 51.0 76.4









#8
Diethyl Ether
Concen: 0.845 ug/l
RT: 2.130 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 74 Resp: 74
Ion Ratio Lower Upper
74 100
45 123.6 54.9 164.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#9
1,1,2-Trichlorotrifluoroethane
Concen: 0.731 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:101 Resp: 1166
Ion Ratio Lower Upper
101 100
85 47.3 38.6 58.0
151 73.1 55.2 82.8





#12

1,1-Dichloroethene

Concen: 0.704 ug/l

RT: 2.313 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:	96	Resp:	109
Ion	Ratio	Lower	Upper
96	100		
61	199.6	146.2	219.2
98	49.6	51.0	76.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 0.660 ug/l




RT: 2.654 min Scan# 257


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:	41	Resp:	1936
Ion	Ratio	Lower	Upper
41	100		
39	85.5	60.6	90.8
76	26.4	24.9	37.3





#15

Acrylonitrile

Concen: 3.424 ug/l

RT: 3.075 min Scan# 31

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 53	Resp: 334
Ion Ratio	Lower Upper
53 100	
52 82.9	65.3 97.9
51 34.4	29.8 44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 3.736 ug/l




RT: 2.380 min Scan# 212


Delta R.T. -0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 43	Resp: 3500
Ion Ratio	Lower Upper
43 100	
58 28.1	21.2 31.8





#17

Carbon Disulfide

Concen: 0.728 ug/l

RT: 2.508 min Scan# 21

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 76 Resp: 2619

Ion Ratio Lower Upper

76 100





78 8.0 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 0.828 ug/l

RT: 2.709 min Scan# 266

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 43 Resp: 1852

Ion Ratio Lower Upper

43 100

74 22.0 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 0.662 ug/l

RT: 3.117 min Scan# 311

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 73 Resp: 3588

Ion Ratio Lower Upper

73 100





57 20.3 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#20

Methylene Chloride

Concen: 0.828 ug/l

RT: 2.782 min Scan# 278

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 84 Resp: 1571




Ion Ratio Lower Upper


84 100

49 140.0 113.9 170.9

51 51.3 33.5 50.3#

86 85.2 53.8 80.8#





#21

trans-1,2-Dichloroethene

Concen: 0.701 ug/l

RT: 3.081 min Scan# 31

Delta R.T. -0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 96 Resp: 111

Ion Ratio Lower Upper

96 100

61 119.0 119.5 179.3





98 30.1 50.0 75.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#22

Diisopropyl ether

Concen: 0.703 ug/l

RT: 3.757 min Scan# 438

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 45 Resp: 3857




Ion Ratio Lower Upper


45 100

43 176.2 66.6 100.0

87 28.9 19.8 29.6


59 9.1 8.6 12.8





#23
Vinyl Acetate
Concen: 3.154 ug/l
RT: 3.727 min Scan# 413
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion: 43 Resp: 1528
Ion Ratio Lower Upper
43 100
86 10.8 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#24
1,1-Dichloroethane
Concen: 0.637 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 63 Resp: 2054
Ion Ratio Lower Upper
63 100
98 6.3 3.6 10.8
100 5.8 2.1 6.3





#25

2-Butanone

Concen: 3.371 ug/l

RT: 4.593 min Scan# 51

Delta R.T. 0.037 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 43 Resp: 4558

Ion Ratio Lower Upper

43 100





72 19.2 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 0.728 ug/l

RT: 4.458 min Scan# 553

Delta R.T. -0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 77 Resp: 1777

Ion Ratio Lower Upper

77 100

97 13.7 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 0.693 ug/l

RT: 4.483 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 96	Resp: 1324
Ion Ratio	Lower Upper
96 100	
61 145.5	0.0 322.8
98 63.1	0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 0.619 ug/l m




RT: 4.910 min Scan# 627


Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 49	Resp: 1061
Ion Ratio	Lower Upper
49 100	
129 0.0	0.0 4.0
130 65.3	56.2 84.2





#29

Tetrahydrofuran

Concen: 3.338 ug/l

RT: 5.019 min Scan# 64

Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 42 Resp: 293

Ion Ratio Lower Upper

42 100

72 41.9 33.7 50.5





71 37.7 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 0.695 ug/l

RT: 5.093 min Scan# 657

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 83 Resp: 2328

Ion Ratio Lower Upper

83 100

85 60.4 49.3 73.9





#32

1,1,1-Trichloroethane

Concen: 0.647 ug/l

RT: 5.379 min Scan# 703

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001

Tgt Ion: 97 Resp: 1869

Ion Ratio Lower Upper

97 100

99 0.0 51.8 77.6





61 55.0 40.1 60.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#33

1,2-Dichloroethane-d4

Concen: 0.000 ug/l

Expected RT: 5.95 min

Lab File: VX046047.D


Acq: 05 May 2025 16:27


Tgt Ion: 65

Sig Exp Ratio

65 100

67 49.5





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 91

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:114 Resp: 16702

Ion Ratio Lower Upper

114 100

63 24.2 0.0 49.2





88 17.1 0.0 33.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#35

Dibromofluoromethane

Concen: 0.000 ug/l

Expected RT: 5.38 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27


Tgt Ion: 113


Sig Exp Ratio

113 100

111 103.9

192 16.6





#36

1,1-Dichloropropene

Concen: 0.758 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:	75	Resp:	164
Ion	Ratio	Lower	Upper
75	100		
110	28.2	16.3	48.9
77	15.8	24.3	36.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#37

Ethyl Acetate

Concen: 0.717 ug/l m




RT: 4.733 min Scan# 598


Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:	43	Resp:	1956
Ion	Ratio	Lower	Upper
43	100		
61	0.0	10.3	15.5#
70	1.0	7.9	11.9#





#38

Carbon Tetrachloride

Concen: 0.724 ug/l

RT: 5.678 min Scan# 752

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:117 Resp: 180

Ion Ratio Lower Upper

117 100

119 67.0 75.2 112.8





121 28.5 24.2 36.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#39

Methylcyclohexane

Concen: 0.770 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion: 83 Resp: 2096


Ion Ratio Lower Upper

83 100

55 80.1 64.7 97.1

98 41.9 37.4 56.2





#40

Benzene

Concen: 0.668 ug/l

RT: 6.044 min Scan# 811

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 78 Resp: 450

Ion Ratio Lower Upper

78 100





77 29.1 19.0 28.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#41

Methacrylonitrile

Concen: 0.513 ug/l

RT: 4.946 min Scan# 633

Delta R.T. 0.030 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 41 Resp: 778




Ion Ratio Lower Upper


41 100

39 51.7 47.2 70.8

67 0.0 50.7 76.1#

52 13.6 24.6 37.0#





#42

1,2-Dichloroethane

Concen: 0.694 ug/l

RT: 6.092 min Scan# 819

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 62 Resp: 1934
Ion Ratio Lower Upper
62 100
98 1.1 0.0 15.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#43

Isopropyl Acetate

Concen: 0.612 ug/l




RT: 6.354 min Scan# 864


Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27


Tgt Ion: 43 Resp: 2553
Ion Ratio Lower Upper
43 100
61 10.7 14.3 21.5#
87 3.2 9.5 14.3#





#44
Trichloroethene
Concen: 0.680 ug/l
RT: 7.135 min Scan# 990
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001




Tgt Ion:130 Resp: 1083
Ion Ratio Lower Upper
130 100
95 117.3 0.0 204.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#45
1,2-Dichloropropane
Concen: 0.633 ug/l
RT: 7.440 min Scan# 1042
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27



Tgt Ion: 63 Resp: 1059
Ion Ratio Lower Upper
63 100
65 36.1 24.1 36.1





#46

Dibromomethane

Concen: 0.667 ug/l

RT: 7.586 min Scan# 1106

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 93	Resp: 880
Ion Ratio	Lower Upper
93	100
95	84.5 65.6 98.4
174	88.5 68.2 102.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#47

Bromodichloromethane

Concen: 0.633 ug/l




RT: 7.818 min Scan# 1104


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 83	Resp: 1620
Ion Ratio	Lower Upper
83	100
85	57.3 50.5 75.7
127	0.0 6.5 9.7





#48

Methyl methacrylate

Concen: 0.577 ug/l

RT: 7.714 min Scan# 1087

Delta R.T. 0.024 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion:	41	Resp:	123
Ion	Ratio	Lower	Upper
41	100		
69	70.6	58.5	87.7
39	84.8	51.7	77.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#49

1,4-Dioxane

Concen: 11.703 ug/l




RT: 7.677 min Scan# 1081


Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:	88	Resp:	498
Ion	Ratio	Lower	Upper
88	100		
43	44.4	33.4	50.2
58	30.3	58.6	88.0





#50
Toluene-d8
Concen: 0.000 ug/l
Expected RT: 8.65 min



Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 98
Sig Exp Ratio
98 100
100 66.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#51
4-Methyl-2-Pentanone
Concen: 3.430 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27


Tgt Ion: 43 Resp: 9378
Ion Ratio Lower Upper
43 100
58 34.4 28.9 43.3





#52
Toluene
Concen: 0.672 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion: 92 Resp: 2684
Ion Ratio Lower Upper
92 100
91 165.7 136.6 205.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#53
t-1,3-Dichloropropene
Concen: 0.576 ug/l
RT: 8.988 min Scan# 1296
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 75 Resp: 1238
Ion Ratio Lower Upper
75 100
77 42.7 25.0 37.6#





#54

cis-1,3-Dichloropropene

Concen: 0.564 ug/l

RT: 8.379 min Scan# 1196

Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 75	Resp: 1414
Ion Ratio	Lower Upper
75 100	
77 44.2	25.4 38.0
39 59.0	52.2 78.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#55

1,1,2-Trichloroethane

Concen: 0.637 ug/l




RT: 9.153 min Scan# 1323


Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 97	Resp: 1029
Ion Ratio	Lower Upper
97 100	
83 74.8	69.0 103.6
85 59.5	44.3 66.5
99 63.0	47.8 71.6





#56

Ethyl methacrylate

Concen: 0.504 ug/l

RT: 9.128 min Scan# 1319

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 69 Resp: 1260

Ion Ratio Lower Upper

69 100

41 86.6 60.8 91.2





39 78.0 39.0 58.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#57

1,3-Dichloropropane

Concen: 0.711 ug/l

RT: 9.311 min Scan# 1349

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 76 Resp: 2036

Ion Ratio Lower Upper

76 100

78 31.4 26.3 39.5









#58
2-Chloroethyl Vinyl ether
Concen: 3.420 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 63 Resp: 3840
Ion Ratio Lower Upper
63 100
106 24.0 21.5 32.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#59
2-Hexanone
Concen: 3.096 ug/l
RT: 9.439 min Scan# 1370
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 43 Resp: 6424
Ion Ratio Lower Upper
43 100
58 41.6 24.9 74.6





#60

Dibromochloromethane

Concen: 0.582 ug/l

RT: 9.525 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:129 Resp: 102

Ion Ratio Lower Upper

129 100





127 72.5 39.3 117.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#61

1,2-Dibromoethane

Concen: 0.650 ug/l

RT: 9.610 min Scan# 1398

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion:107 Resp: 1077

Ion Ratio Lower Upper

107 100

109 94.0 73.3 109.9





#62
4-Bromofluorobenzene
Concen: 0.000 ug/l
Expected RT: 11.08 min



Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 95
Sig Exp Ratio
95 100
174 67.9
176 65.7




Manual Integrations APPROVED


Reviewed By : John Carlone 05/06/2025
Supervised By : Mahesh Dadoda 05/06/2025



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27


Tgt Ion: 117 Resp: 140108
Ion Ratio Lower Upper
117 100
82 61.0 50.6 76.0
119 32.1 25.8 38.6





#64
Tetrachloroethene
Concen: 0.679 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001




Tgt Ion:164 Resp: 972
Ion Ratio Lower Upper
164 100
166 119.1 105.0 157.6
129 103.4 83.5 125.3
131 95.4 76.5 114.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#65
Chlorobenzene
Concen: 0.735 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Tgt Ion:112 Resp: 3170
Ion Ratio Lower Upper
112 100
114 30.6 25.4 38.2





#66
1,1,1,2-Tetrachloroethane
Concen: 0.720 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion:131 Resp: 1035
Ion Ratio Lower Upper
131 100
133 94.9 47.3 141.9
119 0.0 31.6 95.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#67
Ethyl Benzene
Concen: 0.687 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 91 Resp: 5052
Ion Ratio Lower Upper
91 100
106 32.5 23.4 35.2









#68
m/p-Xylenes
Concen: 1.367 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion:106 Resp: 3630
Ion Ratio Lower Upper
106 100
91 209.8 171.2 256.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#69
o-Xylene
Concen: 0.672 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:106 Resp: 1799
Ion Ratio Lower Upper
106 100
91 214.9 112.7 338.1









#70
Styrene
Concen: 0.622 ug/l
RT: 10.659 min Scan# 1569
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001

Tgt Ion:104 Resp: 2664
Ion Ratio Lower Upper
104 100
78 54.3 45.7 68.5
103 53.9 43.7 65.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#71
Bromoform
Concen: 0.608 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27


Tgt Ion:173 Resp: 657
Ion Ratio Lower Upper
173 100
175 46.6 23.4 70.0
254 0.0 0.0 0.0





#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001



Tgt Ion:152 Resp: 5912
Ion Ratio Lower Upper
152 100
115 66.6 46.9 140.7
150 156.3 0.0 351.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#73
Isopropylbenzene
Concen: 0.695 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27


Tgt Ion:105 Resp: 4480
Ion Ratio Lower Upper
105 100
120 23.7 12.8 38.4





#74
N-amy1 acetate
Concen: 0.631 ug/l
RT: 10.854 min Scan# 1602
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion: 43 Resp: 2028
Ion Ratio Lower Upper
43 100
70 31.5 30.9 46.3
55 24.5 18.7 28.1
61 15.4 17.1 25.7#




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#75
1,1,2,2-Tetrachloroethane
Concen: 0.800 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 83 Resp: 1835
Ion Ratio Lower Upper
83 100
131 7.5 5.0 14.9
85 58.7 31.9 95.7









#76
1,2,3-Trichloropropane
Concen: 0.674 ug/l m
RT: 11.244 min Scan# 1665
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 75 Resp: 1665
Ion Ratio Lower Upper
75 100
77 29.7 20.5 61.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#77
Bromobenzene
Concen: 0.734 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:156 Resp: 1095
Ion Ratio Lower Upper
156 100
77 179.5 92.8 278.3
158 93.5 48.8 146.3





#78

n-propylbenzene

Concen: 0.699 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 91 Resp: 5052

Ion Ratio Lower Upper

91 100





120 20.1 10.8 32.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#79

2-Chlorotoluene

Concen: 0.773 ug/l

RT: 11.360 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 91 Resp: 3765

Ion Ratio Lower Upper

91 100

126 28.0 15.6 46.7





#80

1,3,5-Trimethylbenzene

Concen: 0.671 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:105 Resp: 3590

Ion Ratio Lower Upper

105 100





120 43.4 23.1 69.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#82

4-Chlorotoluene

Concen: 0.713 ug/l

RT: 11.457 min Scan# 1701

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 91 Resp: 3815

Ion Ratio Lower Upper

91 100

126 24.8 13.3 39.8





#83

tert-Butylbenzene

Concen: 0.751 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:	119	Resp:	395
Ion	Ratio	Lower	Upper
119	100		
91	63.9	32.9	98.7
134	25.2	11.4	34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#84

1,2,4-Trimethylbenzene

Concen: 0.697 ug/l




RT: 11.756 min Scan# 1750


Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:	105	Resp:	3725
Ion	Ratio	Lower	Upper
105	100		
120	40.3	21.2	63.6





#85

sec-Butylbenzene

Concen: 0.672 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:105 Resp: 438

Ion Ratio Lower Upper

105 100





134 19.2 9.7 29.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#86

p-Isopropyltoluene

Concen: 0.679 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion:119 Resp: 3577


Ion Ratio Lower Upper

119 100

134 25.0 12.5 37.5

91 33.1 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 0.689 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:146	Resp:	1914
Ion Ratio	Lower	Upper
146	100	
111	43.9	22.1 66.3
148	65.3	32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 0.789 ug/l m




RT: 12.036 min Scan# 1796


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:146	Resp:	2149
Ion Ratio	Lower	Upper
146	100	
111	39.1	21.3 63.9
148	58.2	31.9 95.5





#89

n-Butylbenzene

Concen: 0.631 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001





Tgt Ion: 91	Resp: 288
Ion Ratio	Lower Upper
91	100
92	48.1 26.9 80.7
134	20.1 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 0.616 ug/l




RT: 12.536 min Scan# 1878


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27


Tgt Ion: 117	Resp: 604
Ion Ratio	Lower Upper
117	100
201	64.2 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 0.747 ug/l
RT: 12.341 min Scan# 1845
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001




Tgt Ion:146 Resp: 202
Ion Ratio Lower Upper
146 100
111 48.8 22.4 67.3
148 59.4 31.6 94.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 0.586 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27



Tgt Ion: 75 Resp: 306
Ion Ratio Lower Upper
75 100
155 49.3 34.9 104.8
157 78.1 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 0.688 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:180	Resp:	1019
Ion Ratio	Lower	Upper
180	100	
182	100.6	47.9 143.8
145	29.5	17.1 51.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#94

Hexachlorobutadiene

Concen: 0.671 ug/l




RT: 13.725 min Scan# 2073


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27


Tgt Ion:225	Resp:	465
Ion Ratio	Lower	Upper
225	100	
223	69.7	30.8 92.4
227	70.5	30.9 92.7





#95
Naphthalene
Concen: 0.767 ug/l
RT: 13.780 min Scan# 2081
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001




Tgt Ion:128 Resp: 4138
Ion Ratio Lower Upper
128 100
127 12.4 10.4 15.6
129 8.9 8.6 13.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 0.716 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27



Tgt Ion:180 Resp: 1113
Ion Ratio Lower Upper
180 100
182 92.5 47.8 143.3
145 32.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	92588	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	162651	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	145078	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	68229	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83733	48.509	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.020%		
35) Dibromofluoromethane	5.379	113	58192	49.683	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.360%		
50) Toluene-d8	8.647	98	198575	48.984	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.960%		
62) 4-Bromofluorobenzene	11.079	95	75568	48.596	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	97.200%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.167	85	78783	55.594	ug/l	99
3) Chloromethane	1.307	50	72275	52.592	ug/l	100
4) Vinyl Chloride	1.374	62	65994	51.598	ug/l	98
5) Bromomethane	1.593	94	30033	50.627	ug/l	97
6) Chloroethane	1.666	64	35690	52.271	ug/l	94
7) Trichlorofluoromethane	1.874	101	99352	52.560	ug/l	99
8) Diethyl Ether	2.130	74	31875	49.536	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.319	101	59451	50.822	ug/l	99
10) Methyl Iodide	2.447	142	73940	53.419	ug/l	99
11) Tert butyl alcohol	2.977	59	61065	252.025	ug/l	99
12) 1,1-Dichloroethene	2.313	96	56097	51.097	ug/l	98
13) Acrolein	2.233	56	64157	232.505	ug/l	99
14) Allyl chloride	2.660	41	108757	51.833	ug/l	100
15) Acrylonitrile	3.063	53	181808	262.412	ug/l	99
16) Acetone	2.380	43	169923	245.518	ug/l	98
17) Carbon Disulfide	2.502	76	132695	50.981	ug/l	99
18) Methyl Acetate	2.703	43	80251	49.968	ug/l	97
19) Methyl tert-butyl Ether	3.111	73	195442	50.777	ug/l	99
20) Methylene Chloride	2.782	84	63369	47.779	ug/l	99
21) trans-1,2-Dichloroethene	3.087	96	55770	50.513	ug/l	99
22) Diisopropyl ether	3.757	45	209946	51.800	ug/l	97
23) Vinyl Acetate	3.715	43	946746	265.585	ug/l	100
24) 1,1-Dichloroethane	3.605	63	114733	50.825	ug/l	98
25) 2-Butanone	4.556	43	258339	257.110	ug/l	97
26) 2,2-Dichloropropane	4.471	77	88229	49.934	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	67727	50.956	ug/l	100
28) Bromochloromethane	4.891	49	55390	50.975	ug/l	98
29) Tetrahydrofuran	5.001	42	163675	259.958	ug/l	99
30) Chloroform	5.093	83	120782	51.332	ug/l	97
31) Cyclohexane	5.458	56	103158	50.147	ug/l	96
32) 1,1,1-Trichloroethane	5.373	97	104243	51.108	ug/l	100
36) 1,1-Dichloropropene	5.690	75	77931	49.521	ug/l	99
37) Ethyl Acetate	4.715	43	98604	50.714	ug/l	100
38) Carbon Tetrachloride	5.666	117	88100	49.825	ug/l	97
39) Methylcyclohexane	7.373	83	100808	49.757	ug/l	98
40) Benzene	6.031	78	233977	50.759	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	56017	55.076	ug/l	100
42) 1,2-Dichloroethane	6.080	62	101185	50.861	ug/l	98
43) Isopropyl Acetate	6.336	43	154818	52.195	ug/l	99
44) Trichloroethene	7.123	130	55491	50.018	ug/l	100
45) 1,2-Dichloropropane	7.428	63	59582	51.983	ug/l	98
46) Dibromomethane	7.580	93	45181	49.978	ug/l	99
47) Bromodichloromethane	7.818	83	92388	51.889	ug/l	100
48) Methyl methacrylate	7.690	41	82144	54.226	ug/l	98
49) 1,4-Dioxane	7.659	88	30477	1059.599	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	510206	259.133	ug/l	100
52) Toluene	8.714	92	144574	51.150	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	83751	52.920	ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	92517	52.892	ug/l	98
55) 1,1,2-Trichloroethane	9.147	97	56713	50.887	ug/l	96
56) Ethyl methacrylate	9.116	69	95772	53.918	ug/l	99
57) 1,3-Dichloropropane	9.305	76	99663	49.793	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.238	63	240681	265.780	ug/l	99
59) 2-Hexanone	9.427	43	390701	268.218	ug/l	100
60) Dibromochloromethane	9.519	129	63738	52.075	ug/l	99
61) 1,2-Dibromoethane	9.610	107	59134	51.050	ug/l	98
64) Tetrachloroethene	9.269	164	48034	46.795	ug/l	98
65) Chlorobenzene	10.073	112	154640	48.699	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	53735	49.557	ug/l	100
67) Ethyl Benzene	10.189	91	281239	50.245	ug/l	100
68) m/p-Xylenes	10.299	106	206813	101.022	ug/l	99
69) o-Xylene	10.640	106	101469	50.841	ug/l	99
70) Styrene	10.653	104	169703	51.907	ug/l	99
71) Bromoform	10.799	173	40364	49.583	ug/l #	97
73) Isopropylbenzene	10.957	105	269484	50.733	ug/l	100
74) N-amyl acetate	10.842	43	137559	52.407	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	89027	47.827	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	79763m	48.568	ug/l	
77) Bromobenzene	11.195	156	59871	48.549	ug/l	98
78) n-propylbenzene	11.299	91	312285	50.562	ug/l	100
79) 2-Chlorotoluene	11.360	91	195630	49.107	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	228538	51.500	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	24857	49.276	ug/l	98
82) 4-Chlorotoluene	11.451	91	223811	50.661	ug/l	100
83) tert-Butylbenzene	11.713	119	225783	50.511	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	231371	51.485	ug/l	100
85) sec-Butylbenzene	11.890	105	280578	51.122	ug/l	99
86) p-Isopropyltoluene	12.006	119	231942	51.198	ug/l	100
87) 1,3-Dichlorobenzene	11.963	146	113146	50.272	ug/l	99
88) 1,4-Dichlorobenzene	12.037	146	111173	48.368	ug/l	99
89) n-Butylbenzene	12.329	91	205858	51.804	ug/l	100
90) Hexachloroethane	12.536	117	39823	49.895	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	111482	49.361	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21220	51.458	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	64421	49.662	ug/l	100
94) Hexachlorobutadiene	13.719	225	27425	48.408	ug/l	98
95) Naphthalene	13.774	128	239805	50.404	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	64950	48.525	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration



#1
Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.544 min Scan# 731

Delta R.T. -0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion:168 Resp: 9258

Ion Ratio Lower Upper

168 100





99 74.5 54.9 82.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#2

Dichlorodifluoromethane

Concen: 55.594 ug/l

RT: 1.167 min Scan# 13

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 85 Resp: 78783

Ion Ratio Lower Upper

85 100

87 32.2 15.7 47.1









#3
Chloromethane
Concen: 52.592 ug/l
RT: 1.307 min Scan# 30
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion: 50 Resp: 72275
Ion Ratio Lower Upper
50 100
52 31.8 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 51.598 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion: 62 Resp: 65994
Ion Ratio Lower Upper
62 100
64 32.3 25.2 37.8





#5
Bromomethane
Concen: 50.627 ug/l
RT: 1.593 min Scan# 83
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVX050525



Tgt Ion: 94 Resp: 3003
Ion Ratio Lower Upper
94 100
96 93.5 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 52.271 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion: 64 Resp: 35690
Ion Ratio Lower Upper
64 100
66 33.7 24.3 36.5





#7
Trichlorofluoromethane
Concen: 52.560 ug/l
RT: 1.874 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVX050525



Tgt Ion:101 Resp: 99352
Ion Ratio Lower Upper
101 100
103 62.6 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 49.536 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 74 Resp: 31875
Ion Ratio Lower Upper
74 100
45 107.1 54.9 164.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 50.822 ug/l

RT: 2.319 min Scan# 202

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525





Tgt Ion:101	Resp:	5945
Ion Ratio	Lower	Upper
101	100	
85	47.8	38.6 58.0
151	69.4	55.2 82.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#10

Methyl Iodide

Concen: 53.419 ug/l




RT: 2.447 min Scan# 223


Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:142	Resp:	73940
Ion Ratio	Lower	Upper
142	100	
127	52.9	41.7 62.5
141	14.4	11.5 17.3





#11

Tert butyl alcohol

Concen: 252.025 ug/l

RT: 2.977 min Scan# 310

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 59 Resp: 6106

Ion Ratio Lower Upper

59 100





57 11.1 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 51.097 ug/l

RT: 2.313 min Scan# 201

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 96 Resp: 56097


Ion Ratio Lower Upper

96 100

61 179.7 146.2 219.2

98 62.1 51.0 76.6





#13

Acrolein

Concen: 232.505 ug/l

RT: 2.233 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 56 Resp: 6415

Ion Ratio Lower Upper

56 100





55 70.8 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 51.833 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 41 Resp: 108757


Ion Ratio Lower Upper

41 100

39 75.6 60.6 90.8

76 31.1 24.9 37.3





#15

Acrylonitrile

Concen: 262.412 ug/l

RT: 3.063 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 53 Resp: 181808

Ion Ratio Lower Upper

53 100

52 82.9 65.3 97.9





51 36.5 29.8 44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 245.518 ug/l

RT: 2.380 min Scan# 212

Delta R.T. -0.006 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 43 Resp: 169923

Ion Ratio Lower Upper

43 100

58 27.3 21.2 31.8





#17

Carbon Disulfide

Concen: 50.981 ug/l

RT: 2.502 min Scan# 21

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 76 Resp: 13269

Ion Ratio Lower Upper

76 100





78 9.1 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 49.968 ug/l

RT: 2.703 min Scan# 265

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 43 Resp: 80251

Ion Ratio Lower Upper

43 100

74 22.1 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 50.777 ug/l

RT: 3.111 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 73 Resp: 19544

Ion Ratio Lower Upper

73 100





57 21.8 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#20

Methylene Chloride

Concen: 47.779 ug/l

RT: 2.782 min Scan# 278

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion: 84 Resp: 63369




Ion Ratio Lower Upper


84 100

49 142.7 113.9 170.9

51 43.3 33.5 50.3


86 66.8 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 50.513 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion: 96 Resp: 55770
Ion Ratio Lower Upper
96 100
61 149.3 119.5 179.3
98 64.6 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 51.800 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 45 Resp: 209946
Ion Ratio Lower Upper
45 100
43 86.9 66.6 100.0
87 24.8 19.8 29.6
59 10.8 8.6 12.8









#23
Vinyl Acetate
Concen: 265.585 ug/l
RT: 3.715 min Scan# 411
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVX050525

Tgt Ion: 43 Resp: 946740
Ion Ratio Lower Upper
43 100
86 9.4 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#24
1,1-Dichloroethane
Concen: 50.825 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 63 Resp: 114733
Ion Ratio Lower Upper
63 100
98 6.4 3.6 10.8
100 4.8 2.1 6.3





#25

2-Butanone

Concen: 257.110 ug/l

RT: 4.556 min Scan# 50

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 43 Resp: 258339

Ion Ratio Lower Upper

43 100





72 24.3 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 49.934 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 77 Resp: 88229

Ion Ratio Lower Upper

77 100

97 21.0 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 50.956 ug/l

RT: 4.483 min Scan# 51

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 96 Resp: 6772

Ion Ratio Lower Upper

96 100

61 160.7 0.0 322.8





98 64.4 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 50.975 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 49 Resp: 55390


Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 68.7 56.2 84.2





#29

Tetrahydrofuran

Concen: 259.958 ug/l

RT: 5.001 min Scan# 642

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 42 Resp: 163679

Ion Ratio Lower Upper

42 100

72 42.0 33.7 50.5





71 38.5 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 51.332 ug/l

RT: 5.093 min Scan# 657

Delta R.T. 0.006 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 83 Resp: 120782

Ion Ratio Lower Upper

83 100


85 63.6 49.3 73.9





#31
Cyclohexane
Concen: 50.147 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion: 56 Resp: 103158
Ion Ratio Lower Upper
56 100
69 27.8 24.4 36.6
84 80.5 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#32
1,1,1-Trichloroethane
Concen: 51.108 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 97 Resp: 104243
Ion Ratio Lower Upper
97 100
99 64.5 51.8 77.6
61 50.2 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 48.509 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 65 Resp: 8373

Ion Ratio Lower Upper

65 100





67 50.0 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion:114 Resp: 162651


Ion Ratio Lower Upper

114 100

63 23.1 0.0 49.2

88 17.2 0.0 33.6





#35

Dibromofluoromethane

Concen: 49.683 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525





Tgt Ion:	113	Resp:	5819
Ion Ratio	Lower	Upper	
113	100		
111	102.1	83.1	124.7
192	16.2	13.3	19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 49.521 ug/l




RT: 5.690 min Scan# 755


Delta R.T. 0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	75	Resp:	77931
Ion Ratio	Lower	Upper	
75	100		
110	32.7	16.3	48.9
77	31.1	24.3	36.5





#37

Ethyl Acetate

Concen: 50.714 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525





Tgt Ion:	43	Resp:	98604
Ion Ratio	Lower	Upper	
43	100		
61	12.6	10.3	15.5
70	10.0	7.9	11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 49.825 ug/l




RT: 5.666 min Scan# 751


Delta R.T. -0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	117	Resp:	88100
Ion Ratio	Lower	Upper	
117	100		
119	90.7	75.2	112.8
121	29.4	24.2	36.4





#39

Methylcyclohexane

Concen: 49.757 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 83 Resp: 100808

Ion Ratio Lower Upper

83 100

55 82.7 64.7 97.1





98 48.3 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 50.759 ug/l

RT: 6.031 min Scan# 811

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 78 Resp: 233977

Ion Ratio Lower Upper

78 100


77 24.0 19.0 28.4





#41
Methacrylonitrile
Concen: 55.076 ug/l
RT: 4.916 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVX050525



Tgt Ion: 41 Resp: 5601
Ion Ratio Lower Upper
41 100
39 58.7 47.2 70.8
67 63.7 50.7 76.1
52 30.5 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 50.861 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 62 Resp: 101185
Ion Ratio Lower Upper
62 100
98 7.1 0.0 15.2





#43

Isopropyl Acetate

Concen: 52.195 ug/l

RT: 6.336 min Scan# 861

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 43 Resp: 154818

Ion Ratio Lower Upper

43 100

61 17.9 14.3 21.5





87 11.4 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 50.018 ug/l

RT: 7.123 min Scan# 990

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:130 Resp: 55491

Ion Ratio Lower Upper

130 100

95 102.6 0.0 204.2





#45

1,2-Dichloropropane

Concen: 51.983 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 63 Resp: 5958

Ion Ratio Lower Upper

63 100





65 31.4 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 49.978 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 93 Resp: 45181


Ion Ratio Lower Upper

93 100

95 83.0 65.6 98.4

174 84.8 68.2 102.2





#47

Bromodichloromethane

Concen: 51.889 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525





Tgt Ion:	83	Resp:	9238
Ion	Ratio	Lower	Upper
83	100		
85	63.0	50.5	75.7
127	8.8	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 54.226 ug/l




RT: 7.690 min Scan# 1083


Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	41	Resp:	82144
Ion	Ratio	Lower	Upper
41	100		
69	70.4	58.5	87.7
39	64.2	51.7	77.5





#49

1,4-Dioxane

Concen: 1059.599 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion:	88	Resp:	3047
Ion	Ratio	Lower	Upper
88	100		
43	39.9	33.4	50.2
58	76.2	58.6	88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 48.984 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	98	Resp:	198575
Ion	Ratio	Lower	Upper
98	100		
100	65.6	53.5	80.3









#51
4-Methyl-2-Pentanone
Concen: 259.133 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion: 43 Resp: 510200
Ion Ratio Lower Upper
43 100
58 36.1 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 51.150 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 92 Resp: 144574
Ion Ratio Lower Upper
92 100
91 171.4 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 52.920 ug/l

RT: 8.976 min Scan# 1294

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 75 Resp: 8375

Ion Ratio Lower Upper

75 100





77 31.3 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 52.892 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 75 Resp: 92517


Ion Ratio Lower Upper

75 100

77 31.5 25.4 38.0

39 63.2 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 50.887 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 97 Resp: 5671

Ion Ratio Lower Upper

97 100

83 89.8 69.0 103.6

85 58.4 44.3 66.5





99 63.9 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 53.918 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 69 Resp: 95772


Ion Ratio Lower Upper

69 100

41 76.1 60.8 91.2

39 47.4 39.0 58.6





#57

1,3-Dichloropropane

Concen: 49.793 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 76 Resp: 9966

Ion Ratio Lower Upper

76 100





78 31.4 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 265.780 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 63 Resp: 240681

Ion Ratio Lower Upper

63 100

106 26.6 21.5 32.3





#59

2-Hexanone

Concen: 268.218 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 43 Resp: 39070

Ion Ratio Lower Upper

43 100





58 49.6 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 52.075 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:129 Resp: 63738

Ion Ratio Lower Upper

129 100

127 77.9 39.3 117.8









#61
1,2-Dibromoethane
Concen: 51.050 ug/l
RT: 9.610 min Scan# 1397
Delta R.T. 0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVX050525

Tgt Ion:107 Resp: 59134
Ion Ratio Lower Upper
107 100
109 93.8 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#62
4-Bromofluorobenzene
Concen: 48.596 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion: 95 Resp: 75568
Ion Ratio Lower Upper
95 100
174 68.2 0.0 135.8
176 66.5 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:117 Resp: 145078
Ion Ratio Lower Upper
117 100
82 62.1 50.6 76.0
119 32.9 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 46.795 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion:164 Resp: 48034
Ion Ratio Lower Upper
164 100
166 129.9 105.0 157.6
129 105.3 83.5 125.3
131 98.8 76.5 114.7





#65
Chlorobenzene
Concen: 48.699 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:112 Resp: 154640
Ion Ratio Lower Upper
112 100
114 31.1 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 49.557 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:131 Resp: 53735
Ion Ratio Lower Upper
131 100
133 94.9 47.3 141.9
119 63.7 31.6 95.0





#67

Ethyl Benzene

Concen: 50.245 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 91 Resp: 281239

Ion Ratio Lower Upper

91 100





106 29.1 23.4 35.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#68

m/p-Xylenes

Concen: 101.022 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:106 Resp: 206813

Ion Ratio Lower Upper

106 100

91 215.6 171.2 256.8









#69
o-Xylene
Concen: 50.841 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion:106 Resp: 101469
Ion Ratio Lower Upper
106 100
91 227.7 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#70
Styrene
Concen: 51.907 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion:104 Resp: 169703
Ion Ratio Lower Upper
104 100
78 57.6 45.7 68.5
103 55.5 43.7 65.5





#71
Bromoform
Concen: 49.583 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:173 Resp: 40364
Ion Ratio Lower Upper
173 100
175 48.8 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carbone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion:152 Resp: 68229
Ion Ratio Lower Upper
152 100
115 93.3 46.9 140.7
150 175.3 0.0 351.0





#73
Isopropylbenzene
Concen: 50.733 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:105 Resp: 269484
Ion Ratio Lower Upper
105 100
120 25.4 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 52.407 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 43 Resp: 137559
Ion Ratio Lower Upper
43 100
70 37.9 30.9 46.3
55 22.7 18.7 28.1
61 21.8 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 47.827 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion: 83 Resp: 8902
Ion Ratio Lower Upper
83 100
131 9.7 5.0 14.9
85 65.5 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#76
1,2,3-Trichloropropane
Concen: 48.568 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 75 Resp: 79763
Ion Ratio Lower Upper
75 100
77 40.8 20.5 61.5





#77

Bromobenzene

Concen: 48.549 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion:	156	Resp:	5987
Ion	Ratio	Lower	Upper
156	100		
77	189.0	92.8	278.3
158	98.9	48.8	146.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#78

n-propylbenzene

Concen: 50.562 ug/l




RT: 11.299 min Scan# 1675


Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	91	Resp:	312285
Ion	Ratio	Lower	Upper
91	100		
120	21.4	10.8	32.4





#79

2-Chlorotoluene

Concen: 49.107 ug/l

RT: 11.360 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 91 Resp: 195630

Ion Ratio Lower Upper

91 100





126 30.8 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#80

1,3,5-Trimethylbenzene

Concen: 51.500 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:105 Resp: 228538

Ion Ratio Lower Upper

105 100

120 46.3 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 49.276 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion: 75 Resp: 2485
Ion Ratio Lower Upper
75 100
53 122.3 100.5 150.7
89 46.2 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 50.661 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 91 Resp: 223811
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83

tert-Butylbenzene

Concen: 50.511 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion:119 Resp: 22578

Ion Ratio Lower Upper

119 100

91 64.9 32.9 98.7





134 22.6 11.4 34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#84

1,2,4-Trimethylbenzene

Concen: 51.485 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:105 Resp: 231371

Ion Ratio Lower Upper

105 100


120 42.3 21.2 63.6





#85
sec-Butylbenzene
Concen: 51.122 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:105 Resp: 280578
Ion Ratio Lower Upper
105 100
134 19.1 9.7 29.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#86
p-Isopropyltoluene
Concen: 51.198 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:119 Resp: 231942
Ion Ratio Lower Upper
119 100
134 25.0 12.5 37.5
91 27.5 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 50.272 ug/l

RT: 11.963 min Scan# 1784

Delta R.T. -0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion:146 Resp: 113140

Ion Ratio Lower Upper

146 100

111 44.1 22.1 66.3





148 63.5 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 48.368 ug/l

RT: 12.037 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion:146 Resp: 111173


Ion Ratio Lower Upper

146 100

111 43.9 21.3 63.9

148 63.3 31.9 95.5





#89

n-Butylbenzene

Concen: 51.804 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 91 Resp: 20585

Ion Ratio Lower Upper

91 100

92 53.7 26.9 80.7





134 23.7 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 49.895 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:117 Resp: 39823

Ion Ratio Lower Upper

117 100


201 62.4 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 49.361 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:146 Resp: 11148
Ion Ratio Lower Upper
146 100
111 45.8 22.4 67.3
148 64.2 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 51.458 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion: 75 Resp: 21220
Ion Ratio Lower Upper
75 100
155 68.6 34.9 104.8
157 86.7 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 49.662 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:180 Resp: 6442
Ion Ratio Lower Upper
180 100
182 95.5 47.9 143.8
145 34.4 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#94
Hexachlorobutadiene
Concen: 48.408 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion:225 Resp: 27425
Ion Ratio Lower Upper
225 100
223 63.1 30.8 92.4
227 63.7 30.9 92.7





#95
Naphthalene
Concen: 50.404 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:128 Resp: 23980
Ion Ratio Lower Upper
128 100
127 13.0 10.4 15.6
129 10.7 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 48.525 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:180 Resp: 64950
Ion Ratio Lower Upper
180 100
182 97.0 47.8 143.3
145 37.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.765	0.851	-11.2	94	0.00
3 P	Chloromethane	0.742	0.781	-5.3	96	0.00
4 C	Vinyl Chloride	0.691	0.713	-3.2#	96	0.00
5 T	Bromomethane	0.320	0.324	-1.3	95	0.00
6 T	Chloroethane	0.369	0.385	-4.3	97	0.00
7 T	Trichlorofluoromethane	1.021	1.073	-5.1	96	0.00
8 T	Diethyl Ether	0.347	0.344	0.9	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.632	0.642	-1.6	96	0.00
10 T	Methyl Iodide	0.747	0.799	-7.0	94	0.00
11 T	Tert butyl alcohol	0.131	0.132	-0.8	98	0.00
12 CM	1,1-Dichloroethene	0.593	0.606	-2.2#	96	0.00
13 T	Acrolein	0.149	0.139	6.7	87	0.00
14 T	Allyl chloride	1.133	1.175	-3.7	95	0.00
15 T	Acrylonitrile	0.374	0.393	-5.1	97	0.00
16 T	Acetone	0.374	0.367	1.9	97	0.00
17 T	Carbon Disulfide	1.406	1.433	-1.9	94	0.00
18 T	Methyl Acetate	0.867	0.867	0.0	97	0.00
19 T	Methyl tert-butyl Ether	2.079	2.111	-1.5	93	0.00
20 T	Methylene Chloride	0.716	0.684	4.5	95	0.00
21 T	trans-1,2-Dichloroethene	0.596	0.602	-1.0	94	0.00
22 T	Diisopropyl ether	2.189	2.268	-3.6	95	0.00
23 T	Vinyl Acetate	1.925	2.045	-6.2	95	0.00
24 P	1,1-Dichloroethane	1.219	1.239	-1.6	94	0.00
25 T	2-Butanone	0.543	0.558	-2.8	96	0.00
26 T	2,2-Dichloropropane	0.954	0.953	0.1	95	0.00
27 T	cis-1,2-Dichloroethene	0.718	0.731	-1.8	95	0.00
28 T	Bromochloromethane	0.587	0.598	-1.9	99	0.00
29 T	Tetrahydrofuran	0.340	0.354	-4.1	96	0.00
30 C	Chloroform	1.271	1.305	-2.7#	96	0.00
31 T	Cyclohexane	1.111	1.114	-0.3	94	0.00
32 T	1,1,1-Trichloroethane	1.101	1.126	-2.3	95	0.00
33 S	1,2-Dichloroethane-d4	0.932	0.904	3.0	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
35 S	Dibromofluoromethane	0.360	0.358	0.6	98	0.00
36 T	1,1-Dichloropropene	0.484	0.479	1.0	94	0.00
37 T	Ethyl Acetate	0.598	0.606	-1.3	96	0.00
38 T	Carbon Tetrachloride	0.544	0.542	0.4	94	0.00
39 T	Methylcyclohexane	0.623	0.620	0.5	94	0.00
40 TM	Benzene	1.417	1.439	-1.6	95	0.00
41 T	Methacrylonitrile	0.313	0.344	-9.9	96	0.00
42 TM	1,2-Dichloroethane	0.612	0.622	-1.6	96	0.00
43 T	Isopropyl Acetate	0.912	0.952	-4.4	96	0.00
44 TM	Trichloroethene	0.341	0.341	0.0	93	0.00
45 C	1,2-Dichloropropane	0.352	0.366	-4.0#	96	0.00
46 T	Dibromomethane	0.278	0.278	0.0	94	0.00
47 T	Bromodichloromethane	0.547	0.568	-3.8	95	0.00
48 T	Methyl methacrylate	0.466	0.505	-8.4	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	100	0.00
50 S	Toluene-d8	1.246	1.221	2.0	97	0.00
51 T	4-Methyl-2-Pentanone	0.605	0.627	-3.6	96	0.00
52 CM	Toluene	0.869	0.889	-2.3#	96	0.00
53 T	t-1,3-Dichloropropene	0.487	0.515	-5.7	94	0.00
54 T	cis-1,3-Dichloropropene	0.538	0.569	-5.8	95	0.00
55 T	1,1,2-Trichloroethane	0.343	0.349	-1.7	95	0.00
56 T	Ethyl methacrylate	0.546	0.589	-7.9	96	0.00
57 T	1,3-Dichloropropane	0.615	0.613	0.3	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.278	0.296	-6.5	93	0.00
59 T	2-Hexanone	0.448	0.480	-7.1	98	0.00
60 T	Dibromochloromethane	0.376	0.392	-4.3	95	0.00
61 T	1,2-Dibromoethane	0.356	0.364	-2.2	94	0.00
62 S	4-Bromofluorobenzene	0.478	0.465	2.7	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.354	0.331	6.5	88	0.00
65 PM	Chlorobenzene	1.094	1.066	2.6	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.374	0.370	1.1	94	0.00
67 C	Ethyl Benzene	1.929	1.939	-0.5#	95	0.00
68 T	m/p-Xylenes	0.706	0.713	-1.0	95	0.00
69 T	o-Xylene	0.688	0.699	-1.6	95	0.00
70 T	Styrene	1.127	1.170	-3.8	95	0.00
71 P	Bromoform	0.281	0.278	1.1	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.893	3.950	-1.5	96	0.00
74 T	N-amyl acetate	1.924	2.016	-4.8	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.364	1.305	4.3	98	0.00
76 T	1,2,3-Trichloropropane	1.204	1.169	2.9	99	0.00
77 T	Bromobenzene	0.904	0.878	2.9	95	0.00
78 T	n-propylbenzene	4.526	4.577	-1.1	95	0.00
79 T	2-Chlorotoluene	2.919	2.867	1.8	96	0.00
80 T	1,3,5-Trimethylbenzene	3.252	3.350	-3.0	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.370	0.364	1.6	95	0.00
82 T	4-Chlorotoluene	3.238	3.280	-1.3	96	0.00
83 T	tert-Butylbenzene	3.276	3.309	-1.0	97	0.00
84 T	1,2,4-Trimethylbenzene	3.293	3.391	-3.0	97	0.00
85 T	sec-Butylbenzene	4.022	4.112	-2.2	96	0.00
86 T	p-Isopropyltoluene	3.320	3.399	-2.4	96	0.00
87 T	1,3-Dichlorobenzene	1.649	1.658	-0.5	98	0.00
88 T	1,4-Dichlorobenzene	1.684	1.629	3.3	97	0.00
89 T	n-Butylbenzene	2.912	3.017	-3.6	96	0.00
90 T	Hexachloroethane	0.585	0.584	0.2	94	0.00
91 T	1,2-Dichlorobenzene	1.655	1.634	1.3	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.302	0.311	-3.0	97	0.00
93 T	1,2,4-Trichlorobenzene	0.951	0.944	0.7	97	0.00
94 T	Hexachlorobutadiene	0.415	0.402	3.1	95	0.00
95 T	Naphthalene	3.487	3.515	-0.8	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.981	0.952	3.0	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\X050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	50.000	55.594	-11.2	94	0.00
3 P	Chloromethane	50.000	52.592	-5.2	96	0.00
4 C	Vinyl Chloride	50.000	51.598	-3.2#	96	0.00
5 T	Bromomethane	50.000	50.627	-1.3	95	0.00
6 T	Chloroethane	50.000	52.271	-4.5	97	0.00
7 T	Trichlorofluoromethane	50.000	52.560	-5.1	96	0.00
8 T	Diethyl Ether	50.000	49.536	0.9	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.822	-1.6	96	0.00
10 T	Methyl Iodide	50.000	53.419	-6.8	94	0.00
11 T	Tert butyl alcohol	250.000	252.025	-0.8	98	0.00
12 CM	1,1-Dichloroethene	50.000	51.097	-2.2#	96	0.00
13 T	Acrolein	250.000	232.505	7.0	87	0.00
14 T	Allyl chloride	50.000	51.833	-3.7	95	0.00
15 T	Acrylonitrile	250.000	262.412	-5.0	97	0.00
16 T	Acetone	250.000	245.518	1.8	97	0.00
17 T	Carbon Disulfide	50.000	50.981	-2.0	94	0.00
18 T	Methyl Acetate	50.000	49.968	0.1	97	0.00
19 T	Methyl tert-butyl Ether	50.000	50.777	-1.6	93	0.00
20 T	Methylene Chloride	50.000	47.779	4.4	95	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.513	-1.0	94	0.00
22 T	Diisopropyl ether	50.000	51.800	-3.6	95	0.00
23 T	Vinyl Acetate	250.000	265.585	-6.2	95	0.00
24 P	1,1-Dichloroethane	50.000	50.825	-1.7	94	0.00
25 T	2-Butanone	250.000	257.110	-2.8	96	0.00
26 T	2,2-Dichloropropane	50.000	49.934	0.1	95	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.956	-1.9	95	0.00
28 T	Bromochloromethane	50.000	50.975	-2.0	99	0.00
29 T	Tetrahydrofuran	250.000	259.958	-4.0	96	0.00
30 C	Chloroform	50.000	51.332	-2.7#	96	0.00
31 T	Cyclohexane	50.000	50.147	-0.3	94	0.00
32 T	1,1,1-Trichloroethane	50.000	51.108	-2.2	95	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.509	3.0	95	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	97	0.00
35 S	Dibromofluoromethane	50.000	49.683	0.6	98	0.00
36 T	1,1-Dichloropropene	50.000	49.521	1.0	94	0.00
37 T	Ethyl Acetate	50.000	50.714	-1.4	96	0.00
38 T	Carbon Tetrachloride	50.000	49.825	0.3	94	0.00
39 T	Methylcyclohexane	50.000	49.757	0.5	94	0.00
40 TM	Benzene	50.000	50.759	-1.5	95	0.00
41 T	Methacrylonitrile	50.000	55.076	-10.2	96	0.00
42 TM	1,2-Dichloroethane	50.000	50.861	-1.7	96	0.00
43 T	Isopropyl Acetate	50.000	52.195	-4.4	96	0.00
44 TM	Trichloroethene	50.000	50.018	-0.0	93	0.00
45 C	1,2-Dichloropropane	50.000	51.983	-4.0#	96	0.00
46 T	Dibromomethane	50.000	49.978	0.0	94	0.00
47 T	Bromodichloromethane	50.000	51.889	-3.8	95	0.00
48 T	Methyl methacrylate	50.000	54.226	-8.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1059.599	-6.0	100	0.00
50 S	Toluene-d8	50.000	48.984	2.0	97	0.00
51 T	4-Methyl-2-Pentanone	250.000	259.133	-3.7	96	0.00
52 CM	Toluene	50.000	51.150	-2.3#	96	0.00
53 T	t-1,3-Dichloropropene	50.000	52.920	-5.8	94	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.892	-5.8	95	0.00
55 T	1,1,2-Trichloroethane	50.000	50.887	-1.8	95	0.00
56 T	Ethyl methacrylate	50.000	53.918	-7.8	96	0.00
57 T	1,3-Dichloropropane	50.000	49.793	0.4	95	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	265.780	-6.3	93	0.00
59 T	2-Hexanone	250.000	268.218	-7.3	98	0.00
60 T	Dibromochloromethane	50.000	52.075	-4.2	95	0.00
61 T	1,2-Dibromoethane	50.000	51.050	-2.1	94	0.00
62 S	4-Bromofluorobenzene	50.000	48.596	2.8	96	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	99	0.00
64 T	Tetrachloroethene	50.000	46.795	6.4	88	0.00
65 PM	Chlorobenzene	50.000	48.699	2.6	96	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.557	0.9	94	0.00
67 C	Ethyl Benzene	50.000	50.245	-0.5#	95	0.00
68 T	m/p-Xylenes	100.000	101.022	-1.0	95	0.00
69 T	o-Xylene	50.000	50.841	-1.7	95	0.00
70 T	Styrene	50.000	51.907	-3.8	95	0.00
71 P	Bromoform	50.000	49.583	0.8	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	100	0.00
73 T	Isopropylbenzene	50.000	50.733	-1.5	96	0.00
74 T	N-amyl acetate	50.000	52.407	-4.8	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.827	4.3	98	0.00
76 T	1,2,3-Trichloropropane	50.000	48.568	2.9	99	0.00
77 T	Bromobenzene	50.000	48.549	2.9	95	0.00
78 T	n-propylbenzene	50.000	50.562	-1.1	95	0.00
79 T	2-Chlorotoluene	50.000	49.107	1.8	96	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.500	-3.0	96	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.276	1.4	95	0.00
82 T	4-Chlorotoluene	50.000	50.661	-1.3	96	0.00
83 T	tert-Butylbenzene	50.000	50.511	-1.0	97	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.485	-3.0	97	0.00
85 T	sec-Butylbenzene	50.000	51.122	-2.2	96	0.00
86 T	p-Isopropyltoluene	50.000	51.198	-2.4	96	0.00
87 T	1,3-Dichlorobenzene	50.000	50.272	-0.5	98	0.00
88 T	1,4-Dichlorobenzene	50.000	48.368	3.3	97	0.00
89 T	n-Butylbenzene	50.000	51.804	-3.6	96	0.00
90 T	Hexachloroethane	50.000	49.895	0.2	94	0.00
91 T	1,2-Dichlorobenzene	50.000	49.361	1.3	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.458	-2.9	97	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.662	0.7	97	0.00
94 T	Hexachlorobutadiene	50.000	48.408	3.2	95	0.00
95 T	Naphthalene	50.000	50.404	-0.8	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.525	3.0	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1985 SAS No.: Q1985 SDG No.: Q1985

Instrument ID: MSVOA_X Calibration Date/Time: 05/08/2025 10:05

Lab File ID: VX046078.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.742	0.707	0.1	-4.72	20
Vinyl Chloride	0.691	0.643		-6.95	20
Bromomethane	0.320	0.281		-12.19	20
Chloroethane	0.369	0.352		-4.61	20
Trichlorofluoromethane	1.021	0.997		-2.35	20
1,1,2-Trichlorotrifluoroethane	0.632	0.610		-3.48	20
1,1-Dichloroethene	0.593	0.554		-6.58	20
Acetone	0.374	0.352		-5.88	20
Carbon Disulfide	1.406	1.312		-6.69	20
Methyl tert-butyl Ether	2.079	1.974		-5.05	20
Methylene Chloride	0.716	0.617		-13.83	20
trans-1,2-Dichloroethene	0.596	0.560		-6.04	20
1,1-Dichloroethane	1.219	1.143	0.1	-6.24	20
2-Butanone	0.543	0.506		-6.81	20
Carbon Tetrachloride	0.544	0.519		-4.6	20
cis-1,2-Dichloroethene	0.718	0.664		-7.52	20
Chloroform	1.271	1.202		-5.43	20
1,1,1-Trichloroethane	1.101	1.037		-5.81	20
Methylcyclohexane	0.623	0.579		-7.06	20
Benzene	1.417	1.330		-6.14	20
1,2-Dichloroethane	0.612	0.565		-7.68	20
Trichloroethene	0.341	0.325		-4.69	20
1,2-Dichloropropane	0.352	0.336		-4.55	20
Bromodichloromethane	0.547	0.530		-3.11	20
4-Methyl-2-Pentanone	0.605	0.573		-5.29	20
Toluene	0.869	0.808		-7.02	20
t-1,3-Dichloropropene	0.487	0.478		-1.85	20
cis-1,3-Dichloropropene	0.538	0.529		-1.67	20
1,1,2-Trichloroethane	0.343	0.319		-7	20
2-Hexanone	0.448	0.423		-5.58	20
Dibromochloromethane	0.376	0.371		-1.33	20
Tetrachloroethene	0.354	0.343		-3.11	20
Chlorobenzene	1.094	0.988	0.3	-9.69	20
Ethyl Benzene	1.929	1.833		-4.98	20
m/p-Xylenes	0.706	0.665		-5.81	20
o-Xylene	0.688	0.646		-6.11	20
Styrene	1.127	1.099		-2.48	20
Bromoform	0.281	0.273	0.1	-2.85	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1985 SAS No.: Q1985 SDG No.: Q1985

Instrument ID: MSVOA_X Calibration Date/Time: 05/08/2025 10:05

Lab File ID: VX046078.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.893	3.770		-3.16	20
1,1,2,2-Tetrachloroethane	1.364	1.206	0.3	-11.58	20
1,3-Dichlorobenzene	1.649	1.576		-4.43	20
1,4-Dichlorobenzene	1.684	1.554		-7.72	20
1,2-Dichlorobenzene	1.655	1.544		-6.71	20
1,2-Dichloroethane-d4	0.932	0.825		-11.48	20
Dibromofluoromethane	0.360	0.332		-7.78	20
Toluene-d8	1.246	1.149		-7.78	20
4-Bromofluorobenzene	0.478	0.428		-10.46	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046078.D
 Acq On : 08 May 2025 10:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/09/2025
 Supervised By : Mahesh Dadoda 05/09/2025

Quant Time: May 09 01:54:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	105825	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	181991	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	158314	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	73554	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	87330	44.264	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	88.520%		
35) Dibromofluoromethane	5.373	113	60469	46.141	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.280%		
50) Toluene-d8	8.641	98	209081	46.095	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	92.180%		
62) 4-Bromofluorobenzene	11.079	95	77895	44.770	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	89.540%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.167	85	83116	51.315	ug/l	98
3) Chloromethane	1.307	50	74806	47.625	ug/l	99
4) Vinyl Chloride	1.374	62	67994	46.512	ug/l	99
5) Bromomethane	1.593	94	29775	43.914	ug/l	92
6) Chloroethane	1.666	64	37272	47.760	ug/l	96
7) Trichlorofluoromethane	1.874	101	105506	48.834	ug/l	98
8) Diethyl Ether	2.130	74	33259	45.221	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.319	101	64604	48.319	ug/l	98
10) Methyl Iodide	2.441	142	74153	46.872	ug/l	98
11) Tert butyl alcohol	2.971	59	62270	224.852	ug/l	99
12) 1,1-Dichloroethene	2.313	96	58611	46.709	ug/l	99
13) Acrolein	2.233	56	54033	171.322	ug/l	99
14) Allyl chloride	2.654	41	115880	48.320	ug/l	99
15) Acrylonitrile	3.056	53	182115	229.976	ug/l	97
16) Acetone	2.380	43	186025	235.163	ug/l	100
17) Carbon Disulfide	2.502	76	138794	46.655	ug/l	98
18) Methyl Acetate	2.703	43	84794	46.193	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	208870	47.478	ug/l	99
20) Methylene Chloride	2.782	84	65321	43.091	ug/l	98
21) trans-1,2-Dichloroethene	3.081	96	59244	46.948	ug/l	99
22) Diisopropyl ether	3.757	45	219654	47.416	ug/l	94
23) Vinyl Acetate	3.715	43	967835	237.541	ug/l	99
24) 1,1-Dichloroethane	3.599	63	120908	46.860	ug/l	98
25) 2-Butanone	4.550	43	267660	233.066	ug/l	99
26) 2,2-Dichloropropane	4.465	77	95561	47.319	ug/l	99
27) cis-1,2-Dichloroethene	4.477	96	70250	46.243	ug/l	99
28) Bromochloromethane	4.891	49	61656	49.644	ug/l	99
29) Tetrahydrofuran	4.995	42	165473	229.940	ug/l	99
30) Chloroform	5.080	83	127233	47.310	ug/l	99
31) Cyclohexane	5.458	56	109556	46.596	ug/l	97
32) 1,1,1-Trichloroethane	5.373	97	109783	47.092	ug/l	99
36) 1,1-Dichloropropene	5.684	75	81796	46.453	ug/l	100
37) Ethyl Acetate	4.702	43	99719	45.837	ug/l	100
38) Carbon Tetrachloride	5.666	117	94525	47.778	ug/l	94
39) Methylcyclohexane	7.373	83	105443	46.514	ug/l	94
40) Benzene	6.025	78	241966	46.914	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046078.D
 Acq On : 08 May 2025 10:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/09/2025
 Supervised By : Mahesh Dadoda 05/09/2025

Quant Time: May 09 01:54:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	57552	50.572	ug/l	98
42) 1,2-Dichloroethane	6.074	62	102778	46.172	ug/l	99
43) Isopropyl Acetate	6.330	43	158841	47.860	ug/l	99
44) Trichloroethene	7.117	130	59087	47.599	ug/l	98
45) 1,2-Dichloropropane	7.421	63	61126	47.662	ug/l	98
46) Dibromomethane	7.574	93	46530	46.001	ug/l	99
47) Bromodichloromethane	7.818	83	96407	48.392	ug/l	99
48) Methyl methacrylate	7.690	41	81622	48.156	ug/l	99
49) 1,4-Dioxane	7.653	88	30726	954.733	ug/l	99
51) 4-Methyl-2-Pentanone	8.568	43	521204	236.587	ug/l	99
52) Toluene	8.714	92	147100	46.513	ug/l	98
53) t-1,3-Dichloropropene	8.970	75	86965	49.111	ug/l	99
54) cis-1,3-Dichloropropene	8.360	75	96359	49.234	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	58084	46.578	ug/l	97
56) Ethyl methacrylate	9.110	69	96373	48.490	ug/l	99
57) 1,3-Dichloropropane	9.305	76	102289	45.674	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	198555	195.960	ug/l	100
59) 2-Hexanone	9.427	43	384624	235.986	ug/l	100
60) Dibromochloromethane	9.519	129	67444	49.247	ug/l	100
61) 1,2-Dibromoethane	9.604	107	59787	46.129	ug/l	97
64) Tetrachloroethene	9.269	164	54223	48.407	ug/l	97
65) Chlorobenzene	10.073	112	156475	45.157	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	55173	46.629	ug/l	98
67) Ethyl Benzene	10.189	91	290265	47.522	ug/l	99
68) m/p-Xylenes	10.299	106	210404	94.184	ug/l	98
69) o-Xylene	10.640	106	102240	46.944	ug/l	96
70) Styrene	10.653	104	174009	48.774	ug/l	99
71) Bromoform	10.799	173	43259	48.696	ug/l #	98
73) Isopropylbenzene	10.957	105	277319	48.428	ug/l	100
74) N-amyl acetate	10.842	43	135534	47.898	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	88732	44.217	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	80257m	45.331	ug/l	
77) Bromobenzene	11.195	156	62702	47.164	ug/l	97
78) n-propylbenzene	11.299	91	325019	48.814	ug/l	100
79) 2-Chlorotoluene	11.360	91	198350	46.185	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	233158	48.737	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.012	75	25533	46.952	ug/l	99
82) 4-Chlorotoluene	11.451	91	227761	47.823	ug/l	99
83) tert-Butylbenzene	11.713	119	228755	47.471	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	235628	48.637	ug/l	100
85) sec-Butylbenzene	11.890	105	289872	48.992	ug/l	100
86) p-Isopropyltoluene	12.006	119	241375	49.423	ug/l	100
87) 1,3-Dichlorobenzene	11.963	146	115894	47.766	ug/l	99
88) 1,4-Dichlorobenzene	12.037	146	114313	46.133	ug/l	99
89) n-Butylbenzene	12.329	91	217191	50.699	ug/l	99
90) Hexachloroethane	12.536	117	41883	48.677	ug/l	100
91) 1,2-Dichlorobenzene	12.329	146	113569	46.645	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21306	47.926	ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	68569	49.033	ug/l	99
94) Hexachlorobutadiene	13.719	225	29643	48.535	ug/l	97
95) Naphthalene	13.774	128	239114	46.620	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	68552	47.508	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046078.D
Acq On : 08 May 2025 10:05
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025

Quant Time: May 09 01:54:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046078.D
Acq On    : 08 May 2025 10:05
Operator  : JC/MD
Sample    : VSTDCCC050
Misc      : 5.0mL/MSVOA_X/WATER
ALS Vial  : 2    Sample Multiplier: 1
```


Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025


Quant Time: May 09 01:54:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:168 Resp: 105825
Ion Ratio Lower Upper
168 100
99 73.0 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#2
Dichlorodifluoromethane
Concen: 51.315 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 85 Resp: 83116
Ion Ratio Lower Upper
85 100
87 32.3 15.7 47.1









#3
Chloromethane
Concen: 47.625 ug/l
RT: 1.307 min Scan# 30
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 50 Resp: 74800
Ion Ratio Lower Upper
50 100
52 31.4 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#4
Vinyl Chloride
Concen: 46.512 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05


Tgt Ion: 62 Resp: 67994
Ion Ratio Lower Upper
62 100
64 32.1 25.2 37.8





#5
Bromomethane
Concen: 43.914 ug/l
RT: 1.593 min Scan# 83
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 94 Resp: 29775
Ion Ratio Lower Upper
94 100
96 98.4 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#6
Chloroethane
Concen: 47.760 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05


Tgt Ion: 64 Resp: 37272
Ion Ratio Lower Upper
64 100
66 32.8 24.3 36.5





#7
Trichlorofluoromethane
Concen: 48.834 ug/l
RT: 1.874 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:101 Resp: 105500
Ion Ratio Lower Upper
101 100
103 65.5 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#8
Diethyl Ether
Concen: 45.221 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 74 Resp: 33259
Ion Ratio Lower Upper
74 100
45 108.8 54.9 164.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 48.319 ug/l

RT: 2.319 min Scan# 20

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:101 Resp: 64604

Ion Ratio Lower Upper

101 100

85 46.8 38.6 58.0





151 70.0 55.2 82.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#10

Methyl Iodide

Concen: 46.872 ug/l

RT: 2.441 min Scan# 222

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05




Tgt Ion:142 Resp: 74153


Ion Ratio Lower Upper

142 100

127 53.4 41.7 62.5

141 15.1 11.5 17.3





#11

Tert butyl alcohol

Concen: 224.852 ug/l

RT: 2.971 min Scan# 309

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 59 Resp: 62270

Ion Ratio Lower Upper

59 100





57 10.9 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#12

1,1-Dichloroethene

Concen: 46.709 ug/l

RT: 2.313 min Scan# 201

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05




Tgt Ion: 96 Resp: 58611


Ion Ratio Lower Upper

96 100

61 180.8 146.2 219.2


98 64.1 51.0 76.6





#13
Acrolein
Concen: 171.322 ug/l
RT: 2.233 min Scan# 188
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 5403
Ion Ratio Lower Upper
56 100
55 71.3 56.2 84.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#14
Allyl chloride
Concen: 48.320 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05


Tgt Ion: 41 Resp: 115880
Ion Ratio Lower Upper
41 100
39 74.8 60.6 90.8
76 30.5 24.9 37.3





#15
Acrylonitrile
Concen: 229.976 ug/l
RT: 3.056 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 53 Resp: 18211
Ion Ratio Lower Upper
53 100
52 84.8 65.3 97.9
51 37.7 29.8 44.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#16
Acetone
Concen: 235.163 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05


Tgt Ion: 43 Resp: 186025
Ion Ratio Lower Upper
43 100
58 26.7 21.2 31.8





#17
Carbon Disulfide
Concen: 46.655 ug/l
RT: 2.502 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 76 Resp: 138794
Ion Ratio Lower Upper
76 100
78 9.5 7.0 10.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#18
Methyl Acetate
Concen: 46.193 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 43 Resp: 84794
Ion Ratio Lower Upper
43 100
74 21.9 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 47.478 ug/l

RT: 3.111 min Scan# 332

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 73 Resp: 208870

Ion Ratio Lower Upper

73 100





57 21.4 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#20

Methylene Chloride

Concen: 43.091 ug/l

RT: 2.782 min Scan# 278

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Tgt Ion: 84 Resp: 65321




Ion Ratio Lower Upper


84 100

49 140.7 113.9 170.9

51 41.9 33.5 50.3

86 64.9 53.8 80.8





#21

trans-1,2-Dichloroethene

Concen: 46.948 ug/l

RT: 3.081 min Scan# 31

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 96 Resp: 59244

Ion Ratio Lower Upper

96 100

61 148.0 119.5 179.3





98 60.7 50.0 75.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#22

Diisopropyl ether

Concen: 47.416 ug/l

RT: 3.757 min Scan# 438

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Tgt Ion: 45 Resp: 219654




Ion Ratio Lower Upper


45 100

43 75.5 66.6 100.0

87 24.8 19.8 29.6

59 10.9 8.6 12.8





#23

Vinyl Acetate

Concen: 237.541 ug/l

RT: 3.715 min Scan# 411

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 96783

Ion Ratio Lower Upper

43 100





86 9.8 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#24

1,1-Dichloroethane

Concen: 46.860 ug/l

RT: 3.599 min Scan# 412

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05




Tgt Ion: 63 Resp: 120908


Ion Ratio Lower Upper

63 100

98 6.5 3.6 10.8

100 3.9 2.1 6.3





#25

2-Butanone

Concen: 233.066 ug/l

RT: 4.550 min Scan# 50

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 267660

Ion Ratio Lower Upper

43 100





72 22.6 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#26

2,2-Dichloropropane

Concen: 47.319 ug/l

RT: 4.465 min Scan# 554

Delta R.T. 0.000 min

Lab File: VX046078.D




Acq: 08 May 2025 10:05


Tgt Ion: 77 Resp: 95561

Ion Ratio Lower Upper

77 100

97 21.3 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 46.243 ug/l

RT: 4.477 min Scan# 51

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 96 Resp: 70250

Ion Ratio Lower Upper

96 100

61 163.7 0.0 322.8





98 65.0 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#28

Bromochloromethane

Concen: 49.644 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05




Tgt Ion: 49 Resp: 61656


Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 69.1 56.2 84.2





#29

Tetrahydrofuran

Concen: 229.940 ug/l

RT: 4.995 min Scan# 641

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 42 Resp: 16547

Ion Ratio Lower Upper

42 100

72 41.6 33.7 50.5





71 38.8 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#30

Chloroform

Concen: 47.310 ug/l

RT: 5.080 min Scan# 655

Delta R.T. -0.006 min

Lab File: VX046078.D




Acq: 08 May 2025 10:05


Tgt Ion: 83 Resp: 127233

Ion Ratio Lower Upper

83 100


85 62.6 49.3 73.9





#31
Cyclohexane
Concen: 46.596 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 109550
Ion Ratio Lower Upper
56 100
69 29.9 24.4 36.6
84 80.8 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#32
1,1,1-Trichloroethane
Concen: 47.092 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 97 Resp: 109783
Ion Ratio Lower Upper
97 100
99 64.8 51.8 77.6
61 49.2 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 44.264 ug/l

RT: 5.946 min Scan# 79

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 87330

Ion Ratio Lower Upper

65 100





67 49.4 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.751 min Scan# 929

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05




Tgt Ion:114 Resp: 181991


Ion Ratio Lower Upper

114 100

63 22.6 0.0 49.2

88 17.6 0.0 33.6





#35

Dibromofluoromethane

Concen: 46.141 ug/l

RT: 5.373 min Scan# 704

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:113 Resp: 60469

Ion Ratio Lower Upper

113 100

111 103.0 83.1 124.7





192 16.8 13.3 19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#36

1,1-Dichloropropene

Concen: 46.453 ug/l

RT: 5.684 min Scan# 754

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05




Tgt Ion: 75 Resp: 81796


Ion Ratio Lower Upper

75 100

110 32.6 16.3 48.9

77 30.9 24.3 36.5





#37

Ethyl Acetate

Concen: 45.837 ug/l

RT: 4.702 min Scan# 593

Delta R.T. -0.012 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:	43	Resp:	99719
Ion Ratio	Lower	Upper	
43	100		
61	12.8	10.3	15.5
70	9.9	7.9	11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#38

Carbon Tetrachloride

Concen: 47.778 ug/l




RT: 5.666 min Scan# 751


Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Tgt Ion:	117	Resp:	94525
Ion Ratio	Lower	Upper	
117	100		
119	100.7	75.2	112.8
121	31.7	24.2	36.4





#39

Methylcyclohexane

Concen: 46.514 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 10544

Ion Ratio Lower Upper

83 100

55 87.1 64.7 97.1





98 50.2 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#40

Benzene

Concen: 46.914 ug/l

RT: 6.025 min Scan# 810

Delta R.T. -0.006 min

Lab File: VX046078.D




Acq: 08 May 2025 10:05


Tgt Ion: 78 Resp: 241966

Ion Ratio Lower Upper

78 100


77 24.4 19.0 28.4





#41
Methacrylonitrile
Concen: 50.572 ug/l
RT: 4.916 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 41 Resp: 57552
Ion Ratio Lower Upper
41 100
39 57.9 47.2 70.8
67 66.4 50.7 76.1
52 30.2 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#42
1,2-Dichloroethane
Concen: 46.172 ug/l
RT: 6.074 min Scan# 818
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 62 Resp: 102778
Ion Ratio Lower Upper
62 100
98 7.3 0.0 15.2





#43

Isopropyl Acetate

Concen: 47.860 ug/l

RT: 6.330 min Scan# 80

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 15884

Ion Ratio Lower Upper

43

100

61

17.2

87

11.4

14.3

9.5

21.5





14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#44

Trichloroethene

Concen: 47.599 ug/l

RT: 7.117 min Scan# 989

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Tgt Ion:130 Resp: 59087

Ion Ratio Lower Upper

130




100


95

104.3

0.0

204.2





#45

1,2-Dichloropropane

Concen: 47.662 ug/l

RT: 7.421 min Scan# 1039

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 61120

Ion Ratio Lower Upper

63 100





65 29.1 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#46

Dibromomethane

Concen: 46.001 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05




Tgt Ion: 93 Resp: 46530


Ion Ratio Lower Upper

93 100

95 83.1 65.6 98.4

174 85.7 68.2 102.2





#47

Bromodichloromethane

Concen: 48.392 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 9640

Ion Ratio Lower Upper

83 100

85 63.8 50.5 75.7





127 8.2 6.5 9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#48

Methyl methacrylate

Concen: 48.156 ug/l

RT: 7.690 min Scan# 1083

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05




Tgt Ion: 41 Resp: 81622


Ion Ratio Lower Upper

41 100

69 72.1 58.5 87.7

39 64.3 51.7 77.5





#49

1,4-Dioxane

Concen: 954.733 ug/l

RT: 7.653 min Scan# 110

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 88 Resp: 30720

Ion Ratio Lower Upper

88 100

43 41.0 33.4 50.2





58 74.6 58.6 88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#50

Toluene-d8

Concen: 46.095 ug/l

RT: 8.641 min Scan# 1239

Delta R.T. -0.006 min

Lab File: VX046078.D




Acq: 08 May 2025 10:05


Tgt Ion: 98 Resp: 209081

Ion Ratio Lower Upper

98 100

100 64.9 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 236.587 ug/l
RT: 8.568 min Scan# 1227
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 521204
Ion Ratio Lower Upper
43 100
58 35.3 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#52
Toluene
Concen: 46.513 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 92 Resp: 147100
Ion Ratio Lower Upper
92 100
91 173.1 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 49.111 ug/l

RT: 8.970 min Scan# 1129

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 75 Resp: 8696

Ion Ratio Lower Upper

75 100





77 32.0 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#54

cis-1,3-Dichloropropene

Concen: 49.234 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05




Tgt Ion: 75 Resp: 96359


Ion Ratio Lower Upper

75 100

77 29.1 25.4 38.0

39 61.0 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 46.578 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 97 Resp: 58084

Ion Ratio Lower Upper

97 100

83 87.7 69.0 103.6

85 56.5 44.3 66.5





99 63.5 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#56

Ethyl methacrylate

Concen: 48.490 ug/l

RT: 9.110 min Scan# 1316

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05




Tgt Ion: 69 Resp: 96373


Ion Ratio Lower Upper

69 100

41 76.7 60.8 91.2

39 48.5 39.0 58.6





#57

1,3-Dichloropropane

Concen: 45.674 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 102289

Ion Ratio Lower Upper

76 100





78 32.2 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#58

2-Chloroethyl Vinyl ether

Concen: 195.960 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046078.D




Acq: 08 May 2025 10:05


Tgt Ion: 63 Resp: 198555

Ion Ratio Lower Upper

63 100

106 27.0 21.5 32.3





#59

2-Hexanone

Concen: 235.986 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 384624

Ion Ratio Lower Upper

43 100





58 49.6 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#60

Dibromochloromethane

Concen: 49.247 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046078.D




Acq: 08 May 2025 10:05


Tgt Ion:129 Resp: 67444

Ion Ratio Lower Upper

129 100

127 78.1 39.3 117.8





#61

1,2-Dibromoethane

Concen: 46.129 ug/l

RT: 9.604 min Scan# 1397

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:107 Resp: 5978

Ion Ratio Lower Upper

107 100





109 94.4 73.3 109.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#62

4-Bromofluorobenzene

Concen: 44.770 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05




Tgt Ion: 95 Resp: 77895


Ion Ratio Lower Upper

95 100

174 67.6 0.0 135.8


176 65.4 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:117 Resp: 158314
Ion Ratio Lower Upper
117 100
82 60.1 50.6 76.0
119 31.4 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#64
Tetrachloroethene
Concen: 48.407 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05


Tgt Ion:164 Resp: 54223
Ion Ratio Lower Upper
164 100
166 127.0 105.0 157.6
129 100.9 83.5 125.3
131 96.5 76.5 114.7





#65
Chlorobenzene
Concen: 45.157 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tgt Ion:112 Resp: 156475
Ion Ratio Lower Upper
112 100
114 32.1 25.4 38.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 46.629 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Tgt Ion:131 Resp: 55173
Ion Ratio Lower Upper
131 100
133 96.5 47.3 141.9
119 64.4 31.6 95.0





#67
Ethyl Benzene
Concen: 47.522 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 91 Resp: 29026
Ion Ratio Lower Upper
91 100
106 29.7 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#68
m/p-Xylenes
Concen: 94.184 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion:106 Resp: 210404
Ion Ratio Lower Upper
106 100
91 216.6 171.2 256.8









#69
o-Xylene
Concen: 46.944 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 102240
Ion Ratio Lower Upper
106 100
91 232.4 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#70
Styrene
Concen: 48.774 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion:104 Resp: 174009
Ion Ratio Lower Upper
104 100
78 57.7 45.7 68.5
103 55.0 43.7 65.5





#71

Bromoform

Concen: 48.696 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:173 Resp: 43259

Ion Ratio Lower Upper

173 100

175 48.1 23.4 70.0





254 0.0 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05




Tgt Ion:152 Resp: 73554


Ion Ratio Lower Upper

152 100

115 91.0 46.9 140.7

150 174.9 0.0 351.0





#73

Isopropylbenzene

Concen: 48.428 ug/l

RT: 10.957 min Scan# 1619

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:105 Resp: 277319

Ion Ratio Lower Upper

105 100





120 25.5 12.8 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#74

N-ethyl acetate

Concen: 47.898 ug/l

RT: 10.842 min Scan# 1600

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Tgt Ion: 43 Resp: 135534




Ion Ratio Lower Upper


43 100

70 37.6 30.9 46.3

55 23.1 18.7 28.1

61 22.0 17.1 25.7





#75

1,1,2,2-Tetrachloroethane

Concen: 44.217 ug/l

RT: 11.207 min Scan# 1660

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 8873

Ion Ratio Lower Upper

83 100

131 9.7 5.0 14.9





85 64.0 31.9 95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#76

1,2,3-Trichloropropane

Concen: 45.331 ug/l m

RT: 11.238 min Scan# 1665

Delta R.T. 0.000 min

Lab File: VX046078.D




Acq: 08 May 2025 10:05


Tgt Ion: 75 Resp: 80257

Ion Ratio Lower Upper

75 100


77 39.6 20.5 61.5





#77
Bromobenzene
Concen: 47.164 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tgt Ion:156 Resp: 62702
Ion Ratio Lower Upper
156 100
77 181.9 92.8 278.3
158 93.6 48.8 146.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025




#78
n-propylbenzene
Concen: 48.814 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05



Tgt Ion: 91 Resp: 325019
Ion Ratio Lower Upper
91 100
120 21.8 10.8 32.4









#79
2-Chlorotoluene
Concen: 46.185 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 198350
Ion Ratio Lower Upper
91 100
126 31.2 15.6 46.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#80
1,3,5-Trimethylbenzene
Concen: 48.737 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion:105 Resp: 233158
Ion Ratio Lower Upper
105 100
120 47.0 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 46.952 ug/l
RT: 11.012 min Scan# 1629
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 75 Resp: 2553
Ion Ratio Lower Upper
75 100
53 124.6 100.5 150.7
89 47.1 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#82
4-Chlorotoluene
Concen: 47.823 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 91 Resp: 227761
Ion Ratio Lower Upper
91 100
126 26.9 13.3 39.8





#83

tert-Butylbenzene

Concen: 47.471 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:119 Resp: 22875

Ion Ratio Lower Upper

119 100

91 66.0 32.9 98.7





134 22.5 11.4 34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#84

1,2,4-Trimethylbenzene

Concen: 48.637 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046078.D




Acq: 08 May 2025 10:05


Tgt Ion:105 Resp: 235628

Ion Ratio Lower Upper

105 100


120 42.4 21.2 63.6





#85
 sec-Butylbenzene
 Concen: 48.992 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX046078.D
 Acq: 08 May 2025 10:05




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050




Tgt Ion:105 Resp: 289872
 Ion Ratio Lower Upper
 105 100
 134 19.2 9.7 29.1

Manual Integrations
 APPROVED



Reviewed By :John Carlone 05/09/2025
 Supervised By :Mahesh Dadoda 05/09/2025




#86
 p-Isopropyltoluene
 Concen: 49.423 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX046078.D
 Acq: 08 May 2025 10:05



Tgt Ion:119 Resp: 241375
 Ion Ratio Lower Upper
 119 100
 134 25.1 12.5 37.5
 91 27.6 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 47.766 ug/l

RT: 11.963 min Scan# 1784

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:146 Resp: 115894

Ion Ratio Lower Upper

146 100

111 44.6 22.1 66.3





148 63.1 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#88

1,4-Dichlorobenzene

Concen: 46.133 ug/l

RT: 12.037 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05




Tgt Ion:146 Resp: 114313


Ion Ratio Lower Upper

146 100

111 43.2 21.3 63.9

148 64.6 31.9 95.5





#89

n-Butylbenzene

Concen: 50.699 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 21719

Ion Ratio Lower Upper

91 100

92 54.1 26.9 80.7





134 23.9 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#90

Hexachloroethane

Concen: 48.677 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046078.D




Acq: 08 May 2025 10:05


Tgt Ion:117 Resp: 41883

Ion Ratio Lower Upper

117 100


201 63.1 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 46.645 ug/l
RT: 12.329 min Scan# 1845
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:146 Resp: 11356
Ion Ratio Lower Upper
146 100
111 45.8 22.4 67.3
148 63.6 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 47.926 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 75 Resp: 21306
Ion Ratio Lower Upper
75 100
155 70.5 34.9 104.8
157 87.3 43.8 131.4









#93
1,2,4-Trichlorobenzene
Concen: 49.033 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:180 Resp: 68569
Ion Ratio Lower Upper
180 100
182 95.5 47.9 143.8
145 34.6 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#94
Hexachlorobutadiene
Concen: 48.535 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05


Tgt Ion:225 Resp: 29643
Ion Ratio Lower Upper
225 100
223 63.5 30.8 92.4
227 63.9 30.9 92.7





#95
Naphthalene
Concen: 46.620 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tgt Ion:128 Resp: 239114
Ion Ratio Lower Upper
128 100
127 13.2 10.4 15.6
129 11.1 8.6 13.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#96
1,2,3-Trichlorobenzene
Concen: 47.508 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05



Tgt Ion:180 Resp: 68552
Ion Ratio Lower Upper
180 100
182 96.0 47.8 143.3
145 36.3 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046078.D
 Acq On : 08 May 2025 10:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 09 01:54:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	109	0.00
2 T	Dichlorodifluoromethane	0.765	0.785	-2.6	99	0.00
3 P	Chloromethane	0.742	0.707	4.7	99	0.00
4 C	Vinyl Chloride	0.691	0.643	6.9#	99	0.00
5 T	Bromomethane	0.320	0.281	12.2	94	0.00
6 T	Chloroethane	0.369	0.352	4.6	101	0.00
7 T	Trichlorofluoromethane	1.021	0.997	2.4	102	0.00
8 T	Diethyl Ether	0.347	0.314	9.5	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.632	0.610	3.5	104	0.00
10 T	Methyl Iodide	0.747	0.701	6.2	95	0.00
11 T	Tert butyl alcohol	0.131	0.118	9.9	100	0.00
12 CM	1,1-Dichloroethene	0.593	0.554	6.6#	100	0.00
13 T	Acrolein	0.149	0.102	31.5#	73	0.00
14 T	Allyl chloride	1.133	1.095	3.4	101	0.00
15 T	Acrylonitrile	0.374	0.344	8.0	97	0.00
16 T	Acetone	0.374	0.352	5.9	106	0.00
17 T	Carbon Disulfide	1.406	1.312	6.7	98	0.00
18 T	Methyl Acetate	0.867	0.801	7.6	103	0.00
19 T	Methyl tert-butyl Ether	2.079	1.974	5.1	100	0.00
20 T	Methylene Chloride	0.716	0.617	13.8	98	0.00
21 T	trans-1,2-Dichloroethene	0.596	0.560	6.0	100	0.00
22 T	Diisopropyl ether	2.189	2.076	5.2	99	0.00
23 T	Vinyl Acetate	1.925	1.829	5.0	97	0.00
24 P	1,1-Dichloroethane	1.219	1.143	6.2	99	0.00
25 T	2-Butanone	0.543	0.506	6.8	99	0.00
26 T	2,2-Dichloropropane	0.954	0.903	5.3	103	0.00
27 T	cis-1,2-Dichloroethene	0.718	0.664	7.5	98	0.00
28 T	Bromochloromethane	0.587	0.583	0.7	110	0.00
29 T	Tetrahydrofuran	0.340	0.313	7.9	97	0.00
30 C	Chloroform	1.271	1.202	5.4#	101	0.00
31 T	Cyclohexane	1.111	1.035	6.8	100	0.00
32 T	1,1,1-Trichloroethane	1.101	1.037	5.8	100	0.00
33 S	1,2-Dichloroethane-d4	0.932	0.825	11.5	99	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	108	0.00
35 S	Dibromofluoromethane	0.360	0.332	7.8	102	0.00
36 T	1,1-Dichloropropene	0.484	0.449	7.2	98	0.00
37 T	Ethyl Acetate	0.598	0.548	8.4	97	-0.01
38 T	Carbon Tetrachloride	0.544	0.519	4.6	101	0.00
39 T	Methylcyclohexane	0.623	0.579	7.1	98	0.00
40 TM	Benzene	1.417	1.330	6.1	98	0.00
41 T	Methacrylonitrile	0.313	0.316	-1.0	99	0.00
42 TM	1,2-Dichloroethane	0.612	0.565	7.7	98	0.00
43 T	Isopropyl Acetate	0.912	0.873	4.3	98	0.00
44 TM	Trichloroethene	0.341	0.325	4.7	99	0.00
45 C	1,2-Dichloropropane	0.352	0.336	4.5#	98	0.00
46 T	Dibromomethane	0.278	0.256	7.9	97	0.00
47 T	Bromodichloromethane	0.547	0.530	3.1	99	0.00
48 T	Methyl methacrylate	0.466	0.448	3.9	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046078.D
 Acq On : 08 May 2025 10:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 09 01:54:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	101	0.00
50 S	Toluene-d8	1.246	1.149	7.8	102	0.00
51 T	4-Methyl-2-Pentanone	0.605	0.573	5.3	98	0.00
52 CM	Toluene	0.869	0.808	7.0#	98	0.00
53 T	t-1,3-Dichloropropene	0.487	0.478	1.8	98	0.00
54 T	cis-1,3-Dichloropropene	0.538	0.529	1.7	99	0.00
55 T	1,1,2-Trichloroethane	0.343	0.319	7.0	98	0.00
56 T	Ethyl methacrylate	0.546	0.530	2.9	96	0.00
57 T	1,3-Dichloropropane	0.615	0.562	8.6	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.278	0.218	21.6	77	0.00
59 T	2-Hexanone	0.448	0.423	5.6	97	0.00
60 T	Dibromochloromethane	0.376	0.371	1.3	100	0.00
61 T	1,2-Dibromoethane	0.356	0.329	7.6	95	0.00
62 S	4-Bromofluorobenzene	0.478	0.428	10.5	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	108	0.00
64 T	Tetrachloroethene	0.354	0.343	3.1	99	0.00
65 PM	Chlorobenzene	1.094	0.988	9.7	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.374	0.349	6.7	97	0.00
67 C	Ethyl Benzene	1.929	1.833	5.0#	98	0.00
68 T	m/p-Xylenes	0.706	0.665	5.8	97	0.00
69 T	o-Xylene	0.688	0.646	6.1	96	0.00
70 T	Styrene	1.127	1.099	2.5	98	0.00
71 P	Bromoform	0.281	0.273	2.8	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
73 T	Isopropylbenzene	3.893	3.770	3.2	99	0.00
74 T	N-amyl acetate	1.924	1.843	4.2	96	0.00
75 P	1,1,2,2-Tetrachloroethane	1.364	1.206	11.6	98	0.00
76 T	1,2,3-Trichloropropane	1.204	1.091	9.4	99	0.00
77 T	Bromobenzene	0.904	0.852	5.8	99	0.00
78 T	n-propylbenzene	4.526	4.419	2.4	98	0.00
79 T	2-Chlorotoluene	2.919	2.697	7.6	97	0.00
80 T	1,3,5-Trimethylbenzene	3.252	3.170	2.5	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.370	0.347	6.2	98	0.00
82 T	4-Chlorotoluene	3.238	3.097	4.4	98	0.00
83 T	tert-Butylbenzene	3.276	3.110	5.1	98	0.00
84 T	1,2,4-Trimethylbenzene	3.293	3.203	2.7	98	0.00
85 T	sec-Butylbenzene	4.022	3.941	2.0	100	0.00
86 T	p-Isopropyltoluene	3.320	3.282	1.1	100	0.00
87 T	1,3-Dichlorobenzene	1.649	1.576	4.4	100	0.00
88 T	1,4-Dichlorobenzene	1.684	1.554	7.7	99	0.00
89 T	n-Butylbenzene	2.912	2.953	-1.4	102	0.00
90 T	Hexachloroethane	0.585	0.569	2.7	99	0.00
91 T	1,2-Dichlorobenzene	1.655	1.544	6.7	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.302	0.290	4.0	97	0.00
93 T	1,2,4-Trichlorobenzene	0.951	0.932	2.0	103	0.00
94 T	Hexachlorobutadiene	0.415	0.403	2.9	102	0.00
95 T	Naphthalene	3.487	3.251	6.8	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046078.D
Acq On : 08 May 2025 10:05
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 09 01:54:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.981	0.932	5.0	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046078.D
 Acq On : 08 May 2025 10:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 09 01:54:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	109	0.00
2 T	Dichlorodifluoromethane	50.000	51.315	-2.6	99	0.00
3 P	Chloromethane	50.000	47.625	4.8	99	0.00
4 C	Vinyl Chloride	50.000	46.512	7.0#	99	0.00
5 T	Bromomethane	50.000	43.914	12.2	94	0.00
6 T	Chloroethane	50.000	47.760	4.5	101	0.00
7 T	Trichlorofluoromethane	50.000	48.834	2.3	102	0.00
8 T	Diethyl Ether	50.000	45.221	9.6	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.319	3.4	104	0.00
10 T	Methyl Iodide	50.000	46.872	6.3	95	0.00
11 T	Tert butyl alcohol	250.000	224.852	10.1	100	0.00
12 CM	1,1-Dichloroethene	50.000	46.709	6.6#	100	0.00
13 T	Acrolein	250.000	171.322	31.5#	73	0.00
14 T	Allyl chloride	50.000	48.320	3.4	101	0.00
15 T	Acrylonitrile	250.000	229.976	8.0	97	0.00
16 T	Acetone	250.000	235.163	5.9	106	0.00
17 T	Carbon Disulfide	50.000	46.655	6.7	98	0.00
18 T	Methyl Acetate	50.000	46.193	7.6	103	0.00
19 T	Methyl tert-butyl Ether	50.000	47.478	5.0	100	0.00
20 T	Methylene Chloride	50.000	43.091	13.8	98	0.00
21 T	trans-1,2-Dichloroethene	50.000	46.948	6.1	100	0.00
22 T	Diisopropyl ether	50.000	47.416	5.2	99	0.00
23 T	Vinyl Acetate	250.000	237.541	5.0	97	0.00
24 P	1,1-Dichloroethane	50.000	46.860	6.3	99	0.00
25 T	2-Butanone	250.000	233.066	6.8	99	0.00
26 T	2,2-Dichloropropane	50.000	47.319	5.4	103	0.00
27 T	cis-1,2-Dichloroethene	50.000	46.243	7.5	98	0.00
28 T	Bromochloromethane	50.000	49.644	0.7	110	0.00
29 T	Tetrahydrofuran	250.000	229.940	8.0	97	0.00
30 C	Chloroform	50.000	47.310	5.4#	101	0.00
31 T	Cyclohexane	50.000	46.596	6.8	100	0.00
32 T	1,1,1-Trichloroethane	50.000	47.092	5.8	100	0.00
33 S	1,2-Dichloroethane-d4	50.000	44.264	11.5	99	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	108	0.00
35 S	Dibromofluoromethane	50.000	46.141	7.7	102	0.00
36 T	1,1-Dichloropropene	50.000	46.453	7.1	98	0.00
37 T	Ethyl Acetate	50.000	45.837	8.3	97	-0.01
38 T	Carbon Tetrachloride	50.000	47.778	4.4	101	0.00
39 T	Methylcyclohexane	50.000	46.514	7.0	98	0.00
40 TM	Benzene	50.000	46.914	6.2	98	0.00
41 T	Methacrylonitrile	50.000	50.572	-1.1	99	0.00
42 TM	1,2-Dichloroethane	50.000	46.172	7.7	98	0.00
43 T	Isopropyl Acetate	50.000	47.860	4.3	98	0.00
44 TM	Trichloroethene	50.000	47.599	4.8	99	0.00
45 C	1,2-Dichloropropane	50.000	47.662	4.7#	98	0.00
46 T	Dibromomethane	50.000	46.001	8.0	97	0.00
47 T	Bromodichloromethane	50.000	48.392	3.2	99	0.00
48 T	Methyl methacrylate	50.000	48.156	3.7	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046078.D
 Acq On : 08 May 2025 10:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 09 01:54:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	954.733	4.5	101	0.00
50 S	Toluene-d8	50.000	46.095	7.8	102	0.00
51 T	4-Methyl-2-Pentanone	250.000	236.587	5.4	98	0.00
52 CM	Toluene	50.000	46.513	7.0#	98	0.00
53 T	t-1,3-Dichloropropene	50.000	49.111	1.8	98	0.00
54 T	cis-1,3-Dichloropropene	50.000	49.234	1.5	99	0.00
55 T	1,1,2-Trichloroethane	50.000	46.578	6.8	98	0.00
56 T	Ethyl methacrylate	50.000	48.490	3.0	96	0.00
57 T	1,3-Dichloropropane	50.000	45.674	8.7	98	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	195.960	21.6	77	0.00
59 T	2-Hexanone	250.000	235.986	5.6	97	0.00
60 T	Dibromochloromethane	50.000	49.247	1.5	100	0.00
61 T	1,2-Dibromoethane	50.000	46.129	7.7	95	0.00
62 S	4-Bromofluorobenzene	50.000	44.770	10.5	99	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	108	0.00
64 T	Tetrachloroethene	50.000	48.407	3.2	99	0.00
65 PM	Chlorobenzene	50.000	45.157	9.7	97	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	46.629	6.7	97	0.00
67 C	Ethyl Benzene	50.000	47.522	5.0#	98	0.00
68 T	m/p-Xylenes	100.000	94.184	5.8	97	0.00
69 T	o-Xylene	50.000	46.944	6.1	96	0.00
70 T	Styrene	50.000	48.774	2.5	98	0.00
71 P	Bromoform	50.000	48.696	2.6	97	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	108	0.00
73 T	Isopropylbenzene	50.000	48.428	3.1	99	0.00
74 T	N-ethyl acetate	50.000	47.898	4.2	96	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	44.217	11.6	98	0.00
76 T	1,2,3-Trichloropropane	50.000	45.331	9.3	99	0.00
77 T	Bromobenzene	50.000	47.164	5.7	99	0.00
78 T	n-propylbenzene	50.000	48.814	2.4	98	0.00
79 T	2-Chlorotoluene	50.000	46.185	7.6	97	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.737	2.5	98	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	46.952	6.1	98	0.00
82 T	4-Chlorotoluene	50.000	47.823	4.4	98	0.00
83 T	tert-Butylbenzene	50.000	47.471	5.1	98	0.00
84 T	1,2,4-Trimethylbenzene	50.000	48.637	2.7	98	0.00
85 T	sec-Butylbenzene	50.000	48.992	2.0	100	0.00
86 T	p-Isopropyltoluene	50.000	49.423	1.2	100	0.00
87 T	1,3-Dichlorobenzene	50.000	47.766	4.5	100	0.00
88 T	1,4-Dichlorobenzene	50.000	46.133	7.7	99	0.00
89 T	n-Butylbenzene	50.000	50.699	-1.4	102	0.00
90 T	Hexachloroethane	50.000	48.677	2.6	99	0.00
91 T	1,2-Dichlorobenzene	50.000	46.645	6.7	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	47.926	4.1	97	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.033	1.9	103	0.00
94 T	Hexachlorobutadiene	50.000	48.535	2.9	102	0.00
95 T	Naphthalene	50.000	46.620	6.8	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046078.D
Acq On : 08 May 2025 10:05
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 09 01:54:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	47.508	5.0	99	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1985 SAS No.: Q1985 SDG No.: Q1985

Instrument ID: MSVOA_X Calibration Date/Time: 05/08/2025 18:46

Lab File ID: VX046100.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.742	0.761	0.1	2.56	50
Vinyl Chloride	0.691	0.697		0.87	50
Bromomethane	0.320	0.301		-5.94	50
Chloroethane	0.369	0.377		2.17	50
Trichlorofluoromethane	1.021	1.051		2.94	50
1,1,2-Trichlorotrifluoroethane	0.632	0.641		1.42	50
1,1-Dichloroethene	0.593	0.589		-0.68	50
Acetone	0.374	0.389		4.01	50
Carbon Disulfide	1.406	1.508		7.26	50
Methyl tert-butyl Ether	2.079	2.183		5	50
Methylene Chloride	0.716	0.688		-3.91	50
trans-1,2-Dichloroethene	0.596	0.596		0	50
1,1-Dichloroethane	1.219	1.275	0.1	4.59	50
2-Butanone	0.543	0.582		7.18	50
Carbon Tetrachloride	0.544	0.559		2.76	50
cis-1,2-Dichloroethene	0.718	0.741		3.2	50
Chloroform	1.271	1.318		3.7	50
1,1,1-Trichloroethane	1.101	1.150		4.45	50
Methylcyclohexane	0.623	0.607		-2.57	50
Benzene	1.417	1.453		2.54	50
1,2-Dichloroethane	0.612	0.625		2.12	50
Trichloroethene	0.341	0.344		0.88	50
1,2-Dichloropropane	0.352	0.373		5.97	50
Bromodichloromethane	0.547	0.585		6.95	50
4-Methyl-2-Pentanone	0.605	0.658		8.76	50
Toluene	0.869	0.886		1.96	50
t-1,3-Dichloropropene	0.487	0.507		4.11	50
cis-1,3-Dichloropropene	0.538	0.561		4.28	50
1,1,2-Trichloroethane	0.343	0.353		2.91	50
2-Hexanone	0.448	0.496		10.71	50
Dibromochloromethane	0.376	0.409		8.78	50
Tetrachloroethene	0.354	0.354		0	50
Chlorobenzene	1.094	1.092	0.3	-0.18	50
Ethyl Benzene	1.929	2.015		4.46	50
m/p-Xylenes	0.706	0.731		3.54	50
o-Xylene	0.688	0.727		5.67	50
Styrene	1.127	1.222		8.43	50
Bromoform	0.281	0.310	0.1	10.32	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1985 SAS No.: Q1985 SDG No.: Q1985

Instrument ID: MSVOA_X Calibration Date/Time: 05/08/2025 18:46

Lab File ID: VX046100.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.893	3.977		2.16	50
1,1,2,2-Tetrachloroethane	1.364	1.354	0.3	-0.73	50
1,3-Dichlorobenzene	1.649	1.648		-0.06	50
1,4-Dichlorobenzene	1.684	1.636		-2.85	50
1,2-Dichlorobenzene	1.655	1.664		0.54	50
1,2-Dichloroethane-d4	0.932	0.905		-2.9	50
Dibromofluoromethane	0.360	0.350		-2.78	50
Toluene-d8	1.246	1.213		-2.65	50
4-Bromofluorobenzene	0.478	0.464		-2.93	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046100.D
 Acq On : 08 May 2025 18:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :John Carlone 05/09/2025
 Supervised By :Mahesh Dadoda 05/09/2025

Quant Time: May 09 01:59:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	90039	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	158698	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	137076	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66563	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	81456	48.526	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.060%		
35) Dibromofluoromethane	5.379	113	55614	48.665	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.320%		
50) Toluene-d8	8.647	98	192501	48.668	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.340%		
62) 4-Bromofluorobenzene	11.079	95	73670	48.556	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	97.120%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	74341	53.944	ug/l	100
3) Chloromethane	1.307	50	68560	51.301	ug/l	96
4) Vinyl Chloride	1.374	62	62758	50.457	ug/l	98
5) Bromomethane	1.593	94	27111	46.995	ug/l	98
6) Chloroethane	1.666	64	33953	51.135	ug/l	94
7) Trichlorofluoromethane	1.873	101	94604	51.465	ug/l	100
8) Diethyl Ether	2.130	74	31351	50.101	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.319	101	57674	50.699	ug/l	99
10) Methyl Iodide	2.440	142	72888	54.149	ug/l	99
11) Tert butyl alcohol	2.977	59	63473	269.379	ug/l	99
12) 1,1-Dichloroethene	2.306	96	53022	49.663	ug/l	98
13) Acrolein	2.239	56	82589	307.775	ug/l	100
14) Allyl chloride	2.654	41	105974	51.936	ug/l	98
15) Acrylonitrile	3.062	53	179590	266.549	ug/l	99
16) Acetone	2.386	43	175020	260.042	ug/l	99
17) Carbon Disulfide	2.501	76	135805	53.653	ug/l	99
18) Methyl Acetate	2.703	43	86369	55.300	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	196589	52.521	ug/l	99
20) Methylene Chloride	2.782	84	61933	48.019	ug/l	94
21) trans-1,2-Dichloroethene	3.087	96	53687	50.003	ug/l	94
22) Diisopropyl ether	3.763	45	208277	52.843	ug/l	92
23) Vinyl Acetate	3.721	43	931721	268.769	ug/l	100
24) 1,1-Dichloroethane	3.605	63	114758	52.275	ug/l	99
25) 2-Butanone	4.556	43	261893	268.026	ug/l	99
26) 2,2-Dichloropropane	4.471	77	69517	40.458	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	66703	51.607	ug/l	96
28) Bromochloromethane	4.891	49	52572	49.751	ug/l	98
29) Tetrahydrofuran	5.001	42	165543	270.368	ug/l	99
30) Chloroform	5.086	83	118685	51.869	ug/l	99
31) Cyclohexane	5.458	56	100717	50.347	ug/l	99
32) 1,1,1-Trichloroethane	5.379	97	103512	52.187	ug/l	98
36) 1,1-Dichloropropene	5.684	75	75619	49.249	ug/l	98
37) Ethyl Acetate	4.714	43	97419	51.353	ug/l	100
38) Carbon Tetrachloride	5.672	117	88790	51.466	ug/l	97
39) Methylcyclohexane	7.379	83	96361	48.747	ug/l	98
40) Benzene	6.031	78	230602	51.273	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046100.D
 Acq On : 08 May 2025 18:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/09/2025
 Supervised By : Mahesh Dadoda 05/09/2025

Quant Time: May 09 01:59:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	55534	55.961	ug/l	98
42) 1,2-Dichloroethane	6.080	62	99228	51.120	ug/l	100
43) Isopropyl Acetate	6.342	43	156399	54.041	ug/l	99
44) Trichloroethene	7.123	130	54530	50.376	ug/l	98
45) 1,2-Dichloropropane	7.427	63	59216	52.950	ug/l	99
46) Dibromomethane	7.574	93	45589	51.686	ug/l	98
47) Bromodichloromethane	7.818	83	92853	53.449	ug/l	98
48) Methyl methacrylate	7.689	41	82409	55.756	ug/l	99
49) 1,4-Dioxane	7.659	88	31325	1116.209	ug/l	98
51) 4-Methyl-2-Pentanone	8.573	43	522471	271.972	ug/l	99
52) Toluene	8.714	92	140581	50.976	ug/l	98
53) t-1,3-Dichloropropene	8.976	75	80387	52.059	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	88964	52.127	ug/l	97
55) 1,1,2-Trichloroethane	9.153	97	55968	51.469	ug/l	98
56) Ethyl methacrylate	9.116	69	96903	55.913	ug/l	99
57) 1,3-Dichloropropane	9.305	76	98930	50.658	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	233750	264.555	ug/l	100
59) 2-Hexanone	9.427	43	393840	277.107	ug/l	100
60) Dibromochloromethane	9.518	129	64916	54.358	ug/l	98
61) 1,2-Dibromoethane	9.610	107	59495	52.641	ug/l	99
64) Tetrachloroethene	9.268	164	48558	50.066	ug/l	98
65) Chlorobenzene	10.073	112	149645	49.877	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	52783	51.521	ug/l	99
67) Ethyl Benzene	10.189	91	276173	52.220	ug/l	100
68) m/p-Xylenes	10.299	106	200363	103.585	ug/l	98
69) o-Xylene	10.640	106	99641	52.839	ug/l	98
70) Styrene	10.652	104	167552	54.240	ug/l	99
71) Bromoform	10.799	173	42429	55.162	ug/l #	97
73) Isopropylbenzene	10.957	105	264733	51.086	ug/l	100
74) N-amyl acetate	10.841	43	138217	53.976	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	90125	49.628	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	78610m	49.064	ug/l	
77) Bromobenzene	11.195	156	59559	49.505	ug/l	99
78) n-propylbenzene	11.305	91	309136	51.305	ug/l	100
79) 2-Chlorotoluene	11.360	91	190049	48.900	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	220577	50.950	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	23359	47.466	ug/l	99
82) 4-Chlorotoluene	11.451	91	219982	51.040	ug/l	100
83) tert-Butylbenzene	11.713	119	220925	50.661	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	223626	51.007	ug/l	99
85) sec-Butylbenzene	11.890	105	273948	51.163	ug/l	100
86) p-Isopropyltoluene	12.006	119	227600	51.497	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	109705	49.964	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	108913	48.570	ug/l	99
89) n-Butylbenzene	12.329	91	198814	51.283	ug/l	100
90) Hexachloroethane	12.536	117	40761	52.349	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	110746	50.262	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.939	75	22030	54.760	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	63121	49.878	ug/l	98
94) Hexachlorobutadiene	13.725	225	26797	48.483	ug/l	98
95) Naphthalene	13.774	128	236240	50.897	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	65100	49.854	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046100.D
Acq On : 08 May 2025 18:46
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025

Quant Time: May 09 01:59:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046100.D
Acq On : 08 May 2025 18:46
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Manual Integrations

Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025

Quant Time: May 09 01:59:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:168 Resp: 90039
Ion Ratio Lower Upper
168 100
99 73.3 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#2
Dichlorodifluoromethane
Concen: 53.944 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 85 Resp: 74341
Ion Ratio Lower Upper
85 100
87 31.6 15.7 47.1









#3
Chloromethane
Concen: 51.301 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 50 Resp: 68560
Ion Ratio Lower Upper
50 100
52 34.1 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#4
Vinyl Chloride
Concen: 50.457 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46


Tgt Ion: 62 Resp: 62758
Ion Ratio Lower Upper
62 100
64 32.9 25.2 37.8





#5
Bromomethane
Concen: 46.995 ug/l
RT: 1.593 min Scan# 83
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 94 Resp: 2711
Ion Ratio Lower Upper
94 100
96 92.6 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#6
Chloroethane
Concen: 51.135 ug/l
RT: 1.666 min Scan# 95
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46


Tgt Ion: 64 Resp: 33953
Ion Ratio Lower Upper
64 100
66 33.6 24.3 36.5





#7
Trichlorofluoromethane
Concen: 51.465 ug/l
RT: 1.873 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:101 Resp: 94604
Ion Ratio Lower Upper
101 100
103 63.4 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#8
Diethyl Ether
Concen: 50.101 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 74 Resp: 31351
Ion Ratio Lower Upper
74 100
45 105.3 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.699 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:101 Resp: 57674
Ion Ratio Lower Upper
101 100
85 46.9 38.6 58.0
151 68.6 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#10
Methyl Iodide
Concen: 54.149 ug/l
RT: 2.440 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion:142 Resp: 72888
Ion Ratio Lower Upper
142 100
127 53.2 41.7 62.5
141 14.6 11.5 17.3





#11

Tert butyl alcohol

Concen: 269.379 ug/l

RT: 2.977 min Scan# 310

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 59 Resp: 6347

Ion Ratio Lower Upper

59 100





57 11.0 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#12

1,1-Dichloroethene

Concen: 49.663 ug/l

RT: 2.306 min Scan# 200

Delta R.T. -0.006 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46




Tgt Ion: 96 Resp: 53022


Ion Ratio Lower Upper

96 100

61 187.1 146.2 219.2


98 63.8 51.0 76.6





#13
Acrolein
Concen: 307.775 ug/l
RT: 2.239 min Scan# 188
Delta R.T. 0.006 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 56 Resp: 82589
Ion Ratio Lower Upper
56 100
55 70.6 56.2 84.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#14
Allyl chloride
Concen: 51.936 ug/l
RT: 2.654 min Scan# 257
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 41 Resp: 105974
Ion Ratio Lower Upper
41 100
39 73.2 60.6 90.8
76 30.4 24.9 37.3





#15

Acrylonitrile

Concen: 266.549 ug/l

RT: 3.062 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:	53	Resp:	179590
Ion	Ratio	Lower	Upper
53	100		
52	82.8	65.3	97.9
51	37.3	29.8	44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#16

Acetone

Concen: 260.042 ug/l




RT: 2.386 min Scan# 213


Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46


Tgt Ion:	43	Resp:	175020
Ion	Ratio	Lower	Upper
43	100		
58	25.9	21.2	31.8





#17
Carbon Disulfide
Concen: 53.653 ug/l
RT: 2.501 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 76 Resp: 13580
Ion Ratio Lower Upper
76 100
78 9.1 7.0 10.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#18
Methyl Acetate
Concen: 55.300 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 43 Resp: 86369
Ion Ratio Lower Upper
43 100
74 22.1 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 52.521 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 73 Resp: 196589
Ion Ratio Lower Upper
73 100
57 22.3 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#20
Methylene Chloride
Concen: 48.019 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 84 Resp: 61933
Ion Ratio Lower Upper
84 100
49 135.8 113.9 170.9
51 41.4 33.5 50.3
86 60.3 53.8 80.8





#21

trans-1,2-Dichloroethene

Concen: 50.003 ug/l

RT: 3.087 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046100.D


Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 96 Resp: 5368

Ion Ratio Lower Upper

96 100

61 159.1 119.5 179.3




98 63.5 50.0 75.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#22

Diisopropyl ether

Concen: 52.843 ug/l

RT: 3.763 min Scan# 439

Delta R.T. 0.006 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Tgt Ion: 45 Resp: 208277




Ion Ratio Lower Upper


45 100

43 74.3 66.6 100.0

87 26.3 19.8 29.6

59 11.0 8.6 12.8









#23
Vinyl Acetate
Concen: 268.769 ug/l
RT: 3.721 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 93172
Ion Ratio Lower Upper
43 100
86 9.4 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#24
1,1-Dichloroethane
Concen: 52.275 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 63 Resp: 114758
Ion Ratio Lower Upper
63 100
98 6.6 3.6 10.8
100 4.4 2.1 6.3





#25

2-Butanone

Concen: 268.026 ug/l

RT: 4.556 min Scan# 569

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 26189

Ion Ratio Lower Upper

43 100





72 22.6 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#26

2,2-Dichloropropane

Concen: 40.458 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046100.D




Acq: 08 May 2025 18:46


Tgt Ion: 77 Resp: 69517

Ion Ratio Lower Upper

77 100

97 21.3 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 51.607 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 96 Resp: 6670

Ion Ratio Lower Upper

96 100

61 155.4 0.0 322.8





98 63.5 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#28

Bromochloromethane

Concen: 49.751 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46




Tgt Ion: 49 Resp: 52572


Ion Ratio Lower Upper

49 100

129 1.9 0.0 4.0

130 68.3 56.2 84.2





#29

Tetrahydrofuran

Concen: 270.368 ug/l

RT: 5.001 min Scan# 642

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 42 Resp: 16554

Ion Ratio Lower Upper

42 100

72 41.7 33.7 50.5





71 38.7 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#30

Chloroform

Concen: 51.869 ug/l

RT: 5.086 min Scan# 656

Delta R.T. -0.000 min

Lab File: VX046100.D




Acq: 08 May 2025 18:46


Tgt Ion: 83 Resp: 118685

Ion Ratio Lower Upper

83 100

85 62.7 49.3 73.9





#31

Cyclohexane

Concen: 50.347 ug/l

RT: 5.458 min Scan# 717

Delta R.T. -0.006 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 56 Resp: 10071

Ion Ratio Lower Upper

56 100

69 30.2 24.4 36.6





84 82.4 66.9 100.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#32

1,1,1-Trichloroethane

Concen: 52.187 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46




Tgt Ion: 97 Resp: 103512


Ion Ratio Lower Upper

97 100

99 63.7 51.8 77.6

61 48.7 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 48.526 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 81450

Ion Ratio Lower Upper

65 100





67 50.1 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46




Tgt Ion:114 Resp: 158698


Ion Ratio Lower Upper

114 100

63 24.0 0.0 49.2

88 17.3 0.0 33.6





#35

Dibromofluoromethane

Concen: 48.665 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:113 Resp: 55614

Ion Ratio Lower Upper

113 100

111 104.4 83.1 124.7





192 17.1 13.3 19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#36

1,1-Dichloropropene

Concen: 49.249 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46




Tgt Ion: 75 Resp: 75619


Ion Ratio Lower Upper

75 100

110 33.0 16.3 48.9

77 31.9 24.3 36.5





#37

Ethyl Acetate

Concen: 51.353 ug/l

RT: 4.714 min Scan# 595

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 97419





Ion	Ratio	Lower	Upper
43	100		
61	12.7	10.3	15.5
70	9.9	7.9	11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#38

Carbon Tetrachloride

Concen: 51.466 ug/l

RT: 5.672 min Scan# 752




Delta R.T. -0.000 min


Lab File: VX046100.D

Acq: 08 May 2025 18:46

Tgt Ion:117 Resp: 88790

Ion	Ratio	Lower	Upper
117	100		
119	96.0	75.2	112.8
121	32.4	24.2	36.4





#39

Methylcyclohexane

Concen: 48.747 ug/l

RT: 7.379 min Scan# 1031

Delta R.T. 0.006 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 9636

Ion Ratio Lower Upper

83 100

55 83.9 64.7 97.1





98 46.7 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#40

Benzene

Concen: 51.273 ug/l

RT: 6.031 min Scan# 811

Delta R.T. -0.000 min

Lab File: VX046100.D




Acq: 08 May 2025 18:46


Tgt Ion: 78 Resp: 230602

Ion Ratio Lower Upper

78 100

77 23.6 19.0 28.4





#41

Methacrylonitrile

Concen: 55.961 ug/l

RT: 4.922 min Scan# 61

Delta R.T. 0.006 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 41 Resp: 55534

Ion Ratio Lower Upper

41 100

39 57.0 47.2 70.8

67 64.6 50.7 76.1





52 30.3 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#42

1,2-Dichloroethane

Concen: 51.120 ug/l

RT: 6.080 min Scan# 819

Delta R.T. -0.000 min

Lab File: VX046100.D




Acq: 08 May 2025 18:46


Tgt Ion: 62 Resp: 99228

Ion Ratio Lower Upper

62 100

98 7.7 0.0 15.2





#43

Isopropyl Acetate

Concen: 54.041 ug/l

RT: 6.342 min Scan# 861

Delta R.T. 0.006 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 156399

Ion Ratio Lower Upper

43 100

61 17.5 14.3 21.5





87 11.7 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#44

Trichloroethene

Concen: 50.376 ug/l

RT: 7.123 min Scan# 990

Delta R.T. -0.000 min

Lab File: VX046100.D




Acq: 08 May 2025 18:46


Tgt Ion:130 Resp: 54530

Ion Ratio Lower Upper

130 100

95 104.0 0.0 204.2





#45

1,2-Dichloropropane

Concen: 52.950 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 59210

Ion Ratio Lower Upper

63 100





65 30.5 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#46

Dibromomethane

Concen: 51.686 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46




Tgt Ion: 93 Resp: 45589


Ion Ratio Lower Upper

93 100

95 80.8 65.6 98.4

174 82.6 68.2 102.2





#47

Bromodichloromethane

Concen: 53.449 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 83	Resp: 9285
Ion Ratio	Lower Upper
83 100	
85 61.5	50.5 75.7
127 8.5	6.5 9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#48

Methyl methacrylate

Concen: 55.756 ug/l




RT: 7.689 min Scan# 1083


Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Tgt Ion: 41	Resp: 82409
Ion Ratio	Lower Upper
41 100	
69 71.7	58.5 87.7
39 64.0	51.7 77.5





#49

1,4-Dioxane

Concen: 1116.209 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:	88	Resp:	3132
Ion	Ratio	Lower	Upper
88	100		
43	41.3	33.4	50.2
58	75.3	58.6	88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#50

Toluene-d8

Concen: 48.668 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Tgt Ion:	98	Resp:	192501
Ion	Ratio	Lower	Upper
98	100		
100	66.2	53.5	80.3









#51
4-Methyl-2-Pentanone
Concen: 271.972 ug/l
RT: 8.573 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 52247
Ion Ratio Lower Upper
43 100
58 35.6 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#52
Toluene
Concen: 50.976 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 92 Resp: 140581
Ion Ratio Lower Upper
92 100
91 173.5 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 52.059 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 75 Resp: 8038

Ion Ratio Lower Upper

75 100





77 31.5 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#54

cis-1,3-Dichloropropene

Concen: 52.127 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46




Tgt Ion: 75 Resp: 88964


Ion Ratio Lower Upper

75 100

77 31.4 25.4 38.0

39 62.1 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 51.469 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. 0.006 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 97 Resp: 55968

Ion Ratio Lower Upper

97 100

83 85.4 69.0 103.6

85 54.2 44.3 66.5





99 63.3 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#56

Ethyl methacrylate

Concen: 55.913 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46




Tgt Ion: 69 Resp: 96903


Ion Ratio Lower Upper

69 100

41 75.9 60.8 91.2

39 47.7 39.0 58.6





#57

1,3-Dichloropropane

Concen: 50.658 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 98930

Ion Ratio Lower Upper

76 100





78 32.4 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#58

2-Chloroethyl Vinyl ether

Concen: 264.555 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046100.D




Acq: 08 May 2025 18:46


Tgt Ion: 63 Resp: 233750

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3





#59

2-Hexanone

Concen: 277.107 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 393840

Ion Ratio Lower Upper

43 100





58 49.8 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#60

Dibromochloromethane

Concen: 54.358 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046100.D




Acq: 08 May 2025 18:46


Tgt Ion:129 Resp: 64916

Ion Ratio Lower Upper

129 100

127 77.2 39.3 117.8









#61
1,2-Dibromoethane
Concen: 52.641 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.006 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 107 Resp: 5949
Ion Ratio Lower Upper
107 100
109 92.6 73.3 109.9




Manual Integrations
APPROVED


Reviewed By : John Carlone 05/09/2025
Supervised By : Mahesh Dadoda 05/09/2025



#62
4-Bromofluorobenzene
Concen: 48.556 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46


Tgt Ion: 95 Resp: 73670
Ion Ratio Lower Upper
95 100
174 67.8 0.0 135.8
176 66.3 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:117 Resp: 137070
Ion Ratio Lower Upper
117 100
82 65.0 50.6 76.0
119 32.8 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#64
Tetrachloroethene
Concen: 50.066 ug/l
RT: 9.268 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46


Tgt Ion:164 Resp: 48558
Ion Ratio Lower Upper
164 100
166 125.5 105.0 157.6
129 103.8 83.5 125.3
131 95.3 76.5 114.7





#65
Chlorobenzene
Concen: 49.877 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:112 Resp: 14964
Ion Ratio Lower Upper
112 100
114 31.6 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 51.521 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46


Tgt Ion:131 Resp: 52783
Ion Ratio Lower Upper
131 100
133 95.2 47.3 141.9
119 64.8 31.6 95.0





#67
Ethyl Benzene
Concen: 52.220 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 276173
Ion Ratio Lower Upper
91 100
106 29.5 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#68
m/p-Xylenes
Concen: 103.585 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion:106 Resp: 200363
Ion Ratio Lower Upper
106 100
91 217.0 171.2 256.8









#69
o-Xylene
Concen: 52.839 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:106 Resp: 9964
Ion Ratio Lower Upper
106 100
91 228.9 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#70
Styrene
Concen: 54.240 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46


Tgt Ion:104 Resp: 167552
Ion Ratio Lower Upper
104 100
78 57.9 45.7 68.5
103 54.8 43.7 65.5





#71
Bromoform
Concen: 55.162 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:173 Resp: 42429
Ion Ratio Lower Upper
173 100
175 48.5 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46


Tgt Ion:152 Resp: 66563
Ion Ratio Lower Upper
152 100
115 91.2 46.9 140.7
150 171.5 0.0 351.0





#73
Isopropylbenzene
Concen: 51.086 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 105 Resp: 264733
Ion Ratio Lower Upper
105 100
120 25.4 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#74
N-ethyl acetate
Concen: 53.976 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 43 Resp: 138217
Ion Ratio Lower Upper
43 100
70 38.4 30.9 46.3
55 23.1 18.7 28.1
61 22.0 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 49.628 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 83 Resp: 90125
Ion Ratio Lower Upper
83 100
131 9.7 5.0 14.9
85 63.5 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#76
1,2,3-Trichloropropane
Concen: 49.064 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 75 Resp: 78610
Ion Ratio Lower Upper
75 100
77 40.3 20.5 61.5





#77

Bromobenzene

Concen: 49.505 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:	156	Resp:	5955
Ion Ratio	100	Lower	Upper
156	100		
77	187.5	92.8	278.3
158	96.2	48.8	146.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#78

n-propylbenzene

Concen: 51.305 ug/l




RT: 11.305 min Scan# 1676


Delta R.T. 0.006 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Tgt Ion:	91	Resp:	309136
Ion Ratio	100	Lower	Upper
91	100		
120	21.7	10.8	32.4









#79
2-Chlorotoluene
Concen: 48.900 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 190049
Ion Ratio Lower Upper
91 100
126 31.6 15.6 46.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#80
1,3,5-Trimethylbenzene
Concen: 50.950 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion:105 Resp: 220577
Ion Ratio Lower Upper
105 100
120 46.7 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 47.466 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 75 Resp: 23359
Ion Ratio Lower Upper
75 100
53 127.0 100.5 150.7
89 46.3 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#82
4-Chlorotoluene
Concen: 51.040 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46


Tgt Ion: 91 Resp: 219982
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83
tert-Butylbenzene
Concen: 50.661 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion:119 Resp: 220925
Ion Ratio Lower Upper
119 100
91 66.0 32.9 98.7
134 22.7 11.4 34.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025




#84
1,2,4-Trimethylbenzene
Concen: 51.007 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Tgt Ion:105 Resp: 223626
Ion Ratio Lower Upper
105 100
120 43.3 21.2 63.6





#85
 sec-Butylbenzene
 Concen: 51.163 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046100.D
 Acq: 08 May 2025 18:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion:105 Resp: 273948
 Ion Ratio Lower Upper
 105 100
 134 19.3 9.7 29.1

Manual Integrations
 APPROVED



Reviewed By :John Carlone 05/09/2025
 Supervised By :Mahesh Dadoda 05/09/2025




#86
 p-Isopropyltoluene
 Concen: 51.497 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046100.D
 Acq: 08 May 2025 18:46




Tgt Ion:119 Resp: 227600
 Ion Ratio Lower Upper
 119 100
 134 24.7 12.5 37.5
 91 27.4 13.8 41.4





#87
1,3-Dichlorobenzene
Concen: 49.964 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion:146 Resp: 10970
Ion Ratio Lower Upper
146 100
111 44.4 22.1 66.3
148 63.4 32.1 96.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025




#88
1,4-Dichlorobenzene
Concen: 48.570 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46



Tgt Ion:146 Resp: 108913
Ion Ratio Lower Upper
146 100
111 43.4 21.3 63.9
148 64.3 31.9 95.5





#89

n-Butylbenzene

Concen: 51.283 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 91 Resp: 198814

Ion Ratio Lower Upper

91 100

92 53.6 26.9 80.7





134 23.9 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#90

Hexachloroethane

Concen: 52.349 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046100.D




Acq: 08 May 2025 18:46


Tgt Ion:117 Resp: 40761

Ion Ratio Lower Upper

117 100


201 61.9 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 50.262 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:146 Resp: 110740
Ion Ratio Lower Upper
146 100
111 46.4 22.4 67.3
148 64.3 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 54.760 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46


Tgt Ion: 75 Resp: 22030
Ion Ratio Lower Upper
75 100
155 68.8 34.9 104.8
157 88.3 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 49.878 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion:180 Resp: 6312
Ion Ratio Lower Upper
180 100
182 93.5 47.9 143.8
145 34.8 17.1 51.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025




#94
Hexachlorobutadiene
Concen: 48.483 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46



Tgt Ion:225 Resp: 26797
Ion Ratio Lower Upper
225 100
223 64.6 30.8 92.4
227 62.4 30.9 92.7





#95

Naphthalene

Concen: 50.897 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:128 Resp: 236240

Ion Ratio Lower Upper

128 100

127 12.6 10.4 15.6





129 10.8 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#96

1,2,3-Trichlorobenzene

Concen: 49.854 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46




Tgt Ion:180 Resp: 65100

Ion Ratio Lower Upper

180 100

182 95.0 47.8 143.3

145 36.8 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046100.D
 Acq On : 08 May 2025 18:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 09 01:59:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.765	0.826	-8.0	89	0.00
3 P	Chloromethane	0.742	0.761	-2.6	91	0.00
4 C	Vinyl Chloride	0.691	0.697	-0.9#	91	0.00
5 T	Bromomethane	0.320	0.301	5.9	86	0.00
6 T	Chloroethane	0.369	0.377	-2.2	92	0.00
7 T	Trichlorofluoromethane	1.021	1.051	-2.9	91	0.00
8 T	Diethyl Ether	0.347	0.348	-0.3	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.632	0.641	-1.4	93	0.00
10 T	Methyl Iodide	0.747	0.810	-8.4	93	0.00
11 T	Tert butyl alcohol	0.131	0.141	-7.6	101	0.00
12 CM	1,1-Dichloroethene	0.593	0.589	0.7#	91	0.00
13 T	Acrolein	0.149	0.183	-22.8	112	0.00
14 T	Allyl chloride	1.133	1.177	-3.9	93	0.00
15 T	Acrylonitrile	0.374	0.399	-6.7	95	0.00
16 T	Acetone	0.374	0.389	-4.0	100	0.00
17 T	Carbon Disulfide	1.406	1.508	-7.3	96	0.00
18 T	Methyl Acetate	0.867	0.959	-10.6	105	0.00
19 T	Methyl tert-butyl Ether	2.079	2.183	-5.0	94	0.00
20 T	Methylene Chloride	0.716	0.688	3.9	93	0.00
21 T	trans-1,2-Dichloroethene	0.596	0.596	0.0	91	0.00
22 T	Diisopropyl ether	2.189	2.313	-5.7	94	0.00
23 T	Vinyl Acetate	1.925	2.070	-7.5	94	0.00
24 P	1,1-Dichloroethane	1.219	1.275	-4.6	94	0.00
25 T	2-Butanone	0.543	0.582	-7.2	97	0.00
26 T	2,2-Dichloropropane	0.954	0.772	19.1	75	0.00
27 T	cis-1,2-Dichloroethene	0.718	0.741	-3.2	93	0.00
28 T	Bromochloromethane	0.587	0.584	0.5	94	0.00
29 T	Tetrahydrofuran	0.340	0.368	-8.2	97	0.00
30 C	Chloroform	1.271	1.318	-3.7#	94	0.00
31 T	Cyclohexane	1.111	1.119	-0.7	92	0.00
32 T	1,1,1-Trichloroethane	1.101	1.150	-4.5	94	0.00
33 S	1,2-Dichloroethane-d4	0.932	0.905	2.9	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.360	0.350	2.8	93	0.00
36 T	1,1-Dichloropropene	0.484	0.476	1.7	91	0.00
37 T	Ethyl Acetate	0.598	0.614	-2.7	95	0.00
38 T	Carbon Tetrachloride	0.544	0.559	-2.8	95	0.00
39 T	Methylcyclohexane	0.623	0.607	2.6	90	0.00
40 TM	Benzene	1.417	1.453	-2.5	93	0.00
41 T	Methacrylonitrile	0.313	0.350	-11.8	96	0.00
42 TM	1,2-Dichloroethane	0.612	0.625	-2.1	94	0.00
43 T	Isopropyl Acetate	0.912	0.986	-8.1	97	0.00
44 TM	Trichloroethene	0.341	0.344	-0.9	91	0.00
45 C	1,2-Dichloropropane	0.352	0.373	-6.0#	95	0.00
46 T	Dibromomethane	0.278	0.287	-3.2	95	0.00
47 T	Bromodichloromethane	0.547	0.585	-6.9	96	0.00
48 T	Methyl methacrylate	0.466	0.519	-11.4	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046100.D
 Acq On : 08 May 2025 18:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 09 01:59:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.010	-11.1	103	0.00
50 S	Toluene-d8	1.246	1.213	2.6	94	0.00
51 T	4-Methyl-2-Pentanone	0.605	0.658	-8.8	98	0.00
52 CM	Toluene	0.869	0.886	-2.0#	93	0.00
53 T	t-1,3-Dichloropropene	0.487	0.507	-4.1	91	0.00
54 T	cis-1,3-Dichloropropene	0.538	0.561	-4.3	92	0.00
55 T	1,1,2-Trichloroethane	0.343	0.353	-2.9	94	0.00
56 T	Ethyl methacrylate	0.546	0.611	-11.9	97	0.00
57 T	1,3-Dichloropropane	0.615	0.623	-1.3	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.278	0.295	-6.1	91	0.00
59 T	2-Hexanone	0.448	0.496	-10.7	99	0.00
60 T	Dibromochloromethane	0.376	0.409	-8.8	97	0.00
61 T	1,2-Dibromoethane	0.356	0.375	-5.3	95	0.00
62 S	4-Bromofluorobenzene	0.478	0.464	2.9	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.354	0.354	0.0	89	0.00
65 PM	Chlorobenzene	1.094	1.092	0.2	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.374	0.385	-2.9	92	0.00
67 C	Ethyl Benzene	1.929	2.015	-4.5#	93	0.00
68 T	m/p-Xylenes	0.706	0.731	-3.5	93	0.00
69 T	o-Xylene	0.688	0.727	-5.7	94	0.00
70 T	Styrene	1.127	1.222	-8.4	94	0.00
71 P	Bromoform	0.281	0.310	-10.3	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	3.893	3.977	-2.2	94	0.00
74 T	N-amyl acetate	1.924	2.076	-7.9	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.364	1.354	0.7	99	0.00
76 T	1,2,3-Trichloropropane	1.204	1.181	1.9	97	0.00
77 T	Bromobenzene	0.904	0.895	1.0	94	0.00
78 T	n-propylbenzene	4.526	4.644	-2.6	94	0.00
79 T	2-Chlorotoluene	2.919	2.855	2.2	93	0.00
80 T	1,3,5-Trimethylbenzene	3.252	3.314	-1.9	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.370	0.351	5.1	89	0.00
82 T	4-Chlorotoluene	3.238	3.305	-2.1	94	0.00
83 T	tert-Butylbenzene	3.276	3.319	-1.3	95	0.00
84 T	1,2,4-Trimethylbenzene	3.293	3.360	-2.0	93	0.00
85 T	sec-Butylbenzene	4.022	4.116	-2.3	94	0.00
86 T	p-Isopropyltoluene	3.320	3.419	-3.0	94	0.00
87 T	1,3-Dichlorobenzene	1.649	1.648	0.1	95	0.00
88 T	1,4-Dichlorobenzene	1.684	1.636	2.9	95	0.00
89 T	n-Butylbenzene	2.912	2.987	-2.6	93	0.00
90 T	Hexachloroethane	0.585	0.612	-4.6	96	0.00
91 T	1,2-Dichlorobenzene	1.655	1.664	-0.5	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.302	0.331	-9.6	101	0.00
93 T	1,2,4-Trichlorobenzene	0.951	0.948	0.3	95	0.00
94 T	Hexachlorobutadiene	0.415	0.403	2.9	92	0.00
95 T	Naphthalene	3.487	3.549	-1.8	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046100.D
Acq On : 08 May 2025 18:46
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 09 01:59:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.981	0.978	0.3	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046100.D
 Acq On : 08 May 2025 18:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 09 01:59:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	50.000	53.944	-7.9	89	0.00
3 P	Chloromethane	50.000	51.301	-2.6	91	0.00
4 C	Vinyl Chloride	50.000	50.457	-0.9#	91	0.00
5 T	Bromomethane	50.000	46.995	6.0	86	0.00
6 T	Chloroethane	50.000	51.135	-2.3	92	0.00
7 T	Trichlorofluoromethane	50.000	51.465	-2.9	91	0.00
8 T	Diethyl Ether	50.000	50.101	-0.2	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.699	-1.4	93	0.00
10 T	Methyl Iodide	50.000	54.149	-8.3	93	0.00
11 T	Tert butyl alcohol	250.000	269.379	-7.8	101	0.00
12 CM	1,1-Dichloroethene	50.000	49.663	0.7#	91	0.00
13 T	Acrolein	250.000	307.775	-23.1	112	0.00
14 T	Allyl chloride	50.000	51.936	-3.9	93	0.00
15 T	Acrylonitrile	250.000	266.549	-6.6	95	0.00
16 T	Acetone	250.000	260.042	-4.0	100	0.00
17 T	Carbon Disulfide	50.000	53.653	-7.3	96	0.00
18 T	Methyl Acetate	50.000	55.300	-10.6	105	0.00
19 T	Methyl tert-butyl Ether	50.000	52.521	-5.0	94	0.00
20 T	Methylene Chloride	50.000	48.019	4.0	93	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.003	-0.0	91	0.00
22 T	Diisopropyl ether	50.000	52.843	-5.7	94	0.00
23 T	Vinyl Acetate	250.000	268.769	-7.5	94	0.00
24 P	1,1-Dichloroethane	50.000	52.275	-4.5	94	0.00
25 T	2-Butanone	250.000	268.026	-7.2	97	0.00
26 T	2,2-Dichloropropane	50.000	40.458	19.1	75	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.607	-3.2	93	0.00
28 T	Bromochloromethane	50.000	49.751	0.5	94	0.00
29 T	Tetrahydrofuran	250.000	270.368	-8.1	97	0.00
30 C	Chloroform	50.000	51.869	-3.7#	94	0.00
31 T	Cyclohexane	50.000	50.347	-0.7	92	0.00
32 T	1,1,1-Trichloroethane	50.000	52.187	-4.4	94	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.526	2.9	92	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	94	0.00
35 S	Dibromofluoromethane	50.000	48.665	2.7	93	0.00
36 T	1,1-Dichloropropene	50.000	49.249	1.5	91	0.00
37 T	Ethyl Acetate	50.000	51.353	-2.7	95	0.00
38 T	Carbon Tetrachloride	50.000	51.466	-2.9	95	0.00
39 T	Methylcyclohexane	50.000	48.747	2.5	90	0.00
40 TM	Benzene	50.000	51.273	-2.5	93	0.00
41 T	Methacrylonitrile	50.000	55.961	-11.9	96	0.00
42 TM	1,2-Dichloroethane	50.000	51.120	-2.2	94	0.00
43 T	Isopropyl Acetate	50.000	54.041	-8.1	97	0.00
44 TM	Trichloroethene	50.000	50.376	-0.8	91	0.00
45 C	1,2-Dichloropropane	50.000	52.950	-5.9#	95	0.00
46 T	Dibromomethane	50.000	51.686	-3.4	95	0.00
47 T	Bromodichloromethane	50.000	53.449	-6.9	96	0.00
48 T	Methyl methacrylate	50.000	55.756	-11.5	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046100.D
 Acq On : 08 May 2025 18:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 09 01:59:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1116.209	-11.6	103	0.00
50 S	Toluene-d8	50.000	48.668	2.7	94	0.00
51 T	4-Methyl-2-Pentanone	250.000	271.972	-8.8	98	0.00
52 CM	Toluene	50.000	50.976	-2.0#	93	0.00
53 T	t-1,3-Dichloropropene	50.000	52.059	-4.1	91	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.127	-4.3	92	0.00
55 T	1,1,2-Trichloroethane	50.000	51.469	-2.9	94	0.00
56 T	Ethyl methacrylate	50.000	55.913	-11.8	97	0.00
57 T	1,3-Dichloropropane	50.000	50.658	-1.3	95	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	264.555	-5.8	91	0.00
59 T	2-Hexanone	250.000	277.107	-10.8	99	0.00
60 T	Dibromochloromethane	50.000	54.358	-8.7	97	0.00
61 T	1,2-Dibromoethane	50.000	52.641	-5.3	95	0.00
62 S	4-Bromofluorobenzene	50.000	48.556	2.9	93	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	94	0.00
64 T	Tetrachloroethene	50.000	50.066	-0.1	89	0.00
65 PM	Chlorobenzene	50.000	49.877	0.2	93	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.521	-3.0	92	0.00
67 C	Ethyl Benzene	50.000	52.220	-4.4#	93	0.00
68 T	m/p-Xylenes	100.000	103.585	-3.6	93	0.00
69 T	o-Xylene	50.000	52.839	-5.7	94	0.00
70 T	Styrene	50.000	54.240	-8.5	94	0.00
71 P	Bromoform	50.000	55.162	-10.3	95	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	98	0.00
73 T	Isopropylbenzene	50.000	51.086	-2.2	94	0.00
74 T	N-ethyl acetate	50.000	53.976	-8.0	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.628	0.7	99	0.00
76 T	1,2,3-Trichloropropane	50.000	49.064	1.9	97	0.00
77 T	Bromobenzene	50.000	49.505	1.0	94	0.00
78 T	n-propylbenzene	50.000	51.305	-2.6	94	0.00
79 T	2-Chlorotoluene	50.000	48.900	2.2	93	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.950	-1.9	93	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.466	5.1	89	0.00
82 T	4-Chlorotoluene	50.000	51.040	-2.1	94	0.00
83 T	tert-Butylbenzene	50.000	50.661	-1.3	95	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.007	-2.0	93	0.00
85 T	sec-Butylbenzene	50.000	51.163	-2.3	94	0.00
86 T	p-Isopropyltoluene	50.000	51.497	-3.0	94	0.00
87 T	1,3-Dichlorobenzene	50.000	49.964	0.1	95	0.00
88 T	1,4-Dichlorobenzene	50.000	48.570	2.9	95	0.00
89 T	n-Butylbenzene	50.000	51.283	-2.6	93	0.00
90 T	Hexachloroethane	50.000	52.349	-4.7	96	0.00
91 T	1,2-Dichlorobenzene	50.000	50.262	-0.5	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	54.760	-9.5	101	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.878	0.2	95	0.00
94 T	Hexachlorobutadiene	50.000	48.483	3.0	92	0.00
95 T	Naphthalene	50.000	50.897	-1.8	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046100.D
Acq On : 08 May 2025 18:46
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 09 01:59:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.854	0.3	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6





QC SAMPLE DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046038.D
Acq On : 05 May 2025 09:37
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Title : SW846 8260
Last Update : Tue May 06 07:12:22 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.1	8676	PASS
75	95	30	60	56.2	22052	PASS
95	95	100	100	100.0	39213	PASS
96	95	5	9	6.4	2505	PASS
173	174	0.00	2	0.7	185	PASS
174	95	50	100	68.8	26963	PASS
175	174	5	9	7.3	1980	PASS
176	174	95	101	97.0	26147	PASS
177	176	5	9	6.9	1802	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046077.D
Acq On : 08 May 2025 09:36
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Title : SW846 8260
Last Update : Tue May 06 07:12:22 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.3	8665	PASS
75	95	30	60	56.5	21989	PASS
95	95	100	100	100.0	38896	PASS
96	95	5	9	6.6	2565	PASS
173	174	0.00	2	1.2	305	PASS
174	95	50	100	67.8	26381	PASS
175	174	5	9	7.5	1967	PASS
176	174	95	101	95.5	25189	PASS
177	176	5	9	7.0	1755	PASS

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0508WBL01		SDG No.:	Q1985
Lab Sample ID:	VX0508WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046080.D	1		05/08/25 10:56	VX050825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	
Client Sample ID:	VX0508WBL01			SDG No.:	Q1985
Lab Sample ID:	VX0508WBL01			Matrix:	Water
Analytical Method:	SW8260			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID :	0.18	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046080.D	1		05/08/25 10:56	VX050825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.5		81 - 118		105%	SPK: 50
1868-53-7	Dibromofluoromethane	49.3		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	49.8		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.3		85 - 114		97%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	66100	5.544				
540-36-3	1,4-Difluorobenzene	132000	6.757				
3114-55-4	Chlorobenzene-d5	121000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	49500	12.018				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046080.D
 Acq On : 08 May 2025 10:56
 Operator : JC/MD
 Sample : VX0508WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBL01

Quant Time: May 09 01:55:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	66087	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	132156	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	121254	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49493	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64719	52.529	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.060%
35) Dibromofluoromethane	5.379	113	46934	49.318	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.640%
50) Toluene-d8	8.647	98	164126	49.828	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.660%
62) 4-Bromofluorobenzene	11.079	95	61080	48.343	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.680%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046080.D
Acq On : 08 May 2025 10:56
Operator : JC/MD
Sample : VX0508WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBL01

Quant Time: May 09 01:55:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration



#1
Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.544 min Scan# 71

Delta R.T. -0.000 min

Lab File: VX046080.D

Acq: 08 May 2025 10:56

Instrument :

MSVOA_X

ClientSampleId :





VX0508WBL01

Tgt Ion:168 Resp: 66087

Ion Ratio Lower Upper

168 100

99 75.3 54.9 82.3



#33

1,2-Dichloroethane-d4

Concen: 52.529 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046080.D




Acq: 08 May 2025 10:56


Tgt Ion: 65 Resp: 64719

Ion Ratio Lower Upper

65 100

67 48.0 0.0 99.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX046080.D
Acq: 08 May 2025 10:56




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBL01


Tgt Ion:114 Resp: 132156
Ion Ratio Lower Upper
114 100
63 22.0 0.0 49.2
88 17.0 0.0 33.6



#35
Dibromofluoromethane
Concen: 49.318 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.000 min
Lab File: VX046080.D
Acq: 08 May 2025 10:56

Tgt Ion:113 Resp: 46934
Ion Ratio Lower Upper
113 100
111 103.5 83.1 124.7
192 17.2 13.3 19.9





#50

Toluene-d8

Concen: 49.828 ug/l

RT: 8.647 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046080.D

Acq: 08 May 2025 10:56

Instrument :

MSVOA_X

ClientSampleId :





VX0508WBL01

Tgt Ion: 98 Resp: 164126

Ion Ratio Lower Upper

98 100

100 64.6 53.5 80.3



#62

4-Bromofluorobenzene

Concen: 48.343 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.000 min

Lab File: VX046080.D

Acq: 08 May 2025 10:56




Tgt Ion: 95 Resp: 61080


Ion Ratio Lower Upper

95 100

174 66.7 0.0 135.8

176 65.4 0.0 131.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.049 min Scan# 1470

Delta R.T. -0.000 min

Lab File: VX046080.D

Acq: 08 May 2025 10:56

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBL01





Tgt Ion:117 Resp: 121254

Ion Ratio Lower Upper

117 100

82 63.1 50.6 76.0

119 31.4 25.8 38.6



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX046080.D

Acq: 08 May 2025 10:56




Tgt Ion:152 Resp: 49493

Ion Ratio Lower Upper

152 100

115 67.2 46.9 140.7

150 160.8 0.0 351.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046080.D
Acq On : 08 May 2025 10:56
Operator : JC/MD
Sample : VX0508WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBL01

Integration Parameters: RTEINT.P
Integrator: RTE
Smoothing : ON Filtering: 5
Sampling : 1 Min Area: 3 % of largest Peak
Start Thrs: 0.2 Max Peaks: 100
Stop Thrs : 0 Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >
Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Title : SW846 8260

Signal : TIC: VX046080.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.569	69	79	81	rBV6	4339	9856	2.16%	0.410%
2	1.593	81	83	87	rVB4	4930	6472	1.42%	0.269%
3	5.385	693	705	720	rBV	53181	159559	34.89%	6.633%
4	5.544	720	731	749	rVV	77071	221822	48.50%	9.221%
5	5.952	789	798	813	rBV2	60491	168126	36.76%	6.989%
6	6.757	921	930	942	rBV	136434	331044	72.38%	13.761%
7	8.647	1234	1240	1261	rBV	284266	457350	100.00%	19.011%
8	10.049	1465	1470	1489	rBV	287672	409247	89.48%	17.012%
9	11.079	1634	1639	1656	rBV	232200	299938	65.58%	12.468%
10	12.018	1788	1793	1809	rBV	274106	342262	74.84%	14.227%




Sum of corrected areas: 2405676

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046080.D
Acq On : 08 May 2025 10:56
Operator : JC/MD
Sample : VX0508WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046080.D
Acq On : 08 May 2025 10:56
Operator : JC/MD
Sample : VX0508WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046080.D
Acq On : 08 May 2025 10:56
Operator : JC/MD
Sample : VX0508WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX0508WBS01		SDG No.:	Q1985
Lab Sample ID:	VX0508WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046081.D	1		05/08/25 11:20	VX050825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	18.6		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	18.5		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	17.9		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	19.3		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.3		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.5		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.4		0.23	0.75	1.00	ug/L
67-64-1	Acetone	93.0		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	17.3		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.3		0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	17.5		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.8		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.4		0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	96.5		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.9		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.5		0.19	0.75	1.00	ug/L
67-66-3	Chloroform	19.8		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.2		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	18.2		0.16	0.50	1.00	ug/L
71-43-2	Benzene	19.4		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	18.9		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	18.6		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.5		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	19.2		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	97.8		0.68	2.50	5.00	ug/L
108-88-3	Toluene	19.1		0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	17.8		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	18.6		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	19.3		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	98.2		0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	
Client Sample ID:	VX0508WBS01	SDG No.:	Q1985
Lab Sample ID:	VX0508WBS01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046081.D	1		05/08/25 11:20	VX050825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	19.3		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	18.9		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	18.8		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.6		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	39.2		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	20.1		0.12	0.50	1.00	ug/L
100-42-5	Styrene	19.9		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	18.7		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	20.2		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	19.5		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.5		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.1		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.1		0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	49.8		81 - 118		100%	SPK: 50
1868-53-7	Dibromofluoromethane	50.2		80 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	49.1		89 - 112		98%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.2		85 - 114		96%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	97700	5.544				
540-36-3	1,4-Difluorobenzene	173000	6.757				
3114-55-4	Chlorobenzene-d5	148000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	68700	12.018				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046081.D
 Acq On : 08 May 2025 11:20
 Operator : JC/MD
 Sample : VX0508WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/09/2025
 Supervised By : Mahesh Dadoda 05/09/2025

Quant Time: May 09 01:55:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	97732	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	172839	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	147787	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	68652	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	90728	49.795	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.580%		
35) Dibromofluoromethane	5.379	113	62462	50.185	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	100.380%		
50) Toluene-d8	8.647	98	211573	49.114	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.220%		
62) 4-Bromofluorobenzene	11.079	95	79658	48.207	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	96.420%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	29711	19.862	ug/l	96
3) Chloromethane	1.313	50	26941	18.572	ug/l	98
4) Vinyl Chloride	1.374	62	24955	18.484	ug/l	99
5) Bromomethane	1.599	94	11222	17.921	ug/l	100
6) Chloroethane	1.672	64	13881	19.260	ug/l	95
7) Trichlorofluoromethane	1.874	101	38496	19.294	ug/l	100
8) Diethyl Ether	2.130	74	12418	18.283	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.319	101	22890	18.538	ug/l	97
10) Methyl Iodide	2.447	142	26968	18.458	ug/l	99
11) Tert butyl alcohol	2.971	59	23936	93.588	ug/l	100
12) 1,1-Dichloroethene	2.312	96	21341	18.416	ug/l	95
13) Acrolein	2.233	56	27935	95.908	ug/l	98
14) Allyl chloride	2.654	41	41968	18.949	ug/l	98
15) Acrylonitrile	3.062	53	71056	97.160	ug/l	99
16) Acetone	2.380	43	67906	92.952	ug/l	98
17) Carbon Disulfide	2.508	76	47650	17.344	ug/l	100
18) Methyl Acetate	2.703	43	33048	19.494	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	78579	19.341	ug/l	98
20) Methylene Chloride	2.788	84	24521	17.515	ug/l	91
21) trans-1,2-Dichloroethene	3.087	96	21957	18.841	ug/l	96
22) Diisopropyl ether	3.763	45	82904	19.378	ug/l	97
23) Vinyl Acetate	3.721	43	357968	95.133	ug/l	99
24) 1,1-Dichloroethane	3.605	63	46330	19.443	ug/l	97
25) 2-Butanone	4.556	43	102390	96.539	ug/l	97
26) 2,2-Dichloropropane	4.465	77	33519	17.972	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	27420	19.544	ug/l	97
28) Bromochloromethane	4.891	49	23913	20.849	ug/l	99
29) Tetrahydrofuran	5.007	42	64605	97.209	ug/l	100
30) Chloroform	5.086	83	49183	19.803	ug/l	99
31) Cyclohexane	5.464	56	41483	19.104	ug/l	95
32) 1,1,1-Trichloroethane	5.379	97	41440	19.248	ug/l	99
36) 1,1-Dichloropropene	5.684	75	30344	18.145	ug/l	99
37) Ethyl Acetate	4.715	43	37823	18.307	ug/l	99
38) Carbon Tetrachloride	5.672	117	35426	18.854	ug/l	98
39) Methylcyclohexane	7.373	83	39142	18.181	ug/l	97
40) Benzene	6.031	78	94792	19.352	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046081.D
 Acq On : 08 May 2025 11:20
 Operator : JC/MD
 Sample : VX0508WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/09/2025
 Supervised By : Mahesh Dadoda 05/09/2025

Quant Time: May 09 01:55:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	22072	20.422	ug/l	97
42) 1,2-Dichloroethane	6.086	62	40044	18.942	ug/l	100
43) Isopropyl Acetate	6.342	43	61195	19.415	ug/l	98
44) Trichloroethene	7.123	130	21873	18.553	ug/l	100
45) 1,2-Dichloropropane	7.427	63	23731	19.484	ug/l	100
46) Dibromomethane	7.580	93	18545	19.305	ug/l	97
47) Bromodichloromethane	7.818	83	36386	19.231	ug/l	96
48) Methyl methacrylate	7.696	41	31721	19.706	ug/l	99
49) 1,4-Dioxane	7.659	88	12727	416.399	ug/l	99
51) 4-Methyl-2-Pentanone	8.574	43	204540	97.762	ug/l	98
52) Toluene	8.714	92	57389	19.107	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	29985	17.830	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	34522	18.573	ug/l	96
55) 1,1,2-Trichloroethane	9.147	97	22879	19.319	ug/l	98
56) Ethyl methacrylate	9.116	69	36760	19.475	ug/l	98
57) 1,3-Dichloropropane	9.305	76	41307	19.421	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	91264	94.841	ug/l	98
59) 2-Hexanone	9.427	43	151942	98.160	ug/l	100
60) Dibromochloromethane	9.519	129	25083	19.285	ug/l	98
61) 1,2-Dibromoethane	9.610	107	23506	19.096	ug/l	96
64) Tetrachloroethene	9.269	164	19802	18.937	ug/l	96
65) Chlorobenzene	10.079	112	60865	18.816	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	20947	18.964	ug/l	98
67) Ethyl Benzene	10.189	91	111800	19.608	ug/l	100
68) m/p-Xylenes	10.299	106	81709	39.181	ug/l	98
69) o-Xylene	10.640	106	40954	20.144	ug/l	98
70) Styrene	10.652	104	66398	19.937	ug/l	99
71) Bromoform	10.799	173	15535	18.733	ug/l #	96
73) Isopropylbenzene	10.957	105	108018	20.210	ug/l	99
74) N-amyl acetate	10.841	43	51097	19.347	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	36452	19.462	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	31838m	19.267	ug/l	
77) Bromobenzene	11.195	156	24434	19.691	ug/l	97
78) n-propylbenzene	11.299	91	123628	19.893	ug/l	99
79) 2-Chlorotoluene	11.360	91	77625	19.365	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	90424	20.251	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	8871	17.477	ug/l	95
82) 4-Chlorotoluene	11.451	91	87403	19.662	ug/l	99
83) tert-Butylbenzene	11.713	119	90081	20.028	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	90858	20.093	ug/l	100
85) sec-Butylbenzene	11.890	105	109461	19.821	ug/l	99
86) p-Isopropyltoluene	12.006	119	89406	19.614	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	44061	19.456	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	44060	19.051	ug/l	99
89) n-Butylbenzene	12.329	91	75618	18.912	ug/l	99
90) Hexachloroethane	12.536	117	15534	19.343	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	45764	20.138	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.939	75	7740	18.654	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	24066	18.438	ug/l	98
94) Hexachlorobutadiene	13.725	225	10658	18.697	ug/l	97
95) Naphthalene	13.774	128	87722	18.324	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	25237	18.739	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046081.D
Acq On : 08 May 2025 11:20
Operator : JC/MD
Sample : VX0508WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025

Quant Time: May 09 01:55:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046081.D
Acq On : 08 May 2025 11:20
Operator : JC/MD
Sample : VX0508WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025

Quant Time: May 09 01:55:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01

Tgt Ion:168 Resp: 97732
Ion Ratio Lower Upper
168 100
99 70.8 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#2
Dichlorodifluoromethane
Concen: 19.862 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 85 Resp: 29711
Ion Ratio Lower Upper
85 100
87 33.4 15.7 47.1









#3
Chloromethane
Concen: 18.572 ug/l
RT: 1.313 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01

Tgt Ion: 50 Resp: 2694
Ion Ratio Lower Upper
50 100
52 32.9 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#4
Vinyl Chloride
Concen: 18.484 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20


Tgt Ion: 62 Resp: 24955
Ion Ratio Lower Upper
62 100
64 31.0 25.2 37.8





#5
Bromomethane
Concen: 17.921 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01



Tgt Ion: 94 Resp: 1122
Ion Ratio Lower Upper
94 100
96 90.7 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#6
Chloroethane
Concen: 19.260 ug/l
RT: 1.672 min Scan# 96
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20


Tgt Ion: 64 Resp: 13881
Ion Ratio Lower Upper
64 100
66 33.1 24.3 36.5





#7
Trichlorofluoromethane
Concen: 19.294 ug/l
RT: 1.874 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20



Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01



Tgt Ion:101 Resp: 38490
Ion Ratio Lower Upper
101 100
103 64.0 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#8
Diethyl Ether
Concen: 18.283 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 74 Resp: 12418
Ion Ratio Lower Upper
74 100
45 108.5 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.538 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01

Tgt Ion:101 Resp: 22890
Ion Ratio Lower Upper
101 100
85 48.2 38.6 58.0
151 73.4 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#10
Methyl Iodide
Concen: 18.458 ug/l
RT: 2.447 min Scan# 223
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion:142 Resp: 26968
Ion Ratio Lower Upper
142 100
127 52.7 41.7 62.5
141 13.8 11.5 17.3





#11

Tert butyl alcohol

Concen: 93.588 ug/l

RT: 2.971 min Scan# 309

Delta R.T. -0.006 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01

Tgt Ion: 59 Resp: 23930

Ion Ratio Lower Upper

59 100





57 10.7 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#12

1,1-Dichloroethene

Concen: 18.416 ug/l

RT: 2.312 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20




Tgt Ion: 96 Resp: 21341


Ion Ratio Lower Upper

96 100

61 190.8 146.2 219.2

98 67.7 51.0 76.6





#13

Acrolein

Concen: 95.908 ug/l

RT: 2.233 min Scan# 188

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01

Tgt Ion: 56 Resp: 27938

Ion Ratio Lower Upper

56 100





55 71.9 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#14

Allyl chloride

Concen: 18.949 ug/l

RT: 2.654 min Scan# 257

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20




Tgt Ion: 41 Resp: 41968


Ion Ratio Lower Upper

41 100

39 74.5 60.6 90.8

76 29.7 24.9 37.3





#15

Acrylonitrile

Concen: 97.160 ug/l

RT: 3.062 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01





Tgt Ion:	53	Resp:	71050
Ion	Ratio	Lower	Upper
53	100		
52	82.8	65.3	97.9
51	37.9	29.8	44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#16

Acetone

Concen: 92.952 ug/l




RT: 2.380 min Scan# 212


Delta R.T. -0.006 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Tgt Ion:	43	Resp:	67906
Ion	Ratio	Lower	Upper
43	100		
58	25.7	21.2	31.8









#17
Carbon Disulfide
Concen: 17.344 ug/l
RT: 2.508 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01

Tgt Ion: 76 Resp: 47650
Ion Ratio Lower Upper
76 100
78 8.6 7.0 10.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#18
Methyl Acetate
Concen: 19.494 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 43 Resp: 33048
Ion Ratio Lower Upper
43 100
74 21.3 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 19.341 ug/l

RT: 3.111 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01

Tgt Ion: 73 Resp: 78579

Ion Ratio Lower Upper

73 100





57 21.3 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#20

Methylene Chloride

Concen: 17.515 ug/l

RT: 2.788 min Scan# 279

Delta R.T. 0.006 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Tgt Ion: 84 Resp: 24521




Ion Ratio Lower Upper


84 100

49 126.7 113.9 170.9

51 38.1 33.5 50.3


86 65.1 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 18.841 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01



Tgt Ion: 96 Resp: 2195
Ion Ratio Lower Upper
96 100
61 154.1 119.5 179.3
98 66.6 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#22
Diisopropyl ether
Concen: 19.378 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20


Tgt Ion: 45 Resp: 82904
Ion Ratio Lower Upper
45 100
43 80.3 66.6 100.0
87 24.8 19.8 29.6
59 11.5 8.6 12.8





#23
Vinyl Acetate
Concen: 95.133 ug/l
RT: 3.721 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01




Tgt Ion: 43 Resp: 35796
Ion Ratio Lower Upper
43 100
86 9.7 7.5 11.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025




#24
1,1-Dichloroethane
Concen: 19.443 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20



Tgt Ion: 63 Resp: 46330
Ion Ratio Lower Upper
63 100
98 5.6 3.6 10.8
100 4.4 2.1 6.3





#25

2-Butanone

Concen: 96.539 ug/l

RT: 4.556 min Scan# 569

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01

Tgt Ion: 43 Resp: 102390

Ion Ratio Lower Upper

43 100





72 21.6 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#26

2,2-Dichloropropane

Concen: 17.972 ug/l

RT: 4.465 min Scan# 554

Delta R.T. -0.000 min

Lab File: VX046081.D




Acq: 08 May 2025 11:20


Tgt Ion: 77 Resp: 33519

Ion Ratio Lower Upper

77 100


97 21.4 10.5 31.5





#27
 cis-1,2-Dichloroethene
 Concen: 19.544 ug/l
 RT: 4.483 min Scan# 557
 Delta R.T. -0.000 min
 Lab File: VX046081.D
 Acq: 08 May 2025 11:20




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBS01



Tgt Ion: 96 Resp: 27420
 Ion Ratio Lower Upper
 96 100
 61 157.7 0.0 322.8
 98 61.6 0.0 129.0




Manual Integrations
 APPROVED


Reviewed By :John Carlone 05/09/2025
 Supervised By :Mahesh Dadoda 05/09/2025



#28
 Bromochloromethane
 Concen: 20.849 ug/l
 RT: 4.891 min Scan# 624
 Delta R.T. -0.006 min
 Lab File: VX046081.D
 Acq: 08 May 2025 11:20

Tgt Ion: 49 Resp: 23913
 Ion Ratio Lower Upper
 49 100
 129 1.7 0.0 4.0
 130 69.4 56.2 84.2





#29

Tetrahydrofuran

Concen: 97.209 ug/l

RT: 5.007 min Scan# 642

Delta R.T. 0.006 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01





Tgt Ion:	42	Resp:	6460
Ion	Ratio	Lower	Upper
42	100		
72	41.8	33.7	50.5
71	39.2	31.4	47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#30

Chloroform

Concen: 19.803 ug/l




RT: 5.086 min Scan# 656


Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Tgt Ion:	83	Resp:	49183
Ion	Ratio	Lower	Upper
83	100		
85	62.7	49.3	73.9





#31

Cyclohexane

Concen: 19.104 ug/l

RT: 5.464 min Scan# 718

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01

Tgt Ion: 56 Resp: 4148

Ion Ratio Lower Upper

56 100

69 31.5 24.4 36.6





84 78.6 66.9 100.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#32

1,1,1-Trichloroethane

Concen: 19.248 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20




Tgt Ion: 97 Resp: 41440


Ion Ratio Lower Upper

97 100

99 64.0 51.8 77.6

61 51.2 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 49.795 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01

Tgt Ion: 65 Resp: 90728

Ion Ratio Lower Upper

65 100





67 50.3 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20




Tgt Ion:114 Resp: 172839


Ion Ratio Lower Upper

114 100

63 24.7 0.0 49.2

88 17.8 0.0 33.6





#35

Dibromofluoromethane

Concen: 50.185 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01

Tgt Ion:113 Resp: 62462

Ion Ratio Lower Upper

113 100

111 102.1 83.1 124.7





192 17.2 13.3 19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#36

1,1-Dichloropropene

Concen: 18.145 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20




Tgt Ion: 75 Resp: 30344


Ion Ratio Lower Upper

75 100

110 32.6 16.3 48.9

77 32.0 24.3 36.5





#37

Ethyl Acetate

Concen: 18.307 ug/l

RT: 4.715 min Scan# 595

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01





Tgt Ion:	43	Resp:	3782
Ion	Ratio	Lower	Upper
43	100		
61	12.8	10.3	15.5
70	10.5	7.9	11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#38

Carbon Tetrachloride

Concen: 18.854 ug/l




RT: 5.672 min Scan# 752


Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Tgt Ion:	117	Resp:	35426
Ion	Ratio	Lower	Upper
117	100		
119	96.0	75.2	112.8
121	29.9	24.2	36.4





#39

Methylcyclohexane

Concen: 18.181 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01





Tgt Ion:	83	Resp:	3914
Ion	Ratio	Lower	Upper
83	100		
55	84.7	64.7	97.1
98	47.8	37.4	56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#40

Benzene

Concen: 19.352 ug/l




RT: 6.031 min Scan# 811


Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20


Tgt Ion:	78	Resp:	94792
Ion	Ratio	Lower	Upper
78	100		
77	23.9	19.0	28.4





#41
Methacrylonitrile
Concen: 20.422 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01



Tgt Ion: 41 Resp: 22072
Ion Ratio Lower Upper
41 100
39 56.0 47.2 70.8
67 65.5 50.7 76.1
52 30.4 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#42
1,2-Dichloroethane
Concen: 18.942 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 62 Resp: 40044
Ion Ratio Lower Upper
62 100
98 7.6 0.0 15.2





#43

Isopropyl Acetate

Concen: 19.415 ug/l

RT: 6.342 min Scan# 861

Delta R.T. 0.006 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01





Tgt Ion:	43	Resp:	6119
Ion	Ratio	Lower	Upper
43	100		
61	17.5	14.3	21.5
87	10.9	9.5	14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#44

Trichloroethene

Concen: 18.553 ug/l




RT: 7.123 min Scan# 990


Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Tgt Ion:	130	Resp:	21873
Ion	Ratio	Lower	Upper
130	100		
95	102.0	0.0	204.2





#45

1,2-Dichloropropane

Concen: 19.484 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01

Tgt Ion: 63 Resp: 2373

Ion Ratio Lower Upper

63 100





65 30.0 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#46

Dibromomethane

Concen: 19.305 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20




Tgt Ion: 93 Resp: 18545


Ion Ratio Lower Upper

93 100

95 80.6 65.6 98.4

174 81.5 68.2 102.2





#47

Bromodichloromethane

Concen: 19.231 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01





Tgt Ion:	83	Resp:	3638
Ion	Ratio	Lower	Upper
83	100		
85	66.4	50.5	75.7
127	8.3	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#48

Methyl methacrylate

Concen: 19.706 ug/l




RT: 7.696 min Scan# 1084


Delta R.T. 0.006 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Tgt Ion:	41	Resp:	31721
Ion	Ratio	Lower	Upper
41	100		
69	72.4	58.5	87.7
39	63.7	51.7	77.5





#49

1,4-Dioxane

Concen: 416.399 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01





Tgt Ion:	88	Resp:	1272
Ion	Ratio	Lower	Upper
88	100		
43	41.1	33.4	50.2
58	73.7	58.6	88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#50

Toluene-d8

Concen: 49.114 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Tgt Ion:	98	Resp:	211573
Ion	Ratio	Lower	Upper
98	100		
100	64.4	53.5	80.3









#51
4-Methyl-2-Pentanone
Concen: 97.762 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01

Tgt Ion: 43 Resp: 204540
Ion Ratio Lower Upper
43 100
58 35.2 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#52
Toluene
Concen: 19.107 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 92 Resp: 57389
Ion Ratio Lower Upper
92 100
91 172.1 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 17.830 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01

Tgt Ion: 75 Resp: 2998

Ion Ratio Lower Upper

75 100





77 30.1 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#54

cis-1,3-Dichloropropene

Concen: 18.573 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20




Tgt Ion: 75 Resp: 34522


Ion Ratio Lower Upper

75 100

77 33.4 25.4 38.0

39 62.1 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 19.319 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01

Tgt Ion: 97 Resp: 22879

Ion Ratio Lower Upper

97 100

83 85.3 69.0 103.6

85 57.4 44.3 66.5





99 61.4 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#56

Ethyl methacrylate

Concen: 19.475 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20




Tgt Ion: 69 Resp: 36760


Ion Ratio Lower Upper

69 100

41 78.1 60.8 91.2

39 48.9 39.0 58.6









#57
1,3-Dichloropropane
Concen: 19.421 ug/l
RT: 9.305 min Scan# 1130
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01

Tgt Ion: 76 Resp: 4130
Ion Ratio Lower Upper
76 100
78 31.8 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#58
2-Chloroethyl Vinyl ether
Concen: 94.841 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 63 Resp: 91264
Ion Ratio Lower Upper
63 100
106 26.1 21.5 32.3





#59

2-Hexanone

Concen: 98.160 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01

Tgt Ion: 43 Resp: 15194

Ion Ratio Lower Upper

43 100





58 49.5 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#60

Dibromochloromethane

Concen: 19.285 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046081.D




Acq: 08 May 2025 11:20


Tgt Ion:129 Resp: 25083

Ion Ratio Lower Upper

129 100

127 77.2 39.3 117.8









#61
1,2-Dibromoethane
Concen: 19.096 ug/l
RT: 9.610 min Scan# 1397
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01

Tgt Ion: 107 Resp: 23500
Ion Ratio Lower Upper
107 100
109 95.4 73.3 109.9




Manual Integrations
APPROVED


Reviewed By : John Carlone 05/09/2025
Supervised By : Mahesh Dadoda 05/09/2025



#62
4-Bromofluorobenzene
Concen: 48.207 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 95 Resp: 79658
Ion Ratio Lower Upper
95 100
174 68.4 0.0 135.8
176 65.4 0.0 131.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.049 min Scan# 1470

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01





Tgt Ion:	117	Resp:	14778
Ion Ratio	Lower	Upper	
117	100		
82	64.6	50.6	76.0
119	32.2	25.8	38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#64

Tetrachloroethene

Concen: 18.937 ug/l




RT: 9.269 min Scan# 1342


Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Tgt Ion:	164	Resp:	19802
Ion Ratio	Lower	Upper	
164	100		
166	124.3	105.0	157.6
129	99.9	83.5	125.3
131	95.2	76.5	114.7









#65
Chlorobenzene
Concen: 18.816 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01

Tgt Ion:112 Resp: 6086
Ion Ratio Lower Upper
112 100
114 31.5 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 18.964 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion:131 Resp: 20947
Ion Ratio Lower Upper
131 100
133 95.7 47.3 141.9
119 65.7 31.6 95.0





#67

Ethyl Benzene

Concen: 19.608 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01

Tgt Ion: 91 Resp: 111800

Ion Ratio Lower Upper

91 100





106 29.3 23.4 35.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#68

m/p-Xylenes

Concen: 39.181 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046081.D




Acq: 08 May 2025 11:20


Tgt Ion:106 Resp: 81709

Ion Ratio Lower Upper

106 100

91 216.4 171.2 256.8









#69
o-Xylene
Concen: 20.144 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01

Tgt Ion:106 Resp: 40954
Ion Ratio Lower Upper
106 100
91 228.6 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#70
Styrene
Concen: 19.937 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion:104 Resp: 66398
Ion Ratio Lower Upper
104 100
78 57.6 45.7 68.5
103 55.5 43.7 65.5





#71

Bromoform

Concen: 18.733 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01

Tgt Ion:173 Resp: 1553

Ion Ratio Lower Upper

173 100

175 49.4 23.4 70.0





254 0.0 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20




Tgt Ion:152 Resp: 68652


Ion Ratio Lower Upper

152 100

115 75.6 46.9 140.7

150 162.3 0.0 351.0





#73

Isopropylbenzene

Concen: 20.210 ug/l

RT: 10.957 min Scan# 1619

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01

Tgt Ion:105 Resp: 108018

Ion Ratio Lower Upper

105 100





120 25.2 12.8 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#74

N-ethyl acetate

Concen: 19.347 ug/l

RT: 10.841 min Scan# 1600

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Tgt Ion: 43 Resp: 51097




Ion Ratio Lower Upper


43 100

70 38.1 30.9 46.3

55 23.4 18.7 28.1

61 22.0 17.1 25.7





#75

1,1,2,2-Tetrachloroethane

Concen: 19.462 ug/l

RT: 11.207 min Scan# 1660

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01





Tgt Ion	Ratio	Lower	Upper
83	100		
131	10.0	5.0	14.9
85	63.8	31.9	95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#76

1,2,3-Trichloropropane

Concen: 19.267 ug/l m




RT: 11.238 min Scan# 1665


Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20


Tgt Ion	Ratio	Lower	Upper
75	100		
77	40.6	20.5	61.5





#77
Bromobenzene
Concen: 19.691 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01




Tgt Ion:156 Resp: 24434
Ion Ratio Lower Upper
156 100
77 181.1 92.8 278.3
158 94.6 48.8 146.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025




#78
n-propylbenzene
Concen: 19.893 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20



Tgt Ion: 91 Resp: 123628
Ion Ratio Lower Upper
91 100
120 22.2 10.8 32.4









#79
2-Chlorotoluene
Concen: 19.365 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01

Tgt Ion: 91 Resp: 77625
Ion Ratio Lower Upper
91 100
126 30.9 15.6 46.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#80
1,3,5-Trimethylbenzene
Concen: 20.251 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20


Tgt Ion:105 Resp: 90424
Ion Ratio Lower Upper
105 100
120 45.7 23.1 69.2





#81
trans-1,4-Dichloro-2-butene
Concen: 17.477 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01



Tgt Ion: 75 Resp: 887
Ion Ratio Lower Upper
75 100
53 132.4 100.5 150.7
89 44.8 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#82
4-Chlorotoluene
Concen: 19.662 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20


Tgt Ion: 91 Resp: 87403
Ion Ratio Lower Upper
91 100
126 26.8 13.3 39.8





#83
tert-Butylbenzene
Concen: 20.028 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01



Tgt Ion:119 Resp: 90083
Ion Ratio Lower Upper
119 100
91 65.4 32.9 98.7
134 23.2 11.4 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#84
1,2,4-Trimethylbenzene
Concen: 20.093 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20


Tgt Ion:105 Resp: 90858
Ion Ratio Lower Upper
105 100
120 42.4 21.2 63.6





#85
sec-Butylbenzene
Concen: 19.821 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01



Tgt Ion:105 Resp: 10946
Ion Ratio Lower Upper
105 100
134 19.1 9.7 29.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#86
p-Isopropyltoluene
Concen: 19.614 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion:119 Resp: 89406
Ion Ratio Lower Upper
119 100
134 25.3 12.5 37.5
91 28.2 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 19.456 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01

Tgt Ion:146 Resp: 44060

Ion Ratio Lower Upper

146 100

111 44.6 22.1 66.3





148 63.2 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#88

1,4-Dichlorobenzene

Concen: 19.051 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20




Tgt Ion:146 Resp: 44060


Ion Ratio Lower Upper

146 100

111 44.3 21.3 63.9

148 63.5 31.9 95.5





#89

n-Butylbenzene

Concen: 18.912 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01

Tgt Ion: 91 Resp: 75618

Ion Ratio Lower Upper

91 100

92 54.3 26.9 80.7





134 23.7 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#90

Hexachloroethane

Concen: 19.343 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046081.D




Acq: 08 May 2025 11:20


Tgt Ion:117 Resp: 15534

Ion Ratio Lower Upper

117 100


201 63.1 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 20.138 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01



Tgt Ion:146 Resp: 45764
Ion Ratio Lower Upper
146 100
111 43.5 22.4 67.3
148 62.4 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 18.654 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 75 Resp: 7740
Ion Ratio Lower Upper
75 100
155 74.7 34.9 104.8
157 88.9 43.8 131.4









#93
1,2,4-Trichlorobenzene
Concen: 18.438 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01

Tgt Ion:180 Resp: 2406
Ion Ratio Lower Upper
180 100
182 93.8 47.9 143.8
145 34.0 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#94
Hexachlorobutadiene
Concen: 18.697 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20


Tgt Ion:225 Resp: 10658
Ion Ratio Lower Upper
225 100
223 60.0 30.8 92.4
227 64.0 30.9 92.7





#95
Naphthalene
Concen: 18.324 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01



Tgt Ion:128 Resp: 8772
Ion Ratio Lower Upper
128 100
127 12.7 10.4 15.6
129 10.9 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#96
1,2,3-Trichlorobenzene
Concen: 18.739 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion:180 Resp: 25237
Ion Ratio Lower Upper
180 100
182 94.2 47.8 143.3
145 35.7 18.1 54.3



Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	
Client Sample ID:	VX0508WBSD01			SDG No.:	Q1985
Lab Sample ID:	VX0508WBSD01			Matrix:	Water
Analytical Method:	SW8260			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID :	0.18	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046082.D	1		05/08/25 11:45	VX050825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	19.2		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	18.0		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	18.0		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	19.0		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.2		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.3		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.3		0.23	0.75	1.00	ug/L
67-64-1	Acetone	98.2		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	17.4		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	20.0		0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	18.4		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	19.1		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.8		0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	100		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.9		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.8		0.19	0.75	1.00	ug/L
67-66-3	Chloroform	19.9		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.6		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	18.5		0.16	0.50	1.00	ug/L
71-43-2	Benzene	19.3		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.2		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.3		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.3		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	19.8		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.68	2.50	5.00	ug/L
108-88-3	Toluene	19.5		0.14	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	19.1		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.9		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.6		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	
Client Sample ID:	VX0508WBSD01	SDG No.:	Q1985
Lab Sample ID:	VX0508WBSD01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046082.D	1		05/08/25 11:45	VX050825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	20.4		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	19.0		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	18.8		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.5		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	39.4		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	19.8		0.12	0.50	1.00	ug/L
100-42-5	Styrene	20.0		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	19.4		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	20.1		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	20.4		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.7		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.5		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.3		0.16	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.9		81 - 118		102%	SPK: 50
1868-53-7	Dibromofluoromethane	50.1		80 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	49.9		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.4		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	89100	5.544				
540-36-3	1,4-Difluorobenzene	158000	6.757				
3114-55-4	Chlorobenzene-d5	138000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	63600	12.018				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046082.D
 Acq On : 08 May 2025 11:45
 Operator : JC/MD
 Sample : VX0508WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBSD01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/09/2025
 Supervised By : Mahesh Dadoda 05/09/2025

Quant Time: May 09 01:55:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	89125	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	157859	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	138205	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	63610	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	84500	50.855	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 101.720%			
35) Dibromofluoromethane	5.379	113	56900	50.055	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.100%			
50) Toluene-d8	8.647	98	196138	49.851	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.700%			
62) 4-Bromofluorobenzene	11.079	95	74503	49.366	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 98.740%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	27046	19.827	ug/l	97
3) Chloromethane	1.306	50	25411	19.209	ug/l	100
4) Vinyl Chloride	1.374	62	22115	17.963	ug/l	94
5) Bromomethane	1.593	94	10264	17.974	ug/l	98
6) Chloroethane	1.672	64	12495	19.011	ug/l	99
7) Trichlorofluoromethane	1.873	101	35016	19.244	ug/l	99
8) Diethyl Ether	2.130	74	11766	18.996	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.325	101	21725	19.293	ug/l	98
10) Methyl Iodide	2.447	142	25437	19.091	ug/l	98
11) Tert butyl alcohol	2.971	59	24333	104.328	ug/l	99
12) 1,1-Dichloroethene	2.312	96	19388	18.346	ug/l	98
13) Acrolein	2.233	56	28048	105.595	ug/l	97
14) Allyl chloride	2.660	41	39668	19.640	ug/l	98
15) Acrylonitrile	3.062	53	70056	105.044	ug/l	99
16) Acetone	2.379	43	65406	98.176	ug/l	99
17) Carbon Disulfide	2.501	76	43556	17.384	ug/l	99
18) Methyl Acetate	2.703	43	32347	20.923	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	74256	20.042	ug/l	98
20) Methylene Chloride	2.782	84	23482	18.393	ug/l	93
21) trans-1,2-Dichloroethene	3.087	96	20339	19.138	ug/l	94
22) Diisopropyl ether	3.757	45	78706	20.174	ug/l	94
23) Vinyl Acetate	3.721	43	350984	102.285	ug/l	99
24) 1,1-Dichloroethane	3.605	63	42920	19.752	ug/l	98
25) 2-Butanone	4.556	43	101125	104.554	ug/l	100
26) 2,2-Dichloropropane	4.471	77	31214	18.352	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	25319	19.790	ug/l	98
28) Bromochloromethane	4.891	49	22526	21.536	ug/l	99
29) Tetrahydrofuran	5.007	42	64907	107.095	ug/l	99
30) Chloroform	5.080	83	45164	19.940	ug/l	97
31) Cyclohexane	5.458	56	38562	19.474	ug/l	97
32) 1,1,1-Trichloroethane	5.373	97	38497	19.608	ug/l	99
36) 1,1-Dichloropropene	5.690	75	28723	18.806	ug/l	98
37) Ethyl Acetate	4.714	43	37116	19.669	ug/l	99
38) Carbon Tetrachloride	5.672	117	32427	18.896	ug/l	98
39) Methylcyclohexane	7.372	83	36280	18.451	ug/l	99
40) Benzene	6.031	78	86138	19.254	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046082.D
 Acq On : 08 May 2025 11:45
 Operator : JC/MD
 Sample : VX0508WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBSD01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/09/2025
 Supervised By : Mahesh Dadoda 05/09/2025

Quant Time: May 09 01:55:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	21603	21.885	ug/l	99
42) 1,2-Dichloroethane	6.080	62	38922	20.158	ug/l	100
43) Isopropyl Acetate	6.342	43	59168	20.553	ug/l	99
44) Trichloroethene	7.123	130	20766	19.286	ug/l	98
45) 1,2-Dichloropropane	7.427	63	22527	20.250	ug/l	99
46) Dibromomethane	7.580	93	17952	20.461	ug/l	96
47) Bromodichloromethane	7.818	83	34186	19.783	ug/l	99
48) Methyl methacrylate	7.689	41	30850	20.983	ug/l	99
49) 1,4-Dioxane	7.659	88	12525	448.677	ug/l	98
51) 4-Methyl-2-Pentanone	8.567	43	201237	105.311	ug/l	98
52) Toluene	8.714	92	53498	19.502	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	29261	19.050	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	33824	19.924	ug/l	97
55) 1,1,2-Trichloroethane	9.147	97	22314	20.629	ug/l	99
56) Ethyl methacrylate	9.116	69	35662	20.686	ug/l	99
57) 1,3-Dichloropropane	9.305	76	39386	20.275	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	87699	99.784	ug/l	99
59) 2-Hexanone	9.427	43	149616	105.830	ug/l	100
60) Dibromochloromethane	9.518	129	24175	20.351	ug/l	94
61) 1,2-Dibromoethane	9.610	107	22397	19.922	ug/l	97
64) Tetrachloroethene	9.268	164	18567	18.987	ug/l	97
65) Chlorobenzene	10.079	112	56955	18.828	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	19827	19.195	ug/l	98
67) Ethyl Benzene	10.189	91	103863	19.478	ug/l	99
68) m/p-Xylenes	10.299	106	76787	39.374	ug/l	99
69) o-Xylene	10.640	106	37621	19.787	ug/l	98
70) Styrene	10.652	104	62355	20.021	ug/l	100
71) Bromoform	10.799	173	15034	19.386	ug/l #	96
73) Isopropylbenzene	10.957	105	99423	20.076	ug/l	99
74) N-amyl acetate	10.841	43	49759	20.334	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	35488	20.449	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	29936m	19.552	ug/l	
77) Bromobenzene	11.195	156	22147	19.263	ug/l	99
78) n-propylbenzene	11.299	91	115197	20.006	ug/l	100
79) 2-Chlorotoluene	11.360	91	73340	19.747	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	84507	20.426	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	8472	18.014	ug/l	94
82) 4-Chlorotoluene	11.451	91	81988	19.906	ug/l	99
83) tert-Butylbenzene	11.713	119	84407	20.254	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	85289	20.357	ug/l	98
85) sec-Butylbenzene	11.890	105	102105	19.955	ug/l	99
86) p-Isopropyltoluene	12.006	119	84675	20.048	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	41242	19.655	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	41693	19.456	ug/l	97
89) n-Butylbenzene	12.329	91	72472	19.562	ug/l	99
90) Hexachloroethane	12.536	117	14264	19.170	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	42845	20.348	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.939	75	8274	21.521	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	23756	19.643	ug/l	98
94) Hexachlorobutadiene	13.719	225	10435	19.756	ug/l	98
95) Naphthalene	13.774	128	89648	20.211	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	24789	19.865	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046082.D
Acq On : 08 May 2025 11:45
Operator : JC/MD
Sample : VX0508WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025

Quant Time: May 09 01:55:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046082.D
 Acq On : 08 May 2025 11:45
 Operator : JC/MD
 Sample : VX0508WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1


Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBSD01

Manual Integrations APPROVED

Reviewed By : John Carlone 05/09/2025
 Supervised By : Mahesh Dadoda 05/09/2025

Quant Time: May 09 01:55:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 731
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01

Tgt Ion:168 Resp: 89125
Ion Ratio Lower Upper
168 100
99 72.8 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#2
Dichlorodifluoromethane
Concen: 19.827 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 85 Resp: 27046
Ion Ratio Lower Upper
85 100
87 32.9 15.7 47.1









#3
Chloromethane
Concen: 19.209 ug/l
RT: 1.306 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01

Tgt Ion: 50 Resp: 2541
Ion Ratio Lower Upper
50 100
52 31.9 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#4
Vinyl Chloride
Concen: 17.963 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45


Tgt Ion: 62 Resp: 22115
Ion Ratio Lower Upper
62 100
64 34.6 25.2 37.8





#5
Bromomethane
Concen: 17.974 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01



Tgt Ion: 94 Resp: 10264
Ion Ratio Lower Upper
94 100
96 93.3 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#6
Chloroethane
Concen: 19.011 ug/l
RT: 1.672 min Scan# 96
Delta R.T. 0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45


Tgt Ion: 64 Resp: 12495
Ion Ratio Lower Upper
64 100
66 31.0 24.3 36.5





#7
Trichlorofluoromethane
Concen: 19.244 ug/l
RT: 1.873 min Scan# 110
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01



Tgt Ion:101 Resp: 35010
Ion Ratio Lower Upper
101 100
103 64.7 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#8
Diethyl Ether
Concen: 18.996 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 74 Resp: 11766
Ion Ratio Lower Upper
74 100
45 109.6 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.293 ug/l
RT: 2.325 min Scan# 203
Delta R.T. 0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01

Tgt Ion:101 Resp: 21725
Ion Ratio Lower Upper
101 100
85 46.6 38.6 58.0
151 69.7 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#10
Methyl Iodide
Concen: 19.091 ug/l
RT: 2.447 min Scan# 223
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion:142 Resp: 25437
Ion Ratio Lower Upper
142 100
127 53.3 41.7 62.5
141 14.0 11.5 17.3





#11

Tert butyl alcohol

Concen: 104.328 ug/l

RT: 2.971 min Scan# 309

Delta R.T. -0.006 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 59 Resp: 2433

Ion Ratio Lower Upper

59 100





57 11.2 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#12

1,1-Dichloroethene

Concen: 18.346 ug/l

RT: 2.312 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45




Tgt Ion: 96 Resp: 19388


Ion Ratio Lower Upper

96 100

61 183.0 146.2 219.2

98 58.7 51.0 76.6





#13

Acrolein

Concen: 105.595 ug/l

RT: 2.233 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 56 Resp: 2804

Ion Ratio Lower Upper

56 100





55 67.9 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#14

Allyl chloride

Concen: 19.640 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45




Tgt Ion: 41 Resp: 39668


Ion Ratio Lower Upper

41 100

39 74.0 60.6 90.8


76 29.5 24.9 37.3





#15
Acrylonitrile
Concen: 105.044 ug/l
RT: 3.062 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01



Tgt Ion: 53 Resp: 70050
Ion Ratio Lower Upper
53 100
52 81.2 65.3 97.9
51 38.0 29.8 44.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#16
Acetone
Concen: 98.176 ug/l
RT: 2.379 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45


Tgt Ion: 43 Resp: 65406
Ion Ratio Lower Upper
43 100
58 27.0 21.2 31.8





#17
Carbon Disulfide
Concen: 17.384 ug/l
RT: 2.501 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01



Tgt Ion: 76 Resp: 43550
Ion Ratio Lower Upper
76 100
78 8.9 7.0 10.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#18
Methyl Acetate
Concen: 20.923 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 43 Resp: 32347
Ion Ratio Lower Upper
43 100
74 21.8 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 20.042 ug/l

RT: 3.111 min Scan# 311

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 73 Resp: 74250

Ion Ratio Lower Upper

73 100





57 23.1 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#20

Methylene Chloride

Concen: 18.393 ug/l

RT: 2.782 min Scan# 278

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Tgt Ion: 84 Resp: 23482




Ion Ratio Lower Upper


84 100

49 133.7 113.9 170.9

51 40.0 33.5 50.3

86 58.7 53.8 80.8





#21

trans-1,2-Dichloroethene

Concen: 19.138 ug/l

RT: 3.087 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 96 Resp: 20339

Ion Ratio Lower Upper

96 100

61 139.1 119.5 179.3





98 61.2 50.0 75.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#22

Diisopropyl ether

Concen: 20.174 ug/l

RT: 3.757 min Scan# 438

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Tgt Ion: 45 Resp: 78706




Ion Ratio Lower Upper


45 100

43 90.4 66.6 100.0

87 23.5 19.8 29.6

59 11.1 8.6 12.8









#23
Vinyl Acetate
Concen: 102.285 ug/l
RT: 3.721 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01

Tgt Ion: 43 Resp: 350984
Ion Ratio Lower Upper
43 100
86 9.8 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#24
1,1-Dichloroethane
Concen: 19.752 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 63 Resp: 42920
Ion Ratio Lower Upper
63 100
98 6.6 3.6 10.8
100 3.7 2.1 6.3





#25

2-Butanone

Concen: 104.554 ug/l

RT: 4.556 min Scan# 569

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 43 Resp: 10112

Ion Ratio Lower Upper

43 100





72 23.2 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#26

2,2-Dichloropropane

Concen: 18.352 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046082.D




Acq: 08 May 2025 11:45


Tgt Ion: 77 Resp: 31214

Ion Ratio Lower Upper

77 100

97 21.3 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 19.790 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 96 Resp: 25319

Ion Ratio Lower Upper

96 100

61 157.4 0.0 322.8





98 64.2 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#28

Bromochloromethane

Concen: 21.536 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45




Tgt Ion: 49 Resp: 22526


Ion Ratio Lower Upper

49 100

129 1.5 0.0 4.0

130 69.6 56.2 84.2





#29

Tetrahydrofuran

Concen: 107.095 ug/l

RT: 5.007 min Scan# 642

Delta R.T. 0.006 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01





Tgt Ion:	42	Resp:	6490
Ion Ratio	Lower	Upper	
42	100		
72	41.0	33.7	50.5
71	38.8	31.4	47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#30

Chloroform

Concen: 19.940 ug/l




RT: 5.080 min Scan# 655


Delta R.T. -0.006 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45


Tgt Ion:	83	Resp:	45164
Ion Ratio	Lower	Upper	
83	100		
85	59.7	49.3	73.9





#31
Cyclohexane
Concen: 19.474 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01



Tgt Ion: 56 Resp: 3856
Ion Ratio Lower Upper
56 100
69 29.2 24.4 36.6
84 86.6 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#32
1,1,1-Trichloroethane
Concen: 19.608 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 97 Resp: 38497
Ion Ratio Lower Upper
97 100
99 63.4 51.8 77.6
61 50.4 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 50.855 ug/l

RT: 5.946 min Scan# 798

Delta R.T. -0.006 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 65 Resp: 84500

Ion Ratio Lower Upper

65 100





67 49.9 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45




Tgt Ion:114 Resp: 157859


Ion Ratio Lower Upper

114 100

63 23.9 0.0 49.2

88 16.9 0.0 33.6





#35

Dibromofluoromethane

Concen: 50.055 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion:113 Resp: 56900

Ion Ratio Lower Upper

113 100

111 103.8 83.1 124.7





192 16.8 13.3 19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#36

1,1-Dichloropropene

Concen: 18.806 ug/l

RT: 5.690 min Scan# 755

Delta R.T. 0.006 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45




Tgt Ion: 75 Resp: 28723


Ion Ratio Lower Upper

75 100

110 31.9 16.3 48.9

77 31.5 24.3 36.5





#37

Ethyl Acetate

Concen: 19.669 ug/l

RT: 4.714 min Scan# 595

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 43 Resp: 37110





Ion	Ratio	Lower	Upper
43	100		
61	13.4	10.3	15.5
70	9.9	7.9	11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#38

Carbon Tetrachloride

Concen: 18.896 ug/l

RT: 5.672 min Scan# 752




Delta R.T. -0.000 min


Lab File: VX046082.D

Acq: 08 May 2025 11:45

Tgt Ion: 117 Resp: 32427

Ion	Ratio	Lower	Upper
117	100		
119	95.5	75.2	112.8
121	28.2	24.2	36.4





#39

Methylcyclohexane

Concen: 18.451 ug/l

RT: 7.372 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 83 Resp: 36280

Ion Ratio Lower Upper

83 100

55 81.7 64.7 97.1





98 47.9 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#40

Benzene

Concen: 19.254 ug/l

RT: 6.031 min Scan# 811

Delta R.T. -0.000 min

Lab File: VX046082.D




Acq: 08 May 2025 11:45


Tgt Ion: 78 Resp: 86138

Ion Ratio Lower Upper

78 100

77 22.9 19.0 28.4





#41

Methacrylonitrile

Concen: 21.885 ug/l

RT: 4.916 min Scan# 61

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 41 Resp: 2160

Ion Ratio Lower Upper

41 100

39 58.4 47.2 70.8

67 64.9 50.7 76.1





52 30.4 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#42

1,2-Dichloroethane

Concen: 20.158 ug/l

RT: 6.080 min Scan# 819

Delta R.T. -0.000 min

Lab File: VX046082.D




Acq: 08 May 2025 11:45


Tgt Ion: 62 Resp: 38922

Ion Ratio Lower Upper

62 100

98 7.7 0.0 15.2





#43

Isopropyl Acetate

Concen: 20.553 ug/l

RT: 6.342 min Scan# 861

Delta R.T. 0.006 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01





Tgt Ion: 43	Resp: 5916
Ion Ratio	Lower Upper
43 100	
61 17.6	14.3 21.5
87 11.8	9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#44

Trichloroethene

Concen: 19.286 ug/l




RT: 7.123 min Scan# 990


Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Tgt Ion:130	Resp: 20766
Ion Ratio	Lower Upper
130 100	
95 104.4	0.0 204.2





#45

1,2-Dichloropropane

Concen: 20.250 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 63 Resp: 2252

Ion Ratio Lower Upper

63 100





65 29.5 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#46

Dibromomethane

Concen: 20.461 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45




Tgt Ion: 93 Resp: 17952


Ion Ratio Lower Upper

93 100

95 79.0 65.6 98.4

174 80.4 68.2 102.2





#47

Bromodichloromethane

Concen: 19.783 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01





Tgt Ion: 83	Resp: 3418
Ion Ratio	Lower Upper
83 100	
85 62.3	50.5 75.7
127 8.3	6.5 9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#48

Methyl methacrylate

Concen: 20.983 ug/l




RT: 7.689 min Scan# 1083


Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Tgt Ion: 41	Resp: 30850
Ion Ratio	Lower Upper
41 100	
69 72.6	58.5 87.7
39 65.1	51.7 77.5





#49

1,4-Dioxane

Concen: 448.677 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 88 Resp: 1252

Ion Ratio Lower Upper

88 100

43 42.4 33.4 50.2





58 75.7 58.6 88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#50

Toluene-d8

Concen: 49.851 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046082.D




Acq: 08 May 2025 11:45


Tgt Ion: 98 Resp: 196138

Ion Ratio Lower Upper

98 100

100 64.5 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 105.311 ug/l
RT: 8.567 min Scan# 1123
Delta R.T. -0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01

Tgt Ion: 43 Resp: 20123
Ion Ratio Lower Upper
43 100
58 35.0 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#52
Toluene
Concen: 19.502 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 92 Resp: 53498
Ion Ratio Lower Upper
92 100
91 171.6 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 19.050 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 75 Resp: 2926

Ion Ratio Lower Upper

75 100





77 31.9 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#54

cis-1,3-Dichloropropene

Concen: 19.924 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45




Tgt Ion: 75 Resp: 33824


Ion Ratio Lower Upper

75 100

77 32.9 25.4 38.0

39 62.9 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 20.629 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 97 Resp: 22314

Ion Ratio Lower Upper

97 100

83 85.5 69.0 103.6

85 55.9 44.3 66.5





99 59.8 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#56

Ethyl methacrylate

Concen: 20.686 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45




Tgt Ion: 69 Resp: 35662


Ion Ratio Lower Upper

69 100

41 77.4 60.8 91.2

39 48.8 39.0 58.6





#57

1,3-Dichloropropane

Concen: 20.275 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 76 Resp: 3938

Ion Ratio Lower Upper

76 100





78 31.6 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#58

2-Chloroethyl Vinyl ether

Concen: 99.784 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046082.D




Acq: 08 May 2025 11:45


Tgt Ion: 63 Resp: 87699

Ion Ratio Lower Upper

63 100

106 27.5 21.5 32.3





#59

2-Hexanone

Concen: 105.830 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 43 Resp: 149610

Ion Ratio Lower Upper

43 100





58 49.7 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#60

Dibromochloromethane

Concen: 20.351 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046082.D




Acq: 08 May 2025 11:45


Tgt Ion:129 Resp: 24175

Ion Ratio Lower Upper

129 100

127 73.6 39.3 117.8









#61
1,2-Dibromoethane
Concen: 19.922 ug/l
RT: 9.610 min Scan# 1397
Delta R.T. 0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01

Tgt Ion:107 Resp: 2239
Ion Ratio Lower Upper
107 100
109 94.7 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#62
4-Bromofluorobenzene
Concen: 49.366 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45


Tgt Ion: 95 Resp: 74503
Ion Ratio Lower Upper
95 100
174 67.0 0.0 135.8
176 64.0 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01



Tgt Ion:117 Resp: 13820
Ion Ratio Lower Upper
117 100
82 61.6 50.6 76.0
119 31.5 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#64
Tetrachloroethene
Concen: 18.987 ug/l
RT: 9.268 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45


Tgt Ion:164 Resp: 18567
Ion Ratio Lower Upper
164 100
166 124.3 105.0 157.6
129 103.7 83.5 125.3
131 97.1 76.5 114.7





#65
Chlorobenzene
Concen: 18.828 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01



Tgt Ion:112 Resp: 5695
Ion Ratio Lower Upper
112 100
114 32.1 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 19.195 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45


Tgt Ion:131 Resp: 19827
Ion Ratio Lower Upper
131 100
133 95.8 47.3 141.9
119 65.5 31.6 95.0





#67
Ethyl Benzene
Concen: 19.478 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01



Tgt Ion: 91 Resp: 10386
Ion Ratio Lower Upper
91 100
106 28.9 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#68
m/p-Xylenes
Concen: 39.374 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion:106 Resp: 76787
Ion Ratio Lower Upper
106 100
91 215.6 171.2 256.8









#69
o-Xylene
Concen: 19.787 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01

Tgt Ion:106 Resp: 3762
Ion Ratio Lower Upper
106 100
91 228.7 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#70
Styrene
Concen: 20.021 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion:104 Resp: 62355
Ion Ratio Lower Upper
104 100
78 57.8 45.7 68.5
103 54.6 43.7 65.5





#71

Bromoform

Concen: 19.386 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01





Tgt Ion:	173	Resp:	15034
Ion Ratio	Lower	Upper	
173	100		
175	49.3	23.4	70.0
254	0.0	0.0	0.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.018 min Scan# 1793


Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Tgt Ion:	152	Resp:	63610
Ion Ratio	Lower	Upper	
152	100		
115	77.1	46.9	140.7
150	164.6	0.0	351.0









#73
Isopropylbenzene
Concen: 20.076 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01

Tgt Ion:105 Resp: 9942
Ion Ratio Lower Upper
105 100
120 26.0 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#74
N-ethyl acetate
Concen: 20.334 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 43 Resp: 49759
Ion Ratio Lower Upper
43 100
70 38.0 30.9 46.3
55 23.7 18.7 28.1
61 21.4 17.1 25.7





#75

1,1,2,2-Tetrachloroethane

Concen: 20.449 ug/l

RT: 11.207 min Scan# 1660

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 83 Resp: 3548

Ion Ratio Lower Upper

83 100

131 9.5 5.0 14.9





85 65.6 31.9 95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#76

1,2,3-Trichloropropane

Concen: 19.552 ug/l m

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX046082.D




Acq: 08 May 2025 11:45


Tgt Ion: 75 Resp: 29936

Ion Ratio Lower Upper

75 100

77 42.3 20.5 61.5





#77

Bromobenzene

Concen: 19.263 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01





Tgt Ion:	156	Resp:	2214
Ion	Ratio	Lower	Upper
156	100		
77	186.2	92.8	278.3
158	100.9	48.8	146.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#78

n-propylbenzene

Concen: 20.006 ug/l




RT: 11.299 min Scan# 1675


Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Tgt Ion:	91	Resp:	115197
Ion	Ratio	Lower	Upper
91	100		
120	21.5	10.8	32.4





#79

2-Chlorotoluene

Concen: 19.747 ug/l

RT: 11.360 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 91 Resp: 73340

Ion Ratio Lower Upper

91 100





126 30.7 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#80

1,3,5-Trimethylbenzene

Concen: 20.426 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046082.D




Acq: 08 May 2025 11:45


Tgt Ion:105 Resp: 84507

Ion Ratio Lower Upper

105 100

120 46.1 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 18.014 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01

Tgt Ion: 75 Resp: 8472
Ion Ratio Lower Upper
75 100
53 134.7 100.5 150.7
89 47.5 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#82
4-Chlorotoluene
Concen: 19.906 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45


Tgt Ion: 91 Resp: 81988
Ion Ratio Lower Upper
91 100
126 26.8 13.3 39.8





#83
tert-Butylbenzene
Concen: 20.254 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01



Tgt Ion:119 Resp: 8440
Ion Ratio Lower Upper
119 100
91 66.9 32.9 98.7
134 22.5 11.4 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#84
1,2,4-Trimethylbenzene
Concen: 20.357 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45


Tgt Ion:105 Resp: 85289
Ion Ratio Lower Upper
105 100
120 41.3 21.2 63.6





#85
 sec-Butylbenzene
 Concen: 19.955 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046082.D
 Acq: 08 May 2025 11:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBSD01




Tgt Ion:105 Resp: 10210
 Ion Ratio Lower Upper
 105 100
 134 19.1 9.7 29.1

Manual Integrations
 APPROVED



Reviewed By :John Carlone 05/09/2025
 Supervised By :Mahesh Dadoda 05/09/2025




#86
 p-Isopropyltoluene
 Concen: 20.048 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046082.D
 Acq: 08 May 2025 11:45




Tgt Ion:119 Resp: 84675
 Ion Ratio Lower Upper
 119 100
 134 25.4 12.5 37.5
 91 27.7 13.8 41.4





#87
1,3-Dichlorobenzene
Concen: 19.655 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01



Tgt Ion:146 Resp: 41242
Ion Ratio Lower Upper
146 100
111 44.3 22.1 66.3
148 63.9 32.1 96.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#88
1,4-Dichlorobenzene
Concen: 19.456 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion:146 Resp: 41693
Ion Ratio Lower Upper
146 100
111 44.4 21.3 63.9
148 66.1 31.9 95.5





#89

n-Butylbenzene

Concen: 19.562 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01

Tgt Ion: 91 Resp: 7247

Ion Ratio Lower Upper

91 100

92 53.0 26.9 80.7





134 24.6 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/09/2025

Supervised By :Mahesh Dadoda 05/09/2025



#90

Hexachloroethane

Concen: 19.170 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046082.D




Acq: 08 May 2025 11:45


Tgt Ion:117 Resp: 14264

Ion Ratio Lower Upper

117 100


201 62.0 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 20.348 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01



Tgt Ion:146 Resp: 4284
Ion Ratio Lower Upper
146 100
111 46.7 22.4 67.3
148 63.8 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 21.521 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45


Tgt Ion: 75 Resp: 8274
Ion Ratio Lower Upper
75 100
155 68.1 34.9 104.8
157 85.6 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 19.643 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01




Tgt Ion:180 Resp: 23750
Ion Ratio Lower Upper
180 100
182 93.9 47.9 143.8
145 34.9 17.1 51.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025




#94
Hexachlorobutadiene
Concen: 19.756 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Tgt Ion:225 Resp: 10435
Ion Ratio Lower Upper
225 100
223 63.7 30.8 92.4
227 62.9 30.9 92.7





#95
Naphthalene
Concen: 20.211 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01



Tgt Ion:128 Resp: 89648
Ion Ratio Lower Upper
128 100
127 13.2 10.4 15.6
129 10.8 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/09/2025
Supervised By :Mahesh Dadoda 05/09/2025



#96
1,2,3-Trichlorobenzene
Concen: 19.865 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion:180 Resp: 24789
Ion Ratio Lower Upper
180 100
182 94.3 47.8 143.3
145 36.0 18.1 54.3



Manual Integration Report

Sequence:	VX050525	Instrument	MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC020	VX046041.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:13 AM	MMDadoda	5/6/2025 12:42:46 PM	Peak Integrated by Software
VSTDICCC050	VX046042.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:18 AM	MMDadoda	5/6/2025 12:42:48 PM	Peak Integrated by Software
VSTDICC100	VX046043.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:22 AM	MMDadoda	5/6/2025 12:42:50 PM	Peak Integrated by Software
VSTDICC150	VX046044.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:27 AM	MMDadoda	5/6/2025 12:42:53 PM	Peak Integrated by Software
VSTDICC005	VX046046.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:32 AM	MMDadoda	5/6/2025 12:42:56 PM	Peak Integrated by Software
VSTDICC005	VX046046.D	Ethyl Acetate	JOHN	5/6/2025 9:53:32 AM	MMDadoda	5/6/2025 12:42:56 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	1,4-Dichlorobenzene	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	Bromochloromethane	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	Ethyl Acetate	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	Methyl methacrylate	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICV050	VX046048.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:45 AM	MMDadoda	5/6/2025 12:41:37 PM	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

VX050525

Instrument

MSVOA_x

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:

vx050825

Instrument

MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VX046078.D	1,2,3-Trichloropropane	JOHN	5/9/2025 9:51:48 AM	MMDadoda	5/9/2025 2:53:27 PM	Peak Integrated by Software
VX0508WBS01	VX046081.D	1,2,3-Trichloropropane	JOHN	5/9/2025 9:51:53 AM	MMDadoda	5/9/2025 2:53:28 PM	Peak Integrated by Software
VX0508WBSD0 1	VX046082.D	1,2,3-Trichloropropane	JOHN	5/9/2025 9:51:57 AM	MMDadoda	5/9/2025 2:53:30 PM	Peak Integrated by Software
VSTDCCC050	VX046100.D	1,2,3-Trichloropropane	JOHN	5/9/2025 9:52:41 AM	MMDadoda	5/9/2025 2:53:41 PM	Peak Integrated by Software

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX050525

Review By	John Carlone	Review On	5/6/2025 9:53:58 AM
Supervise By	Maresh Dadoda	Supervise On	5/6/2025 12:43:00 PM
SubDirectory	VX050525	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133811		
Initial Calibration Stds	VP133832,VP133833,VP133834,VP133835,VP133836,VP133837		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP133838		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX046038.D	05 May 2025 09:37	JC/MD	Ok
2	VSTDIC001	VX046039.D	05 May 2025 10:49	JC/MD	Not Ok
3	VSTDIC005	VX046040.D	05 May 2025 11:12	JC/MD	Not Ok
4	VSTDIC020	VX046041.D	05 May 2025 11:35	JC/MD	Ok,M
5	VSTDIC050	VX046042.D	05 May 2025 11:58	JC/MD	Ok,M
6	VSTDIC100	VX046043.D	05 May 2025 12:21	JC/MD	Ok,M
7	VSTDIC150	VX046044.D	05 May 2025 12:45	JC/MD	Ok,M
8	IBLK	VX046045.D	05 May 2025 13:08	JC/MD	Ok
9	VSTDIC005	VX046046.D	05 May 2025 16:04	JC/MD	Ok,M
10	VSTDIC001	VX046047.D	05 May 2025 16:27	JC/MD	Ok,M
11	VSTDICV050	VX046048.D	05 May 2025 16:50	JC/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX050825

Review By	John Carlone	Review On	5/9/2025 9:54:47 AM
Supervise By	Maresh Dadoda	Supervise On	5/9/2025 2:53:48 PM
SubDirectory	VX050825	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133850		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133851,VP133852		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX046077.D	08 May 2025 09:36	JC/MD	Ok
2	VSTDCCC050	VX046078.D	08 May 2025 10:05	JC/MD	Ok,M
3	VX0508MBL01	VX046079.D	08 May 2025 10:33	JC/MD	Ok
4	VX0508WBL01	VX046080.D	08 May 2025 10:56	JC/MD	Ok
5	VX0508WBS01	VX046081.D	08 May 2025 11:20	JC/MD	Ok,M
6	VX0508WBSD01	VX046082.D	08 May 2025 11:45	JC/MD	Ok,M
7	IBLK	VX046083.D	08 May 2025 12:08	JC/MD	Ok
8	Q1985-02	VX046084.D	08 May 2025 12:32	JC/MD	Ok
9	Q1985-01	VX046085.D	08 May 2025 12:55	JC/MD	Ok
10	PB167895TB	VX046086.D	08 May 2025 13:18	JC/MD	Ok,M
11	Q1964-04	VX046087.D	08 May 2025 13:42	JC/MD	Ok
12	Q1968-04	VX046088.D	08 May 2025 14:05	JC/MD	Ok,M
13	Q1968-08	VX046089.D	08 May 2025 14:29	JC/MD	Ok
14	Q1972-04	VX046090.D	08 May 2025 14:52	JC/MD	Ok,M
15	Q1972-08	VX046091.D	08 May 2025 15:15	JC/MD	Ok
16	Q1972-12	VX046092.D	08 May 2025 15:39	JC/MD	Ok
17	Q1972-16	VX046093.D	08 May 2025 16:02	JC/MD	Ok,M
18	Q1972-20	VX046094.D	08 May 2025 16:26	JC/MD	Ok
19	Q1972-24	VX046095.D	08 May 2025 16:49	JC/MD	Ok
20	VX0508MBS01	VX046096.D	08 May 2025 17:12	JC/MD	Ok,M
21	Q1966-03	VX046097.D	08 May 2025 17:36	JC/MD	Ok

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX050825

Review By	John Carlone	Review On	5/9/2025 9:54:47 AM
Supervise By	Mahesh Dadoda	Supervise On	5/9/2025 2:53:48 PM
SubDirectory	VX050825	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133850		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133851,VP133852		

22	IBLK	VX046098.D	08 May 2025 17:59	JC/MD	Ok
23	IBLK	VX046099.D	08 May 2025 18:23	JC/MD	Ok
24	VSTDCCC050	VX046100.D	08 May 2025 18:46	JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX050525

Review By	John Carlone	Review On	5/6/2025 9:53:58 AM
Supervise By	Maresh Dadoda	Supervise On	5/6/2025 12:43:00 PM
SubDirectory	VX050525	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133811		
Initial Calibration Stds	VP133832,VP133833,VP133834,VP133835,VP133836,VP133837		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP133838		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046038.D	05 May 2025 09:37		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX046039.D	05 May 2025 10:49	Not used	JC/MD	Not Ok
3	VSTDIC005	VSTDIC005	VX046040.D	05 May 2025 11:12	Not used	JC/MD	Not Ok
4	VSTDIC020	VSTDIC020	VX046041.D	05 May 2025 11:35		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX046042.D	05 May 2025 11:58		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX046043.D	05 May 2025 12:21		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX046044.D	05 May 2025 12:45		JC/MD	Ok,M
8	IBLK	IBLK	VX046045.D	05 May 2025 13:08		JC/MD	Ok
9	VSTDIC005	VSTDIC005	VX046046.D	05 May 2025 16:04		JC/MD	Ok,M
10	VSTDIC001	VSTDIC001	VX046047.D	05 May 2025 16:27		JC/MD	Ok,M
11	VSTDICV050	ICVVX050525	VX046048.D	05 May 2025 16:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX050825

Review By	John Carlone	Review On	5/9/2025 9:54:47 AM
Supervise By	Maresh Dadoda	Supervise On	5/9/2025 2:53:48 PM
SubDirectory	VX050825	HP Acquire Method	HP Processing Method 82X050525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133850
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133851,VP133852

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046077.D	08 May 2025 09:36		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046078.D	08 May 2025 10:05	pH#Lot#V12668	JC/MD	Ok,M
3	VX0508MBL01	VX0508MBL01	VX046079.D	08 May 2025 10:33		JC/MD	Ok
4	VX0508WBL01	VX0508WBL01	VX046080.D	08 May 2025 10:56		JC/MD	Ok
5	VX0508WBS01	VX0508WBS01	VX046081.D	08 May 2025 11:20		JC/MD	Ok,M
6	VX0508WBSD01	VX0508WBSD01	VX046082.D	08 May 2025 11:45		JC/MD	Ok,M
7	IBLK	IBLK	VX046083.D	08 May 2025 12:08		JC/MD	Ok
8	Q1985-02	TB-20250507	VX046084.D	08 May 2025 12:32	vial A pH<2 TB	JC/MD	Ok
9	Q1985-01	RW8-BW-20250507	VX046085.D	08 May 2025 12:55	vial A pH<2	JC/MD	Ok
10	PB167895TB	PB167895TB	VX046086.D	08 May 2025 13:18		JC/MD	Ok,M
11	Q1964-04	MH-LL	VX046087.D	08 May 2025 13:42	vial A pH#5.0	JC/MD	Ok
12	Q1968-04	MH-M	VX046088.D	08 May 2025 14:05	vial A pH#5.0	JC/MD	Ok,M
13	Q1968-08	MH-M	VX046089.D	08 May 2025 14:29	vial A pH#5.0	JC/MD	Ok
14	Q1972-04	SUB-WC	VX046090.D	08 May 2025 14:52	vial A pH#5.0	JC/MD	Ok,M
15	Q1972-08	WC-B-10	VX046091.D	08 May 2025 15:15	vial A pH#5.0	JC/MD	Ok
16	Q1972-12	WC-B-2	VX046092.D	08 May 2025 15:39	vial A pH#5.0	JC/MD	Ok
17	Q1972-16	WC-B-7	VX046093.D	08 May 2025 16:02	vial A pH#5.0	JC/MD	Ok,M
18	Q1972-20	SUB-WC-2	VX046094.D	08 May 2025 16:26	vial A pH#5.0	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX050825

Review By	John Carlone	Review On	5/9/2025 9:54:47 AM
Supervise By	Maresh Dadoda	Supervise On	5/9/2025 2:53:48 PM
SubDirectory	VX050825	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133850		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133851,VP133852		

19	Q1972-24	WC-B-1	VX046095.D	08 May 2025 16:49	vial A pH#5.0	JC/MD	Ok
20	VX0508MBS01	VX0508MBS01	VX046096.D	08 May 2025 17:12		JC/MD	Ok,M
21	Q1966-03	SILICA-GEL	VX046097.D	08 May 2025 17:36		JC/MD	Ok
22	IBLK	IBLK	VX046098.D	08 May 2025 17:59		JC/MD	Ok
23	IBLK	IBLK	VX046099.D	08 May 2025 18:23		JC/MD	Ok
24	VSTDCCC050	VSTDCCC050EC	VX046100.D	08 May 2025 18:46		JC/MD	Ok,M

M : Manual Integration

Prep Standard - Chemical Standard Summary

Order ID : Q1985

Test : VOCMS Group1

Prepbatch ID :

Sequence ID/Qc Batch ID: vx050825,

Standard ID :

VP131746,VP131767,VP132035,VP132096,VP133174,VP133544,VP133758,VP133850,VP133851,VP133852,

Chemical ID :

V13391,V13457,V13460,V13465,V13466,V13706,V14154,V14289,V14431,V14435,V14501,V14502,V14523,V14524,V14613,V14614,V14615,V14630,V14631,V14632,V14633,V14719,V14720,V14726,V14744,V14753,V14804,V14805,V14842,V14915,V14916,V14917,V14918,W3112,

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
247	8260 Internal Standard, 250PPM	VP131746	11/22/2024	05/18/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								11/23/2024

FROM 0.50000ml of V14289 + 49.50000ml of V14154 = Final Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
218	BFB, 25PPM	VP131767	11/22/2024	05/18/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								11/27/2024

FROM 0.50000ml of V13391 + 49.50000ml of V14154 = Final Quantity: 50.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1810	8260 Working Std(2-CVE)-800ppm	VP132035	12/10/2024	06/10/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								12/12/2024

FROM 1.00000ml of V14630 + 1.00000ml of V14631 + 1.00000ml of V14632 + 1.00000ml of V14633 + 46.00000ml of V14614 = Final
Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
719	8260 Working STD (BCM)-First source, 400PPM	VP132096	12/12/2024	06/10/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								12/19/2024

FROM 1.00000ml of V13465 + 1.00000ml of V13466 + 1.50000ml of V13457 + 1.50000ml of V13460 + 20.00000ml of V14614 = Final
Quantity: 25.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
617	8260 Surrogate, 400PPM	VP133174	02/27/2025	08/27/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								03/04/2025

FROM 0.40000ml of V13706 + 24.60000ml of V14613 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
257	8260 Calibration Working STD Mix-First source, 160PPM	VP133544	04/01/2025	05/10/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								04/03/2025

FROM 0.40000ml of V14842 + 1.00000ml of V14431 + 1.00000ml of V14435 + 1.00000ml of V14501 + 1.00000ml of V14502 + 1.00000ml of V14523 + 1.00000ml of V14524 + 1.00000ml of V14726 + 1.00000ml of V14744 + 1.00000ml of V14753 + 1.00000ml of V14804 + 1.00000ml of V14805 + 1.50000ml of V14719 + 1.50000ml of V14720 + 10.60000ml of V14615 = Final Quantity: 25.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
51	8260 Working STD (Acrolein) -first source, 800PPM	VP133758	04/24/2025	05/23/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								05/07/2025

FROM 1.00000ml of V14915 + 1.00000ml of V14916 + 1.00000ml of V14917 + 1.00000ml of V14918 + 21.00000ml of V14615 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
589	BFB TUNE CHECK	VP133850	05/08/2025	05/09/2025	John Carlone	None	None	Maresh Dadoda
								05/12/2025

FROM 39.98400ml of W3112 + 0.01600ml of VP131767 = Final Quantity: 40.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	VP133851	05/08/2025	05/09/2025	John Carlone	None	None	Mahesh Dadoda 05/12/2025
<u>FROM</u> 39.94450ml of W3112 + 0.00500ml of VP132096 + 0.00500ml of VP133174 + 0.00800ml of VP131746 + 0.01250ml of VP132035 + 0.01250ml of VP133544 + 0.01250ml of VP133758 = Final Quantity: 40.000 ml								

[illegible]

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30067 / BFB tuneing solution	A0191805	11/22/2025	11/22/2024 / SAM	01/13/2023 / SAM	V13391

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13457

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13460

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13465

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	06/12/2025	12/12/2024 / SAM	01/27/2023 / SAM	V13466

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555582 / Custom Mixture, 8260 A/B Surrogate Mix [CS 5179-2]	A0196865	02/27/2026	02/27/2025 / SAM	04/12/2023 / SAM	V13706

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	05/18/2025	11/18/2024 / pedro	02/06/2024 / SAM	V14154

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555581 / Custom Standard, 8260 Internal Std [CS 5179-1]	A0210184	11/22/2025	11/22/2024 / SAM	04/15/2024 / SAM	V14289

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/20/2025	03/20/2025 / SAM	08/15/2024 / SAM	V14431

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/20/2025	03/20/2025 / SAM	08/15/2024 / SAM	V14435

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	09/20/2025	03/20/2025 / SAM	09/17/2024 / SAM	V14501

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	09/20/2025	03/20/2025 / SAM	09/17/2024 / SAM	V14502

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	09/20/2025	03/20/2025 / SAM	09/18/2024 / SAM	V14523

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	09/20/2025	03/20/2025 / SAM	09/18/2024 / SAM	V14524

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	08/27/2025	02/27/2025 / SAM	11/26/2024 / SAM	V14613

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	06/10/2025	12/10/2024 / SAM	11/26/2024 / SAM	V14614

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	09/19/2025	03/19/2025 / SAM	11/26/2024 / SAM	V14615

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14630

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14631

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14632

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	06/10/2025	12/10/2024 / SAM	12/06/2024 / SAM	V14633

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	09/20/2025	03/20/2025 / SAM	12/17/2024 / SAM	V14719

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	09/20/2025	03/20/2025 / SAM	12/17/2024 / SAM	V14720

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	07/30/2025	01/30/2025 / SAM	12/17/2024 / SAM	V14726

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	08/27/2025	02/27/2025 / SAM	12/17/2024 / SAM	V14744

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	07/30/2025	01/30/2025 / SAM	12/17/2024 / SAM	V14753

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0220471	09/20/2025	03/20/2025 / SAM	01/08/2025 / SAM	V14804

LOTS

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0220471	09/20/2025	03/20/2025 / SAM	01/08/2025 / SAM	V14805

LOTS

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30470 / VOA Stock Solution, tert-butanol std, 1mL, P&TM	A0217535	08/27/2025	02/27/2025 / SAM	01/21/2025 / SAM	V14842

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	042325	05/23/2025	04/24/2025 / SAM	04/24/2025 / SAM	V14915

CHEMICAL RECEIPT LOG BOOK


Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	042325	05/23/2025	04/24/2025 / SAM	04/24/2025 / SAM	V14916

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	042325	05/23/2025	04/24/2025 / SAM	04/24/2025 / SAM	V14917

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	042325	05/23/2025	04/24/2025 / SAM	04/24/2025 / SAM	V14918

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis




Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis


Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis


Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier
Vice President Global Quality

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis




Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality



Ree 03117/24

CERTIFIED WEIGHT REPORT

Part Number: 95317

Lot Number: 021624

Description: Universal VOA Megamix

69 components

Expiration Date: 021627

Recommended Storage: Freezer (0 °C)

Nominal Concentration (µg/mL): 2000

NIST Test ID#: BUTB

Solvent(s): Lot#
Methanol EG359-USQ12Weight(s) shown below were combined and diluted to (mL): 100.0 0.021 Balance Uncertainty
Flask Uncertainty

Formulated By:	Prashant Chauhan	021624
Reviewed By:	Pedro L. Renteria	021624
DATE		

Compound	(K049)	Lot	Dir.	Initial	Initial	Nominal	Purity	Purity	Uncertainty	Target	Actual	Actual	Expanded	SDS Information				
	Part Number	Number	Factor	Vol. (mL)	Conc.(µg/mL)	Conc (µg/mL)	(%)	Uncertainty	Pipette (mL)	Weight(g)	Weight(g)	Conc (µg/mL)	Uncertainty	(+/-) (µg/mL)	(Solvent Safety Info. On Attached pg.)	CASE	OSHA PEL (TWA)	LD50
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)			or-rat 2450mg/kg
2. Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)			or-rat 700mg/kg
3. Carbon disulfide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)			or-rat 1200mg/kg
4. cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21056	0.21069	2001.1	8.5	1478-11-5	NA			N/A
5. trans-1,4-Dichloro-2-butene	(0486)	MKBP6041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20746	2001.7	8.4	110-57-6	NA			N/A
6. Diethyl ether	(0153)	IK18CAS000C	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	NA			N/A
7. Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	NA			or-rat 14800mg/kg
8. Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm(28mg/m3/8H)(skin)			or-rat 76mg/kg
9. 2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)			or-rat 2460mg/kg
10. Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	128-98-7	1 ppm (3mg/m3/8H)(skin)			or-rat 120mg/kg
11. Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm(35mg/m3/8H)(skin)			or-rat 277mg/kg
12. Methyl methacrylate	(0404)	MKGW6137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)			or-rat 7872mg/kg
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H)(skin)			or-rat 780mg/kg
14. 2-Nitropropane	(0461)	14002JK	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	78-46-9	10 ppm (30mg/m3/8H)			or-rat 720mg/kg
15. Perchloroethane	(0460)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	78-01-7	NA			N/A
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7600mg/m3/8H)			or-rat 43g/kg
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	NA			or-rat 915mg/kg
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	NA			or-rat 848mg/kg
19. cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	NA			N/A
20. trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	NA			N/A
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm			or-rat 1235mg/kg
22. 1,1-Dichloroethane	32261	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.8	20.5	78-25-2	0.5 ppm (5mg/m3) (skin)			or-rat 933mg/kg
23. Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)			or-rat 2350mg/kg
24. Carbon tetrachloride	95321	020724	0.10	10.00	20024.0	2000	NA	NA	0.042	NA	NA	2001.3	20.5	67-68-3	60 ppm (240mg/m3) (CL)			or-rat 908mg/kg
25. Chloroform	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	100 ppm			or-rat 725mg/kg
26. Dibromomethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	NA			N/A
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	NA			N/A
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm			or-rat 108mg/kg
29. Trichloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.8	20.8	127-18-4	25 ppm (170mg/m3/8H)(final)			or-rat 2629mg/kg
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)			or-rat 10300mg/kg
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40018.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	86-12-8	0.001 ppm			or-rat 170mg/kg
32. 1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)			or-rat 108mg/kg
33. 1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)			or-rat 670mg/kg
34. 1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m3/8H)			or-rat 1947mg/kg
35. 1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	NA			or-rat 3600mg/kg
36. 1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	28.7	563-58-6	NA			N/A
37. cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	NA			N/A
38. trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	NA			N/A
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)			or-rat 82mg/kg
40. 1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	830-20-6	NA			or-rat 670mg/kg
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H)(skin)			or-rat 800mg/kg
42. 1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m3/8H)(skin)			or-rat 936mg/kg
43. Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)			or-rat 3402mg/kg
44. 1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (60mg/m3/8H)			or-rat 149.8mg/kg
45. Benzene	35162	050823	0.05	5.00	40006.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm			or-rat 4894mg/kg
46. Bromobenzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	NA			or-rat 2699mg/kg
47. n-Butyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)			or-rat >2000mg/kg
48. Ethyl benzene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	NA			or-rat 4750mg/kg
49. p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)			or-rat 490mg/kg
50. Naphthalene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm			or-rat 5000mg/kg
51. Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-98-3	200 ppm			or-rat 5000mg/kg
52. Toluene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-81-6	NA			or-rat 1390mg/kg
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)			or-rat 756mg/kg
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.6	23.0	95-63-6	NA			or-rat 5g/kg
55. 1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-87-8	NA			or-rat 5000mg/kg
56. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	23.0	95-63-6	NA			or-rat 756mg/kg
57. m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-87-8	NA			or-rat 5000mg/kg
58. tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)			or-rat 5g/kg
59. sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.8	22.9	96-06-6	NA			N/A
60. Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	135-98-8	NA			or-rat 2240mg/kg
61. 2-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.5	22.9	108-90-7	75 ppm (350mg/m3/8H)			or-rat 2290mg/kg
62. 4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.									




Run 16, "P95317 L021624 (2000µg/mL in MeOH)"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments

GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.53
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.60
5	Iodomethane	12.91
6	Allyl chloride	12.96
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethene	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethene	18.00
12	Methoxyvinylbenzene/Methyl acrylate/Chloroform	18.49
13	Isobutene/1,1,1-Trichloroethane	18.91
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethene	23.88
18	1,2-Dichloropropane	24.52
19	Methyl methacrylate	25.13
20	Bromochloropropane	25.46
21	Dibromomethane/2,2-Dichloropropane	27.07
22	cis-1,2-Dichloroethene	28.03
23	Toluene	28.73
24	Ethyl methacrylate/Trans-1,2-Dichloropropane	29.34
25	1,1,2-Trichloroethane	30.34
26	Tetrachloroethane/1,2-Dichloropropane	31.16
27	Dibromochloromethane	32.84
28	1,2-Dibromomethane	33.26
29	Chlorobenzene	33.40
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.85
31	m-Xylene/p-Xylene	35.33
32	o-Xylene	35.70
33	Styrene	35.70
34	Isopropylbenzene/Bromofarm	36.48
35	cis-1,4-Dichloro-3-butene	36.80
36	1,1,2-Trichloropropane	37.23
37	n-Propylbenzene	37.77
38	trans-1,4-Dichloro-3-butene	37.92
39	Bromobenzene	38.05
40	1,3,5-Trimethylbenzene	38.14
41	1,3,5-Trimethylbenzene	38.50
42	Chlorobenzene	38.63
43	4-Chlorobenzene	38.77
44	tert-Butylbenzene	39.76
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropylbenzene	41.02
49	1,3-Trichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.52
52	1,2-Dichlorobenzene	42.58
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.40
55	1,2,4-Trifluorobenzene	49.28
56	Hexachlorocyclopentadiene	49.72
57	Naphthalene	50.56
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225 H370 P271 P302,332	Highly Flammable Liquid and Vapor Cause damage to organs Use in ventilated area If on skin, wash with soap and water	H301, 311, 331 H351 P280 P305,351,338	Toxic if swallowed, skin contact, inhaled Suspected of causing cancer Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#	% (optional)
Methanol METHYL ALCOHOL	67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption	ingestion and inhalation.
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing.	Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Ree 03117/24

CERTIFIED WEIGHT REPORT

Part Number: 95317

Lot Number: 021624

Description: Universal VOA Megamix

69 components

Expiration Date: 021627

Recommended Storage: Freezer (0 °C)

Nominal Concentration (µg/mL): 2000

NIST Test ID#: BUTB

Solvent(s): Lot#
Methanol EG359-USQ12Weight(s) shown below were combined and diluted to (mL): 100.0 0.021 Balance Uncertainty
Flask Uncertainty

Formulated By:	Prashant Chauhan	021624
Reviewed By:	Pedro L. Renteria	021624
DATE		

Compound	(KME)	Lot	DIL	Initial Vol. (mL)	Initial Conc. (µg/mL)	Nominal Conc. (µg/mL)	Purity (%)	Purity Uncertainty	Uncertainty Pipette (mL)	Target Weight(g)	Actual Weight(g)	Actual Conc. (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information		
														(Solvent Safety Info. On Attached pg.)	OSHA PEL (TWA)	LD50
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2450mg/kg
2. Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg
3. Carbon disulfide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg
4. cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21056	0.21069	2001.1	8.5	1478-11-5	N/A	N/A
5. trans-1,4-Dichloro-2-butene	(0486)	MKBP9041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20746	2001.7	8.4	110-57-6	N/A	N/A
6. Diethyl ether	(0153)	IK18CASA000C	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A
7. Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	or-rat 14800mg/kg
8. Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm(28mg/m3/8H)(skin)	or-rat 75mg/kg
9. 2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 2460mg/kg
10. Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	128-98-7	1 ppm (3mg/m3/8H)(skin)	or-rat 120mg/kg
11. Methyl acrylate	(0755)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm(35mg/m3/8H)(skin)	or-rat 277mg/kg
12. Methyl methacrylate	(0404)	MKGW6137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 7872mg/kg
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H)(skin)	or-rat 780mg/kg
14. 2-Nitropropane	(0461)	14002JX	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	78-46-9	10 ppm (35mg/m3/8H)	or-rat 720mg/kg
15. Perchloroethane	(0460)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	78-01-7	N/A	N/A
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7600mg/m3/8H)	or-rat 435g/kg
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	N/A	or-rat 915mg/kg
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	or-rat 848mg/kg
19. cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	158-59-2	N/A	N/A
20. trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	158-60-5	N/A	or-rat 1235mg/kg
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 820mg/kg
22. 1,1-Dichloroethane	32261	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	or-rat 200mg/kg
23. Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 933mg/kg
24. Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	58-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg
25. Chloroform	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	67-68-3	60 ppm (240mg/m3) (CL)	or-rat 908mg/kg
26. Dibromomethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	N/A	or-rat 108mg/kg
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	100 ppm	or-rat 725mg/kg
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A
29. Trichloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.8	20.8	127-18-4	25 ppm (170mg/m3/8H)(final)	or-rat 2629mg/kg
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10300mg/kg
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40018.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	86-12-8	0.001 ppm	or-rat 170mg/kg
32. 1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	108-83-4	20 ppm (8H)	or-rat 108mg/kg
33. 1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)	or-rat 670mg/kg
34. 1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m3/8H)	or-rat 1947mg/kg
35. 1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	or-rat 3600mg/kg
36. 1,1-Dichloropropane	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	28.7	563-58-6	N/A	N/A
37. cis-1,3-Dichloropropane	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	N/A	N/A
38. trans-1,3-Dichloropropane	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	N/A	N/A
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg
40. 1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	830-20-6	N/A	or-rat 670mg/kg
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H)(skin)	or-rat 800mg/kg
42. 1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m3/8H)(skin)	or-rat 936mg/kg
43. Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-mus 2402mg/kg
44. 1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	98-18-4	10 ppm (60mg/m3/8H)	or-rat 149.8mg/kg
45. Benzene	35162	050823	0.05	5.00	40006.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	or-rat 4894mg/kg
46. Bromobenzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-1	N/A	or-rat 2699mg/kg
47. n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	N/A
48. Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg
49. p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	N/A	or-rat 4750mg/kg
50. Naphthalene	35162	050823	0.05	5.00	40005.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg
51. Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	or-rat 5000mg/kg
52. Toluene	35162	050823	0.05	5.00	40008.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-3	200 ppm	or-rat 5000mg/kg
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-81-6	N/A	or-rat 1390mg/kg
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 756mg/kg
55. 1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.8	23.0	95-63-6	N/A	or-rat 5g/kg
56. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-87-8	N/A	or-rat 5000mg/kg
57. m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	or-rat 5g/kg
58. tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	98-06-6	N/A	N/A
59. sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.8	22.9	135-98-8	N/A	or-rat 2240mg/kg
60. Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2290mg/kg
61. 2-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	or-rat 3900mg/kg
62. 4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	or-rat 2100mg/kg
63. 1,2-Dichlorobenzene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.6	23.0	95-50-1	50 ppm (300mg/m3) (CL)	or-rat 500mg/kg
64. 1,3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	541-73-1	N/A	or-mus 1060mg/kg
65. 1,4-Dichlorobenzene	35163	101923	0.05	5.00	40001.8	2000</										




Run 16, "P95317 L021624 (2000µg/mL in MeOH)"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments

GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.53
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.60
5	Iodomethane	12.91
6	Allyl chloride	12.96
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethene	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethene	18.00
12	Methoxyvinylbenzene/Methyl acrylate/Chloroform	18.49
13	Isobutene/1,1,1-Trichloroethane	18.91
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethene	23.88
18	1,2-Dichloropropane	24.52
19	Methyl methacrylate	25.13
20	Bromochloropropane	25.46
21	Dibromomethane/2-Pentene	27.07
22	cis-1,2-Dichloroethene	28.03
23	Toluene	28.73
24	Ethyl methacrylate/Trans-1,2-Dichloroethene	29.34
25	1,1,2-Trichloroethane	30.34
26	Tetrachloroethane/1,2-Dichloropropane	31.16
27	Dibromochloromethane	32.84
28	1,2-Dibromomethane	33.26
29	Chlorobenzene	33.40
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.85
31	m-Xylene/p-Xylene	35.33
32	o-Xylene	35.70
33	Styrene	35.70
34	Isopropylbenzene/Bromofarm	36.48
35	cis-1,4-Dichloro-3-butene	36.80
36	1,1,2-Trichloropropane	37.23
37	n-Propylbenzene	37.77
38	trans-1,4-Dichloro-3-butene	37.92
39	Bromobenzene	38.05
40	1,3,5-Trimethylbenzene	38.14
41	1,3,5-Trimethylbenzene	38.50
42	Chlorobenzene	38.63
43	4-Chlorobenzene	38.77
44	tert-Butylbenzene	39.76
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropylbenzene	41.02
49	1,3-Dichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.52
52	1,2-Dichlorobenzene	42.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.40
55	1,2,4-Trifluorobenzene	49.28
56	Hexachlorocyclopentadiene	49.72
57	Naphthalene	50.56
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225 H370 P271 P302,332	Highly Flammable Liquid and Vapor Cause damage to organs Use in ventilated area If on skin, wash with soap and water	H301, 311, 331 H351 P280 P305,351,338	Toxic if swallowed, skin contact, inhaled Suspected of causing cancer Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		
Methanol	METHYL ALCOHOL	CAS#: 67-56-1
		% (optional) > 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.	
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.	

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM

Rec 04/24/25 5 N19



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 042325
Description: Acrolein

Solvent(s):
Water
Lot#: 072324Q

V14914 to
V14918



<i>Eli Aliaga</i>		042325
Formulated By:	Eli Aliaga	DATE
<i>Pedro L. Rentas</i>		042325
Reviewed By:	Pedro L. Rentas	DATE

Expiration Date: 052325
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB
Weight(s) shown below were combined and diluted to (mL): 10.0

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LDSO
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05178	5011.8	52.6	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC/MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.)
Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately
Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Rev 1.0, 2/25/2025

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN WATER

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2025

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271 P302,332	Use in ventilated area If on skin, wash with soap and water	H315 P280 P305,351,338	Causes skin and eye irritation. Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))
Water

CAS#: 7732-18-5

%(optional)
> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water	CAS#: 7732-18-5	TWA: 500 ppm
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H ₂ O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

Vapor Density (AIR = 1)	NA	Evaporation rate (Butyl Acetate = 1)	0°C
Solubility in Water	Completely miscible	NA	NA
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	NA
Conditions to avoid	NA
Materials to avoid	NA
Hazardous decomposition products - No data available	

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat	NA
LC50 Inhalation - Rat	NA
LD50 Dermal - Guinea pig	NA
Causes skin irritation.	
Eye irritation	

Section XII. ECOLOGICAL INFORMATION

LC50	NA
EC50	NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
Not dangerous goods	Not dangerous goods
Proper shipping name: Water	Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM

Rec 04/24/25 5 N19



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 042325
Description: Acrolein

Solvent(s):
Water
Lot# 072324Q

V14914 to
V14918



<i>Eli Aliaga</i>		042325
Formulated By:	Eli Aliaga	DATE
<i>Pedro L. Rentas</i>		042325
Reviewed By:	Pedro L. Rentas	DATE

Expiration Date: 052325
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB
Weight(s) shown below were combined and diluted to (mL): 10.0

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LDSO
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05178	5011.8	52.6	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC/MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.)
Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately
Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Rev 1.0, 2/25/2025

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN WATER

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossetto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2025

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271 P302,332	Use in ventilated area If on skin, wash with soap and water	H315 P280 P305,351,338	Causes skin and eye irritation. Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))
Water

CAS#: 7732-18-5

%(optional)
> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water	CAS#: 7732-18-5	TWA: 500 ppm
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H ₂ O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

Vapor Density (AIR = 1)	NA	Evaporation rate (Butyl Acetate = 1)	0°C
Solubility in Water	Completely miscible	NA	NA
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	NA
Conditions to avoid	NA
Materials to avoid	NA
Hazardous decomposition products - No data available	

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat	NA
LC50 Inhalation - Rat	NA
LD50 Dermal - Guinea pig	NA
Causes skin irritation.	
Eye irritation	

Section XII. ECOLOGICAL INFORMATION

LC50	NA
EC50	NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
Not dangerous goods	Not dangerous goods
Proper shipping name: Water	Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM

Rec 04/24/25 5 N109



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 042325
Description: Acrolein

Solvent(s):
Water
Lot# 072324Q

V14914 to
V14918



<i>Eli Aliaga</i>		042325
Formulated By:	Eli Aliaga	DATE
<i>Pedro L. Rentas</i>		042325
Reviewed By:	Pedro L. Rentas	DATE

Expiration Date: 052325
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB
Weight(s) shown below were combined and diluted to (mL): 10.0

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LDSO
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05178	5011.8	52.6	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC/MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.)
Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately
Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Rev 1.0, 2/25/2025

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN WATER

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2025

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271 P302,332	Use in ventilated area If on skin, wash with soap and water	H315 P280 P305,351,338	Causes skin and eye irritation. Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))
Water

CAS#: 7732-18-5

%(optional)
> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water	CAS#: 7732-18-5	TWA: 500 ppm
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H ₂ O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

Vapor Density (AIR = 1)	NA	Evaporation rate (Butyl Acetate = 1)	0°C
Solubility in Water	Completely miscible	NA	NA
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	NA
Conditions to avoid	NA
Materials to avoid	NA
Hazardous decomposition products - No data available	

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat	NA
LC50 Inhalation - Rat	NA
LD50 Dermal - Guinea pig	NA
Causes skin irritation.	
Eye irritation	

Section XII. ECOLOGICAL INFORMATION

LC50	NA
EC50	NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
Not dangerous goods	Not dangerous goods
Proper shipping name: Water	Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM

Rec 04/24/25 5 N109



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 042325
Description: Acrolein

Solvent(s):
Water
Lot#: 072324Q

V14914 to
V14918



<i>Eli Aliaga</i>		042325
Formulated By:	Eli Aliaga	DATE
<i>Pedro L. Rentas</i>		042325
Reviewed By:	Pedro L. Rentas	DATE

Expiration Date: 052325
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB
Weight(s) shown below were combined and diluted to (mL): 10.0

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LDSO
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05178	5011.8	52.6	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC/MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.)
Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately
Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Rev 1.0, 2/25/2025

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN WATER

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2025

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271 P302,332	Use in ventilated area If on skin, wash with soap and water	H315 P280 P305,351,338	Causes skin and eye irritation. Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s)) Water	CAS#: 7732-18-5	% (optional) > 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water	CAS#: 7732-18-5	TWA: 500 ppm
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H ₂ O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

Vapor Density (AIR = 1)	NA	Evaporation rate (Butyl Acetate = 1)	0°C
Solubility in Water	Completely miscible	NA	NA
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	NA
Conditions to avoid	NA
Materials to avoid	NA
Hazardous decomposition products - No data available	

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat	NA
LC50 Inhalation - Rat	NA
LD50 Dermal - Guinea pig	NA
Causes skin irritation.	
Eye irritation	

Section XII. ECOLOGICAL INFORMATION

LC50	NA
EC50	NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
Not dangerous goods	Not dangerous goods
Proper shipping name: Water	Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM

Dec 09/17/24

2 vials

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **091424**
Description: **Acrolein**Solvent(s):
Water Lot#
072324Q

<i>Justin Dippold</i>		091424
Formulated By:	Justin Dippold	DATE
<i>Pedro L. Rentas</i>		091424
Reviewed By:	Pedro L. Rentas	DATE

Expiration Date: 101424

Recommended Storage: Refrigerate (4 °C)

Nominal Concentration (µg/mL): 5000

NIST Test ID#: 6UTB

5E-05 Balance Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0 0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.

Abundance TIC: [BSB2]79005.D

Abundance

Scan 232 (8.927 min): [BSB2]79005.D

250000 8.93

200000

150000

100000

50000

Time--> 0

10.00 15.00 20.00 25.00 30.00 35.00 40.00 45.00 50.00 55.00 60.00

m/z--> 0

20 30 40 50 60 70 80 90 100 110 120 130 140 150 160 170

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Certified Reference Material CRM

Dec 09/17/24

2 vial

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **091424**
Description: **Acrolein**Solvent(s):
Water Lot#
072324Q

<i>Justin Dippold</i>		091424
Formulated By:	Justin Dippold	DATE
<i>Pedro L. Rentas</i>		091424
Reviewed By:	Pedro L. Rentas	DATE

Expiration Date: 101424

Recommended Storage: Refrigerate (4 °C)

Nominal Concentration (µg/mL): 5000

NIST Test ID#: 6UTB

5E-05 Balance Uncertainty

Weight(s) shown below were combined and diluted to (mL): 10.0 0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.

Abundance TIC: [BSB2]79005.D

Abundance

Scan 232 (8.927 min): [BSB2]79005.D

250000 8.93

200000

150000

100000

50000

Time--> 0

10.00 15.00 20.00 25.00 30.00 35.00 40.00 45.00 50.00 55.00 60.00



m/z--> 0

20 30 40 50 60 70 80 90 100 110 120 130 140 150 160 170

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	67-56-1	TWA 200 ppm
Skin notation			TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	67-56-1	TWA 200 ppm
Skin notation			TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



ANAB ISO 17034 Accredited
AR-1539 Certificate Number
<https://absolutestandards.com>

CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	67-56-1	TWA 200 ppm
Skin notation			TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	67-56-1	TWA 200 ppm
Skin notation			TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30067 **Lot No.:** A0191805

Description : 4-Bromofluorobenzene Standard

4-Bromofluorobenzene Standard 2,500µg/mL, P&T Methanol,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	2,483.9 µg/mL	+/- 139.5488

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Alicia Leathers - Operation Technician I

Date Mixed: 17-Nov-2022

Balance Serial # B251644995


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 21-Nov-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : December 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555582 **Lot No.:** A0196865

Description : Custom 8260A/B Surrogate Mix
Custom 8260A/B Surrogate Mix 25,000µg/mL, P&T Methanol,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2-Dichloroethane-d4	17060-07-0	PR-32845	99%	25,036.0 µg/mL	+/- 1,417.9179
2	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	25,132.0 µg/mL	+/- 1,423.3549
3	Dibromofluoromethane	1868-53-7	022013	99%	25,040.0 µg/mL	+/- 1,418.1445
4	Toluene-d8	2037-26-5	PR-33397	99%	25,028.0 µg/mL	+/- 1,417.4648

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer - Operations Technician I

Date Mixed: 11-Apr-2023

Balance: 1127510105

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024

Balance Serial # B707717271

Dillon Murphy
Dillon Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024

Balance Serial # B707717271

Dylan Murphy
Dylan Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555581 **Lot No.:** A0210184

Description : Custom 8260 Internal Standard Mix

Custom 8260 Internal Standard Mix 25,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL


Expiration Date : April 30, 2027 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	25,212.0 µg/mL	+/- 1,427.8857
2	1,4-Difluorobenzene	540-36-3	MKCS8657	99%	25,220.0 µg/mL	+/- 1,428.3388
3	Chlorobenzene-d5	3114-55-4	PR-31132	99%	25,116.0 µg/mL	+/- 1,422.4487
4	Pentafluorobenzene	363-72-4	MKCR9383	99%	25,180.0 µg/mL	+/- 1,426.0734

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%


John Friedline - Operations Technician I

Date Mixed: 11-Apr-2024

Balance: 1127510105

REVISED
By Restek Review of SDS per April 18, 2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus

V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24
CERTIFIED REFERENCE MATERIAL

30 ml
Certificate of Analysis
chromatographic plus

V14727 to
V14756



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826
Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : May 31, 2031 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Tom Suckal - Mix Technician

Date Mixed: 23-Sep-2024

Balance Serial # B707717271



Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24
30 ml
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus

V14727 to
V14756



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826
Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : May 31, 2031 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckal - Mix Technician

Date Mixed: 23-Sep-2024

Balance Serial # B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

✓ 14842 to 14846



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30470 **Lot No.:** A0217535

Description : tert-Butanol Standard

tert-Butanol Std 50,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	tert-Butanol (TBA)	75-65-0	SHBQ8002-1	99%	50,007.5 µg/mL	+/- 717.6137

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A. O. E.
Aaron Enyart - Operations Tech I

Date Mixed: 07-Oct-2024

Balance Serial # B251644995

Brittany Federinko
Brittany Federinko - Operations Tech I

Date Passed: 09-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2014 Dec 01/08/21
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024

Balance Serial # 1127510105

Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED
By Jennifer Pollock at 7:12 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2014 Dec 01/08/21
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024

Balance Serial # 1127510105

Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED
By Jennifer Pollock at 7:12 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (µeq/g)	≤ 0.3	0.2
Titration Base (µeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis - Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier
Vice President Global Quality



SHIPPING DOCUMENTS

CHEMTECH

CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 Fax: (908) 78-8922
www.chemtech.net

Chemtech Project Number:

Q1985

COC Number:

CLIENT INFORMATION

COMPANY: Tetra Tech
ADDRESS: 4433 Corporation Ln, Suite 300
CITY: Virginia Beach STATE: VA ZIP: 23462
ATTENTION: Ernie Wu
PHONE: 757-466-4901 FAX: 757-461-4148

PROJECT INFORMATION

PROJECT NAME: NWIRP Bethpage
PROJECT #: 112G08005-WE13 LOCATION: RW8
PROJECT MANAGER: Ernie Wu
E-MAIL: ernie.wu@tetratech.com
PHONE: 757-466-4901 FAX: 757-461-4148

BILLING INFORMATION

BILL TO: PO#
ADDRESS:
CITY: STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

1,4-Dioxane SW646 8270	VOC SW846 8260	TAL Metals	pH	TSS	SVOC 8270			
1	2	3	4	5	6	7	8	9

PRESERVATIVES

COMMENTS

<-- Specify Preservatives
A-HCl B-HNO3
C-H2SO4 D-NaOH
E-ICE F-Other

DATA TURNAROUND INFORMATION

FAX: 5 DAYS*
HARD COPY: 5 DAYS*
EDD 5 DAYS*
* TO BE APPROVED BY CHEMTECH
STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

- ☐ RESEULTS ONLY ☐ USEPA CLP
☐ RESULTS + QC ☐ New York State ASP "B"
☐ New Jersey REDUCED ☐ New York State ASP "A"
☐ New Jersey CLP ☐ Other
☐ EDD Format

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles										
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	RW8-BW-20250507	GW		X	5/7/25	12:30	8	1	3	1	1	1	1				
2.	TB-20250507	WQ		X	5/7/25	9:00	8	1	3	1	1	1	1				
3.																	
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER 1. <i>[Signature]</i>	DATE/TIME 5/7/25/1700	RECEIVED BY 1. <i>[Signature]</i>	Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant ☐ Cooler Temp <u>2.3°C</u> MeOH extraction requires an additional 4oz. Jar for percent solid Comments: Adjust Factor #1 ILL run #1
RELINQUISHED BY 2. FedEx	DATE/TIME 5-8-25 0950	RECEIVED BY 2. <i>[Signature]</i>	
RELINQUISHED BY 3.	DATE/TIME	RECEIVED FOR LAB BY 3.	

Page _____ of _____

SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight
CHEMTECH: ☐ Picked Up ☐ Overnight

Shipment Complete
☐ YES ☐ NO

WHITE - CHEMTECH COPY FOR RETURN TO CLIENT YELLOW - CHEMTECH COPY PINK - SAMPLER COPY

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1985	TETR06	Order Date : 5/8/2025 10:49:00 AM	Project Mgr :
Client Name : Tetra Tech NUS, Inc.		Project Name : NWIRP Bethpage 112G080	Report Type : Level 4
Client Contact : Ernie Wu		Receive DateTime : 5/8/2025 9:50:00 AM	EDD Type : ADAPT
Invoice Name : Tetra Tech NUS, Inc.		Purchase Order :	Hard Copy Date :
Invoice Contact : Ernie Wu			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1985-01	RW8-BW-20250507	Water	05/07/2025	12:30					
					VOCMS Group1		8260-Low	5 Bus. Days	
Q1985-02	TB-20250507	Water	05/07/2025	09:00					
					VOCMS Group1		8260-Low	5 Bus. Days	

Relinquished By :

Date / Time :


5/8/25 1143

Received By :

Date / Time :


5/8/25 1143

Storage Area : VOA Refridgerator Room

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046041.D
 Acq On : 05 May 2025 11:35
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Quant Time: May 06 06:08:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	83671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	147096	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127829	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	31909	12.820	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	25.640%#
35) Dibromofluoromethane	5.385	113	21112	12.804	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	25.600%#
50) Toluene-d8	8.646	98	73333	13.200	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	26.400%#
62) 4-Bromofluorobenzene	11.079	95	26760	13.236	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	26.480%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	23341	12.638	ug/l	96
3) Chloromethane	1.306	50	24316	13.371	ug/l	98
4) Vinyl Chloride	1.373	62	22096	13.868	ug/l	94
5) Bromomethane	1.593	94	9920	12.279	ug/l	93
6) Chloroethane	1.666	64	11832	14.414	ug/l	95
7) Trichlorofluoromethane	1.873	101	34636	14.271	ug/l	100
8) Diethyl Ether	2.129	74	11392	14.278	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.324	101	21027	14.499	ug/l	98
10) Methyl Iodide	2.446	142	25661	15.070	ug/l	99
11) Tert butyl alcohol	2.971	59	20370	68.662	ug/l	97
12) 1,1-Dichloroethene	2.312	96	18899	13.385	ug/l	98
13) Acrolein	2.233	56	26427	76.689	ug/l	95
14) Allyl chloride	2.660	41	37708	14.133	ug/l	99
15) Acrylonitrile	3.062	53	63290	71.312	ug/l	98
16) Acetone	2.379	43	60363	70.884	ug/l	99
17) Carbon Disulfide	2.501	76	43348	13.251	ug/l	100
18) Methyl Acetate	2.702	43	27234	13.397	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	68402	13.890	ug/l	100
20) Methylene Chloride	2.782	84	23050	13.362	ug/l	96
21) trans-1,2-Dichloroethene	3.086	96	19161	13.294	ug/l	96
22) Diisopropyl ether	3.763	45	74257	14.896	ug/l	95
23) Vinyl Acetate	3.720	43	322598	73.230	ug/l	100
24) 1,1-Dichloroethane	3.605	63	41257	14.082	ug/l	98
25) 2-Butanone	4.556	43	90331	73.489	ug/l	98
26) 2,2-Dichloropropane	4.470	77	30472	13.729	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	23970	13.791	ug/l	98
28) Bromochloromethane	4.897	49	21028	13.497	ug/l	99
29) Tetrahydrofuran	5.007	42	56819	71.175	ug/l	100
30) Chloroform	5.086	83	43065	14.144	ug/l	95
31) Cyclohexane	5.458	56	36470	14.769	ug/l	95
32) 1,1,1-Trichloroethane	5.379	97	37003	14.092	ug/l	98
36) 1,1-Dichloropropene	5.684	75	27242	14.231	ug/l	98
37) Ethyl Acetate	4.720	43	33458	13.931	ug/l	98
38) Carbon Tetrachloride	5.671	117	31091	14.137	ug/l	98
39) Methylcyclohexane	7.372	83	35077	14.623	ug/l	99
40) Benzene	6.031	78	83898	14.140	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046041.D
 Acq On : 05 May 2025 11:35
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Quant Time: May 06 06:08:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.915	41	18732	14.038	ug/l	95
42) 1,2-Dichloroethane	6.080	62	37203	15.165	ug/l	100
43) Isopropyl Acetate	6.342	43	53275	14.512	ug/l	99
44) Trichloroethene	7.116	130	20239	14.429	ug/l	99
45) 1,2-Dichloropropane	7.427	63	20940	14.204	ug/l	98
46) Dibromomethane	7.580	93	16781	14.439	ug/l	99
47) Bromodichloromethane	7.817	83	32762	14.545	ug/l	94
48) Methyl methacrylate	7.689	41	27374	14.521	ug/l	100
49) 1,4-Dioxane	7.659	88	10585	282.442	ug/l	98
51) 4-Methyl-2-Pentanone	8.573	43	182455	75.775	ug/l	98
52) Toluene	8.713	92	52030	14.788	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	27549	14.566	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	31226	14.154	ug/l	98
55) 1,1,2-Trichloroethane	9.152	97	20536	14.442	ug/l	96
56) Ethyl methacrylate	9.116	69	31787	14.425	ug/l	98
57) 1,3-Dichloropropane	9.305	76	36341	14.417	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	79531	80.312	ug/l	100
59) 2-Hexanone	9.427	43	137054	74.997	ug/l	99
60) Dibromochloromethane	9.518	129	22267	14.387	ug/l	99
61) 1,2-Dibromoethane	9.610	107	21113	14.458	ug/l	96
64) Tetrachloroethene	9.268	164	19924	15.263	ug/l	97
65) Chlorobenzene	10.079	112	55863	14.202	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.158	131	18671	14.231	ug/l	98
67) Ethyl Benzene	10.189	91	98113	14.631	ug/l	100
68) m/p-Xylenes	10.299	106	72202	29.806	ug/l	99
69) o-Xylene	10.640	106	35178	14.408	ug/l	96
70) Styrene	10.652	104	58034	14.853	ug/l	98
71) Bromoform	10.799	173	13825	14.030	ug/l #	100
73) Isopropylbenzene	10.957	105	93013	14.096	ug/l	99
74) N-amyl acetate	10.841	43	44684	13.581	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	31830	13.560	ug/l	99
76) 1,2,3-Trichloropropane	11.237	75	34837	13.811	ug/l	89
77) Bromobenzene	11.195	156	21695	14.212	ug/l	96
78) n-propylbenzene	11.298	91	106337	14.376	ug/l	99
79) 2-Chlorotoluene	11.359	91	68534	13.748	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	79261	14.468	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	8104	13.371	ug/l	91
82) 4-Chlorotoluene	11.451	91	77338	14.116	ug/l	100
83) tert-Butylbenzene	11.713	119	75382	14.011	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	79232	14.480	ug/l	99
85) sec-Butylbenzene	11.890	105	95271	14.263	ug/l	100
86) p-Isopropyltoluene	12.006	119	77599	14.387	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	39520	13.898	ug/l	98
88) 1,4-Dichlorobenzene	12.036	146	39419	14.146	ug/l	99
89) n-Butylbenzene	12.329	91	66506	14.184	ug/l	99
90) Hexachloroethane	12.536	117	13332	13.287	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	39031	14.092	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.938	75	7232	13.535	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	20846	13.745	ug/l	99
94) Hexachlorobutadiene	13.725	225	9524	13.426	ug/l	97
95) Naphthalene	13.774	128	77542	14.040	ug/l	99
96) 1,2,3-Trichlorobenzene	13.956	180	22301	14.023	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046041.D
Acq On : 05 May 2025 11:35
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Quant Time: May 06 06:08:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 06 06:08:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.543 min Scan# 71
Delta R.T. -0.001 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:168 Resp: 83671
Ion Ratio Lower Upper
168 100
99 72.7 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 12.638 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 85 Resp: 23341
Ion Ratio Lower Upper
85 100
87 33.5 15.7 47.1









#3
Chloromethane
Concen: 13.371 ug/l
RT: 1.306 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 50 Resp: 24316
Ion Ratio Lower Upper
50 100
52 33.0 25.4 38.2



#4
Vinyl Chloride
Concen: 13.868 ug/l
RT: 1.373 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 62 Resp: 22096
Ion Ratio Lower Upper
62 100
64 34.9 25.2 37.8









#5
Bromomethane
Concen: 12.279 ug/l
RT: 1.593 min Scan# 83
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 94 Resp: 9920
Ion Ratio Lower Upper
94 100
96 98.1 72.8 109.2



#6
Chloroethane
Concen: 14.414 ug/l
RT: 1.666 min Scan# 95
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 64 Resp: 11832
Ion Ratio Lower Upper
64 100
66 32.9 24.3 36.5









#7
Trichlorofluoromethane
Concen: 14.271 ug/l
RT: 1.873 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:101 Resp: 34636
Ion Ratio Lower Upper
101 100
103 63.5 51.0 76.4



#8
Diethyl Ether
Concen: 14.278 ug/l
RT: 2.129 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 74 Resp: 11392
Ion Ratio Lower Upper
74 100
45 103.7 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 14.499 ug/l
RT: 2.324 min Scan# 203
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:101 Resp: 21027
Ion Ratio Lower Upper
101 100
85 46.6 38.6 58.0
151 70.4 55.2 82.8



#10
Methyl Iodide
Concen: 15.070 ug/l
RT: 2.446 min Scan# 223
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:142 Resp: 25661
Ion Ratio Lower Upper
142 100
127 53.0 41.7 62.5
141 14.7 11.5 17.3





#11

Tert butyl alcohol

Concen: 68.662 ug/l

RT: 2.971 min Scan# 310

Delta R.T. -0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion: 59 Resp: 20370

Ion Ratio Lower Upper

59 100

57 11.7 8.6 12.8



#12

1,1-Dichloroethene

Concen: 13.385 ug/l

RT: 2.312 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 96 Resp: 18899


Ion Ratio Lower Upper

96 100

61 179.8 146.2 219.2

98 61.8 51.0 76.6





#13

Acrolein

Concen: 76.689 ug/l

RT: 2.233 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion: 56 Resp: 26427

Ion Ratio Lower Upper

56 100

55 74.3 56.2 84.4



#14

Allyl chloride

Concen: 14.133 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 41 Resp: 37708


Ion Ratio Lower Upper

41 100

39 75.0 60.6 90.8

76 30.6 24.9 37.3





#15

Acrylonitrile

Concen: 71.312 ug/l

RT: 3.062 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion: 53 Resp: 63290

Ion Ratio Lower Upper

53 100

52 83.5 65.3 97.9

51 38.4 29.8 44.8



#16

Acetone

Concen: 70.884 ug/l

RT: 2.379 min Scan# 212

Delta R.T. -0.006 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 43 Resp: 60363

Ion Ratio Lower Upper

43 100

58 26.0 21.2 31.8









#17
Carbon Disulfide
Concen: 13.251 ug/l
RT: 2.501 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 76 Resp: 43348
Ion Ratio Lower Upper
76 100
78 8.8 7.0 10.4



#18
Methyl Acetate
Concen: 13.397 ug/l
RT: 2.702 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 43 Resp: 27234
Ion Ratio Lower Upper
43 100
74 20.7 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 13.890 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 73 Resp: 68402
Ion Ratio Lower Upper
73 100
57 22.2 17.7 26.5



#20
Methylene Chloride
Concen: 13.362 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 84 Resp: 23050
Ion Ratio Lower Upper
84 100
49 146.0 113.9 170.9
51 43.8 33.5 50.3
86 62.2 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 13.294 ug/l
RT: 3.086 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 96 Resp: 19161
Ion Ratio Lower Upper
96 100
61 155.6 119.5 179.3
98 63.2 50.0 75.0



#22
Diisopropyl ether
Concen: 14.896 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 45 Resp: 74257
Ion Ratio Lower Upper
45 100
43 78.2 66.6 100.0
87 24.0 19.8 29.6
59 11.7 8.6 12.8









#23
 Vinyl Acetate
 Concen: 73.230 ug/l
 RT: 3.720 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046041.D
 Acq: 05 May 2025 11:35




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020


Tgt Ion: 43 Resp: 322598
 Ion Ratio Lower Upper
 43 100
 86 9.3 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 14.082 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046041.D
 Acq: 05 May 2025 11:35

Tgt Ion: 63 Resp: 41257
 Ion Ratio Lower Upper
 63 100
 98 7.9 3.6 10.8
 100 5.0 2.1 6.3









#25
2-Butanone
Concen: 73.489 ug/l
RT: 4.556 min Scan# 569
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 90331
Ion Ratio Lower Upper
43 100
72 23.8 18.4 27.6



#26
2,2-Dichloropropane
Concen: 13.729 ug/l
RT: 4.470 min Scan# 555
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 77 Resp: 30472
Ion Ratio Lower Upper
77 100
97 20.9 10.5 31.5









#27
 cis-1,2-Dichloroethene
 Concen: 13.791 ug/l
 RT: 4.483 min Scan# 557
 Delta R.T. -0.000 min
 Lab File: VX046041.D
 Acq: 05 May 2025 11:35

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC020

Tgt Ion: 96 Resp: 23970




Ion	Ratio	Lower	Upper
96	100		
61	158.3	0.0	322.8
98	63.5	0.0	129.0




#28
 Bromochloromethane
 Concen: 13.497 ug/l
 RT: 4.897 min Scan# 625
 Delta R.T. -0.000 min
 Lab File: VX046041.D
 Acq: 05 May 2025 11:35

Tgt Ion: 49 Resp: 21028

Ion	Ratio	Lower	Upper
49	100		
129	1.9	0.0	4.0
130	70.9	56.2	84.2





#29

Tetrahydrofuran

Concen: 71.175 ug/l

RT: 5.007 min Scan# 642

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion: 42 Resp: 56819

Ion Ratio Lower Upper

42 100

72 42.3 33.7 50.5

71 39.2 31.4 47.2



#30

Chloroform

Concen: 14.144 ug/l

RT: 5.086 min Scan# 656

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 83 Resp: 43065

Ion Ratio Lower Upper

83 100

85 65.7 49.3 73.9









#31
Cyclohexane
Concen: 14.769 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 56 Resp: 36470
Ion Ratio Lower Upper
56 100
69 28.8 24.4 36.6
84 78.5 66.9 100.3



#32
1,1,1-Trichloroethane
Concen: 14.092 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 97 Resp: 37003
Ion Ratio Lower Upper
97 100
99 62.8 51.8 77.6
61 49.5 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 12.820 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion: 65 Resp: 31909

Ion Ratio Lower Upper

65 100

67 49.9 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 114 Resp: 147096


Ion Ratio Lower Upper

114 100

63 25.2 0.0 49.2

88 18.0 0.0 33.6









#35
Dibromofluoromethane
Concen: 12.804 ug/l
RT: 5.385 min Scan# 704
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:113 Resp: 21112
Ion Ratio Lower Upper
113 100
111 103.0 83.1 124.7
192 17.0 13.3 19.9



#36
1,1-Dichloropropene
Concen: 14.231 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 75 Resp: 27242
Ion Ratio Lower Upper
75 100
110 31.6 16.3 48.9
77 31.9 24.3 36.5





#37

Ethyl Acetate

Concen: 13.931 ug/l

RT: 4.720 min Scan# 596

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion: 43 Resp: 33458

Ion Ratio Lower Upper

43 100

61 12.6 10.3 15.5

70 10.9 7.9 11.9



#38

Carbon Tetrachloride

Concen: 14.137 ug/l

RT: 5.671 min Scan# 752

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 117 Resp: 31091


Ion Ratio Lower Upper

117 100

119 95.6 75.2 112.8

121 32.1 24.2 36.4





#39

Methylcyclohexane

Concen: 14.623 ug/l

RT: 7.372 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion: 83 Resp: 35077

Ion Ratio Lower Upper

83 100

55 80.7 64.7 97.1

98 44.8 37.4 56.2



#40

Benzene

Concen: 14.140 ug/l

RT: 6.031 min Scan# 811

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 78 Resp: 83898

Ion Ratio Lower Upper

78 100

77 22.3 19.0 28.4









#41
Methacrylonitrile
Concen: 14.038 ug/l
RT: 4.915 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion:	41	Resp:	18732
Ion	Ratio	Lower	Upper
41	100		
39	62.6	47.2	70.8
67	68.3	50.7	76.1
52	32.5	24.6	37.0



#42
1,2-Dichloroethane
Concen: 15.165 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:	62	Resp:	37203
Ion	Ratio	Lower	Upper
62	100		
98	7.6	0.0	15.2









#43
Isopropyl Acetate
Concen: 14.512 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 53275
Ion Ratio Lower Upper
43 100
61 17.3 14.3 21.5
87 11.6 9.5 14.3



#44
Trichloroethene
Concen: 14.429 ug/l
RT: 7.116 min Scan# 989
Delta R.T. -0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:130 Resp: 20239
Ion Ratio Lower Upper
130 100
95 102.8 0.0 204.2





#45

1,2-Dichloropropane

Concen: 14.204 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion: 63 Resp: 20940

Ion Ratio Lower Upper

63 100

65 29.1 24.1 36.1



#46

Dibromomethane

Concen: 14.439 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 93 Resp: 16781


Ion Ratio Lower Upper

93 100

95 81.5 65.6 98.4

174 83.1 68.2 102.2









#47
Bromodichloromethane
Concen: 14.545 ug/l
RT: 7.817 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 83 Resp: 32762
Ion Ratio Lower Upper
83 100
85 68.4 50.5 75.7
127 8.3 6.5 9.7



#48
Methyl methacrylate
Concen: 14.521 ug/l
RT: 7.689 min Scan# 1083
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 41 Resp: 27374
Ion Ratio Lower Upper
41 100
69 73.5 58.5 87.7
39 64.9 51.7 77.5









#49
1,4-Dioxane
Concen: 282.442 ug/l
RT: 7.659 min Scan# 1078
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 88 Resp: 10585
Ion Ratio Lower Upper
88 100
43 45.5 33.4 50.2
58 73.6 58.6 88.0



#50
Toluene-d8
Concen: 13.200 ug/l
RT: 8.646 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 98 Resp: 73333
Ion Ratio Lower Upper
98 100
100 64.8 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 75.775 ug/l
RT: 8.573 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 182455
Ion Ratio Lower Upper
43 100
58 34.8 28.9 43.3



#52
Toluene
Concen: 14.788 ug/l
RT: 8.713 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 92 Resp: 52030
Ion Ratio Lower Upper
92 100
91 172.2 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 14.566 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion: 75 Resp: 27549

Ion Ratio Lower Upper

75 100

77 31.4 25.0 37.6



#54

cis-1,3-Dichloropropene

Concen: 14.154 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 75 Resp: 31226


Ion Ratio Lower Upper

75 100

77 30.8 25.4 38.0

39 64.0 52.2 78.4









#55
1,1,2-Trichloroethane
Concen: 14.442 ug/l
RT: 9.152 min Scan# 1323
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:	97	Resp:	20536
Ion	Ratio	Lower	Upper
97	100		
83	87.0	69.0	103.6
85	58.7	44.3	66.5
99	66.8	47.8	71.6



#56
Ethyl methacrylate
Concen: 14.425 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:	69	Resp:	31787
Ion	Ratio	Lower	Upper
69	100		
41	78.2	60.8	91.2
39	49.3	39.0	58.6





#57

1,3-Dichloropropane

Concen: 14.417 ug/l

RT: 9.305 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC020

Tgt Ion: 76 Resp: 36341

Ion Ratio Lower Upper

76 100

78 31.8 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 80.312 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 63 Resp: 79531

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3









#59
2-Hexanone
Concen: 74.997 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 137054
Ion Ratio Lower Upper
43 100
58 48.7 24.9 74.6



#60
Dibromochloromethane
Concen: 14.387 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 129 Resp: 22267
Ion Ratio Lower Upper
129 100
127 78.0 39.3 117.8









#61
1,2-Dibromoethane
Concen: 14.458 ug/l
RT: 9.610 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:107 Resp: 21113
Ion Ratio Lower Upper
107 100
109 95.3 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 13.236 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 95 Resp: 26760
Ion Ratio Lower Upper
95 100
174 67.4 0.0 135.8
176 67.0 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:117 Resp: 127829
Ion Ratio Lower Upper
117 100
82 57.7 50.6 76.0
119 32.7 25.8 38.6



#64
Tetrachloroethene
Concen: 15.263 ug/l
RT: 9.268 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:164 Resp: 19924
Ion Ratio Lower Upper
164 100
166 127.2 105.0 157.6
129 107.6 83.5 125.3
131 97.7 76.5 114.7





#65

Chlorobenzene

Concen: 14.202 ug/l

RT: 10.079 min Scan# 1475

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC020

Tgt Ion:112 Resp: 55863

Ion Ratio Lower Upper

112 100

114 32.0 25.4 38.2



#66

1,1,1,2-Tetrachloroethane

Concen: 14.231 ug/l

RT: 10.158 min Scan# 1488

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:131 Resp: 18671


Ion Ratio Lower Upper

131 100

133 94.5 47.3 141.9

119 66.7 31.6 95.0









#67
Ethyl Benzene
Concen: 14.631 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion: 91 Resp: 98113
Ion Ratio Lower Upper
91 100
106 29.3 23.4 35.2



#68
m/p-Xylenes
Concen: 29.806 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 106 Resp: 72202
Ion Ratio Lower Upper
106 100
91 216.1 171.2 256.8









#69
o-Xylene
Concen: 14.408 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:106 Resp: 35178
Ion Ratio Lower Upper
106 100
91 231.7 112.7 338.1



#70
Styrene
Concen: 14.853 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:104 Resp: 58034
Ion Ratio Lower Upper
104 100
78 58.7 45.7 68.5
103 55.6 43.7 65.5





#71

Bromoform

Concen: 14.030 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion:173 Resp: 13825

Ion Ratio Lower Upper

173 100

175 46.4 23.4 70.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:152 Resp: 60503


Ion Ratio Lower Upper

152 100

115 76.6 46.9 140.7

150 163.6 0.0 351.0









#73
Isopropylbenzene
Concen: 14.096 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:105 Resp: 93013
Ion Ratio Lower Upper
105 100
120 26.1 12.8 38.4



#74
N-ethyl acetate
Concen: 13.581 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 43 Resp: 44684
Ion Ratio Lower Upper
43 100
70 38.5 30.9 46.3
55 24.1 18.7 28.1
61 22.3 17.1 25.7





#75

1,1,2-Tetrachloroethane

Concen: 13.560 ug/l

RT: 11.207 min Scan# 1660

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion: 83 Resp: 31830

Ion Ratio Lower Upper

83 100

131 10.1 5.0 14.9

85 63.3 31.9 95.7



#76

1,2,3-Trichloropropane

Concen: 13.811 ug/l

RT: 11.237 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 75 Resp: 34837

Ion Ratio Lower Upper

75 100

77 34.4 20.5 61.5





#77

Bromobenzene

Concen: 14.212 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:156 Resp: 21695

Ion Ratio Lower Upper

156 100

77 179.9 92.8 278.3

158 93.1 48.8 146.3



#78

n-propylbenzene

Concen: 14.376 ug/l

RT: 11.298 min Scan# 1675

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 91 Resp: 106337

Ion Ratio Lower Upper

91 100

120 22.0 10.8 32.4









#79
2-Chlorotoluene
Concen: 13.748 ug/l
RT: 11.359 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 91 Resp: 68534
Ion Ratio Lower Upper
91 100
126 31.4 15.6 46.7



#80
1,3,5-Trimethylbenzene
Concen: 14.468 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 105 Resp: 79261
Ion Ratio Lower Upper
105 100
120 45.9 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 13.371 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020


Tgt Ion: 75 Resp: 8104
Ion Ratio Lower Upper
75 100
53 138.3 100.5 150.7
89 44.6 37.0 55.6



#82
4-Chlorotoluene
Concen: 14.116 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 91 Resp: 77338
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83

tert-Butylbenzene

Concen: 14.011 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion:119 Resp: 75382

Ion Ratio Lower Upper

119 100

91 66.8 32.9 98.7

134 23.7 11.4 34.1



#84

1,2,4-Trimethylbenzene

Concen: 14.480 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion:105 Resp: 79232

Ion Ratio Lower Upper

105 100

120 42.0 21.2 63.6









#85
sec-Butylbenzene
Concen: 14.263 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:105 Resp: 95271
Ion Ratio Lower Upper
105 100
134 19.3 9.7 29.1



#86
p-Isopropyltoluene
Concen: 14.387 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:119 Resp: 77599
Ion Ratio Lower Upper
119 100
134 25.9 12.5 37.5
91 28.8 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 13.898 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:146 Resp: 39520

Ion Ratio Lower Upper

146 100

111 42.7 22.1 66.3

148 62.9 32.1 96.5



#88

1,4-Dichlorobenzene

Concen: 14.146 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:146 Resp: 39419


Ion Ratio Lower Upper

146 100

111 42.8 21.3 63.9

148 64.2 31.9 95.5





#89

n-Butylbenzene

Concen: 14.184 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion: 91 Resp: 66506

Ion Ratio Lower Upper

91 100

92 52.9 26.9 80.7

134 23.1 11.8 35.3



#90

Hexachloroethane

Concen: 13.287 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 117 Resp: 13332

Ion Ratio Lower Upper

117 100

201 62.9 31.6 94.7









#91
1,2-Dichlorobenzene
Concen: 14.092 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020


Tgt Ion:146 Resp: 39031
Ion Ratio Lower Upper
146 100
111 46.6 22.4 67.3
148 64.2 31.6 94.8



#92
1,2-Dibromo-3-Chloropropane
Concen: 13.535 ug/l
RT: 12.938 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 75 Resp: 7232
Ion Ratio Lower Upper
75 100
155 69.4 34.9 104.8
157 88.4 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 13.745 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:180 Resp: 20846

Ion Ratio Lower Upper

180 100

182 96.6 47.9 143.8

145 34.1 17.1 51.2



#94

Hexachlorobutadiene

Concen: 13.426 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:225 Resp: 9524


Ion Ratio Lower Upper

225 100

223 65.6 30.8 92.4

227 61.4 30.9 92.7





#95

Naphthalene

Concen: 14.040 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:128 Resp: 77542

Ion Ratio Lower Upper

128 100

127 12.9 10.4 15.6

129 11.1 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 14.023 ug/l

RT: 13.956 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:180 Resp: 22301

Ion Ratio Lower Upper

180 100

182 92.1 47.8 143.3

145 36.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046042.D
 Acq On : 05 May 2025 11:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Quant Time: May 06 06:09:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	97076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	167947	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	146257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67976	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	88371	30.601	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	61.200%#		
35) Dibromofluoromethane	5.379	113	59550	31.633	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	63.260%#		
50) Toluene-d8	8.647	98	205479	32.395	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	64.780%#		
62) 4-Bromofluorobenzene	11.079	95	79012	34.229	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	68.460%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	83826	39.120	ug/l	100
3) Chloromethane	1.307	50	75238	35.658	ug/l	100
4) Vinyl Chloride	1.374	62	68901	37.273	ug/l	100
5) Bromomethane	1.593	94	31648	33.765	ug/l	100
6) Chloroethane	1.666	64	36722	38.558	ug/l	100
7) Trichlorofluoromethane	1.874	101	103693	36.824	ug/l	100
8) Diethyl Ether	2.136	74	32724	35.350	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	62251	36.997	ug/l	100
10) Methyl Iodide	2.447	142	78277	39.622	ug/l	100
11) Tert butyl alcohol	2.977	59	62557	181.745	ug/l	100
12) 1,1-Dichloroethene	2.313	96	58334	35.610	ug/l	100
13) Acrolein	2.233	56	73927	184.907	ug/l	100
14) Allyl chloride	2.654	41	114445	36.971	ug/l	100
15) Acrylonitrile	3.062	53	188304	182.874	ug/l	100
16) Acetone	2.386	43	175777	177.910	ug/l	100
17) Carbon Disulfide	2.502	76	141281	37.223	ug/l	100
18) Methyl Acetate	2.703	43	82347	34.915	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	209665	36.695	ug/l	100
20) Methylene Chloride	2.782	84	66412	33.182	ug/l	100
21) trans-1,2-Dichloroethene	3.087	96	59234	35.421	ug/l	100
22) Diisopropyl ether	3.757	45	221172	38.241	ug/l	100
23) Vinyl Acetate	3.721	43	994236	194.528	ug/l	100
24) 1,1-Dichloroethane	3.605	63	122601	36.067	ug/l	100
25) 2-Butanone	4.556	43	269224	188.783	ug/l	100
26) 2,2-Dichloropropane	4.465	77	92857	36.059	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	71518	35.466	ug/l	100
28) Bromochloromethane	4.898	49	56091	31.031	ug/l	100
29) Tetrahydrofuran	5.001	42	170124	183.679	ug/l	100
30) Chloroform	5.087	83	125850	35.626	ug/l	100
31) Cyclohexane	5.464	56	109459	38.205	ug/l	100
32) 1,1,1-Trichloroethane	5.373	97	109781	36.034	ug/l	100
36) 1,1-Dichloropropene	5.684	75	83215	38.073	ug/l	100
37) Ethyl Acetate	4.715	43	102537	37.394	ug/l	100
38) Carbon Tetrachloride	5.672	117	93712	37.320	ug/l	100
39) Methylcyclohexane	7.373	83	107651	39.307	ug/l	100
40) Benzene	6.031	78	247476	36.530	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046042.D
 Acq On : 05 May 2025 11:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Quant Time: May 06 06:09:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	58082	38.123	ug/l	100
42) 1,2-Dichloroethane	6.080	62	105332	37.605	ug/l	100
43) Isopropyl Acetate	6.336	43	161787	38.598	ug/l	100
44) Trichloroethene	7.123	130	59623	37.230	ug/l	100
45) 1,2-Dichloropropane	7.428	63	62387	37.064	ug/l	100
46) Dibromomethane	7.574	93	48201	36.326	ug/l	100
47) Bromodichloromethane	7.818	83	96916	37.685	ug/l	100
48) Methyl methacrylate	7.690	41	84277	39.157	ug/l	100
49) 1,4-Dioxane	7.659	88	30529	713.477	ug/l	100
51) 4-Methyl-2-Pentanone	8.574	43	532388	193.653	ug/l	100
52) Toluene	8.714	92	150757	37.530	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	88655	41.056	ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	97031	38.521	ug/l	100
55) 1,1,2-Trichloroethane	9.147	97	59499	36.648	ug/l	100
56) Ethyl methacrylate	9.116	69	99947	39.726	ug/l	100
57) 1,3-Dichloropropane	9.305	76	104606	36.347	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	257678	227.902	ug/l	100
59) 2-Hexanone	9.427	43	396784	190.168	ug/l	100
60) Dibromochloromethane	9.519	129	67198	38.027	ug/l	100
61) 1,2-Dibromoethane	9.604	107	62633	37.567	ug/l	100
64) Tetrachloroethene	9.269	164	54853	36.726	ug/l	100
65) Chlorobenzene	10.079	112	160600	35.686	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	57066	38.017	ug/l	100
67) Ethyl Benzene	10.189	91	295712	38.543	ug/l	100
68) m/p-Xylenes	10.299	106	216578	78.142	ug/l	100
69) o-Xylene	10.640	106	106335	38.064	ug/l	100
70) Styrene	10.653	104	178313	39.888	ug/l	100
71) Bromoform	10.799	173	44486	39.457	ug/l #	100
73) Isopropylbenzene	10.957	105	280719	37.865	ug/l	100
74) N-amyl acetate	10.842	43	140509	38.012	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.207	83	90975	34.495	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	102335	36.110	ug/l	86
77) Bromobenzene	11.195	156	63053	36.763	ug/l	100
78) n-propylbenzene	11.299	91	329981	39.707	ug/l	100
79) 2-Chlorotoluene	11.360	91	203509	36.336	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	237066	38.516	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	26179	38.445	ug/l	100
82) 4-Chlorotoluene	11.451	91	233172	37.880	ug/l	100
83) tert-Butylbenzene	11.713	119	233468	38.622	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	239433	38.946	ug/l	100
85) sec-Butylbenzene	11.890	105	291074	38.786	ug/l	100
86) p-Isopropyltoluene	12.006	119	241658	39.879	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	115651	36.199	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	115097	36.763	ug/l	100
89) n-Butylbenzene	12.329	91	213899	40.603	ug/l	100
90) Hexachloroethane	12.536	117	42314	37.534	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	115286	37.047	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21909	36.496	ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	66655	39.117	ug/l	100
94) Hexachlorobutadiene	13.725	225	29007	36.395	ug/l	100
95) Naphthalene	13.774	128	245568	39.575	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	69262	38.765	ug/l	100


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046042.D
Acq On : 05 May 2025 11:58
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Quant Time: May 06 06:09:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:168 Resp: 97076
Ion Ratio Lower Upper
168 100
99 68.6 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 39.120 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 85 Resp: 83826
Ion Ratio Lower Upper
85 100
87 31.4 15.7 47.1









#3
Chloromethane
Concen: 35.658 ug/l
RT: 1.307 min Scan# 30
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 50 Resp: 75238
Ion Ratio Lower Upper
50 100
52 31.8 25.4 38.2



#4
Vinyl Chloride
Concen: 37.273 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 62 Resp: 68901
Ion Ratio Lower Upper
62 100
64 31.5 25.2 37.8









#5
Bromomethane
Concen: 33.765 ug/l
RT: 1.593 min Scan# 83
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 94 Resp: 31648
Ion Ratio Lower Upper
94 100
96 91.0 72.8 109.2



#6
Chloroethane
Concen: 38.558 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 64 Resp: 36722
Ion Ratio Lower Upper
64 100
66 30.4 24.3 36.5









#7
 Trichlorofluoromethane
 Concen: 36.824 ug/l
 RT: 1.874 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion:101 Resp: 103693
 Ion Ratio Lower Upper
 101 100
 103 63.7 51.0 76.4



#8
 Diethyl Ether
 Concen: 35.350 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58

Tgt Ion: 74 Resp: 32724
 Ion Ratio Lower Upper
 74 100
 45 109.9 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 36.997 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:101 Resp: 62251
Ion Ratio Lower Upper
101 100
85 48.3 38.6 58.0
151 69.0 55.2 82.8



#10
Methyl Iodide
Concen: 39.622 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:142 Resp: 78277
Ion Ratio Lower Upper
142 100
127 52.1 41.7 62.5
141 14.4 11.5 17.3









#11
Tert butyl alcohol
Concen: 181.745 ug/l
RT: 2.977 min Scan# 310
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 59 Resp: 62557
Ion Ratio Lower Upper
59 100
57 10.7 8.6 12.8



#12
1,1-Dichloroethene
Concen: 35.610 ug/l
RT: 2.313 min Scan# 201
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 96 Resp: 58334
Ion Ratio Lower Upper
96 100
61 182.7 146.2 219.2
98 63.8 51.0 76.6









#13
Acrolein
Concen: 184.907 ug/l
RT: 2.233 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 56 Resp: 73927
Ion Ratio Lower Upper
56 100
55 70.3 56.2 84.4



#14
Allyl chloride
Concen: 36.971 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 41 Resp: 114445
Ion Ratio Lower Upper
41 100
39 75.7 60.6 90.8
76 31.1 24.9 37.3









#15
Acrylonitrile
Concen: 182.874 ug/l
RT: 3.062 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 53 Resp: 188304
Ion Ratio Lower Upper
53 100
52 81.6 65.3 97.9
51 37.3 29.8 44.8



#16
Acetone
Concen: 177.910 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 43 Resp: 175777
Ion Ratio Lower Upper
43 100
58 26.5 21.2 31.8









#17
Carbon Disulfide
Concen: 37.223 ug/l
RT: 2.502 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 76 Resp: 141281
Ion Ratio Lower Upper
76 100
78 8.7 7.0 10.4



#18
Methyl Acetate
Concen: 34.915 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 43 Resp: 82347
Ion Ratio Lower Upper
43 100
74 20.9 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 36.695 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 73 Resp: 209665
Ion Ratio Lower Upper
73 100
57 22.1 17.7 26.5



#20
Methylene Chloride
Concen: 33.182 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 84 Resp: 66412
Ion Ratio Lower Upper
84 100
49 142.4 113.9 170.9
51 41.9 33.5 50.3
86 67.3 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 35.421 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 96 Resp: 59234
Ion Ratio Lower Upper
96 100
61 149.4 119.5 179.3
98 62.5 50.0 75.0



#22
Diisopropyl ether
Concen: 38.241 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 45 Resp: 221172
Ion Ratio Lower Upper
45 100
43 83.3 66.6 100.0
87 24.7 19.8 29.6
59 10.7 8.6 12.8









#23
 Vinyl Acetate
 Concen: 194.528 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 43 Resp: 994236
 Ion Ratio Lower Upper
 43 100
 86 9.4 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 36.067 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58

Tgt Ion: 63 Resp: 122601
 Ion Ratio Lower Upper
 63 100
 98 7.2 3.6 10.8
 100 4.2 2.1 6.3









#25
2-Butanone
Concen: 188.783 ug/l
RT: 4.556 min Scan# 569
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 269224
Ion Ratio Lower Upper
43 100
72 23.0 18.4 27.6



#26
2,2-Dichloropropane
Concen: 36.059 ug/l
RT: 4.465 min Scan# 554
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 77 Resp: 92857
Ion Ratio Lower Upper
77 100
97 21.0 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 35.466 ug/l

RT: 4.483 min Scan# 557

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 96 Resp: 71518

Ion Ratio Lower Upper

96 100

61 161.4 0.0 322.8

98 64.5 0.0 129.0



#28

Bromochloromethane

Concen: 31.031 ug/l

RT: 4.898 min Scan# 625

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 49 Resp: 56091


Ion Ratio Lower Upper

49 100

129 2.0 0.0 4.0

130 70.2 56.2 84.2





#29

Tetrahydrofuran

Concen: 183.679 ug/l

RT: 5.001 min Scan# 642

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 42 Resp: 170124

Ion Ratio Lower Upper

42 100

72 42.1 33.7 50.5

71 39.3 31.4 47.2



#30

Chloroform

Concen: 35.626 ug/l

RT: 5.087 min Scan# 656

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 83 Resp: 125850

Ion Ratio Lower Upper

83 100

85 61.6 49.3 73.9









#31
Cyclohexane
Concen: 38.205 ug/l
RT: 5.464 min Scan# 718
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 56 Resp: 109459
Ion Ratio Lower Upper
56 100
69 30.5 24.4 36.6
84 83.6 66.9 100.3



#32
1,1,1-Trichloroethane
Concen: 36.034 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 97 Resp: 109781
Ion Ratio Lower Upper
97 100
99 64.7 51.8 77.6
61 50.1 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 30.601 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 65 Resp: 88371

Ion Ratio Lower Upper

65 100

67 49.5 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 114 Resp: 167947


Ion Ratio Lower Upper

114 100

63 24.6 0.0 49.2

88 16.8 0.0 33.6









#35
Dibromofluoromethane
Concen: 31.633 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:113 Resp: 59550
Ion Ratio Lower Upper
113 100
111 103.9 83.1 124.7
192 16.6 13.3 19.9



#36
1,1-Dichloropropene
Concen: 38.073 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 75 Resp: 83215
Ion Ratio Lower Upper
75 100
110 32.6 16.3 48.9
77 30.4 24.3 36.5









#37
Ethyl Acetate
Concen: 37.394 ug/l
RT: 4.715 min Scan# 595
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 102537
Ion Ratio Lower Upper
43 100
61 12.9 10.3 15.5
70 9.9 7.9 11.9



#38
Carbon Tetrachloride
Concen: 37.320 ug/l
RT: 5.672 min Scan# 752
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 117 Resp: 93712
Ion Ratio Lower Upper
117 100
119 94.0 75.2 112.8
121 30.3 24.2 36.4









#39
Methylcyclohexane
Concen: 39.307 ug/l
RT: 7.373 min Scan# 1031
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 83 Resp: 107651
Ion Ratio Lower Upper
83 100
55 80.9 64.7 97.1
98 46.8 37.4 56.2



#40
Benzene
Concen: 36.530 ug/l
RT: 6.031 min Scan# 811
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 78 Resp: 247476
Ion Ratio Lower Upper
78 100
77 23.7 19.0 28.4









#41
Methacrylonitrile
Concen: 38.123 ug/l
RT: 4.916 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:	41	Resp:	58082
Ion	Ratio	Lower	Upper
41	100		
39	59.0	47.2	70.8
67	63.4	50.7	76.1
52	30.8	24.6	37.0



#42
1,2-Dichloroethane
Concen: 37.605 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:	62	Resp:	105332
Ion	Ratio	Lower	Upper
62	100		
98	7.6	0.0	15.2





#43

Isopropyl Acetate

Concen: 38.598 ug/l

RT: 6.336 min Scan# 861

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 43 Resp: 161787

Ion Ratio Lower Upper

43 100

61 17.9 14.3 21.5

87 11.9 9.5 14.3



#44

Trichloroethene

Concen: 37.230 ug/l

RT: 7.123 min Scan# 990

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:130 Resp: 59623

Ion Ratio Lower Upper

130 100

95 102.1 0.0 204.2





#45

1,2-Dichloropropane

Concen: 37.064 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 63 Resp: 62387

Ion Ratio Lower Upper

63 100

65 30.1 24.1 36.1



#46

Dibromomethane

Concen: 36.326 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 93 Resp: 48201


Ion Ratio Lower Upper

93 100

95 82.0 65.6 98.4

174 85.2 68.2 102.2









#47
Bromodichloromethane
Concen: 37.685 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 83 Resp: 96916
Ion Ratio Lower Upper
83 100
85 63.1 50.5 75.7
127 8.1 6.5 9.7



#48
Methyl methacrylate
Concen: 39.157 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 41 Resp: 84277
Ion Ratio Lower Upper
41 100
69 73.1 58.5 87.7
39 64.6 51.7 77.5





#49

1,4-Dioxane

Concen: 713.477 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 88 Resp: 30529

Ion Ratio Lower Upper

88 100

43 41.8 33.4 50.2

58 73.3 58.6 88.0



#50

Toluene-d8

Concen: 32.395 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 98 Resp: 205479

Ion Ratio Lower Upper

98 100

100 66.9 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 193.653 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 532388
Ion Ratio Lower Upper
43 100
58 36.1 28.9 43.3



#52
Toluene
Concen: 37.530 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 92 Resp: 150757
Ion Ratio Lower Upper
92 100
91 170.8 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 41.056 ug/l

RT: 8.976 min Scan# 1294

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 75 Resp: 88655

Ion Ratio Lower Upper

75 100

77 31.3 25.0 37.6



#54

cis-1,3-Dichloropropene

Concen: 38.521 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 75 Resp: 97031


Ion Ratio Lower Upper

75 100

77 31.7 25.4 38.0

39 65.3 52.2 78.4









#55
 1,1,2-Trichloroethane
 Concen: 36.648 ug/l
 RT: 9.147 min Scan# 1322
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion:	97	Resp:	59499
Ion	Ratio	Lower	Upper
97	100		
83	86.3	69.0	103.6
85	55.4	44.3	66.5
99	59.7	47.8	71.6



#56
 Ethyl methacrylate
 Concen: 39.726 ug/l
 RT: 9.116 min Scan# 1317
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58

Tgt Ion:	69	Resp:	99947
Ion	Ratio	Lower	Upper
69	100		
41	76.0	60.8	91.2
39	48.8	39.0	58.6





#57

1,3-Dichloropropane

Concen: 36.347 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 76 Resp: 104606

Ion Ratio Lower Upper

76 100

78 32.9 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 227.902 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 63 Resp: 257678

Ion Ratio Lower Upper

63 100

106 26.9 21.5 32.3





#59

2-Hexanone

Concen: 190.168 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 43 Resp: 396784

Ion Ratio Lower Upper

43 100

58 49.7 24.9 74.6



#60

Dibromochloromethane

Concen: 38.027 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 129 Resp: 67198

Ion Ratio Lower Upper

129 100

127 78.5 39.3 117.8









#61
1,2-Dibromoethane
Concen: 37.567 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:107 Resp: 62633
Ion Ratio Lower Upper
107 100
109 91.6 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 34.229 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 95 Resp: 79012
Ion Ratio Lower Upper
95 100
174 67.9 0.0 135.8
176 65.7 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:117 Resp: 146257
Ion Ratio Lower Upper
117 100
82 63.3 50.6 76.0
119 32.2 25.8 38.6



#64
Tetrachloroethene
Concen: 36.726 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:164 Resp: 54853
Ion Ratio Lower Upper
164 100
166 131.3 105.0 157.6
129 104.4 83.5 125.3
131 95.6 76.5 114.7





#65

Chlorobenzene

Concen: 35.686 ug/l

RT: 10.079 min Scan# 1475

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:112 Resp: 160600

Ion Ratio Lower Upper

112 100

114 31.8 25.4 38.2



#66

1,1,1,2-Tetrachloroethane

Concen: 38.017 ug/l

RT: 10.159 min Scan# 1488

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:131 Resp: 57066


Ion Ratio Lower Upper

131 100

133 94.6 47.3 141.9

119 63.3 31.6 95.0









#67
Ethyl Benzene
Concen: 38.543 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 295712
Ion Ratio Lower Upper
91 100
106 29.3 23.4 35.2



#68
m/p-Xylenes
Concen: 78.142 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 106 Resp: 216578
Ion Ratio Lower Upper
106 100
91 214.0 171.2 256.8









#69
o-Xylene
Concen: 38.064 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:106 Resp: 106335
Ion Ratio Lower Upper
106 100
91 225.4 112.7 338.1



#70
Styrene
Concen: 39.888 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:104 Resp: 178313
Ion Ratio Lower Upper
104 100
78 57.1 45.7 68.5
103 54.6 43.7 65.5





#71

Bromoform

Concen: 39.457 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:173 Resp: 44486

Ion Ratio Lower Upper

173 100

175 46.7 23.4 70.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:152 Resp: 67976


Ion Ratio Lower Upper

152 100

115 93.8 46.9 140.7

150 175.5 0.0 351.0









#73
Isopropylbenzene
Concen: 37.865 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:105 Resp: 280719
Ion Ratio Lower Upper
105 100
120 25.6 12.8 38.4



#74
N-ethyl acetate
Concen: 38.012 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 43 Resp: 140509
Ion Ratio Lower Upper
43 100
70 38.6 30.9 46.3
55 23.4 18.7 28.1
61 21.4 17.1 25.7









#75
 1,1,2,2-Tetrachloroethane
 Concen: 34.495 ug/l
 RT: 11.207 min Scan# 1660
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 83 Resp: 90975
 Ion Ratio Lower Upper
 83 100
 131 9.9 5.0 14.9
 85 63.8 31.9 95.7



#76
 1,2,3-Trichloropropane
 Concen: 36.110 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58

Tgt Ion: 75 Resp: 102335
 Ion Ratio Lower Upper
 75 100
 77 32.3 20.5 61.5









#77
Bromobenzene
Concen: 36.763 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:156 Resp: 63053
Ion Ratio Lower Upper
156 100
77 185.5 92.8 278.3
158 97.5 48.8 146.3



#78
n-propylbenzene
Concen: 39.707 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 91 Resp: 329981
Ion Ratio Lower Upper
91 100
120 21.6 10.8 32.4









#79
2-Chlorotoluene
Concen: 36.336 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 203509
Ion Ratio Lower Upper
91 100
126 31.1 15.6 46.7



#80
1,3,5-Trimethylbenzene
Concen: 38.516 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 105 Resp: 237066
Ion Ratio Lower Upper
105 100
120 46.1 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 38.445 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 75 Resp: 26179
Ion Ratio Lower Upper
75 100
53 125.6 100.5 150.7
89 46.3 37.0 55.6



#82
4-Chlorotoluene
Concen: 37.880 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 91 Resp: 233172
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83

tert-Butylbenzene

Concen: 38.622 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:119 Resp: 233468

Ion Ratio Lower Upper

119 100

91 65.8 32.9 98.7

134 22.7 11.4 34.1



#84

1,2,4-Trimethylbenzene

Concen: 38.946 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:105 Resp: 239433

Ion Ratio Lower Upper

105 100

120 42.4 21.2 63.6





#85

sec-Butylbenzene

Concen: 38.786 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:105 Resp: 291074

Ion Ratio Lower Upper

105 100

134 19.4 9.7 29.1



#86

p-Isopropyltoluene

Concen: 39.879 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:119 Resp: 241658


Ion Ratio Lower Upper

119 100

134 25.0 12.5 37.5

91 27.6 13.8 41.4









#87
1,3-Dichlorobenzene
Concen: 36.199 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:146 Resp: 115651
Ion Ratio Lower Upper
146 100
111 44.2 22.1 66.3
148 64.3 32.1 96.5



#88
1,4-Dichlorobenzene
Concen: 36.763 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:146 Resp: 115097
Ion Ratio Lower Upper
146 100
111 42.6 21.3 63.9
148 63.7 31.9 95.5





#89

n-Butylbenzene

Concen: 40.603 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 91 Resp: 213899

Ion Ratio Lower Upper

91 100

92 53.8 26.9 80.7

134 23.5 11.8 35.3



#90

Hexachloroethane

Concen: 37.534 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 117 Resp: 42314

Ion Ratio Lower Upper

117 100

201 63.1 31.6 94.7









#91
1,2-Dichlorobenzene
Concen: 37.047 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:146 Resp: 115286
Ion Ratio Lower Upper
146 100
111 44.9 22.4 67.3
148 63.2 31.6 94.8



#92
1,2-Dibromo-3-Chloropropane
Concen: 36.496 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 75 Resp: 21909
Ion Ratio Lower Upper
75 100
155 69.9 34.9 104.8
157 87.6 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 39.117 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:180 Resp: 66655

Ion Ratio Lower Upper

180 100

182 95.9 47.9 143.8

145 34.1 17.1 51.2



#94

Hexachlorobutadiene

Concen: 36.395 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:225 Resp: 29007


Ion Ratio Lower Upper

225 100

223 61.6 30.8 92.4

227 61.8 30.9 92.7





#95

Naphthalene

Concen: 39.575 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:128 Resp: 245568

Ion Ratio Lower Upper

128 100

127 13.0 10.4 15.6

129 10.8 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 38.765 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:180 Resp: 69262

Ion Ratio Lower Upper

180 100

182 95.5 47.8 143.3

145 36.2 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046043.D
 Acq On : 05 May 2025 12:21
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Quant Time: May 06 06:10:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	87028	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	152958	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	135214	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66369	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	161877	62.526	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 125.060%#	
35) Dibromofluoromethane	5.379	113	111455	65.006	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 130.020%#	
50) Toluene-d8	8.647	98	387230	67.031	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 134.060%#	
62) 4-Bromofluorobenzene	11.079	95	153085	72.818	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 145.640%#	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	149508	77.828	ug/l	99
3) Chloromethane	1.307	50	136937	72.392	ug/l	98
4) Vinyl Chloride	1.374	62	126609	76.398	ug/l	98
5) Bromomethane	1.593	94	59246	70.507	ug/l	99
6) Chloroethane	1.660	64	57332	67.148	ug/l	99
7) Trichlorofluoromethane	1.867	101	171042	67.754	ug/l	99
8) Diethyl Ether	2.136	74	58762	70.807	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	109520	72.604	ug/l	99
10) Methyl Iodide	2.440	142	138045	77.943	ug/l	100
11) Tert butyl alcohol	2.995	59	125094	405.394	ug/l	100
12) 1,1-Dichloroethene	2.306	96	105566	71.884	ug/l	96
13) Acrolein	2.239	56	134456	375.132	ug/l	99
14) Allyl chloride	2.660	41	206519	74.419	ug/l	98
15) Acrylonitrile	3.068	53	331142	358.723	ug/l	97
16) Acetone	2.386	43	314383	354.936	ug/l	98
17) Carbon Disulfide	2.501	76	264956	77.867	ug/l	100
18) Methyl Acetate	2.703	43	147014	69.531	ug/l	96
19) Methyl tert-butyl Ether	3.117	73	378113	73.818	ug/l	100
20) Methylene Chloride	2.782	84	120270	67.030	ug/l	96
21) trans-1,2-Dichloroethene	3.087	96	106606	71.109	ug/l	99
22) Diisopropyl ether	3.763	45	399494	77.048	ug/l	90
23) Vinyl Acetate	3.721	43	1811550	395.363	ug/l	100
24) 1,1-Dichloroethane	3.605	63	219759	72.114	ug/l	99
25) 2-Butanone	4.556	43	485448	379.704	ug/l	100
26) 2,2-Dichloropropane	4.464	77	174579	75.621	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	128454	71.056	ug/l	99
28) Bromochloromethane	4.891	49	103602	63.932	ug/l	100
29) Tetrahydrofuran	5.001	42	305757	368.234	ug/l	99
30) Chloroform	5.086	83	222253	70.180	ug/l	95
31) Cyclohexane	5.458	56	196302	76.428	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	201112	73.635	ug/l	98
36) 1,1-Dichloropropene	5.690	75	147873	74.285	ug/l	99
37) Ethyl Acetate	4.714	43	186215	74.565	ug/l	99
38) Carbon Tetrachloride	5.672	117	168946	73.875	ug/l	99
39) Methylcyclohexane	7.372	83	191941	76.952	ug/l	96
40) Benzene	6.031	78	440885	71.457	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046043.D
 Acq On : 05 May 2025 12:21
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Quant Time: May 06 06:10:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	104863	75.574	ug/l	98
42) 1,2-Dichloroethane	6.080	62	187036	73.318	ug/l	99
43) Isopropyl Acetate	6.342	43	300491	78.714	ug/l	100
44) Trichloroethene	7.123	130	105628	72.420	ug/l	98
45) 1,2-Dichloropropane	7.427	63	112554	73.420	ug/l	100
46) Dibromomethane	7.574	93	85728	70.938	ug/l	100
47) Bromodichloromethane	7.818	83	175301	74.845	ug/l	99
48) Methyl methacrylate	7.689	41	153093	78.101	ug/l	99
49) 1,4-Dioxane	7.659	88	57313	1470.687	ug/l	99
51) 4-Methyl-2-Pentanone	8.573	43	963319	384.740	ug/l	100
52) Toluene	8.714	92	270763	74.010	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	169863	86.372	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	184134	80.264	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	107264	72.543	ug/l	99
56) Ethyl methacrylate	9.116	69	188717	82.359	ug/l	99
57) 1,3-Dichloropropane	9.305	76	187649	71.590	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	463169	449.791	ug/l	100
59) 2-Hexanone	9.433	43	729602	383.945	ug/l	100
60) Dibromochloromethane	9.518	129	127012	78.918	ug/l	98
61) 1,2-Dibromoethane	9.610	107	112669	74.200	ug/l	98
64) Tetrachloroethene	9.268	164	93195	67.493	ug/l	96
65) Chlorobenzene	10.079	112	293284	70.490	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	103177	74.349	ug/l	99
67) Ethyl Benzene	10.189	91	535122	75.444	ug/l	99
68) m/p-Xylenes	10.299	106	389935	152.181	ug/l	99
69) o-Xylene	10.640	106	190833	73.890	ug/l	97
70) Styrene	10.652	104	328195	79.411	ug/l	100
71) Bromoform	10.799	173	84353	80.927	ug/l #	98
73) Isopropylbenzene	10.957	105	514528	71.083	ug/l	100
74) N-amyl acetate	10.841	43	274553	76.073	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	170481	66.207	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	196706	71.091	ug/l	86
77) Bromobenzene	11.195	156	117151	69.959	ug/l	99
78) n-propylbenzene	11.305	91	608332	74.974	ug/l	100
79) 2-Chlorotoluene	11.360	91	372392	68.100	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	432096	71.903	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	54398	81.820	ug/l	93
82) 4-Chlorotoluene	11.451	91	432053	71.890	ug/l	100
83) tert-Butylbenzene	11.713	119	432011	73.197	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	442728	73.758	ug/l	100
85) sec-Butylbenzene	11.890	105	543597	74.188	ug/l	100
86) p-Isopropyltoluene	12.006	119	457898	77.393	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	219753	70.449	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	217613	71.190	ug/l	99
89) n-Butylbenzene	12.329	91	416602	80.996	ug/l	100
90) Hexachloroethane	12.536	117	82528	74.978	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	216834	71.367	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	43673	74.512	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	137430	82.605	ug/l	98
94) Hexachlorobutadiene	13.725	225	55494	71.314	ug/l	98
95) Naphthalene	13.774	128	489778	80.842	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	139525	79.981	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046043.D
Acq On : 05 May 2025 12:21
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Quant Time: May 06 06:10:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046043.D
Acq On : 05 May 2025 12:21
Operator : JC/MD
Sample : VSTDIC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100

Quant Time: May 06 06:10:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:168 Resp: 87028
Ion Ratio Lower Upper
168 100
99 72.1 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 77.828 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 85 Resp: 149508
Ion Ratio Lower Upper
85 100
87 32.0 15.7 47.1









#3
Chloromethane
Concen: 72.392 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100


Tgt Ion: 50 Resp: 136937
Ion Ratio Lower Upper
50 100
52 32.9 25.4 38.2



#4
Vinyl Chloride
Concen: 76.398 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 62 Resp: 126609
Ion Ratio Lower Upper
62 100
64 30.6 25.2 37.8









#5
Bromomethane
Concen: 70.507 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 94 Resp: 59246
Ion Ratio Lower Upper
94 100
96 91.7 72.8 109.2



#6
Chloroethane
Concen: 67.148 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 64 Resp: 57332
Ion Ratio Lower Upper
64 100
66 30.8 24.3 36.5









#7
Trichlorofluoromethane
Concen: 67.754 ug/l
RT: 1.867 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:101 Resp: 171042
Ion Ratio Lower Upper
101 100
103 64.8 51.0 76.4



#8
Diethyl Ether
Concen: 70.807 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 74 Resp: 58762
Ion Ratio Lower Upper
74 100
45 110.1 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 72.604 ug/l
RT: 2.319 min Scan# 20
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:101 Resp: 109520
Ion Ratio Lower Upper
101 100
85 48.1 38.6 58.0
151 69.9 55.2 82.8



#10
Methyl Iodide
Concen: 77.943 ug/l
RT: 2.440 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:142 Resp: 138045
Ion Ratio Lower Upper
142 100
127 52.0 41.7 62.5
141 14.5 11.5 17.3









#11
Tert butyl alcohol
Concen: 405.394 ug/l
RT: 2.995 min Scan# 310
Delta R.T. 0.018 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 59 Resp: 125094
Ion Ratio Lower Upper
59 100
57 10.9 8.6 12.8



#12
1,1-Dichloroethene
Concen: 71.884 ug/l
RT: 2.306 min Scan# 200
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 96 Resp: 105566
Ion Ratio Lower Upper
96 100
61 189.9 146.2 219.2
98 63.3 51.0 76.6









#13
Acrolein
Concen: 375.132 ug/l
RT: 2.239 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 56 Resp: 134456
Ion Ratio Lower Upper
56 100
55 71.3 56.2 84.4



#14
Allyl chloride
Concen: 74.419 ug/l
RT: 2.660 min Scan# 258
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 41 Resp: 206519
Ion Ratio Lower Upper
41 100
39 73.8 60.6 90.8
76 31.4 24.9 37.3









#15
Acrylonitrile
Concen: 358.723 ug/l
RT: 3.068 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100


Tgt Ion: 53 Resp: 331142
Ion Ratio Lower Upper
53 100
52 84.7 65.3 97.9
51 37.9 29.8 44.8



#16
Acetone
Concen: 354.936 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 43 Resp: 314383
Ion Ratio Lower Upper
43 100
58 27.6 21.2 31.8









#17
Carbon Disulfide
Concen: 77.867 ug/l
RT: 2.501 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 76 Resp: 264956
Ion Ratio Lower Upper
76 100
78 8.7 7.0 10.4



#18
Methyl Acetate
Concen: 69.531 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 43 Resp: 147014
Ion Ratio Lower Upper
43 100
74 22.6 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 73.818 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 73 Resp: 378113
Ion Ratio Lower Upper
73 100
57 22.1 17.7 26.5



#20
Methylene Chloride
Concen: 67.030 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 84 Resp: 120270
Ion Ratio Lower Upper
84 100
49 136.4 113.9 170.9
51 41.1 33.5 50.3
86 64.0 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 71.109 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 96 Resp: 106606
Ion Ratio Lower Upper
96 100
61 149.7 119.5 179.3
98 64.4 50.0 75.0



#22
Diisopropyl ether
Concen: 77.048 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 45 Resp: 399494
Ion Ratio Lower Upper
45 100
43 71.2 66.6 100.0
87 25.7 19.8 29.6
59 11.0 8.6 12.8









#23
 Vinyl Acetate
 Concen: 395.363 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046043.D
 Acq: 05 May 2025 12:21




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100


Tgt Ion: 43 Resp: 1811550
 Ion Ratio Lower Upper
 43 100
 86 9.5 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 72.114 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046043.D
 Acq: 05 May 2025 12:21

Tgt Ion: 63 Resp: 219759
 Ion Ratio Lower Upper
 63 100
 98 6.8 3.6 10.8
 100 4.5 2.1 6.3





#25

2-Butanone

Concen: 379.704 ug/l

RT: 4.556 min Scan# 5

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion: 43 Resp: 485448

Ion Ratio Lower Upper

43 100

72 23.1 18.4 27.6



#26

2,2-Dichloropropane

Concen: 75.621 ug/l

RT: 4.464 min Scan# 554

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 77 Resp: 174579

Ion Ratio Lower Upper

77 100

97 21.1 10.5 31.5









#27
 cis-1,2-Dichloroethene
 Concen: 71.056 ug/l
 RT: 4.483 min Scan# 557
 Delta R.T. -0.000 min
 Lab File: VX046043.D
 Acq: 05 May 2025 12:21

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100

Tgt Ion: 96 Resp: 128454




Ion	Ratio	Lower	Upper
96	100		
61	160.1	0.0	322.8
98	64.9	0.0	129.0




#28
 Bromochloromethane
 Concen: 63.932 ug/l
 RT: 4.891 min Scan# 624
 Delta R.T. -0.006 min
 Lab File: VX046043.D
 Acq: 05 May 2025 12:21

Tgt Ion: 49 Resp: 103602

Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	4.0
130	70.1	56.2	84.2





#29

Tetrahydrofuran

Concen: 368.234 ug/l

RT: 5.001 min Scan# 642

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion: 42 Resp: 305757

Ion Ratio Lower Upper

42 100

72 42.1 33.7 50.5

71 38.7 31.4 47.2



#30

Chloroform

Concen: 70.180 ug/l

RT: 5.086 min Scan# 656

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 83 Resp: 222253

Ion Ratio Lower Upper

83 100

85 65.1 49.3 73.9









#31
Cyclohexane
Concen: 76.428 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100


Tgt Ion: 56 Resp: 196302
Ion Ratio Lower Upper
56 100
69 31.2 24.4 36.6
84 82.3 66.9 100.3



#32
1,1,1-Trichloroethane
Concen: 73.635 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 97 Resp: 201112
Ion Ratio Lower Upper
97 100
99 63.8 51.8 77.6
61 48.7 40.1 60.1









#33
 1,2-Dichloroethane-d4
 Concen: 62.526 ug/l
 RT: 5.952 min Scan# 798
 Delta R.T. -0.000 min
 Lab File: VX046043.D
 Acq: 05 May 2025 12:21




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100


Tgt Ion: 65 Resp: 161877
 Ion Ratio Lower Upper
 65 100
 67 49.5 0.0 99.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.757 min Scan# 930
 Delta R.T. -0.000 min
 Lab File: VX046043.D
 Acq: 05 May 2025 12:21

Tgt Ion: 114 Resp: 152958
 Ion Ratio Lower Upper
 114 100
 63 24.3 0.0 49.2
 88 17.1 0.0 33.6









#35
Dibromofluoromethane
Concen: 65.006 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:113 Resp: 111455
Ion Ratio Lower Upper
113 100
111 101.7 83.1 124.7
192 16.6 13.3 19.9



#36
1,1-Dichloropropene
Concen: 74.285 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 75 Resp: 147873
Ion Ratio Lower Upper
75 100
110 33.5 16.3 48.9
77 31.1 24.3 36.5









#37
Ethyl Acetate
Concen: 74.565 ug/l
RT: 4.714 min Scan# 595
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 43 Resp: 186215
Ion Ratio Lower Upper
43 100
61 12.4 10.3 15.5
70 9.9 7.9 11.9



#38
Carbon Tetrachloride
Concen: 73.875 ug/l
RT: 5.672 min Scan# 752
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 117 Resp: 168946
Ion Ratio Lower Upper
117 100
119 95.0 75.2 112.8
121 29.6 24.2 36.4









#39
Methylcyclohexane
Concen: 76.952 ug/l
RT: 7.372 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100


Tgt Ion: 83 Resp: 191941
Ion Ratio Lower Upper
83 100
55 84.2 64.7 97.1
98 49.0 37.4 56.2



#40
Benzene
Concen: 71.457 ug/l
RT: 6.031 min Scan# 811
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 78 Resp: 440885
Ion Ratio Lower Upper
78 100
77 24.6 19.0 28.4









#41
Methacrylonitrile
Concen: 75.574 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100


Tgt Ion: 41 Resp: 104863
Ion Ratio Lower Upper
41 100
39 57.0 47.2 70.8
67 64.6 50.7 76.1
52 29.7 24.6 37.0



#42
1,2-Dichloroethane
Concen: 73.318 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 62 Resp: 187036
Ion Ratio Lower Upper
62 100
98 7.4 0.0 15.2









#43
Isopropyl Acetate
Concen: 78.714 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 43 Resp: 300491
Ion Ratio Lower Upper
43 100
61 18.0 14.3 21.5
87 11.6 9.5 14.3



#44
Trichloroethene
Concen: 72.420 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:130 Resp: 105628
Ion Ratio Lower Upper
130 100
95 103.8 0.0 204.2





#45

1,2-Dichloropropane

Concen: 73.420 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC100

Tgt Ion: 63 Resp: 112554

Ion Ratio Lower Upper

63 100

65 30.0 24.1 36.1



#46

Dibromomethane

Concen: 70.938 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 93 Resp: 85728


Ion Ratio Lower Upper

93 100

95 82.4 65.6 98.4

174 85.0 68.2 102.2









#47
Bromodichloromethane
Concen: 74.845 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 83 Resp: 175301
Ion Ratio Lower Upper
83 100
85 63.6 50.5 75.7
127 7.8 6.5 9.7



#48
Methyl methacrylate
Concen: 78.101 ug/l
RT: 7.689 min Scan# 1083
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 41 Resp: 153093
Ion Ratio Lower Upper
41 100
69 74.1 58.5 87.7
39 65.5 51.7 77.5





#49

1,4-Dioxane

Concen: 1470.687 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100





Tgt Ion: 88 Resp: 57313

Ion Ratio Lower Upper

88 100

43 40.1 33.4 50.2

58 73.3 58.6 88.0



#50

Toluene-d8

Concen: 67.031 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 98 Resp: 387230

Ion Ratio Lower Upper

98 100

100 65.2 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 384.740 ug/l
RT: 8.573 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 43 Resp: 963319
Ion Ratio Lower Upper
43 100
58 36.1 28.9 43.3



#52
Toluene
Concen: 74.010 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 92 Resp: 270763
Ion Ratio Lower Upper
92 100
91 171.2 136.6 205.0









#53
 t-1,3-Dichloropropene
 Concen: 86.372 ug/l
 RT: 8.976 min Scan# 1194
 Delta R.T. -0.000 min
 Lab File: VX046043.D
 Acq: 05 May 2025 12:21




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100


Tgt Ion: 75 Resp: 169863
 Ion Ratio Lower Upper
 75 100
 77 31.5 25.0 37.6



#54
 cis-1,3-Dichloropropene
 Concen: 80.264 ug/l
 RT: 8.366 min Scan# 1194
 Delta R.T. 0.006 min
 Lab File: VX046043.D
 Acq: 05 May 2025 12:21

Tgt Ion: 75 Resp: 184134
 Ion Ratio Lower Upper
 75 100
 77 31.9 25.4 38.0
 39 59.6 52.2 78.4









#55
1,1,2-Trichloroethane
Concen: 72.543 ug/l
RT: 9.147 min Scan# 1322
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100

Tgt Ion: 97 Resp: 107264




Ion	Ratio	Lower	Upper
97	100		
83	86.8	69.0	103.6
85	56.4	44.3	66.5
99	61.8	47.8	71.6




#56
Ethyl methacrylate
Concen: 82.359 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 69 Resp: 188717

Ion	Ratio	Lower	Upper
69	100		
41	75.7	60.8	91.2
39	47.7	39.0	58.6





#57

1,3-Dichloropropane

Concen: 71.590 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC100

Tgt Ion: 76 Resp: 187649

Ion Ratio Lower Upper

76 100

78 32.1 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 449.791 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 63 Resp: 463169

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3





#59

2-Hexanone

Concen: 383.945 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion: 43 Resp: 729602

Ion Ratio Lower Upper

43 100

58 49.9 24.9 74.6



#60

Dibromochloromethane

Concen: 78.918 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 129 Resp: 127012

Ion Ratio Lower Upper

129 100

127 77.0 39.3 117.8









#61
1,2-Dibromoethane
Concen: 74.200 ug/l
RT: 9.610 min Scan# 1397
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:107 Resp: 112669
Ion Ratio Lower Upper
107 100
109 93.9 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 72.818 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 95 Resp: 153085
Ion Ratio Lower Upper
95 100
174 67.1 0.0 135.8
176 65.3 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:117 Resp: 135214
Ion Ratio Lower Upper
117 100
82 61.7 50.6 76.0
119 30.6 25.8 38.6



#64
Tetrachloroethene
Concen: 67.493 ug/l
RT: 9.268 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:164 Resp: 93195
Ion Ratio Lower Upper
164 100
166 123.4 105.0 157.6
129 101.4 83.5 125.3
131 98.0 76.5 114.7









#65
Chlorobenzene
Concen: 70.490 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:112 Resp: 293284
Ion Ratio Lower Upper
112 100
114 31.9 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 74.349 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:131 Resp: 103177
Ion Ratio Lower Upper
131 100
133 95.3 47.3 141.9
119 64.5 31.6 95.0









#67
Ethyl Benzene
Concen: 75.444 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 91 Resp: 535122
Ion Ratio Lower Upper
91 100
106 29.6 23.4 35.2



#68
m/p-Xylenes
Concen: 152.181 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:106 Resp: 389935
Ion Ratio Lower Upper
106 100
91 215.8 171.2 256.8









#69
o-Xylene
Concen: 73.890 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:106 Resp: 190833
Ion Ratio Lower Upper
106 100
91 230.3 112.7 338.1



#70
Styrene
Concen: 79.411 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:104 Resp: 328195
Ion Ratio Lower Upper
104 100
78 57.4 45.7 68.5
103 54.8 43.7 65.5





#71

Bromoform

Concen: 80.927 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion:173 Resp: 84353

Ion Ratio Lower Upper

173 100

175 48.3 23.4 70.0

254 0.0 0.0 0.0#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:152 Resp: 66369


Ion Ratio Lower Upper

152 100

115 120.4 46.9 140.7

150 190.4 0.0 351.0









#73
Isopropylbenzene
Concen: 71.083 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:105 Resp: 514528
Ion Ratio Lower Upper
105 100
120 25.4 12.8 38.4



#74
N-ethyl acetate
Concen: 76.073 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 43 Resp: 274553
Ion Ratio Lower Upper
43 100
70 38.2 30.9 46.3
55 22.5 18.7 28.1
61 22.2 17.1 25.7









#75
 1,1,2,2-Tetrachloroethane
 Concen: 66.207 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. 0.006 min
 Lab File: VX046043.D
 Acq: 05 May 2025 12:21




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC100


Tgt Ion: 83 Resp: 170481
 Ion Ratio Lower Upper
 83 100
 131 9.7 5.0 14.9
 85 64.9 31.9 95.7



#76
 1,2,3-Trichloropropane
 Concen: 71.091 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. -0.000 min
 Lab File: VX046043.D
 Acq: 05 May 2025 12:21

Tgt Ion: 75 Resp: 196706
 Ion Ratio Lower Upper
 75 100
 77 32.2 20.5 61.5









#77
Bromobenzene
Concen: 69.959 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:156 Resp: 117151
Ion Ratio Lower Upper
156 100
77 187.4 92.8 278.3
158 96.1 48.8 146.3



#78
n-propylbenzene
Concen: 74.974 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 91 Resp: 608332
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4





#79

2-Chlorotoluene

Concen: 68.100 ug/l

RT: 11.360 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC100

Tgt Ion: 91 Resp: 372392

Ion Ratio Lower Upper

91 100

126 30.9 15.6 46.7



#80

1,3,5-Trimethylbenzene

Concen: 71.903 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion:105 Resp: 432096

Ion Ratio Lower Upper

105 100

120 46.3 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 81.820 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion: 75 Resp: 54398
Ion Ratio Lower Upper
75 100
53 115.8 100.5 150.7
89 47.0 37.0 55.6



#82
4-Chlorotoluene
Concen: 71.890 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 91 Resp: 432053
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83

tert-Butylbenzene

Concen: 73.197 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion:119 Resp: 432011

Ion Ratio Lower Upper

119 100

91 64.4 32.9 98.7

134 22.8 11.4 34.1



#84

1,2,4-Trimethylbenzene

Concen: 73.758 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion:105 Resp: 442728

Ion Ratio Lower Upper

105 100

120 42.4 21.2 63.6





#85

sec-Butylbenzene

Concen: 74.188 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC100

Tgt Ion:105 Resp: 543597

Ion Ratio Lower Upper

105 100

134 19.4 9.7 29.1



#86

p-Isopropyltoluene

Concen: 77.393 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:119 Resp: 457898


Ion Ratio Lower Upper

119 100

134 24.6 12.5 37.5

91 26.8 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 70.449 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100





Tgt Ion:146 Resp: 219753

Ion Ratio Lower Upper

146 100

111 44.1 22.1 66.3

148 63.5 32.1 96.5



#88

1,4-Dichlorobenzene

Concen: 71.190 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:146 Resp: 217613


Ion Ratio Lower Upper

146 100

111 43.4 21.3 63.9

148 63.2 31.9 95.5





#89

n-Butylbenzene

Concen: 80.996 ug/l

RT: 12.329 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100





Tgt Ion: 91 Resp: 416602

Ion Ratio Lower Upper

91 100

92 54.0 26.9 80.7

134 24.0 11.8 35.3



#90

Hexachloroethane

Concen: 74.978 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 117 Resp: 82528

Ion Ratio Lower Upper

117 100

201 63.3 31.6 94.7









#91
1,2-Dichlorobenzene
Concen: 71.367 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100


Tgt Ion:146 Resp: 216834
Ion Ratio Lower Upper
146 100
111 46.2 22.4 67.3
148 64.0 31.6 94.8



#92
1,2-Dibromo-3-Chloropropane
Concen: 74.512 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 75 Resp: 43673
Ion Ratio Lower Upper
75 100
155 69.4 34.9 104.8
157 90.1 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 82.605 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100





Tgt Ion:180 Resp: 137430

Ion Ratio Lower Upper

180 100

182 93.7 47.9 143.8

145 34.6 17.1 51.2



#94

Hexachlorobutadiene

Concen: 71.314 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:225 Resp: 55494


Ion Ratio Lower Upper

225 100

223 62.3 30.8 92.4

227 64.9 30.9 92.7





#95

Naphthalene

Concen: 80.842 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100





Tgt Ion:128 Resp: 489778

Ion Ratio Lower Upper

128 100

127 12.9 10.4 15.6

129 10.9 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 79.981 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:180 Resp: 139525

Ion Ratio Lower Upper

180 100

182 93.7 47.8 143.3

145 35.3 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX050525\
 Data File : VX046044.D
 Acq On : 05 May 2025 12:45
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Quant Time: May 06 06:11:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	82963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	144975	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	128557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60345	50.000	ug/l	# 0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	231952	93.983	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 187.960%#	
35) Dibromofluoromethane	5.385	113	160168	98.562	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 197.120%#	
50) Toluene-d8	8.647	98	554390	101.252	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 202.500%#	
62) 4-Bromofluorobenzene	11.079	95	217536	109.174	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 218.340%#	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	217793	118.929	ug/l	98
3) Chloromethane	1.307	50	196843	109.161	ug/l	99
4) Vinyl Chloride	1.374	62	187924	118.953	ug/l	100
5) Bromomethane	1.593	94	83008	103.625	ug/l	95
6) Chloroethane	1.660	64	78806	96.822	ug/l	98
7) Trichlorofluoromethane	1.868	101	245108	101.851	ug/l	97
8) Diethyl Ether	2.136	74	88441	111.791	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.313	101	161404	112.243	ug/l	99
10) Methyl Iodide	2.441	142	189843	112.441	ug/l	100
11) Tert butyl alcohol	3.002	59	181238	616.119	ug/l	100
12) 1,1-Dichloroethene	2.307	96	155451	111.039	ug/l	99
13) Acrolein	2.239	56	203257	594.872	ug/l	98
14) Allyl chloride	2.654	41	297776	112.561	ug/l	98
15) Acrylonitrile	3.069	53	484854	550.973	ug/l	98
16) Acetone	2.392	43	460820	545.754	ug/l	98
17) Carbon Disulfide	2.502	76	397353	122.499	ug/l	100
18) Methyl Acetate	2.709	43	217885	108.099	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	557210	114.112	ug/l	100
20) Methylene Chloride	2.782	84	172035	100.579	ug/l	96
21) trans-1,2-Dichloroethene	3.081	96	154694	108.241	ug/l	96
22) Diisopropyl ether	3.764	45	577610	116.858	ug/l	90
23) Vinyl Acetate	3.721	43	2655751	608.005	ug/l	100
24) 1,1-Dichloroethane	3.605	63	320083	110.182	ug/l	99
25) 2-Butanone	4.562	43	708463	581.292	ug/l	98
26) 2,2-Dichloropropane	4.471	77	258698	117.548	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	187835	108.994	ug/l	99
28) Bromochloromethane	4.898	49	146923	95.108	ug/l	100
29) Tetrahydrofuran	5.007	42	450456	569.081	ug/l	99
30) Chloroform	5.087	83	323659	107.208	ug/l	95
31) Cyclohexane	5.458	56	286312	116.934	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	295702	113.572	ug/l	98
36) 1,1-Dichloropropene	5.684	75	219731	116.462	ug/l	99
37) Ethyl Acetate	4.715	43	274312	115.889	ug/l	99
38) Carbon Tetrachloride	5.666	117	250916	115.760	ug/l	99
39) Methylcyclohexane	7.373	83	286311	121.107	ug/l	97
40) Benzene	6.031	78	642178	109.813	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046044.D
 Acq On : 05 May 2025 12:45
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Quant Time: May 06 06:11:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	151424	115.139	ug/l	98
42) 1,2-Dichloroethane	6.080	62	271992	112.492	ug/l	99
43) Isopropyl Acetate	6.342	43	447871	123.780	ug/l	100
44) Trichloroethene	7.117	130	157641	114.032	ug/l	96
45) 1,2-Dichloropropane	7.428	63	164505	113.217	ug/l	99
46) Dibromomethane	7.574	93	125846	109.869	ug/l	100
47) Bromodichloromethane	7.818	83	258326	116.365	ug/l	98
48) Methyl methacrylate	7.690	41	231040	124.356	ug/l	99
49) 1,4-Dioxane	7.665	88	82800	2241.695	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	1371275	577.831	ug/l	100
52) Toluene	8.714	92	393323	113.430	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	257233	138.001	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	271054	124.658	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	154932	110.551	ug/l	97
56) Ethyl methacrylate	9.116	69	277921	127.968	ug/l	99
57) 1,3-Dichloropropane	9.305	76	272814	109.813	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	680608	697.344	ug/l	100
59) 2-Hexanone	9.433	43	1028597	571.094	ug/l	99
60) Dibromochloromethane	9.519	129	187497	122.915	ug/l	98
61) 1,2-Dibromoethane	9.604	107	165634	115.087	ug/l	97
64) Tetrachloroethene	9.269	164	132488	100.918	ug/l	96
65) Chlorobenzene	10.080	112	429656	108.615	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	152173	115.333	ug/l	99
67) Ethyl Benzene	10.189	91	785121	116.421	ug/l	99
68) m/p-Xylenes	10.299	106	570913	234.349	ug/l	100
69) o-Xylene	10.640	106	279906	113.990	ug/l	99
70) Styrene	10.653	104	474254	120.694	ug/l	99
71) Bromoform	10.799	173	125986	127.128	ug/l #	99
73) Isopropylbenzene	10.957	105	752337	114.312	ug/l	99
74) N-amyl acetate	10.842	43	396877	120.944	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	243532	104.018	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	284571	113.113	ug/l	85
77) Bromobenzene	11.195	156	167934	110.297	ug/l	99
78) n-propylbenzene	11.305	91	881218	119.448	ug/l	100
79) 2-Chlorotoluene	11.360	91	534560	107.514	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	616477	112.825	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	81372	134.609	ug/l	93
82) 4-Chlorotoluene	11.451	91	611736	111.948	ug/l	100
83) tert-Butylbenzene	11.713	119	617473	115.065	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	623456	114.236	ug/l	99
85) sec-Butylbenzene	11.890	105	786293	118.023	ug/l	99
86) p-Isopropyltoluene	12.006	119	651531	121.113	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	313100	110.394	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	311752	112.167	ug/l	100
89) n-Butylbenzene	12.329	91	605783	129.533	ug/l	100
90) Hexachloroethane	12.536	117	123188	123.091	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	308039	111.506	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	64438	120.915	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	203255	134.366	ug/l	98
94) Hexachlorobutadiene	13.719	225	80558	113.857	ug/l	99
95) Naphthalene	13.774	128	721236	130.929	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	200379	126.331	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046044.D
Acq On : 05 May 2025 12:45
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Quant Time: May 06 06:11:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed





Page: 4



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:168 Resp: 82963
Ion Ratio Lower Upper
168 100
99 67.6 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 118.929 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 85 Resp: 217793
Ion Ratio Lower Upper
85 100
87 32.5 15.7 47.1









#3
Chloromethane
Concen: 109.161 ug/l
RT: 1.307 min Scan# 30
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150


Tgt Ion: 50 Resp: 196843
Ion Ratio Lower Upper
50 100
52 32.5 25.4 38.2



#4
Vinyl Chloride
Concen: 118.953 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 62 Resp: 187924
Ion Ratio Lower Upper
62 100
64 31.7 25.2 37.8









#5
Bromomethane
Concen: 103.625 ug/l
RT: 1.593 min Scan# 81
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 94 Resp: 83008
Ion Ratio Lower Upper
94 100
96 95.8 72.8 109.2



#6
Chloroethane
Concen: 96.822 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 64 Resp: 78806
Ion Ratio Lower Upper
64 100
66 31.3 24.3 36.5









#7
Trichlorofluoromethane
Concen: 101.851 ug/l
RT: 1.868 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:101 Resp: 245108
Ion Ratio Lower Upper
101 100
103 66.4 51.0 76.4



#8
Diethyl Ether
Concen: 111.791 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 74 Resp: 88441
Ion Ratio Lower Upper
74 100
45 104.7 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 112.243 ug/l
RT: 2.313 min Scan# 202
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:101 Resp: 161404
Ion Ratio Lower Upper
101 100
85 46.8 38.6 58.0
151 69.8 55.2 82.8



#10
Methyl Iodide
Concen: 112.441 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:142 Resp: 189843
Ion Ratio Lower Upper
142 100
127 52.4 41.7 62.5
141 14.4 11.5 17.3









#11
Tert butyl alcohol
Concen: 616.119 ug/l
RT: 3.002 min Scan# 31
Delta R.T. 0.024 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 59 Resp: 181238
Ion Ratio Lower Upper
59 100
57 10.8 8.6 12.8



#12
1,1-Dichloroethene
Concen: 111.039 ug/l
RT: 2.307 min Scan# 200
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 96 Resp: 155451
Ion Ratio Lower Upper
96 100
61 184.5 146.2 219.2
98 64.3 51.0 76.6









#13
Acrolein
Concen: 594.872 ug/l
RT: 2.239 min Scan# 188
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 56 Resp: 203257
Ion Ratio Lower Upper
56 100
55 72.1 56.2 84.4



#14
Allyl chloride
Concen: 112.561 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 41 Resp: 297776
Ion Ratio Lower Upper
41 100
39 73.7 60.6 90.8
76 31.8 24.9 37.3









#15
Acrylonitrile
Concen: 550.973 ug/l
RT: 3.069 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150


Tgt Ion: 53 Resp: 484854
Ion Ratio Lower Upper
53 100
52 84.2 65.3 97.9
51 37.5 29.8 44.8



#16
Acetone
Concen: 545.754 ug/l
RT: 2.392 min Scan# 214
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 43 Resp: 460820
Ion Ratio Lower Upper
43 100
58 27.3 21.2 31.8









#17
Carbon Disulfide
Concen: 122.499 ug/l
RT: 2.502 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 76 Resp: 397353
Ion Ratio Lower Upper
76 100
78 8.8 7.0 10.4



#18
Methyl Acetate
Concen: 108.099 ug/l
RT: 2.709 min Scan# 266
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 43 Resp: 217885
Ion Ratio Lower Upper
43 100
74 21.9 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 114.112 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 73 Resp: 557210
Ion Ratio Lower Upper
73 100
57 22.0 17.7 26.5



#20
Methylene Chloride
Concen: 100.579 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 84 Resp: 172035
Ion Ratio Lower Upper
84 100
49 136.0 113.9 170.9
51 41.5 33.5 50.3
86 63.7 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 108.241 ug/l
RT: 3.081 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 96 Resp: 154694
Ion Ratio Lower Upper
96 100
61 154.6 119.5 179.3
98 64.3 50.0 75.0



#22
Diisopropyl ether
Concen: 116.858 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 45 Resp: 577610
Ion Ratio Lower Upper
45 100
43 70.3 66.6 100.0
87 24.7 19.8 29.6
59 10.6 8.6 12.8









#23
 Vinyl Acetate
 Concen: 608.005 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150


Tgt Ion: 43 Resp: 2655751
 Ion Ratio Lower Upper
 43 100
 86 9.6 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 110.182 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45

Tgt Ion: 63 Resp: 320083
 Ion Ratio Lower Upper
 63 100
 98 7.0 3.6 10.8
 100 4.7 2.1 6.3





#25

2-Butanone

Concen: 581.292 ug/l

RT: 4.562 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 43 Resp: 708463

Ion Ratio Lower Upper

43 100

72 24.0 18.4 27.6



#26

2,2-Dichloropropane

Concen: 117.548 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 77 Resp: 258698

Ion Ratio Lower Upper

77 100

97 21.1 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 108.994 ug/l

RT: 4.483 min Scan# 51

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 96 Resp: 187835

Ion Ratio Lower Upper

96 100

61 159.8 0.0 322.8

98 64.8 0.0 129.0



#28

Bromochloromethane

Concen: 95.108 ug/l

RT: 4.898 min Scan# 625

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 49 Resp: 146923


Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 70.5 56.2 84.2





#29

Tetrahydrofuran

Concen: 569.081 ug/l

RT: 5.007 min Scan# 64

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 42 Resp: 450456

Ion Ratio Lower Upper

42 100

72 41.6 33.7 50.5

71 38.7 31.4 47.2



#30

Chloroform

Concen: 107.208 ug/l

RT: 5.087 min Scan# 656

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 83 Resp: 323659

Ion Ratio Lower Upper

83 100

85 65.1 49.3 73.9









#31
Cyclohexane
Concen: 116.934 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 56 Resp: 286312
Ion Ratio Lower Upper
56 100
69 31.2 24.4 36.6
84 82.8 66.9 100.3



#32
1,1,1-Trichloroethane
Concen: 113.572 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 97 Resp: 295702
Ion Ratio Lower Upper
97 100
99 64.3 51.8 77.6
61 48.0 40.1 60.1









#33
 1,2-Dichloroethane-d4
 Concen: 93.983 ug/l
 RT: 5.952 min Scan# 798
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150


Tgt Ion: 65 Resp: 231952
 Ion Ratio Lower Upper
 65 100
 67 49.4 0.0 99.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.757 min Scan# 930
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45

Tgt Ion: 114 Resp: 144975
 Ion Ratio Lower Upper
 114 100
 63 23.9 0.0 49.2
 88 18.3 0.0 33.6





#35

Dibromofluoromethane

Concen: 98.562 ug/l

RT: 5.385 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:113 Resp: 160168

Ion Ratio Lower Upper

113 100

111 103.0 83.1 124.7

192 16.9 13.3 19.9



#36

1,1-Dichloropropene

Concen: 116.462 ug/l

RT: 5.684 min Scan# 754

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 75 Resp: 219731


Ion Ratio Lower Upper

75 100

110 32.9 16.3 48.9

77 30.8 24.3 36.5









#37
Ethyl Acetate
Concen: 115.889 ug/l
RT: 4.715 min Scan# 595
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 274312
Ion Ratio Lower Upper
43 100
61 12.5 10.3 15.5
70 10.2 7.9 11.9



#38
Carbon Tetrachloride
Concen: 115.760 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 117 Resp: 250916
Ion Ratio Lower Upper
117 100
119 94.3 75.2 112.8
121 29.7 24.2 36.4





#39

Methylcyclohexane

Concen: 121.107 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 83 Resp: 286311

Ion Ratio Lower Upper

83 100

55 83.8 64.7 97.1

98 47.8 37.4 56.2



#40

Benzene

Concen: 109.813 ug/l

RT: 6.031 min Scan# 811

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 78 Resp: 642178

Ion Ratio Lower Upper

78 100

77 23.4 19.0 28.4









#41
Methacrylonitrile
Concen: 115.139 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150


Tgt Ion: 41 Resp: 151424
Ion Ratio Lower Upper
41 100
39 60.0 47.2 70.8
67 65.9 50.7 76.1
52 30.9 24.6 37.0



#42
1,2-Dichloroethane
Concen: 112.492 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 62 Resp: 271992
Ion Ratio Lower Upper
62 100
98 7.4 0.0 15.2









#43
Isopropyl Acetate
Concen: 123.780 ug/l
RT: 6.342 min Scan# 861
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 447871
Ion Ratio Lower Upper
43 100
61 18.1 14.3 21.5
87 11.7 9.5 14.3



#44
Trichloroethene
Concen: 114.032 ug/l
RT: 7.117 min Scan# 989
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:130 Resp: 157641
Ion Ratio Lower Upper
130 100
95 105.9 0.0 204.2





#45

1,2-Dichloropropane

Concen: 113.217 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 63 Resp: 164505

Ion Ratio Lower Upper

63 100

65 30.8 24.1 36.1



#46

Dibromomethane

Concen: 109.869 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 93 Resp: 125846


Ion Ratio Lower Upper

93 100

95 82.4 65.6 98.4

174 85.0 68.2 102.2









#47
Bromodichloromethane
Concen: 116.365 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 83 Resp: 258326
Ion Ratio Lower Upper
83 100
85 64.8 50.5 75.7
127 8.0 6.5 9.7



#48
Methyl methacrylate
Concen: 124.356 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 41 Resp: 231040
Ion Ratio Lower Upper
41 100
69 73.0 58.5 87.7
39 63.8 51.7 77.5





#49

1,4-Dioxane

Concen: 2241.695 ug/l

RT: 7.665 min Scan# 110

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion: 88 Resp: 82800

Ion Ratio Lower Upper

88 100

43 40.0 33.4 50.2

58 74.8 58.6 88.0



#50

Toluene-d8

Concen: 101.252 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 98 Resp: 554390

Ion Ratio Lower Upper

98 100

100 65.4 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 577.831 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 1371275
Ion Ratio Lower Upper
43 100
58 36.2 28.9 43.3



#52
Toluene
Concen: 113.430 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 92 Resp: 393323
Ion Ratio Lower Upper
92 100
91 171.0 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 138.001 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 75 Resp: 257233

Ion Ratio Lower Upper

75 100

77 31.7 25.0 37.6



#54

cis-1,3-Dichloropropene

Concen: 124.658 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 75 Resp: 271054


Ion Ratio Lower Upper

75 100

77 31.4 25.4 38.0

39 60.0 52.2 78.4









#55
 1,1,2-Trichloroethane
 Concen: 110.551 ug/l
 RT: 9.147 min Scan# 1322
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150

Tgt Ion: 97 Resp: 154932




Ion	Ratio	Lower	Upper
97	100		
83	88.5	69.0	103.6
85	56.0	44.3	66.5
99	64.5	47.8	71.6




#56
 Ethyl methacrylate
 Concen: 127.968 ug/l
 RT: 9.116 min Scan# 1317
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45

Tgt Ion: 69 Resp: 277921

Ion	Ratio	Lower	Upper
69	100		
41	75.4	60.8	91.2
39	47.2	39.0	58.6





#57

1,3-Dichloropropane

Concen: 109.813 ug/l

RT: 9.305 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 76 Resp: 272814

Ion Ratio Lower Upper

76 100

78 32.3 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 697.344 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.006 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 63 Resp: 680608

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3









#59
2-Hexanone
Concen: 571.094 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 43 Resp: 1028597
Ion Ratio Lower Upper
43 100
58 50.2 24.9 74.6



#60
Dibromochloromethane
Concen: 122.915 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 129 Resp: 187497
Ion Ratio Lower Upper
129 100
127 77.1 39.3 117.8









#61
1,2-Dibromoethane
Concen: 115.087 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:107 Resp: 165634
Ion Ratio Lower Upper
107 100
109 94.6 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 109.174 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 95 Resp: 217536
Ion Ratio Lower Upper
95 100
174 67.5 0.0 135.8
176 64.4 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:117 Resp: 128557
Ion Ratio Lower Upper
117 100
82 61.4 50.6 76.0
119 31.9 25.8 38.6



#64
Tetrachloroethene
Concen: 100.918 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:164 Resp: 132488
Ion Ratio Lower Upper
164 100
166 123.3 105.0 157.6
129 101.6 83.5 125.3
131 97.3 76.5 114.7









#65
Chlorobenzene
Concen: 108.615 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:112 Resp: 429656
Ion Ratio Lower Upper
112 100
114 31.5 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 115.333 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:131 Resp: 152173
Ion Ratio Lower Upper
131 100
133 95.7 47.3 141.9
119 64.0 31.6 95.0





#67

Ethyl Benzene

Concen: 116.421 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion: 91 Resp: 785121

Ion Ratio Lower Upper

91 100

106 29.9 23.4 35.2



#68

m/p-Xylenes

Concen: 234.349 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 106 Resp: 570913

Ion Ratio Lower Upper

106 100

91 214.2 171.2 256.8









#69
o-Xylene
Concen: 113.990 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150


Tgt Ion:106 Resp: 279906
Ion Ratio Lower Upper
106 100
91 226.9 112.7 338.1



#70
Styrene
Concen: 120.694 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:104 Resp: 474254
Ion Ratio Lower Upper
104 100
78 57.2 45.7 68.5
103 55.5 43.7 65.5





#71

Bromoform

Concen: 127.128 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:173 Resp: 125986

Ion Ratio Lower Upper

173 100

175 47.3 23.4 70.0

254 0.1 0.0 0.0#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:152 Resp: 60345


Ion Ratio Lower Upper

152 100

115 150.7 46.9 140.7#

150 211.6 0.0 351.0





#73

Isopropylbenzene

Concen: 114.312 ug/l

RT: 10.957 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC150

Tgt Ion:105 Resp: 752337

Ion Ratio Lower Upper

105 100

120 25.3 12.8 38.4



#74

N-ethyl acetate

Concen: 120.944 ug/l

RT: 10.842 min Scan# 1600

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Tgt Ion: 43 Resp: 396877




Ion Ratio Lower Upper


43 100

70 38.5 30.9 46.3

55 22.5 18.7 28.1

61 22.5 17.1 25.7









#75
 1,1,2,2-Tetrachloroethane
 Concen: 104.018 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. 0.006 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150


Tgt Ion: 83 Resp: 243532
 Ion Ratio Lower Upper
 83 100
 131 9.7 5.0 14.9
 85 64.4 31.9 95.7



#76
 1,2,3-Trichloropropane
 Concen: 113.113 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45

Tgt Ion: 75 Resp: 284571
 Ion Ratio Lower Upper
 75 100
 77 31.8 20.5 61.5









#77
Bromobenzene
Concen: 110.297 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion:156 Resp: 167934
Ion Ratio Lower Upper
156 100
77 187.1 92.8 278.3
158 97.3 48.8 146.3



#78
n-propylbenzene
Concen: 119.448 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 91 Resp: 881218
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4









#79
2-Chlorotoluene
Concen: 107.514 ug/l
RT: 11.360 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150


Tgt Ion: 91 Resp: 534560
Ion Ratio Lower Upper
91 100
126 31.0 15.6 46.7



#80
1,3,5-Trimethylbenzene
Concen: 112.825 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 105 Resp: 616477
Ion Ratio Lower Upper
105 100
120 46.0 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 134.609 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150


Tgt Ion: 75 Resp: 81372
Ion Ratio Lower Upper
75 100
53 114.9 100.5 150.7
89 46.5 37.0 55.6



#82
4-Chlorotoluene
Concen: 111.948 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 91 Resp: 611736
Ion Ratio Lower Upper
91 100
126 26.7 13.3 39.8





#83

tert-Butylbenzene

Concen: 115.065 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:119 Resp: 617473

Ion Ratio Lower Upper

119 100

91 65.6 32.9 98.7

134 22.6 11.4 34.1



#84

1,2,4-Trimethylbenzene

Concen: 114.236 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:105 Resp: 623456

Ion Ratio Lower Upper

105 100

120 43.1 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 118.023 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150


Tgt Ion:105 Resp: 786293
 Ion Ratio Lower Upper
 105 100
 134 19.1 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 121.113 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45

Tgt Ion:119 Resp: 651531
 Ion Ratio Lower Upper
 119 100
 134 25.4 12.5 37.5
 91 27.0 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 110.394 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:146 Resp: 313100

Ion Ratio Lower Upper

146 100

111 44.5 22.1 66.3

148 64.0 32.1 96.5



#88

1,4-Dichlorobenzene

Concen: 112.167 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:146 Resp: 311752


Ion Ratio Lower Upper

146 100

111 43.0 21.3 63.9

148 63.9 31.9 95.5





#89

n-Butylbenzene

Concen: 129.533 ug/l

RT: 12.329 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150





Tgt Ion: 91 Resp: 605783

Ion Ratio Lower Upper

91 100

92 53.6 26.9 80.7

134 24.0 11.8 35.3



#90

Hexachloroethane

Concen: 123.091 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 117 Resp: 123188

Ion Ratio Lower Upper

117 100

201 63.2 31.6 94.7









#91
 1,2-Dichlorobenzene
 Concen: 111.506 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150


Tgt Ion:146 Resp: 308039
 Ion Ratio Lower Upper
 146 100
 111 46.3 22.4 67.3
 148 63.6 31.6 94.8



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 120.915 ug/l
 RT: 12.939 min Scan# 1944
 Delta R.T. 0.000 min
 Lab File: VX046044.D
 Acq: 05 May 2025 12:45

Tgt Ion: 75 Resp: 64438
 Ion Ratio Lower Upper
 75 100
 155 69.6 34.9 104.8
 157 88.4 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 134.366 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:180 Resp: 203255

Ion Ratio Lower Upper

180 100

182 94.1 47.9 143.8

145 34.5 17.1 51.2



#94

Hexachlorobutadiene

Concen: 113.857 ug/l

RT: 13.719 min Scan# 2072

Delta R.T. -0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:225 Resp: 80558


Ion Ratio Lower Upper

225 100

223 62.5 30.8 92.4

227 63.1 30.9 92.7





#95

Naphthalene

Concen: 130.929 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150





Tgt Ion:128 Resp: 721236

Ion Ratio Lower Upper

128 100

127 12.9 10.4 15.6

129 11.0 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 126.331 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:180 Resp: 200379

Ion Ratio Lower Upper

180 100

182 96.5 47.8 143.3

145 36.7 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046046.D
 Acq On : 05 May 2025 16:04
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Quant Time: May 06 06:12:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	96964	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	168484	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	147167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67939	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	9067	3.143	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	6.280%#
35) Dibromofluoromethane	5.373	113	5969	3.161	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	6.320%#
50) Toluene-d8	8.647	98	20566	3.232	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	6.460%#
62) 4-Bromofluorobenzene	11.079	95	7822	3.378	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	6.760%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	6195	2.894	ug/l	95
3) Chloromethane	1.307	50	6587	3.125	ug/l	99
4) Vinyl Chloride	1.374	62	5999	3.249	ug/l	96
5) Bromomethane	1.593	94	2962	3.164	ug/l	88
6) Chloroethane	1.666	64	3567	3.750	ug/l	94
7) Trichlorofluoromethane	1.874	101	9599	3.413	ug/l	88
8) Diethyl Ether	2.130	74	3020	3.266	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.325	101	5911	3.517	ug/l	99
10) Methyl Iodide	2.441	142	5900	2.990	ug/l	99
11) Tert butyl alcohol	2.965	59	5539	16.111	ug/l	98
12) 1,1-Dichloroethene	2.306	96	5496	3.359	ug/l	97
13) Acrolein	2.233	56	5673	14.206	ug/l	95
14) Allyl chloride	2.654	41	10262	3.319	ug/l	98
15) Acrylonitrile	3.062	53	16746	16.282	ug/l	97
16) Acetone	2.380	43	19773	20.036	ug/l	98
17) Carbon Disulfide	2.501	76	11068	2.919	ug/l	96
18) Methyl Acetate	2.703	43	7909	3.357	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	18497	3.241	ug/l	97
20) Methylene Chloride	2.782	84	6680	3.341	ug/l	97
21) trans-1,2-Dichloroethene	3.081	96	5402	3.234	ug/l	100
22) Diisopropyl ether	3.751	45	18657	3.230	ug/l #	65
23) Vinyl Acetate	3.715	43	82328	16.127	ug/l	99
24) 1,1-Dichloroethane	3.605	63	11193	3.297	ug/l	99
25) 2-Butanone	4.562	43	26125	18.340	ug/l	98
26) 2,2-Dichloropropane	4.465	77	8245	3.205	ug/l	96
27) cis-1,2-Dichloroethene	4.483	96	6222	3.089	ug/l	95
28) Bromochloromethane	4.891	49	5360	2.969	ug/l #	100
29) Tetrahydrofuran	5.001	42	15431	16.680	ug/l	98
30) Chloroform	5.080	83	11624	3.294	ug/l	89
31) Cyclohexane	5.458	56	10269	3.588	ug/l	96
32) 1,1,1-Trichloroethane	5.367	97	9827	3.229	ug/l	97
36) 1,1-Dichloropropene	5.678	75	7787	3.551	ug/l	98
37) Ethyl Acetate	4.721	43	9021	3.279	ug/l	98
38) Carbon Tetrachloride	5.659	117	8505	3.376	ug/l	97
39) Methylcyclohexane	7.373	83	9882	3.597	ug/l	93
40) Benzene	6.031	78	22528	3.315	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046046.D
 Acq On : 05 May 2025 16:04
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Quant Time: May 06 06:12:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	4851	3.174	ug/l #	65
42) 1,2-Dichloroethane	6.080	62	10011	3.563	ug/l	100
43) Isopropyl Acetate	6.348	43	13914	3.309	ug/l	96
44) Trichloroethene	7.123	130	5313	3.307	ug/l	94
45) 1,2-Dichloropropane	7.427	63	5451	3.228	ug/l #	89
46) Dibromomethane	7.580	93	4417	3.318	ug/l	98
47) Bromodichloromethane	7.818	83	8394	3.254	ug/l	98
48) Methyl methacrylate	7.696	41	7171	3.321	ug/l	97
49) 1,4-Dioxane	7.659	88	2907	67.721	ug/l	96
51) 4-Methyl-2-Pentanone	8.567	43	46794	16.967	ug/l	98
52) Toluene	8.714	92	14126	3.505	ug/l	97
53) t-1,3-Dichloropropene	8.976	75	6834	3.155	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	7908	3.129	ug/l	91
55) 1,1,2-Trichloroethane	9.153	97	5685	3.490	ug/l	98
56) Ethyl methacrylate	9.116	69	8555	3.390	ug/l	98
57) 1,3-Dichloropropane	9.305	76	10127	3.508	ug/l	94
58) 2-Chloroethyl Vinyl ether	8.238	63	20809	18.346	ug/l	99
59) 2-Hexanone	9.427	43	34853	16.651	ug/l	97
60) Dibromochloromethane	9.518	129	5500	3.102	ug/l	98
61) 1,2-Dibromoethane	9.604	107	5614	3.356	ug/l	99
64) Tetrachloroethene	9.269	164	4752	3.162	ug/l	96
65) Chlorobenzene	10.079	112	15391	3.399	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	5024	3.326	ug/l	99
67) Ethyl Benzene	10.195	91	26732	3.463	ug/l	99
68) m/p-Xylenes	10.299	106	19957	7.156	ug/l	98
69) o-Xylene	10.640	106	9398	3.343	ug/l	95
70) Styrene	10.652	104	14900	3.312	ug/l	99
71) Bromoform	10.799	173	3470	3.059	ug/l #	99
73) Isopropylbenzene	10.957	105	24201	3.266	ug/l	98
74) N-amyl acetate	10.841	43	11224	3.038	ug/l	96
75) 1,1,2,2-Tetrachloroethane	11.207	83	9170	3.479	ug/l	97
76) 1,2,3-Trichloropropane	11.238	75	8548	3.018	ug/l	95
77) Bromobenzene	11.195	156	5857	3.417	ug/l	96
78) n-propylbenzene	11.299	91	28436	3.424	ug/l	100
79) 2-Chlorotoluene	11.360	91	18672	3.336	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	20742	3.372	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	1828	2.686	ug/l	99
82) 4-Chlorotoluene	11.451	91	19966	3.245	ug/l	98
83) tert-Butylbenzene	11.713	119	21049	3.484	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	20613	3.355	ug/l	97
85) sec-Butylbenzene	11.890	105	25590	3.412	ug/l	99
86) p-Isopropyltoluene	12.006	119	20955	3.460	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	10585	3.315	ug/l	97
88) 1,4-Dichlorobenzene	12.036	146	10908	3.486	ug/l	90
89) n-Butylbenzene	12.329	91	18004	3.419	ug/l	94
90) Hexachloroethane	12.536	117	3553	3.153	ug/l	94
91) 1,2-Dichlorobenzene	12.335	146	10712	3.444	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	1687	2.812	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	5720	3.359	ug/l	96
94) Hexachlorobutadiene	13.719	225	2816	3.535	ug/l	96
95) Naphthalene	13.774	128	19902	3.209	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	5745	3.217	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046046.D
Acq On : 05 May 2025 16:04
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Quant Time: May 06 06:12:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 06 06:12:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.538 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:168 Resp: 96964
Ion Ratio Lower Upper
168 100
99 74.0 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 2.894 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 85 Resp: 6195
Ion Ratio Lower Upper
85 100
87 28.7 15.7 47.1









#3
Chloromethane
Concen: 3.125 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 50 Resp: 6587
Ion Ratio Lower Upper
50 100
52 32.2 25.4 38.2



#4
Vinyl Chloride
Concen: 3.249 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 62 Resp: 5999
Ion Ratio Lower Upper
62 100
64 29.1 25.2 37.8









#5
Bromomethane
Concen: 3.164 ug/l
RT: 1.593 min Scan# 83
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 94 Resp: 2962
Ion Ratio Lower Upper
94 100
96 102.2 72.8 109.2



#6
Chloroethane
Concen: 3.750 ug/l
RT: 1.666 min Scan# 95
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 64 Resp: 3567
Ion Ratio Lower Upper
64 100
66 33.7 24.3 36.5









#7
 Trichlorofluoromethane
 Concen: 3.413 ug/l
 RT: 1.874 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion:101 Resp: 9599
 Ion Ratio Lower Upper
 101 100
 103 54.0 51.0 76.4



#8
 Diethyl Ether
 Concen: 3.266 ug/l
 RT: 2.130 min Scan# 171
 Delta R.T. -0.006 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04

Tgt Ion: 74 Resp: 3020
 Ion Ratio Lower Upper
 74 100
 45 115.1 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 3.517 ug/l
RT: 2.325 min Scan# 20
Delta R.T. 0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:101 Resp: 5911
Ion Ratio Lower Upper
101 100
85 46.3 38.6 58.0
151 68.8 55.2 82.8



#10
Methyl Iodide
Concen: 2.990 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:142 Resp: 5900
Ion Ratio Lower Upper
142 100
127 52.8 41.7 62.5
141 14.3 11.5 17.3









#11
Tert butyl alcohol
Concen: 16.111 ug/l
RT: 2.965 min Scan# 30
Delta R.T. -0.012 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 59 Resp: 5539
Ion Ratio Lower Upper
59 100
57 10.0 8.6 12.8



#12
1,1-Dichloroethene
Concen: 3.359 ug/l
RT: 2.306 min Scan# 200
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 96 Resp: 5496
Ion Ratio Lower Upper
96 100
61 178.7 146.2 219.2
98 59.5 51.0 76.6





#13

Acrolein

Concen: 14.206 ug/l

RT: 2.233 min Scan# 188

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 56 Resp: 5673

Ion Ratio Lower Upper

56 100

55 74.5 56.2 84.4



#14

Allyl chloride

Concen: 3.319 ug/l

RT: 2.654 min Scan# 257

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 41 Resp: 10262


Ion Ratio Lower Upper

41 100

39 74.1 60.6 90.8

76 29.1 24.9 37.3





#15

Acrylonitrile

Concen: 16.282 ug/l

RT: 3.062 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 53 Resp: 16746

Ion Ratio Lower Upper

53 100

52 84.3 65.3 97.9

51 38.5 29.8 44.8



#16

Acetone

Concen: 20.036 ug/l

RT: 2.380 min Scan# 212

Delta R.T. -0.006 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 43 Resp: 19773

Ion Ratio Lower Upper

43 100

58 27.7 21.2 31.8





#17

Carbon Disulfide

Concen: 2.919 ug/l

RT: 2.501 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 76 Resp: 11068

Ion Ratio Lower Upper

76 100

78 10.2 7.0 10.4



#18

Methyl Acetate

Concen: 3.357 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 43 Resp: 7909

Ion Ratio Lower Upper

43 100

74 20.9 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 3.241 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 73 Resp: 18497
Ion Ratio Lower Upper
73 100
57 20.5 17.7 26.5



#20
Methylene Chloride
Concen: 3.341 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 84 Resp: 6680
Ion Ratio Lower Upper
84 100
49 137.0 113.9 170.9
51 43.8 33.5 50.3
86 67.6 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 3.234 ug/l
RT: 3.081 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 96 Resp: 5402
Ion Ratio Lower Upper
96 100
61 149.9 119.5 179.3
98 62.9 50.0 75.0



#22
Diisopropyl ether
Concen: 3.230 ug/l
RT: 3.751 min Scan# 437
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 45 Resp: 18657
Ion Ratio Lower Upper
45 100
43 128.0 66.6 100.0#
87 24.7 19.8 29.6
59 10.7 8.6 12.8









#23
 Vinyl Acetate
 Concen: 16.127 ug/l
 RT: 3.715 min Scan# 411
 Delta R.T. -0.006 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion: 43 Resp: 82328
 Ion Ratio Lower Upper
 43 100
 86 9.1 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 3.297 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04

Tgt Ion: 63 Resp: 11193
 Ion Ratio Lower Upper
 63 100
 98 6.9 3.6 10.8
 100 4.7 2.1 6.3





#25

2-Butanone

Concen: 18.340 ug/l

RT: 4.562 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC005

Tgt Ion: 43 Resp: 26125

Ion Ratio Lower Upper

43 100

72 21.8 18.4 27.6



#26

2,2-Dichloropropane

Concen: 3.205 ug/l

RT: 4.465 min Scan# 554

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 77 Resp: 8245

Ion Ratio Lower Upper

77 100

97 23.1 10.5 31.5









#27
 cis-1,2-Dichloroethene
 Concen: 3.089 ug/l
 RT: 4.483 min Scan# 557
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005

Tgt Ion: 96 Resp: 6222




Ion	Ratio	Lower	Upper
96	100		
61	168.5	0.0	322.8
98	67.6	0.0	129.0




#28
 Bromochloromethane
 Concen: 2.969 ug/l
 RT: 4.891 min Scan# 624
 Delta R.T. -0.006 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04

Tgt Ion: 49 Resp: 5360

Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	4.0
130	70.5	56.2	84.2





#29

Tetrahydrofuran

Concen: 16.680 ug/l

RT: 5.001 min Scan# 642

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 42 Resp: 15431

Ion Ratio Lower Upper

42 100

72 42.2 33.7 50.5

71 37.4 31.4 47.2



#30

Chloroform

Concen: 3.294 ug/l

RT: 5.080 min Scan# 655

Delta R.T. -0.006 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 83 Resp: 11624

Ion Ratio Lower Upper

83 100

85 69.9 49.3 73.9









#31
Cyclohexane
Concen: 3.588 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 56 Resp: 10269
Ion Ratio Lower Upper
56 100
69 31.6 24.4 36.6
84 79.1 66.9 100.3



#32
1,1,1-Trichloroethane
Concen: 3.229 ug/l
RT: 5.367 min Scan# 702
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 97 Resp: 9827
Ion Ratio Lower Upper
97 100
99 62.6 51.8 77.6
61 51.8 40.1 60.1









#33
 1,2-Dichloroethane-d4
 Concen: 3.143 ug/l
 RT: 5.946 min Scan# 798
 Delta R.T. -0.006 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion: 65 Resp: 9067
 Ion Ratio Lower Upper
 65 100
 67 50.1 0.0 99.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.751 min Scan# 929
 Delta R.T. -0.006 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04

Tgt Ion: 114 Resp: 168484
 Ion Ratio Lower Upper
 114 100
 63 24.5 0.0 49.2
 88 17.7 0.0 33.6









#35
Dibromofluoromethane
Concen: 3.161 ug/l
RT: 5.373 min Scan# 704
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:113 Resp: 5969
Ion Ratio Lower Upper
113 100
111 104.4 83.1 124.7
192 17.5 13.3 19.9



#36
1,1-Dichloropropene
Concen: 3.551 ug/l
RT: 5.678 min Scan# 753
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 75 Resp: 7787
Ion Ratio Lower Upper
75 100
110 30.6 16.3 48.9
77 30.5 24.3 36.5





#37

Ethyl Acetate

Concen: 3.279 ug/l

RT: 4.721 min Scan# 595

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 43 Resp: 9021

Ion Ratio Lower Upper

43 100

61 11.9 10.3 15.5

70 9.2 7.9 11.9



#38

Carbon Tetrachloride

Concen: 3.376 ug/l

RT: 5.659 min Scan# 750

Delta R.T. -0.012 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 117 Resp: 8505


Ion Ratio Lower Upper

117 100

119 90.6 75.2 112.8

121 29.5 24.2 36.4





#39

Methylcyclohexane

Concen: 3.597 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005





Tgt Ion: 83 Resp: 9882

Ion Ratio Lower Upper

83 100

55 73.3 64.7 97.1

98 50.3 37.4 56.2



#40

Benzene

Concen: 3.315 ug/l

RT: 6.031 min Scan# 811

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 78 Resp: 22528

Ion Ratio Lower Upper

78 100

77 23.9 19.0 28.4









#41
Methacrylonitrile
Concen: 3.174 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Tgt Ion:	41	Resp:	4851
Ion	Ratio	Lower	Upper
41	100		
39	59.7	47.2	70.8
67	0.0	50.7	76.1#
52	28.8	24.6	37.0



#42
1,2-Dichloroethane
Concen: 3.563 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:	62	Resp:	10011
Ion	Ratio	Lower	Upper
62	100		
98	7.8	0.0	15.2





#43

Isopropyl Acetate

Concen: 3.309 ug/l

RT: 6.348 min Scan# 863

Delta R.T. 0.012 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 43 Resp: 13914

Ion Ratio Lower Upper

43 100

61 16.2 14.3 21.5

87 10.6 9.5 14.3



#44

Trichloroethene

Concen: 3.307 ug/l

RT: 7.123 min Scan# 990

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 130 Resp: 5313

Ion Ratio Lower Upper

130 100

95 108.1 0.0 204.2





#45

1,2-Dichloropropane

Concen: 3.228 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC005

Tgt Ion: 63 Resp: 5451

Ion Ratio Lower Upper

63 100

65 36.2 24.1 36.1#



#46

Dibromomethane

Concen: 3.318 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 93 Resp: 4417


Ion Ratio Lower Upper

93 100

95 80.1 65.6 98.4

174 87.6 68.2 102.2









#47
Bromodichloromethane
Concen: 3.254 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 83	Resp:	8394
Ion	Ratio	Lower Upper
83	100	
85	61.9	50.5 75.7
127	9.0	6.5 9.7



#48
Methyl methacrylate
Concen: 3.321 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 41	Resp:	7171
Ion	Ratio	Lower Upper
41	100	
69	68.4	58.5 87.7
39	65.1	51.7 77.5





#49

1,4-Dioxane

Concen: 67.721 ug/l

RT: 7.659 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion: 88 Resp: 2907

Ion Ratio Lower Upper

88 100

43 45.3 33.4 50.2

58 71.0 58.6 88.0



#50

Toluene-d8

Concen: 3.232 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 98 Resp: 20566

Ion Ratio Lower Upper

98 100

100 66.1 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 16.967 ug/l
RT: 8.567 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 46794
Ion Ratio Lower Upper
43 100
58 34.8 28.9 43.3



#52
Toluene
Concen: 3.505 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 92 Resp: 14126
Ion Ratio Lower Upper
92 100
91 167.2 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 3.155 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Tgt Ion: 75 Resp: 6834

Ion Ratio Lower Upper

75 100

77 31.1 25.0 37.6



#54

cis-1,3-Dichloropropene

Concen: 3.129 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 75 Resp: 7908


Ion Ratio Lower Upper

75 100

77 31.9 25.4 38.0

39 55.3 52.2 78.4









#55
 1,1,2-Trichloroethane
 Concen: 3.490 ug/l
 RT: 9.153 min Scan# 1323
 Delta R.T. 0.006 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion:	97	Resp:	5685
Ion	Ratio	Lower	Upper
97	100		
83	88.0	69.0	103.6
85	54.4	44.3	66.5
99	62.2	47.8	71.6



#56
 Ethyl methacrylate
 Concen: 3.390 ug/l
 RT: 9.116 min Scan# 1317
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04

Tgt Ion:	69	Resp:	8555
Ion	Ratio	Lower	Upper
69	100		
41	73.7	60.8	91.2
39	49.3	39.0	58.6









#57
1,3-Dichloropropane
Concen: 3.508 ug/l
RT: 9.305 min Scan# 1348
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005


Tgt Ion: 76 Resp: 10127
Ion Ratio Lower Upper
76 100
78 29.5 26.3 39.5



#58
2-Chloroethyl Vinyl ether
Concen: 18.346 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 63 Resp: 20809
Ion Ratio Lower Upper
63 100
106 26.4 21.5 32.3









#59
2-Hexanone
Concen: 16.651 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 34853
Ion Ratio Lower Upper
43 100
58 47.5 24.9 74.6



#60
Dibromochloromethane
Concen: 3.102 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 129 Resp: 5500
Ion Ratio Lower Upper
129 100
127 80.4 39.3 117.8









#61
1,2-Dibromoethane
Concen: 3.356 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:107 Resp: 5614
Ion Ratio Lower Upper
107 100
109 91.0 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 3.378 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 95 Resp: 7822
Ion Ratio Lower Upper
95 100
174 65.0 0.0 135.8
176 66.6 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:117 Resp: 147167
Ion Ratio Lower Upper
117 100
82 61.1 50.6 76.0
119 32.1 25.8 38.6



#64
Tetrachloroethene
Concen: 3.162 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:164 Resp: 4752
Ion Ratio Lower Upper
164 100
166 127.8 105.0 157.6
129 105.8 83.5 125.3
131 104.1 76.5 114.7









#65
Chlorobenzene
Concen: 3.399 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:112 Resp: 15391
Ion Ratio Lower Upper
112 100
114 32.6 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 3.326 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:131 Resp: 5024
Ion Ratio Lower Upper
131 100
133 95.0 47.3 141.9
119 64.8 31.6 95.0





#67

Ethyl Benzene

Concen: 3.463 ug/l

RT: 10.195 min Scan# 1493

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC005

Tgt Ion: 91 Resp: 26732

Ion Ratio Lower Upper

91 100

106 29.9 23.4 35.2



#68

m/p-Xylenes

Concen: 7.156 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion:106 Resp: 19957

Ion Ratio Lower Upper

106 100

91 210.1 171.2 256.8









#69
o-Xylene
Concen: 3.343 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:106 Resp: 9398
Ion Ratio Lower Upper
106 100
91 233.6 112.7 338.1



#70
Styrene
Concen: 3.312 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:104 Resp: 14900
Ion Ratio Lower Upper
104 100
78 57.9 45.7 68.5
103 55.6 43.7 65.5





#71

Bromoform

Concen: 3.059 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005





Tgt Ion:173 Resp: 3470

Ion Ratio Lower Upper

173 100

175 45.8 23.4 70.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion:152 Resp: 67939


Ion Ratio Lower Upper

152 100

115 66.8 46.9 140.7

150 157.5 0.0 351.0









#73
Isopropylbenzene
Concen: 3.266 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:105 Resp: 24201
Ion Ratio Lower Upper
105 100
120 26.4 12.8 38.4



#74
N-ethyl acetate
Concen: 3.038 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 43 Resp: 11224
Ion Ratio Lower Upper
43 100
70 35.4 30.9 46.3
55 25.0 18.7 28.1
61 19.6 17.1 25.7









#75
 1,1,2,2-Tetrachloroethane
 Concen: 3.479 ug/l
 RT: 11.207 min Scan# 1660
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion: 83 Resp: 9170
 Ion Ratio Lower Upper
 83 100
 131 8.3 5.0 14.9
 85 61.7 31.9 95.7



#76
 1,2,3-Trichloropropane
 Concen: 3.018 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04

Tgt Ion: 75 Resp: 8548
 Ion Ratio Lower Upper
 75 100
 77 37.7 20.5 61.5









#77
Bromobenzene
Concen: 3.417 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion:156 Resp: 5857
Ion Ratio Lower Upper
156 100
77 177.0 92.8 278.3
158 97.7 48.8 146.3



#78
n-propylbenzene
Concen: 3.424 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 91 Resp: 28436
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4









#79
2-Chlorotoluene
Concen: 3.336 ug/l
RT: 11.360 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 91 Resp: 18672
Ion Ratio Lower Upper
91 100
126 29.7 15.6 46.7



#80
1,3,5-Trimethylbenzene
Concen: 3.372 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 105 Resp: 20742
Ion Ratio Lower Upper
105 100
120 45.9 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 2.686 ug/l
RT: 11.018 min Scan# 1100
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005


Tgt Ion: 75 Resp: 1828
Ion Ratio Lower Upper
75 100
53 125.8 100.5 150.7
89 47.9 37.0 55.6



#82
4-Chlorotoluene
Concen: 3.245 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 91 Resp: 19966
Ion Ratio Lower Upper
91 100
126 27.3 13.3 39.8





#83

tert-Butylbenzene

Concen: 3.484 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005





Tgt Ion:119 Resp: 21049

Ion Ratio Lower Upper

119 100

91 67.0 32.9 98.7

134 23.8 11.4 34.1



#84

1,2,4-Trimethylbenzene

Concen: 3.355 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion:105 Resp: 20613

Ion Ratio Lower Upper

105 100

120 44.5 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 3.412 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion:105 Resp: 25590
 Ion Ratio Lower Upper
 105 100
 134 19.1 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 3.460 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04

Tgt Ion:119 Resp: 20955
 Ion Ratio Lower Upper
 119 100
 134 25.6 12.5 37.5
 91 28.0 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 3.315 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:146 Resp: 10585

Ion Ratio Lower Upper

146 100

111 46.2 22.1 66.3

148 62.4 32.1 96.5



#88

1,4-Dichlorobenzene

Concen: 3.486 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion:146 Resp: 10908


Ion Ratio Lower Upper

146 100

111 52.1 21.3 63.9

148 69.4 31.9 95.5





#89

n-Butylbenzene

Concen: 3.419 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005





Tgt Ion: 91 Resp: 18004

Ion Ratio Lower Upper

91 100

92 49.5 26.9 80.7

134 21.2 11.8 35.3



#90

Hexachloroethane

Concen: 3.153 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 117 Resp: 3553

Ion Ratio Lower Upper

117 100

201 58.5 31.6 94.7





#91

1,2-Dichlorobenzene

Concen: 3.444 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:146 Resp: 10712

Ion Ratio Lower Upper

146 100

111 44.7 22.4 67.3

148 64.2 31.6 94.8



#92

1,2-Dibromo-3-Chloropropane

Concen: 2.812 ug/l

RT: 12.945 min Scan# 1945

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 75 Resp: 1687


Ion Ratio Lower Upper

75 100

155 71.7 34.9 104.8

157 87.0 43.8 131.4









#93
 1,2,4-Trichlorobenzene
 Concen: 3.359 ug/l
 RT: 13.585 min Scan# 2050
 Delta R.T. -0.000 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC005


Tgt Ion:180 Resp: 5720
 Ion Ratio Lower Upper
 180 100
 182 91.6 47.9 143.8
 145 35.7 17.1 51.2



#94
 Hexachlorobutadiene
 Concen: 3.535 ug/l
 RT: 13.719 min Scan# 2072
 Delta R.T. -0.006 min
 Lab File: VX046046.D
 Acq: 05 May 2025 16:04

Tgt Ion:225 Resp: 2816
 Ion Ratio Lower Upper
 225 100
 223 65.1 30.8 92.4
 227 58.8 30.9 92.7





#95

Naphthalene

Concen: 3.209 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005





Tgt Ion:128 Resp: 19902

Ion Ratio Lower Upper

128 100

127 13.0 10.4 15.6

129 11.5 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 3.217 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion:180 Resp: 5745

Ion Ratio Lower Upper

180 100

182 99.1 47.8 143.3

145 34.8 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046047.D
 Acq On : 05 May 2025 16:27
 Operator : JC/MD
 Sample : VSTDIC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001

Quant Time: May 06 06:13:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	92040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	167022	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	140108	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59123	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	2188	0.799	ug/l	0.02
Spiked Amount	50.000	Range	74 - 125	Recovery	=	1.600%#
35) Dibromofluoromethane	5.391	113	1350	0.721	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	1.440%#
50) Toluene-d8	8.647	98	3964	0.628	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	1.260%#
62) 4-Bromofluorobenzene	11.085	95	1250	0.545	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	1.080%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	1211	0.596	ug/l	98
3) Chloromethane	1.307	50	1278	0.639	ug/l	97
4) Vinyl Chloride	1.374	62	1239	0.707	ug/l #	82
5) Bromomethane	1.593	94	770	0.866	ug/l	88
6) Chloroethane	1.678	64	859	0.951	ug/l	91
7) Trichlorofluoromethane	1.880	101	1959	0.734	ug/l	95
8) Diethyl Ether	2.130	74	742	0.845	ug/l	87
9) 1,1,2-Trichlorotrifluo...	2.319	101	1166	0.731	ug/l	96
10) Methyl Iodide	2.441	142	1297	0.692	ug/l #	87
11) Tert butyl alcohol	2.965	59	1176	3.604	ug/l #	78
12) 1,1-Dichloroethene	2.313	96	1093	0.704	ug/l #	87
13) Acrolein	2.233	56	1374	3.625	ug/l	83
14) Allyl chloride	2.654	41	1936	0.660	ug/l	89
15) Acrylonitrile	3.075	53	3343	3.424	ug/l	98
16) Acetone	2.380	43	3500	3.736	ug/l	97
17) Carbon Disulfide	2.508	76	2619	0.728	ug/l	98
18) Methyl Acetate	2.709	43	1852	0.828	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	3588	0.662	ug/l	96
20) Methylene Chloride	2.782	84	1571	0.828	ug/l #	90
21) trans-1,2-Dichloroethene	3.081	96	1111	0.701	ug/l #	71
22) Diisopropyl ether	3.757	45	3857	0.703	ug/l #	26
23) Vinyl Acetate	3.727	43	15283	3.154	ug/l	96
24) 1,1-Dichloroethane	3.605	63	2054	0.637	ug/l	97
25) 2-Butanone	4.593	43	4558	3.371	ug/l	92
26) 2,2-Dichloropropane	4.458	77	1777	0.728	ug/l	84
27) cis-1,2-Dichloroethene	4.483	96	1324	0.693	ug/l	91
28) Bromochloromethane	4.910	49	682	0.398	ug/l #	63
29) Tetrahydrofuran	5.019	42	2931	3.338	ug/l	99
30) Chloroform	5.093	83	2328	0.695	ug/l	98
31) Cyclohexane	5.458	56	2072	0.763	ug/l	88
32) 1,1,1-Trichloroethane	5.379	97	1869	0.647	ug/l #	51
36) 1,1-Dichloropropene	5.684	75	1648	0.758	ug/l #	83
37) Ethyl Acetate	4.733	43	1447	0.531	ug/l #	72
38) Carbon Tetrachloride	5.678	117	1807	0.724	ug/l #	78
39) Methylcyclohexane	7.373	83	2096	0.770	ug/l	97
40) Benzene	6.044	78	4502	0.668	ug/l #	89

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046047.D
 Acq On : 05 May 2025 16:27
 Operator : JC/MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC001

Quant Time: May 06 06:13:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.946	41	778	0.513	ug/l #	56
42) 1,2-Dichloroethane	6.092	62	1934	0.694	ug/l	81
43) Isopropyl Acetate	6.354	43	2553	0.612	ug/l #	81
44) Trichloroethene	7.135	130	1083	0.680	ug/l	85
45) 1,2-Dichloropropane	7.440	63	1059	0.633	ug/l	89
46) Dibromomethane	7.586	93	880	0.667	ug/l	97
47) Bromodichloromethane	7.818	83	1620	0.633	ug/l #	91
48) Methyl methacrylate	7.714	41	1033	0.483	ug/l #	71
49) 1,4-Dioxane	7.677	88	498	11.703	ug/l #	66
51) 4-Methyl-2-Pentanone	8.580	43	9378	3.430	ug/l	97
52) Toluene	8.720	92	2684	0.672	ug/l	96
53) t-1,3-Dichloropropene	8.988	75	1238	0.576	ug/l #	79
54) cis-1,3-Dichloropropene	8.379	75	1414	0.564	ug/l #	87
55) 1,1,2-Trichloroethane	9.153	97	1029	0.637	ug/l	92
56) Ethyl methacrylate	9.128	69	1260	0.504	ug/l #	76
57) 1,3-Dichloropropane	9.311	76	2036	0.711	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.244	63	3846	3.420	ug/l	94
59) 2-Hexanone	9.439	43	6424	3.096	ug/l	88
60) Dibromochloromethane	9.525	129	1023	0.582	ug/l	93
61) 1,2-Dibromoethane	9.610	107	1077	0.650	ug/l	98
64) Tetrachloroethene	9.275	164	972	0.679	ug/l	95
65) Chlorobenzene	10.079	112	3170	0.735	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	1035	0.720	ug/l #	67
67) Ethyl Benzene	10.195	91	5052	0.687	ug/l	94
68) m/p-Xylenes	10.299	106	3630	1.367	ug/l	97
69) o-Xylene	10.640	106	1799	0.672	ug/l	94
70) Styrene	10.659	104	2664	0.622	ug/l	98
71) Bromoform	10.799	173	657	0.608	ug/l #	100
73) Isopropylbenzene	10.963	105	4480	0.695	ug/l	96
74) N-amyl acetate	10.854	43	2028	0.631	ug/l #	91
75) 1,1,2,2-Tetrachloroethane	11.213	83	1835	0.800	ug/l	94
76) 1,2,3-Trichloropropane	11.244	75	1858	0.754	ug/l	77
77) Bromobenzene	11.201	156	1095	0.734	ug/l	96
78) n-propylbenzene	11.305	91	5052	0.699	ug/l	97
79) 2-Chlorotoluene	11.360	91	3765	0.773	ug/l	94
80) 1,3,5-Trimethylbenzene	11.451	105	3590	0.671	ug/l	96
81) trans-1,4-Dichloro-2-b...	11.024	75	252	0.425	ug/l #	71
82) 4-Chlorotoluene	11.457	91	3815	0.713	ug/l	97
83) tert-Butylbenzene	11.713	119	3951	0.751	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	3725	0.697	ug/l	97
85) sec-Butylbenzene	11.890	105	4385	0.672	ug/l	100
86) p-Isopropyltoluene	12.006	119	3577	0.679	ug/l	94
87) 1,3-Dichlorobenzene	11.969	146	1914	0.689	ug/l	99
88) 1,4-Dichlorobenzene	11.969	146	1914	0.703	ug/l	98
89) n-Butylbenzene	12.335	91	2889	0.631	ug/l	92
90) Hexachloroethane	12.536	117	604	0.616	ug/l	99
91) 1,2-Dichlorobenzene	12.341	146	2022	0.747	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	12.945	75	306	0.586	ug/l	83
93) 1,2,4-Trichlorobenzene	13.591	180	1019	0.688	ug/l	94
94) Hexachlorobutadiene	13.725	225	465	0.671	ug/l	89
95) Naphthalene	13.780	128	4138	0.767	ug/l	97
96) 1,2,3-Trichlorobenzene	13.963	180	1113	0.716	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046047.D
Acq On : 05 May 2025 16:27
Operator : JC/MD
Sample : VSTDICC001
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001


Quant Time: May 06 06:13:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 06 06:13:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:168 Resp: 92040
Ion Ratio Lower Upper
168 100
99 71.5 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 0.596 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 85 Resp: 1211
Ion Ratio Lower Upper
85 100
87 32.4 15.7 47.1









#3
Chloromethane
Concen: 0.639 ug/l
RT: 1.307 min Scan# 36
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 50 Resp: 1278
Ion Ratio Lower Upper
50 100
52 30.3 25.4 38.2



#4
Vinyl Chloride
Concen: 0.707 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 62 Resp: 1239
Ion Ratio Lower Upper
62 100
64 41.7 25.2 37.8#









#5
Bromomethane
Concen: 0.866 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 94 Resp: 770
Ion Ratio Lower Upper
94 100
96 79.9 72.8 109.2



#6
Chloroethane
Concen: 0.951 ug/l
RT: 1.678 min Scan# 97
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 64 Resp: 859
Ion Ratio Lower Upper
64 100
66 35.2 24.3 36.5









#7
 Trichlorofluoromethane
 Concen: 0.734 ug/l
 RT: 1.880 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion:101 Resp: 1959
 Ion Ratio Lower Upper
 101 100
 103 67.8 51.0 76.4



#8
 Diethyl Ether
 Concen: 0.845 ug/l
 RT: 2.130 min Scan# 171
 Delta R.T. -0.006 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27

Tgt Ion: 74 Resp: 742
 Ion Ratio Lower Upper
 74 100
 45 123.6 54.9 164.8









#9
 1,1,2-Trichlorotrifluoroethane
 Concen: 0.731 ug/l
 RT: 2.319 min Scan# 202
 Delta R.T. -0.000 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion:101 Resp: 1166
 Ion Ratio Lower Upper
 101 100
 85 47.3 38.6 58.0
 151 73.1 55.2 82.8



#10
 Methyl Iodide
 Concen: 0.692 ug/l
 RT: 2.441 min Scan# 222
 Delta R.T. -0.006 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27

Tgt Ion:142 Resp: 1297
 Ion Ratio Lower Upper
 142 100
 127 45.5 41.7 62.5
 141 3.9 11.5 17.3#





#11

Tert butyl alcohol

Concen: 3.604 ug/l

RT: 2.965 min Scan# 30

Delta R.T. -0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 59 Resp: 1176

Ion Ratio Lower Upper

59 100

57 2.4 8.6 12.8#



#12

1,1-Dichloroethene

Concen: 0.704 ug/l

RT: 2.313 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion: 96 Resp: 1093


Ion Ratio Lower Upper

96 100

61 199.6 146.2 219.2

98 49.6 51.0 76.6#





#13

Acrolein

Concen: 3.625 ug/l

RT: 2.233 min Scan# 188

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 56 Resp: 1374

Ion Ratio Lower Upper

56 100

55 83.9 56.2 84.4



#14

Allyl chloride

Concen: 0.660 ug/l

RT: 2.654 min Scan# 257

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion: 41 Resp: 1936


Ion Ratio Lower Upper

41 100

39 85.5 60.6 90.8

76 26.4 24.9 37.3









#15
Acrylonitrile
Concen: 3.424 ug/l
RT: 3.075 min Scan# 31
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 53 Resp: 3343
Ion Ratio Lower Upper
53 100
52 82.9 65.3 97.9
51 34.4 29.8 44.8



#16
Acetone
Concen: 3.736 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 43 Resp: 3500
Ion Ratio Lower Upper
43 100
58 28.1 21.2 31.8





#17

Carbon Disulfide

Concen: 0.728 ug/l

RT: 2.508 min Scan# 21

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 76 Resp: 2619

Ion Ratio Lower Upper

76 100

78 8.0 7.0 10.4



#18

Methyl Acetate

Concen: 0.828 ug/l

RT: 2.709 min Scan# 266

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 43 Resp: 1852

Ion Ratio Lower Upper

43 100

74 22.0 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 0.662 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 73 Resp: 3588
Ion Ratio Lower Upper
73 100
57 20.3 17.7 26.5



#20
Methylene Chloride
Concen: 0.828 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 84 Resp: 1571
Ion Ratio Lower Upper
84 100
49 140.0 113.9 170.9
51 51.3 33.5 50.3#
86 85.2 53.8 80.8#









#21
trans-1,2-Dichloroethene
Concen: 0.701 ug/l
RT: 3.081 min Scan# 311
Delta R.T. -0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 96 Resp: 1111
Ion Ratio Lower Upper
96 100
61 119.0 119.5 179.3#
98 30.1 50.0 75.0#



#22
Diisopropyl ether
Concen: 0.703 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 45 Resp: 3857
Ion Ratio Lower Upper
45 100
43 176.2 66.6 100.0#
87 28.9 19.8 29.6
59 9.1 8.6 12.8









#23
 Vinyl Acetate
 Concen: 3.154 ug/l
 RT: 3.727 min Scan# 413
 Delta R.T. 0.006 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion: 43 Resp: 15283
 Ion Ratio Lower Upper
 43 100
 86 10.8 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 0.637 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27

Tgt Ion: 63 Resp: 2054
 Ion Ratio Lower Upper
 63 100
 98 6.3 3.6 10.8
 100 5.8 2.1 6.3









#25
2-Butanone
Concen: 3.371 ug/l
RT: 4.593 min Scan# 51
Delta R.T. 0.037 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 4558
Ion Ratio Lower Upper
43 100
72 19.2 18.4 27.6



#26
2,2-Dichloropropane
Concen: 0.728 ug/l
RT: 4.458 min Scan# 553
Delta R.T. -0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 77 Resp: 1777
Ion Ratio Lower Upper
77 100
97 13.7 10.5 31.5









#27
cis-1,2-Dichloroethene
Concen: 0.693 ug/l
RT: 4.483 min Scan# 51
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 96	Resp: 1324		
Ion Ratio	Lower	Upper	
96	100		
61	145.5	0.0	322.8
98	63.1	0.0	129.0



#28
Bromochloromethane
Concen: 0.398 ug/l
RT: 4.910 min Scan# 627
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 49	Resp: 682		
Ion Ratio	Lower	Upper	
49 100			
129 0.0	0.0	4.0	
130 101.6	56.2	84.2#	









#29
Tetrahydrofuran
Concen: 3.338 ug/l
RT: 5.019 min Scan# 642
Delta R.T. 0.018 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 42 Resp: 2931
Ion Ratio Lower Upper
42 100
72 41.9 33.7 50.5
71 37.7 31.4 47.2



#30
Chloroform
Concen: 0.695 ug/l
RT: 5.093 min Scan# 657
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27


Tgt Ion: 83 Resp: 2328
Ion Ratio Lower Upper
83 100
85 60.4 49.3 73.9








#31
Cyclohexane
Concen: 0.763 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion: 56 Resp: 2072




Ion	Ratio	Lower	Upper
56	100		
69	25.1	24.4	36.6
84	95.5	66.9	100.3




#32
1,1,1-Trichloroethane
Concen: 0.647 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 97 Resp: 1869

Ion	Ratio	Lower	Upper
97	100		
99	0.0	51.8	77.6#
61	55.0	40.1	60.1





#33

1,2-Dichloroethane-d4

Concen: 0.799 ug/l

RT: 5.970 min Scan# 801

Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 65 Resp: 2188

Ion Ratio Lower Upper

65 100

67 42.3 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion: 114 Resp: 167022


Ion Ratio Lower Upper

114 100

63 24.2 0.0 49.2

88 17.1 0.0 33.6





#35

Dibromofluoromethane

Concen: 0.721 ug/l

RT: 5.391 min Scan# 704

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:113 Resp: 1350

Ion Ratio Lower Upper

113 100

111 117.0 83.1 124.7

192 12.8 13.3 19.9#



#36

1,1-Dichloropropene

Concen: 0.758 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion: 75 Resp: 1648


Ion Ratio Lower Upper

75 100

110 28.2 16.3 48.9

77 15.8 24.3 36.5#









#37
Ethyl Acetate
Concen: 0.531 ug/l
RT: 4.733 min Scan# 598
Delta R.T. 0.018 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 1447
Ion Ratio Lower Upper
43 100
61 0.0 10.3 15.5#
70 1.4 7.9 11.9#



#38
Carbon Tetrachloride
Concen: 0.724 ug/l
RT: 5.678 min Scan# 753
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 117 Resp: 1807
Ion Ratio Lower Upper
117 100
119 67.0 75.2 112.8#
121 28.5 24.2 36.4





#39

Methylcyclohexane

Concen: 0.770 ug/l

RT: 7.373 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 83 Resp: 2096

Ion Ratio Lower Upper

83 100

55 80.1 64.7 97.1

98 41.9 37.4 56.2



#40

Benzene

Concen: 0.668 ug/l

RT: 6.044 min Scan# 813

Delta R.T. 0.012 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 78 Resp: 4502

Ion Ratio Lower Upper

78 100

77 29.1 19.0 28.4#









#41
Methacrylonitrile
Concen: 0.513 ug/l
RT: 4.946 min Scan# 61
Delta R.T. 0.030 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:	41	Resp:	778
Ion	Ratio	Lower	Upper
41	100		
39	51.7	47.2	70.8
67	0.0	50.7	76.1#
52	13.6	24.6	37.0#



#42
1,2-Dichloroethane
Concen: 0.694 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:	62	Resp:	1934
Ion	Ratio	Lower	Upper
62	100		
98	1.1	0.0	15.2





#43

Isopropyl Acetate

Concen: 0.612 ug/l

RT: 6.354 min Scan# 864

Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 43 Resp: 2553

Ion Ratio Lower Upper

43 100

61 10.7 14.3 21.5#

87 3.2 9.5 14.3#



#44

Trichloroethene

Concen: 0.680 ug/l

RT: 7.135 min Scan# 992

Delta R.T. 0.012 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion:130 Resp: 1083

Ion Ratio Lower Upper

130 100

95 117.3 0.0 204.2





#45

1,2-Dichloropropane

Concen: 0.633 ug/l

RT: 7.440 min Scan# 1042

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 63 Resp: 1059

Ion Ratio Lower Upper

63 100

65 36.1 24.1 36.1



#46

Dibromomethane

Concen: 0.667 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion: 93 Resp: 880


Ion Ratio Lower Upper

93 100

95 84.5 65.6 98.4

174 88.5 68.2 102.2









#47
Bromodichloromethane
Concen: 0.633 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:	83	Resp:	1620
Ion	Ratio	Lower	Upper
83	100		
85	57.3	50.5	75.7
127	0.0	6.5	9.7#



#48
Methyl methacrylate
Concen: 0.483 ug/l
RT: 7.714 min Scan# 1087
Delta R.T. 0.024 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:	41	Resp:	1033
Ion	Ratio	Lower	Upper
41	100		
69	84.4	58.5	87.7
39	101.4	51.7	77.5#









#49
1,4-Dioxane
Concen: 11.703 ug/l
RT: 7.677 min Scan# 1081
Delta R.T. 0.018 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 88 Resp: 498
Ion Ratio Lower Upper
88 100
43 44.4 33.4 50.2
58 30.3 58.6 88.0#



#50
Toluene-d8
Concen: 0.628 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 98 Resp: 3964
Ion Ratio Lower Upper
98 100
100 73.5 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 3.430 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 9378
Ion Ratio Lower Upper
43 100
58 34.4 28.9 43.3



#52
Toluene
Concen: 0.672 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 92 Resp: 2684
Ion Ratio Lower Upper
92 100
91 165.7 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 0.576 ug/l

RT: 8.988 min Scan# 1196

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 75 Resp: 1238

Ion Ratio Lower Upper

75 100

77 42.7 25.0 37.6#



#54

cis-1,3-Dichloropropene

Concen: 0.564 ug/l

RT: 8.379 min Scan# 1196

Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion: 75 Resp: 1414


Ion Ratio Lower Upper

75 100

77 44.2 25.4 38.0#

39 59.0 52.2 78.4









#55
 1,1,2-Trichloroethane
 Concen: 0.637 ug/l
 RT: 9.153 min Scan# 1323
 Delta R.T. 0.006 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion: 97	Resp: 1029
Ion Ratio	Lower Upper
97 100	
83 74.8	69.0 103.6
85 59.5	44.3 66.5
99 63.0	47.8 71.6



#56
 Ethyl methacrylate
 Concen: 0.504 ug/l
 RT: 9.128 min Scan# 1319
 Delta R.T. 0.012 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27

Tgt Ion: 69	Resp: 1260		
Ion Ratio	Lower	Upper	
69 100			
41 86.6	60.8	91.2	
39 78.0	39.0	58.6	#





#57

1,3-Dichloropropane

Concen: 0.711 ug/l

RT: 9.311 min Scan# 1134

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 76 Resp: 2036

Ion Ratio Lower Upper

76 100

78 31.4 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 3.420 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 63 Resp: 3846

Ion Ratio Lower Upper

63 100

106 24.0 21.5 32.3









#59
2-Hexanone
Concen: 3.096 ug/l
RT: 9.439 min Scan# 1370
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 6424
Ion Ratio Lower Upper
43 100
58 41.6 24.9 74.6



#60
Dibromochloromethane
Concen: 0.582 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 129 Resp: 1023
Ion Ratio Lower Upper
129 100
127 72.5 39.3 117.8





#61

1,2-Dibromoethane

Concen: 0.650 ug/l

RT: 9.610 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion:107 Resp: 1077

Ion Ratio Lower Upper

107 100

109 94.0 73.3 109.9



#62

4-Bromofluorobenzene

Concen: 0.545 ug/l

RT: 11.085 min Scan# 1640

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion: 95 Resp: 1250


Ion Ratio Lower Upper

95 100

174 72.2 0.0 135.8

176 68.9 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:117 Resp: 140108
Ion Ratio Lower Upper
117 100
82 61.0 50.6 76.0
119 32.1 25.8 38.6



#64
Tetrachloroethene
Concen: 0.679 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:164 Resp: 972
Ion Ratio Lower Upper
164 100
166 119.1 105.0 157.6
129 103.4 83.5 125.3
131 95.4 76.5 114.7









#65
Chlorobenzene
Concen: 0.735 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:112 Resp: 3170
Ion Ratio Lower Upper
112 100
114 30.6 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 0.720 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:131 Resp: 1035
Ion Ratio Lower Upper
131 100
133 94.9 47.3 141.9
119 0.0 31.6 95.0#





#67

Ethyl Benzene

Concen: 0.687 ug/l

RT: 10.195 min Scan# 1493

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 91 Resp: 5052

Ion Ratio Lower Upper

91 100

106 32.5 23.4 35.2



#68

m/p-Xylenes

Concen: 1.367 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion:106 Resp: 3630

Ion Ratio Lower Upper

106 100

91 209.8 171.2 256.8









#69
o-Xylene
Concen: 0.672 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:106 Resp: 1799
Ion Ratio Lower Upper
106 100
91 214.9 112.7 338.1



#70
Styrene
Concen: 0.622 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:104 Resp: 2664
Ion Ratio Lower Upper
104 100
78 54.3 45.7 68.5
103 53.9 43.7 65.5





#71

Bromoform

Concen: 0.608 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:173 Resp: 657

Ion Ratio Lower Upper

173 100

175 46.6 23.4 70.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion:152 Resp: 59123


Ion Ratio Lower Upper

152 100

115 66.6 46.9 140.7

150 156.3 0.0 351.0









#73
Isopropylbenzene
Concen: 0.695 ug/l
RT: 10.963 min Scan# 1619
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:105 Resp: 4480
Ion Ratio Lower Upper
105 100
120 23.7 12.8 38.4



#74
N-ethyl acetate
Concen: 0.631 ug/l
RT: 10.854 min Scan# 1602
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 43 Resp: 2028
Ion Ratio Lower Upper
43 100
70 31.5 30.9 46.3
55 24.5 18.7 28.1
61 15.4 17.1 25.7#





#75

1,1,2,2-Tetrachloroethane

Concen: 0.800 ug/l

RT: 11.213 min Scan# 1661

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 83 Resp: 1835

Ion Ratio Lower Upper

83 100

131 7.5 5.0 14.9

85 58.7 31.9 95.7



#76

1,2,3-Trichloropropane

Concen: 0.754 ug/l

RT: 11.244 min Scan# 1666

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 75 Resp: 1858

Ion Ratio Lower Upper

75 100

77 26.5 20.5 61.5









#77
Bromobenzene
Concen: 0.734 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:156 Resp: 1095
Ion Ratio Lower Upper
156 100
77 179.5 92.8 278.3
158 93.5 48.8 146.3



#78
n-propylbenzene
Concen: 0.699 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 91 Resp: 5052
Ion Ratio Lower Upper
91 100
120 20.1 10.8 32.4









#79
2-Chlorotoluene
Concen: 0.773 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion: 91 Resp: 3765
Ion Ratio Lower Upper
91 100
126 28.0 15.6 46.7



#80
1,3,5-Trimethylbenzene
Concen: 0.671 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 105 Resp: 3590
Ion Ratio Lower Upper
105 100
120 43.4 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 0.425 ug/l
RT: 11.024 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001


Tgt Ion: 75 Resp: 252
Ion Ratio Lower Upper
75 100
53 164.7 100.5 150.7#
89 37.3 37.0 55.6



#82
4-Chlorotoluene
Concen: 0.713 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 91 Resp: 3815
Ion Ratio Lower Upper
91 100
126 24.8 13.3 39.8









#83
 tert-Butylbenzene
 Concen: 0.751 ug/l
 RT: 11.713 min Scan# 1743
 Delta R.T. -0.000 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion:119	Resp:	3951
Ion	Ratio	Lower Upper
119	100	
91	63.9	32.9 98.7
134	25.2	11.4 34.1



#84
 1,2,4-Trimethylbenzene
 Concen: 0.697 ug/l
 RT: 11.756 min Scan# 1750
 Delta R.T. 0.006 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27

Tgt Ion:105	Resp:	3725
Ion	Ratio	Lower Upper
105	100	
120	40.3	21.2 63.6









#85
 sec-Butylbenzene
 Concen: 0.672 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001


Tgt Ion:105 Resp: 4385
 Ion Ratio Lower Upper
 105 100
 134 19.2 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 0.679 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27

Tgt Ion:119 Resp: 3577
 Ion Ratio Lower Upper
 119 100
 134 25.0 12.5 37.5
 91 33.1 13.8 41.4









#87
1,3-Dichlorobenzene
Concen: 0.689 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:146 Resp: 1914
Ion Ratio Lower Upper
146 100
111 43.9 22.1 66.3
148 65.3 32.1 96.5



#88
1,4-Dichlorobenzene
Concen: 0.703 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.067 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:146 Resp: 1914
Ion Ratio Lower Upper
146 100
111 43.9 21.3 63.9
148 65.3 31.9 95.5





#89

n-Butylbenzene

Concen: 0.631 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 91 Resp: 2889

Ion Ratio Lower Upper

91 100

92 48.1 26.9 80.7

134 20.1 11.8 35.3



#90

Hexachloroethane

Concen: 0.616 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 117 Resp: 604

Ion Ratio Lower Upper

117 100

201 64.2 31.6 94.7









#91
1,2-Dichlorobenzene
Concen: 0.747 ug/l
RT: 12.341 min Scan# 1846
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001


Tgt Ion:146 Resp: 2022
Ion Ratio Lower Upper
146 100
111 48.8 22.4 67.3
148 59.4 31.6 94.8



#92
1,2-Dibromo-3-Chloropropane
Concen: 0.586 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 75 Resp: 306
Ion Ratio Lower Upper
75 100
155 49.3 34.9 104.8
157 78.1 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 0.688 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:180 Resp: 1019

Ion Ratio Lower Upper

180 100

182 100.6 47.9 143.8

145 29.5 17.1 51.2



#94

Hexachlorobutadiene

Concen: 0.671 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion:225 Resp: 465


Ion Ratio Lower Upper

225 100

223 69.7 30.8 92.4

227 70.5 30.9 92.7





#95

Naphthalene

Concen: 0.767 ug/l

RT: 13.780 min Scan# 2082

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:128 Resp: 4138

Ion Ratio Lower Upper

128 100

127 12.4 10.4 15.6

129 8.9 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 0.716 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion:180 Resp: 1113

Ion Ratio Lower Upper

180 100

182 92.5 47.8 143.3

145 32.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	92588	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	162651	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	145078	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	68229	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83733	48.509	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.020%
35) Dibromofluoromethane	5.379	113	58192	49.683	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.360%
50) Toluene-d8	8.647	98	198575	48.984	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.960%
62) 4-Bromofluorobenzene	11.079	95	75568	48.596	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.200%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.167	85	78783	55.594	ug/l	99
3) Chloromethane	1.307	50	72275	52.592	ug/l	100
4) Vinyl Chloride	1.374	62	65994	51.598	ug/l	98
5) Bromomethane	1.593	94	30033	50.627	ug/l	97
6) Chloroethane	1.666	64	35690	52.271	ug/l	94
7) Trichlorofluoromethane	1.874	101	99352	52.560	ug/l	99
8) Diethyl Ether	2.130	74	31875	49.536	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.319	101	59451	50.822	ug/l	99
10) Methyl Iodide	2.447	142	73940	53.419	ug/l	99
11) Tert butyl alcohol	2.977	59	61065	252.025	ug/l	99
12) 1,1-Dichloroethene	2.313	96	56097	51.097	ug/l	98
13) Acrolein	2.233	56	64157	232.505	ug/l	99
14) Allyl chloride	2.660	41	108757	51.833	ug/l	100
15) Acrylonitrile	3.063	53	181808	262.412	ug/l	99
16) Acetone	2.380	43	169923	245.518	ug/l	98
17) Carbon Disulfide	2.502	76	132695	50.981	ug/l	99
18) Methyl Acetate	2.703	43	80251	49.968	ug/l	97
19) Methyl tert-butyl Ether	3.111	73	195442	50.777	ug/l	99
20) Methylene Chloride	2.782	84	63369	47.779	ug/l	99
21) trans-1,2-Dichloroethene	3.087	96	55770	50.513	ug/l	99
22) Diisopropyl ether	3.757	45	209946	51.800	ug/l	97
23) Vinyl Acetate	3.715	43	946746	265.585	ug/l	100
24) 1,1-Dichloroethane	3.605	63	114733	50.825	ug/l	98
25) 2-Butanone	4.556	43	258339	257.110	ug/l	97
26) 2,2-Dichloropropane	4.471	77	88229	49.934	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	67727	50.956	ug/l	100
28) Bromochloromethane	4.891	49	55390	50.975	ug/l	98
29) Tetrahydrofuran	5.001	42	163675	259.958	ug/l	99
30) Chloroform	5.093	83	120782	51.332	ug/l	97
31) Cyclohexane	5.458	56	103158	50.147	ug/l	96
32) 1,1,1-Trichloroethane	5.373	97	104243	51.108	ug/l	100
36) 1,1-Dichloropropene	5.690	75	77931	49.521	ug/l	99
37) Ethyl Acetate	4.715	43	98604	50.714	ug/l	100
38) Carbon Tetrachloride	5.666	117	88100	49.825	ug/l	97
39) Methylcyclohexane	7.373	83	100808	49.757	ug/l	98
40) Benzene	6.031	78	233977	50.759	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	56017	55.076	ug/l	100
42) 1,2-Dichloroethane	6.080	62	101185	50.861	ug/l	98
43) Isopropyl Acetate	6.336	43	154818	52.195	ug/l	99
44) Trichloroethene	7.123	130	55491	50.018	ug/l	100
45) 1,2-Dichloropropane	7.428	63	59582	51.983	ug/l	98
46) Dibromomethane	7.580	93	45181	49.978	ug/l	99
47) Bromodichloromethane	7.818	83	92388	51.889	ug/l	100
48) Methyl methacrylate	7.690	41	82144	54.226	ug/l	98
49) 1,4-Dioxane	7.659	88	30477	1059.599	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	510206	259.133	ug/l	100
52) Toluene	8.714	92	144574	51.150	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	83751	52.920	ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	92517	52.892	ug/l	98
55) 1,1,2-Trichloroethane	9.147	97	56713	50.887	ug/l	96
56) Ethyl methacrylate	9.116	69	95772	53.918	ug/l	99
57) 1,3-Dichloropropane	9.305	76	99663	49.793	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.238	63	240681	265.780	ug/l	99
59) 2-Hexanone	9.427	43	390701	268.218	ug/l	100
60) Dibromochloromethane	9.519	129	63738	52.075	ug/l	99
61) 1,2-Dibromoethane	9.610	107	59134	51.050	ug/l	98
64) Tetrachloroethene	9.269	164	48034	46.795	ug/l	98
65) Chlorobenzene	10.073	112	154640	48.699	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	53735	49.557	ug/l	100
67) Ethyl Benzene	10.189	91	281239	50.245	ug/l	100
68) m/p-Xylenes	10.299	106	206813	101.022	ug/l	99
69) o-Xylene	10.640	106	101469	50.841	ug/l	99
70) Styrene	10.653	104	169703	51.907	ug/l	99
71) Bromoform	10.799	173	40364	49.583	ug/l #	97
73) Isopropylbenzene	10.957	105	269484	50.733	ug/l	100
74) N-amyl acetate	10.842	43	137559	52.407	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	89027	47.827	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	99886	60.821	ug/l	86
77) Bromobenzene	11.195	156	59871	48.549	ug/l	98
78) n-propylbenzene	11.299	91	312285	50.562	ug/l	100
79) 2-Chlorotoluene	11.360	91	195630	49.107	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	228538	51.500	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	24857	49.276	ug/l	98
82) 4-Chlorotoluene	11.451	91	223811	50.661	ug/l	100
83) tert-Butylbenzene	11.713	119	225783	50.511	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	231371	51.485	ug/l	100
85) sec-Butylbenzene	11.890	105	280578	51.122	ug/l	99
86) p-Isopropyltoluene	12.006	119	231942	51.198	ug/l	100
87) 1,3-Dichlorobenzene	11.963	146	113146	50.272	ug/l	99
88) 1,4-Dichlorobenzene	12.037	146	111173	48.368	ug/l	99
89) n-Butylbenzene	12.329	91	205858	51.804	ug/l	100
90) Hexachloroethane	12.536	117	39823	49.895	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	111482	49.361	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21220	51.458	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	64421	49.662	ug/l	100
94) Hexachlorobutadiene	13.719	225	27425	48.408	ug/l	98
95) Naphthalene	13.774	128	239805	50.404	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	64950	48.525	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion:168 Resp: 92588
Ion Ratio Lower Upper
168 100
99 74.5 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 55.594 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 85 Resp: 78783
Ion Ratio Lower Upper
85 100
87 32.2 15.7 47.1





#3

Chloromethane

Concen: 52.592 ug/l

RT: 1.307 min Scan# 36

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVVX050525

Tgt Ion: 50 Resp: 72275

Ion Ratio Lower Upper

50 100

52 31.8 25.4 38.2



#4

Vinyl Chloride

Concen: 51.598 ug/l

RT: 1.374 min Scan# 47

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 62 Resp: 65994

Ion Ratio Lower Upper

62 100

64 32.3 25.2 37.8









#5
Bromomethane
Concen: 50.627 ug/l
RT: 1.593 min Scan# 81
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 94 Resp: 30033
Ion Ratio Lower Upper
94 100
96 93.5 72.8 109.2



#6
Chloroethane
Concen: 52.271 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 64 Resp: 35690
Ion Ratio Lower Upper
64 100
66 33.7 24.3 36.5









#7
 Trichlorofluoromethane
 Concen: 52.560 ug/l
 RT: 1.874 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50




Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525


Tgt Ion:101 Resp: 99352
 Ion Ratio Lower Upper
 101 100
 103 62.6 51.0 76.4



#8
 Diethyl Ether
 Concen: 49.536 ug/l
 RT: 2.130 min Scan# 171
 Delta R.T. -0.006 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50

Tgt Ion: 74 Resp: 31875
 Ion Ratio Lower Upper
 74 100
 45 107.1 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.822 ug/l
RT: 2.319 min Scan# 20
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion:101 Resp: 59451
Ion Ratio Lower Upper
101 100
85 47.8 38.6 58.0
151 69.4 55.2 82.8



#10
Methyl Iodide
Concen: 53.419 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:142 Resp: 73940
Ion Ratio Lower Upper
142 100
127 52.9 41.7 62.5
141 14.4 11.5 17.3









#11
Tert butyl alcohol
Concen: 252.025 ug/l
RT: 2.977 min Scan# 310
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 59 Resp: 61065
Ion Ratio Lower Upper
59 100
57 11.1 8.6 12.8



#12
1,1-Dichloroethene
Concen: 51.097 ug/l
RT: 2.313 min Scan# 201
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 96 Resp: 56097
Ion Ratio Lower Upper
96 100
61 179.7 146.2 219.2
98 62.1 51.0 76.6





#13

Acrolein

Concen: 232.505 ug/l

RT: 2.233 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVVX050525

Tgt Ion: 56 Resp: 64157

Ion Ratio Lower Upper

56 100

55 70.8 56.2 84.4



#14

Allyl chloride

Concen: 51.833 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 41 Resp: 108757


Ion Ratio Lower Upper

41 100

39 75.6 60.6 90.8

76 31.1 24.9 37.3









#15
Acrylonitrile
Concen: 262.412 ug/l
RT: 3.063 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 53 Resp: 181808
Ion Ratio Lower Upper
53 100
52 82.9 65.3 97.9
51 36.5 29.8 44.8



#16
Acetone
Concen: 245.518 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 43 Resp: 169923
Ion Ratio Lower Upper
43 100
58 27.3 21.2 31.8









#17
Carbon Disulfide
Concen: 50.981 ug/l
RT: 2.502 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 76 Resp: 132695
Ion Ratio Lower Upper
76 100
78 9.1 7.0 10.4



#18
Methyl Acetate
Concen: 49.968 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 43 Resp: 80251
Ion Ratio Lower Upper
43 100
74 22.1 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 50.777 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 73 Resp: 195442
Ion Ratio Lower Upper
73 100
57 21.8 17.7 26.5



#20
Methylene Chloride
Concen: 47.779 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 84 Resp: 63369
Ion Ratio Lower Upper
84 100
49 142.7 113.9 170.9
51 43.3 33.5 50.3
86 66.8 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 50.513 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 96 Resp: 55770
Ion Ratio Lower Upper
96 100
61 149.3 119.5 179.3
98 64.6 50.0 75.0



#22
Diisopropyl ether
Concen: 51.800 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 45 Resp: 209946
Ion Ratio Lower Upper
45 100
43 86.9 66.6 100.0
87 24.8 19.8 29.6
59 10.8 8.6 12.8









#23
 Vinyl Acetate
 Concen: 265.585 ug/l
 RT: 3.715 min Scan# 413
 Delta R.T. -0.006 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50




Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525


Tgt Ion: 43 Resp: 946746
 Ion Ratio Lower Upper
 43 100
 86 9.4 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 50.825 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50

Tgt Ion: 63 Resp: 114733
 Ion Ratio Lower Upper
 63 100
 98 6.4 3.6 10.8
 100 4.8 2.1 6.3





#25

2-Butanone

Concen: 257.110 ug/l

RT: 4.556 min Scan# 5

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVVX050525

Tgt Ion: 43 Resp: 258339

Ion Ratio Lower Upper

43 100

72 24.3 18.4 27.6



#26

2,2-Dichloropropane

Concen: 49.934 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 77 Resp: 88229

Ion Ratio Lower Upper

77 100

97 21.0 10.5 31.5









#27
 cis-1,2-Dichloroethene
 Concen: 50.956 ug/l
 RT: 4.483 min Scan# 51
 Delta R.T. 0.000 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Tgt Ion: 96 Resp: 67727




Ion	Ratio	Lower	Upper
96	100		
61	160.7	0.0	322.8
98	64.4	0.0	129.0




#28
 Bromochloromethane
 Concen: 50.975 ug/l
 RT: 4.891 min Scan# 624
 Delta R.T. -0.006 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50

Tgt Ion: 49 Resp: 55390

Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	4.0
130	68.7	56.2	84.2





#29

Tetrahydrofuran

Concen: 259.958 ug/l

RT: 5.001 min Scan# 64

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion: 42 Resp: 163675

Ion Ratio Lower Upper

42 100

72 42.0 33.7 50.5

71 38.5 31.4 47.2



#30

Chloroform

Concen: 51.332 ug/l

RT: 5.093 min Scan# 657

Delta R.T. 0.006 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 83 Resp: 120782

Ion Ratio Lower Upper

83 100

85 63.6 49.3 73.9









#31
Cyclohexane
Concen: 50.147 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 56 Resp: 103158
Ion Ratio Lower Upper
56 100
69 27.8 24.4 36.6
84 80.5 66.9 100.3



#32
1,1,1-Trichloroethane
Concen: 51.108 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 97 Resp: 104243
Ion Ratio Lower Upper
97 100
99 64.5 51.8 77.6
61 50.2 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 48.509 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVVX050525

Tgt Ion: 65 Resp: 83733

Ion Ratio Lower Upper

65 100

67 50.0 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 114 Resp: 162651


Ion Ratio Lower Upper

114 100

63 23.1 0.0 49.2

88 17.2 0.0 33.6









#35
Dibromofluoromethane
Concen: 49.683 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion:113 Resp: 58192
Ion Ratio Lower Upper
113 100
111 102.1 83.1 124.7
192 16.2 13.3 19.9



#36
1,1-Dichloropropene
Concen: 49.521 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 75 Resp: 77931
Ion Ratio Lower Upper
75 100
110 32.7 16.3 48.9
77 31.1 24.3 36.5









#37
Ethyl Acetate
Concen: 50.714 ug/l
RT: 4.715 min Scan# 595
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 43 Resp: 98604
Ion Ratio Lower Upper
43 100
61 12.6 10.3 15.5
70 10.0 7.9 11.9



#38
Carbon Tetrachloride
Concen: 49.825 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 117 Resp: 88100
Ion Ratio Lower Upper
117 100
119 90.7 75.2 112.8
121 29.4 24.2 36.4









#39
Methylcyclohexane
Concen: 49.757 ug/l
RT: 7.373 min Scan# 1031
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 83 Resp: 100808
Ion Ratio Lower Upper
83 100
55 82.7 64.7 97.1
98 48.3 37.4 56.2



#40
Benzene
Concen: 50.759 ug/l
RT: 6.031 min Scan# 811
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 78 Resp: 233977
Ion Ratio Lower Upper
78 100
77 24.0 19.0 28.4









#41
Methacrylonitrile
Concen: 55.076 ug/l
RT: 4.916 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion:	41	Resp:	56017
Ion	Ratio	Lower	Upper
41	100		
39	58.7	47.2	70.8
67	63.7	50.7	76.1
52	30.5	24.6	37.0



#42
1,2-Dichloroethane
Concen: 50.861 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:	62	Resp:	101185
Ion	Ratio	Lower	Upper
62	100		
98	7.1	0.0	15.2









#43
Isopropyl Acetate
Concen: 52.195 ug/l
RT: 6.336 min Scan# 861
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 43 Resp: 154818
Ion Ratio Lower Upper
43 100
61 17.9 14.3 21.5
87 11.4 9.5 14.3



#44
Trichloroethene
Concen: 50.018 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 130 Resp: 55491
Ion Ratio Lower Upper
130 100
95 102.6 0.0 204.2









#45
1,2-Dichloropropane
Concen: 51.983 ug/l
RT: 7.428 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 63 Resp: 59582
Ion Ratio Lower Upper
63 100
65 31.4 24.1 36.1



#46
Dibromomethane
Concen: 49.978 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 93 Resp: 45181
Ion Ratio Lower Upper
93 100
95 83.0 65.6 98.4
174 84.8 68.2 102.2









#47
Bromodichloromethane
Concen: 51.889 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 83 Resp: 92388
Ion Ratio Lower Upper
83 100
85 63.0 50.5 75.7
127 8.8 6.5 9.7



#48
Methyl methacrylate
Concen: 54.226 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 41 Resp: 82144
Ion Ratio Lower Upper
41 100
69 70.4 58.5 87.7
39 64.2 51.7 77.5





#49

1,4-Dioxane

Concen: 1059.599 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion: 88 Resp: 30477

Ion Ratio Lower Upper

88 100

43 39.9 33.4 50.2

58 76.2 58.6 88.0



#50

Toluene-d8

Concen: 48.984 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 98 Resp: 198575

Ion Ratio Lower Upper

98 100

100 65.6 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 259.133 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 43 Resp: 510206
Ion Ratio Lower Upper
43 100
58 36.1 28.9 43.3



#52
Toluene
Concen: 51.150 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 92 Resp: 144574
Ion Ratio Lower Upper
92 100
91 171.4 136.6 205.0









#53
t-1,3-Dichloropropene
Concen: 52.920 ug/l
RT: 8.976 min Scan# 1193
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 75 Resp: 83751
Ion Ratio Lower Upper
75 100
77 31.3 25.0 37.6



#54
cis-1,3-Dichloropropene
Concen: 52.892 ug/l
RT: 8.360 min Scan# 1193
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 75 Resp: 92517
Ion Ratio Lower Upper
75 100
77 31.5 25.4 38.0
39 63.2 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 50.887 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 97 Resp: 56713





Ion Ratio Lower Upper

97 100

83 89.8 69.0 103.6

85 58.4 44.3 66.5

99 63.9 47.8 71.6



#56

Ethyl methacrylate

Concen: 53.918 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 69 Resp: 95772


Ion Ratio Lower Upper

69 100

41 76.1 60.8 91.2

39 47.4 39.0 58.6





#57

1,3-Dichloropropane

Concen: 49.793 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVVX050525

Tgt Ion: 76 Resp: 99663

Ion Ratio Lower Upper

76 100

78 31.4 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 265.780 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 63 Resp: 240681

Ion Ratio Lower Upper

63 100

106 26.6 21.5 32.3









#59
2-Hexanone
Concen: 268.218 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 43 Resp: 390701
Ion Ratio Lower Upper
43 100
58 49.6 24.9 74.6



#60
Dibromochloromethane
Concen: 52.075 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 129 Resp: 63738
Ion Ratio Lower Upper
129 100
127 77.9 39.3 117.8









#61
1,2-Dibromoethane
Concen: 51.050 ug/l
RT: 9.610 min Scan# 1397
Delta R.T. 0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion:107 Resp: 59134
Ion Ratio Lower Upper
107 100
109 93.8 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 48.596 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 95 Resp: 75568
Ion Ratio Lower Upper
95 100
174 68.2 0.0 135.8
176 66.5 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument : MSVOA_X
ClientSampleId : ICVVX050525


Tgt Ion:117 Resp: 145078
Ion Ratio Lower Upper
117 100
82 62.1 50.6 76.0
119 32.9 25.8 38.6



#64
Tetrachloroethene
Concen: 46.795 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:164 Resp: 48034
Ion Ratio Lower Upper
164 100
166 129.9 105.0 157.6
129 105.3 83.5 125.3
131 98.8 76.5 114.7









#65
Chlorobenzene
Concen: 48.699 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion:112 Resp: 154640
Ion Ratio Lower Upper
112 100
114 31.1 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 49.557 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:131 Resp: 53735
Ion Ratio Lower Upper
131 100
133 94.9 47.3 141.9
119 63.7 31.6 95.0





#67

Ethyl Benzene

Concen: 50.245 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVVX050525

Tgt Ion: 91 Resp: 281239

Ion Ratio Lower Upper

91 100

106 29.1 23.4 35.2



#68

m/p-Xylenes

Concen: 101.022 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:106 Resp: 206813

Ion Ratio Lower Upper

106 100

91 215.6 171.2 256.8





#69

o-Xylene

Concen: 50.841 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVVX050525

Tgt Ion:106 Resp: 101469

Ion Ratio Lower Upper

106 100

91 227.7 112.7 338.1



#70

Styrene

Concen: 51.907 ug/l

RT: 10.653 min Scan# 1569

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion:104 Resp: 169703


Ion Ratio Lower Upper

104 100

78 57.6 45.7 68.5

103 55.5 43.7 65.5





#71

Bromoform

Concen: 49.583 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion:173 Resp: 40364

Ion Ratio Lower Upper

173 100

175 48.8 23.4 70.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion:152 Resp: 68229


Ion Ratio Lower Upper

152 100

115 93.3 46.9 140.7

150 175.3 0.0 351.0









#73
Isopropylbenzene
Concen: 50.733 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion:105 Resp: 269484
Ion Ratio Lower Upper
105 100
120 25.4 12.8 38.4



#74
N-ethyl acetate
Concen: 52.407 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 43 Resp: 137559
Ion Ratio Lower Upper
43 100
70 37.9 30.9 46.3
55 22.7 18.7 28.1
61 21.8 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 47.827 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 83 Resp: 89027
Ion Ratio Lower Upper
83 100
131 9.7 5.0 14.9
85 65.5 31.9 95.7



#76
1,2,3-Trichloropropane
Concen: 60.821 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 75 Resp: 99886
Ion Ratio Lower Upper
75 100
77 32.5 20.5 61.5





#77

Bromobenzene

Concen: 48.549 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion:156 Resp: 59871

Ion Ratio Lower Upper

156 100

77 189.0 92.8 278.3

158 98.9 48.8 146.3



#78

n-propylbenzene

Concen: 50.562 ug/l

RT: 11.299 min Scan# 1675

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 91 Resp: 312285

Ion Ratio Lower Upper

91 100

120 21.4 10.8 32.4





#79

2-Chlorotoluene

Concen: 49.107 ug/l

RT: 11.360 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVVX050525

Tgt Ion: 91 Resp: 195630

Ion Ratio Lower Upper

91 100

126 30.8 15.6 46.7



#80

1,3,5-Trimethylbenzene

Concen: 51.500 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 105 Resp: 228538

Ion Ratio Lower Upper

105 100

120 46.3 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 49.276 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion: 75 Resp: 24857
Ion Ratio Lower Upper
75 100
53 122.3 100.5 150.7
89 46.2 37.0 55.6



#82
4-Chlorotoluene
Concen: 50.661 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 91 Resp: 223811
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83

tert-Butylbenzene

Concen: 50.511 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion:119 Resp: 225783

Ion Ratio Lower Upper

119 100

91 64.9 32.9 98.7

134 22.6 11.4 34.1



#84

1,2,4-Trimethylbenzene

Concen: 51.485 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:105 Resp: 231371

Ion Ratio Lower Upper

105 100

120 42.3 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 51.122 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50




Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525


Tgt Ion:105 Resp: 280578
 Ion Ratio Lower Upper
 105 100
 134 19.1 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 51.198 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50

Tgt Ion:119 Resp: 231942
 Ion Ratio Lower Upper
 119 100
 134 25.0 12.5 37.5
 91 27.5 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 50.272 ug/l

RT: 11.963 min Scan# 1784

Delta R.T. -0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion:146 Resp: 113146

Ion Ratio Lower Upper

146 100

111 44.1 22.1 66.3

148 63.5 32.1 96.5



#88

1,4-Dichlorobenzene

Concen: 48.368 ug/l

RT: 12.037 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion:146 Resp: 111173


Ion Ratio Lower Upper

146 100

111 43.9 21.3 63.9


148 63.3 31.9 95.5








#89
n-Butylbenzene
Concen: 51.804 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525







Tgt Ion: 91 Resp: 205858
Ion Ratio Lower Upper
91 100
92 53.7 26.9 80.7
134 23.7 11.8 35.3



#90
Hexachloroethane
Concen: 49.895 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 117 Resp: 39823
Ion Ratio Lower Upper
117 100
201 62.4 31.6 94.7









#91
1,2-Dichlorobenzene
Concen: 49.361 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525


Tgt Ion:146 Resp: 111482
Ion Ratio Lower Upper
146 100
111 45.8 22.4 67.3
148 64.2 31.6 94.8



#92
1,2-Dibromo-3-Chloropropane
Concen: 51.458 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 75 Resp: 21220
Ion Ratio Lower Upper
75 100
155 68.6 34.9 104.8
157 86.7 43.8 131.4









#93
 1,2,4-Trichlorobenzene
 Concen: 49.662 ug/l
 RT: 13.585 min Scan# 2050
 Delta R.T. 0.000 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50




Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525


Tgt Ion:180 Resp: 64421
 Ion Ratio Lower Upper
 180 100
 182 95.5 47.9 143.8
 145 34.4 17.1 51.2



#94
 Hexachlorobutadiene
 Concen: 48.408 ug/l
 RT: 13.719 min Scan# 2072
 Delta R.T. -0.006 min
 Lab File: VX046048.D
 Acq: 05 May 2025 16:50

Tgt Ion:225 Resp: 27425
 Ion Ratio Lower Upper
 225 100
 223 63.1 30.8 92.4
 227 63.7 30.9 92.7









#95
Naphthalene
Concen: 50.404 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion:128 Resp: 239805
Ion Ratio Lower Upper
128 100
127 13.0 10.4 15.6
129 10.7 8.6 13.0



#96
1,2,3-Trichlorobenzene
Concen: 48.525 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:180 Resp: 64950
Ion Ratio Lower Upper
180 100
182 97.0 47.8 143.3
145 37.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046078.D
 Acq On : 08 May 2025 10:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: May 09 01:54:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	105825	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	181991	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	158314	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	73554	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	87330	44.264	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.520%
35) Dibromofluoromethane	5.373	113	60469	46.141	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.280%
50) Toluene-d8	8.641	98	209081	46.095	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.180%
62) 4-Bromofluorobenzene	11.079	95	77895	44.770	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	89.540%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.167	85	83116	51.315	ug/l	98
3) Chloromethane	1.307	50	74806	47.625	ug/l	99
4) Vinyl Chloride	1.374	62	67994	46.512	ug/l	99
5) Bromomethane	1.593	94	29775	43.914	ug/l	92
6) Chloroethane	1.666	64	37272	47.760	ug/l	96
7) Trichlorofluoromethane	1.874	101	105506	48.834	ug/l	98
8) Diethyl Ether	2.130	74	33259	45.221	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.319	101	64604	48.319	ug/l	98
10) Methyl Iodide	2.441	142	74153	46.872	ug/l	98
11) Tert butyl alcohol	2.971	59	62270	224.852	ug/l	99
12) 1,1-Dichloroethene	2.313	96	58611	46.709	ug/l	99
13) Acrolein	2.233	56	54033	171.322	ug/l	99
14) Allyl chloride	2.654	41	115880	48.320	ug/l	99
15) Acrylonitrile	3.056	53	182115	229.976	ug/l	97
16) Acetone	2.380	43	186025	235.163	ug/l	100
17) Carbon Disulfide	2.502	76	138794	46.655	ug/l	98
18) Methyl Acetate	2.703	43	84794	46.193	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	208870	47.478	ug/l	99
20) Methylene Chloride	2.782	84	65321	43.091	ug/l	98
21) trans-1,2-Dichloroethene	3.081	96	59244	46.948	ug/l	99
22) Diisopropyl ether	3.757	45	219654	47.416	ug/l	94
23) Vinyl Acetate	3.715	43	967835	237.541	ug/l	99
24) 1,1-Dichloroethane	3.599	63	120908	46.860	ug/l	98
25) 2-Butanone	4.550	43	267660	233.066	ug/l	99
26) 2,2-Dichloropropane	4.465	77	95561	47.319	ug/l	99
27) cis-1,2-Dichloroethene	4.477	96	70250	46.243	ug/l	99
28) Bromochloromethane	4.891	49	61656	49.644	ug/l	99
29) Tetrahydrofuran	4.995	42	165473	229.940	ug/l	99
30) Chloroform	5.080	83	127233	47.310	ug/l	99
31) Cyclohexane	5.458	56	109556	46.596	ug/l	97
32) 1,1,1-Trichloroethane	5.373	97	109783	47.092	ug/l	99
36) 1,1-Dichloropropene	5.684	75	81796	46.453	ug/l	100
37) Ethyl Acetate	4.702	43	99719	45.837	ug/l	100
38) Carbon Tetrachloride	5.666	117	94525	47.778	ug/l	94
39) Methylcyclohexane	7.373	83	105443	46.514	ug/l	94
40) Benzene	6.025	78	241966	46.914	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046078.D
 Acq On : 08 May 2025 10:05
 Operator : JC/MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDC0050

Quant Time: May 09 01:54:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	57552	50.572	ug/l	98
42) 1,2-Dichloroethane	6.074	62	102778	46.172	ug/l	99
43) Isopropyl Acetate	6.330	43	158841	47.860	ug/l	99
44) Trichloroethene	7.117	130	59087	47.599	ug/l	98
45) 1,2-Dichloropropane	7.421	63	61126	47.662	ug/l	98
46) Dibromomethane	7.574	93	46530	46.001	ug/l	99
47) Bromodichloromethane	7.818	83	96407	48.392	ug/l	99
48) Methyl methacrylate	7.690	41	81622	48.156	ug/l	99
49) 1,4-Dioxane	7.653	88	30726	954.733	ug/l	99
51) 4-Methyl-2-Pentanone	8.568	43	521204	236.587	ug/l	99
52) Toluene	8.714	92	147100	46.513	ug/l	98
53) t-1,3-Dichloropropene	8.970	75	86965	49.111	ug/l	99
54) cis-1,3-Dichloropropene	8.360	75	96359	49.234	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	58084	46.578	ug/l	97
56) Ethyl methacrylate	9.110	69	96373	48.490	ug/l	99
57) 1,3-Dichloropropane	9.305	76	102289	45.674	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	198555	195.960	ug/l	100
59) 2-Hexanone	9.427	43	384624	235.986	ug/l	100
60) Dibromochloromethane	9.519	129	67444	49.247	ug/l	100
61) 1,2-Dibromoethane	9.604	107	59787	46.129	ug/l	97
64) Tetrachloroethene	9.269	164	54223	48.407	ug/l	97
65) Chlorobenzene	10.073	112	156475	45.157	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	55173	46.629	ug/l	98
67) Ethyl Benzene	10.189	91	290265	47.522	ug/l	99
68) m/p-Xylenes	10.299	106	210404	94.184	ug/l	98
69) o-Xylene	10.640	106	102240	46.944	ug/l	96
70) Styrene	10.653	104	174009	48.774	ug/l	99
71) Bromoform	10.799	173	43259	48.696	ug/l #	98
73) Isopropylbenzene	10.957	105	277319	48.428	ug/l	100
74) N-amyl acetate	10.842	43	135534	47.898	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	88732	44.217	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	100720	56.889	ug/l	85
77) Bromobenzene	11.195	156	62702	47.164	ug/l	97
78) n-propylbenzene	11.299	91	325019	48.814	ug/l	100
79) 2-Chlorotoluene	11.360	91	198350	46.185	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	233158	48.737	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.012	75	25533	46.952	ug/l	99
82) 4-Chlorotoluene	11.451	91	227761	47.823	ug/l	99
83) tert-Butylbenzene	11.713	119	228755	47.471	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	235628	48.637	ug/l	100
85) sec-Butylbenzene	11.890	105	289872	48.992	ug/l	100
86) p-Isopropyltoluene	12.006	119	241375	49.423	ug/l	100
87) 1,3-Dichlorobenzene	11.963	146	115894	47.766	ug/l	99
88) 1,4-Dichlorobenzene	12.037	146	114313	46.133	ug/l	99
89) n-Butylbenzene	12.329	91	217191	50.699	ug/l	99
90) Hexachloroethane	12.536	117	41883	48.677	ug/l	100
91) 1,2-Dichlorobenzene	12.329	146	113569	46.645	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21306	47.926	ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	68569	49.033	ug/l	99
94) Hexachlorobutadiene	13.719	225	29643	48.535	ug/l	97
95) Naphthalene	13.774	128	239114	46.620	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	68552	47.508	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046078.D
Acq On : 08 May 2025 10:05
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Quant Time: May 09 01:54:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046078.D
Acq On : 08 May 2025 10:05
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Quant Time: May 09 01:54:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:168 Resp: 105825
Ion Ratio Lower Upper
168 100
99 73.0 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 51.315 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 85 Resp: 83116
Ion Ratio Lower Upper
85 100
87 32.3 15.7 47.1









#3
Chloromethane
Concen: 47.625 ug/l
RT: 1.307 min Scan# 36
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 50 Resp: 74806
Ion Ratio Lower Upper
50 100
52 31.4 25.4 38.2



#4
Vinyl Chloride
Concen: 46.512 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 62 Resp: 67994
Ion Ratio Lower Upper
62 100
64 32.1 25.2 37.8









#5
Bromomethane
Concen: 43.914 ug/l
RT: 1.593 min Scan# 81
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 94 Resp: 29775
Ion Ratio Lower Upper
94 100
96 98.4 72.8 109.2



#6
Chloroethane
Concen: 47.760 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 64 Resp: 37272
Ion Ratio Lower Upper
64 100
66 32.8 24.3 36.5









#7
 Trichlorofluoromethane
 Concen: 48.834 ug/l
 RT: 1.874 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX046078.D
 Acq: 08 May 2025 10:05




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:101 Resp: 105506
 Ion Ratio Lower Upper
 101 100
 103 65.5 51.0 76.4



#8
 Diethyl Ether
 Concen: 45.221 ug/l
 RT: 2.130 min Scan# 171
 Delta R.T. -0.006 min
 Lab File: VX046078.D
 Acq: 08 May 2025 10:05

Tgt Ion: 74 Resp: 33259
 Ion Ratio Lower Upper
 74 100
 45 108.8 54.9 164.8









#9
 1,1,2-Trichlorotrifluoroethane
 Concen: 48.319 ug/l
 RT: 2.319 min Scan# 20
 Delta R.T. 0.000 min
 Lab File: VX046078.D
 Acq: 08 May 2025 10:05




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:101 Resp: 64604
 Ion Ratio Lower Upper
 101 100
 85 46.8 38.6 58.0
 151 70.0 55.2 82.8



#10
 Methyl Iodide
 Concen: 46.872 ug/l
 RT: 2.441 min Scan# 222
 Delta R.T. -0.006 min
 Lab File: VX046078.D
 Acq: 08 May 2025 10:05

Tgt Ion:142 Resp: 74153
 Ion Ratio Lower Upper
 142 100
 127 53.4 41.7 62.5
 141 15.1 11.5 17.3









#11
Tert butyl alcohol
Concen: 224.852 ug/l
RT: 2.971 min Scan# 309
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 59 Resp: 62270
Ion Ratio Lower Upper
59 100
57 10.9 8.6 12.8



#12
1,1-Dichloroethene
Concen: 46.709 ug/l
RT: 2.313 min Scan# 201
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 96 Resp: 58611
Ion Ratio Lower Upper
96 100
61 180.8 146.2 219.2
98 64.1 51.0 76.6









#13
Acrolein
Concen: 171.322 ug/l
RT: 2.233 min Scan# 188
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 54033
Ion Ratio Lower Upper
56 100
55 71.3 56.2 84.4



#14
Allyl chloride
Concen: 48.320 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 41 Resp: 115880
Ion Ratio Lower Upper
41 100
39 74.8 60.6 90.8
76 30.5 24.9 37.3





#15

Acrylonitrile

Concen: 229.976 ug/l

RT: 3.056 min Scan# 31

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 53 Resp: 182115

Ion Ratio Lower Upper

53 100

52 84.8 65.3 97.9

51 37.7 29.8 44.8



#16

Acetone

Concen: 235.163 ug/l

RT: 2.380 min Scan# 212

Delta R.T. -0.006 min

Lab File: VX046078.D




Acq: 08 May 2025 10:05


Tgt Ion: 43 Resp: 186025

Ion Ratio Lower Upper

43 100

58 26.7 21.2 31.8









#17
Carbon Disulfide
Concen: 46.655 ug/l
RT: 2.502 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 138794
Ion Ratio Lower Upper
76 100
78 9.5 7.0 10.4



#18
Methyl Acetate
Concen: 46.193 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 43 Resp: 84794
Ion Ratio Lower Upper
43 100
74 21.9 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 47.478 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 73 Resp: 208870
Ion Ratio Lower Upper
73 100
57 21.4 17.7 26.5



#20
Methylene Chloride
Concen: 43.091 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 84 Resp: 65321
Ion Ratio Lower Upper
84 100
49 140.7 113.9 170.9
51 41.9 33.5 50.3
86 64.9 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 46.948 ug/l
RT: 3.081 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 96 Resp: 59244
Ion Ratio Lower Upper
96 100
61 148.0 119.5 179.3
98 60.7 50.0 75.0



#22
Diisopropyl ether
Concen: 47.416 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 45 Resp: 219654
Ion Ratio Lower Upper
45 100
43 75.5 66.6 100.0
87 24.8 19.8 29.6
59 10.9 8.6 12.8









#23
 Vinyl Acetate
 Concen: 237.541 ug/l
 RT: 3.715 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: VX046078.D
 Acq: 08 May 2025 10:05




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 43 Resp: 967835
 Ion Ratio Lower Upper
 43 100
 86 9.8 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 46.860 ug/l
 RT: 3.599 min Scan# 412
 Delta R.T. -0.006 min
 Lab File: VX046078.D
 Acq: 08 May 2025 10:05

Tgt Ion: 63 Resp: 120908
 Ion Ratio Lower Upper
 63 100
 98 6.5 3.6 10.8
 100 3.9 2.1 6.3









#25
2-Butanone
Concen: 233.066 ug/l
RT: 4.550 min Scan# 569
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 267660
Ion Ratio Lower Upper
43 100
72 22.6 18.4 27.6



#26
2,2-Dichloropropane
Concen: 47.319 ug/l
RT: 4.465 min Scan# 554
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 77 Resp: 95561
Ion Ratio Lower Upper
77 100
97 21.3 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 46.243 ug/l

RT: 4.477 min Scan# 51

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 96 Resp: 70250

Ion Ratio Lower Upper

96 100

61 163.7 0.0 322.8

98 65.0 0.0 129.0



#28

Bromochloromethane

Concen: 49.644 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05




Tgt Ion: 49 Resp: 61656


Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 69.1 56.2 84.2





#29

Tetrahydrofuran

Concen: 229.940 ug/l

RT: 4.995 min Scan# 641

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 42 Resp: 165473

Ion Ratio Lower Upper

42 100

72 41.6 33.7 50.5

71 38.8 31.4 47.2



#30

Chloroform

Concen: 47.310 ug/l

RT: 5.080 min Scan# 655

Delta R.T. -0.006 min

Lab File: VX046078.D




Acq: 08 May 2025 10:05


Tgt Ion: 83 Resp: 127233

Ion Ratio Lower Upper

83 100

85 62.6 49.3 73.9









#31
Cyclohexane
Concen: 46.596 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 109556
Ion Ratio Lower Upper
56 100
69 29.9 24.4 36.6
84 80.8 66.9 100.3



#32
1,1,1-Trichloroethane
Concen: 47.092 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 97 Resp: 109783
Ion Ratio Lower Upper
97 100
99 64.8 51.8 77.6
61 49.2 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 44.264 ug/l

RT: 5.946 min Scan# 798

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 65 Resp: 87330

Ion Ratio Lower Upper

65 100

67 49.4 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.751 min Scan# 929

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05




Tgt Ion: 114 Resp: 181991


Ion Ratio Lower Upper

114 100

63 22.6 0.0 49.2

88 17.6 0.0 33.6









#35
Dibromofluoromethane
Concen: 46.141 ug/l
RT: 5.373 min Scan# 704
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:113 Resp: 60469
Ion Ratio Lower Upper
113 100
111 103.0 83.1 124.7
192 16.8 13.3 19.9



#36
1,1-Dichloropropene
Concen: 46.453 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 75 Resp: 81796
Ion Ratio Lower Upper
75 100
110 32.6 16.3 48.9
77 30.9 24.3 36.5









#37
Ethyl Acetate
Concen: 45.837 ug/l
RT: 4.702 min Scan# 593
Delta R.T. -0.012 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 99719
Ion Ratio Lower Upper
43 100
61 12.8 10.3 15.5
70 9.9 7.9 11.9



#38
Carbon Tetrachloride
Concen: 47.778 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 117 Resp: 94525
Ion Ratio Lower Upper
117 100
119 100.7 75.2 112.8
121 31.7 24.2 36.4









#39
Methylcyclohexane
Concen: 46.514 ug/l
RT: 7.373 min Scan# 1031
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 105443
Ion Ratio Lower Upper
83 100
55 87.1 64.7 97.1
98 50.2 37.4 56.2



#40
Benzene
Concen: 46.914 ug/l
RT: 6.025 min Scan# 810
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 78 Resp: 241966
Ion Ratio Lower Upper
78 100
77 24.4 19.0 28.4









#41
Methacrylonitrile
Concen: 50.572 ug/l
RT: 4.916 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:	41	Resp:	57552
Ion	Ratio	Lower	Upper
41	100		
39	57.9	47.2	70.8
67	66.4	50.7	76.1
52	30.2	24.6	37.0



#42
1,2-Dichloroethane
Concen: 46.172 ug/l
RT: 6.074 min Scan# 818
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion:	62	Resp:	102778
Ion	Ratio	Lower	Upper
62	100		
98	7.3	0.0	15.2









#43
Isopropyl Acetate
Concen: 47.860 ug/l
RT: 6.330 min Scan# 861
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 158841
Ion Ratio Lower Upper
43 100
61 17.2 14.3 21.5
87 11.4 9.5 14.3



#44
Trichloroethene
Concen: 47.599 ug/l
RT: 7.117 min Scan# 989
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion:130 Resp: 59087
Ion Ratio Lower Upper
130 100
95 104.3 0.0 204.2









#45
1,2-Dichloropropane
Concen: 47.662 ug/l
RT: 7.421 min Scan# 1039
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 63 Resp: 61126
Ion Ratio Lower Upper
63 100
65 29.1 24.1 36.1



#46
Dibromomethane
Concen: 46.001 ug/l
RT: 7.574 min Scan# 1064
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 93 Resp: 46530
Ion Ratio Lower Upper
93 100
95 83.1 65.6 98.4
174 85.7 68.2 102.2









#47
Bromodichloromethane
Concen: 48.392 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 96407
Ion Ratio Lower Upper
83 100
85 63.8 50.5 75.7
127 8.2 6.5 9.7



#48
Methyl methacrylate
Concen: 48.156 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 41 Resp: 81622
Ion Ratio Lower Upper
41 100
69 72.1 58.5 87.7
39 64.3 51.7 77.5









#49
1,4-Dioxane
Concen: 954.733 ug/l
RT: 7.653 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 88 Resp: 30726
Ion Ratio Lower Upper
88 100
43 41.0 33.4 50.2
58 74.6 58.6 88.0



#50
Toluene-d8
Concen: 46.095 ug/l
RT: 8.641 min Scan# 1239
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 98 Resp: 209081
Ion Ratio Lower Upper
98 100
100 64.9 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 236.587 ug/l
RT: 8.568 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 521204
Ion Ratio Lower Upper
43 100
58 35.3 28.9 43.3



#52
Toluene
Concen: 46.513 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 92 Resp: 147100
Ion Ratio Lower Upper
92 100
91 173.1 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 49.111 ug/l

RT: 8.970 min Scan# 1129

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 75 Resp: 86965

Ion Ratio Lower Upper

75 100

77 32.0 25.0 37.6



#54

cis-1,3-Dichloropropene

Concen: 49.234 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05




Tgt Ion: 75 Resp: 96359


Ion Ratio Lower Upper

75 100

77 29.1 25.4 38.0

39 61.0 52.2 78.4









#55
 1,1,2-Trichloroethane
 Concen: 46.578 ug/l
 RT: 9.147 min Scan# 1322
 Delta R.T. 0.000 min
 Lab File: VX046078.D
 Acq: 08 May 2025 10:05




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:	97	Resp:	58084
Ion	Ratio	Lower	Upper
97	100		
83	87.7	69.0	103.6
85	56.5	44.3	66.5
99	63.5	47.8	71.6



#56
 Ethyl methacrylate
 Concen: 48.490 ug/l
 RT: 9.110 min Scan# 1316
 Delta R.T. -0.006 min
 Lab File: VX046078.D
 Acq: 08 May 2025 10:05

Tgt Ion:	69	Resp:	96373
Ion	Ratio	Lower	Upper
69	100		
41	76.7	60.8	91.2
39	48.5	39.0	58.6





#57

1,3-Dichloropropane

Concen: 45.674 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 76 Resp: 102289

Ion Ratio Lower Upper

76 100

78 32.2 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 195.960 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046078.D




Acq: 08 May 2025 10:05


Tgt Ion: 63 Resp: 198555

Ion Ratio Lower Upper

63 100

106 27.0 21.5 32.3





#59

2-Hexanone

Concen: 235.986 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 43 Resp: 384624

Ion Ratio Lower Upper

43 100

58 49.6 24.9 74.6



#60

Dibromochloromethane

Concen: 49.247 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046078.D




Acq: 08 May 2025 10:05


Tgt Ion: 129 Resp: 67444

Ion Ratio Lower Upper

129 100

127 78.1 39.3 117.8









#61
1,2-Dibromoethane
Concen: 46.129 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:107 Resp: 59787
Ion Ratio Lower Upper
107 100
109 94.4 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 44.770 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 95 Resp: 77895
Ion Ratio Lower Upper
95 100
174 67.6 0.0 135.8
176 65.4 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:117 Resp: 158314
Ion Ratio Lower Upper
117 100
82 60.1 50.6 76.0
119 31.4 25.8 38.6



#64
Tetrachloroethene
Concen: 48.407 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion:164 Resp: 54223
Ion Ratio Lower Upper
164 100
166 127.0 105.0 157.6
129 100.9 83.5 125.3
131 96.5 76.5 114.7









#65
Chlorobenzene
Concen: 45.157 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:112 Resp: 156475
Ion Ratio Lower Upper
112 100
114 32.1 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 46.629 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion:131 Resp: 55173
Ion Ratio Lower Upper
131 100
133 96.5 47.3 141.9
119 64.4 31.6 95.0









#67
Ethyl Benzene
Concen: 47.522 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 290265
Ion Ratio Lower Upper
91 100
106 29.7 23.4 35.2



#68
m/p-Xylenes
Concen: 94.184 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 106 Resp: 210404
Ion Ratio Lower Upper
106 100
91 216.6 171.2 256.8





#69

o-Xylene

Concen: 46.944 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion:106 Resp: 102240

Ion Ratio Lower Upper

106 100

91 232.4 112.7 338.1



#70

Styrene

Concen: 48.774 ug/l

RT: 10.653 min Scan# 1569

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05




Tgt Ion:104 Resp: 174009


Ion Ratio Lower Upper

104 100

78 57.7 45.7 68.5

103 55.0 43.7 65.5









#71
Bromoform
Concen: 48.696 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:173 Resp: 43259
Ion Ratio Lower Upper
173 100
175 48.1 23.4 70.0
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion:152 Resp: 73554
Ion Ratio Lower Upper
152 100
115 91.0 46.9 140.7
150 174.9 0.0 351.0









#73
Isopropylbenzene
Concen: 48.428 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:105 Resp: 277319
Ion Ratio Lower Upper
105 100
120 25.5 12.8 38.4



#74
N-ethyl acetate
Concen: 47.898 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 43 Resp: 135534
Ion Ratio Lower Upper
43 100
70 37.6 30.9 46.3
55 23.1 18.7 28.1
61 22.0 17.1 25.7









#75
 1,1,2-Tetrachloroethane
 Concen: 44.217 ug/l
 RT: 11.207 min Scan# 1660
 Delta R.T. 0.000 min
 Lab File: VX046078.D
 Acq: 08 May 2025 10:05




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 83 Resp: 88732
 Ion Ratio Lower Upper
 83 100
 131 9.7 5.0 14.9
 85 64.0 31.9 95.7



#76
 1,2,3-Trichloropropane
 Concen: 56.889 ug/l
 RT: 11.238 min Scan# 1665
 Delta R.T. 0.000 min
 Lab File: VX046078.D
 Acq: 08 May 2025 10:05

Tgt Ion: 75 Resp: 100720
 Ion Ratio Lower Upper
 75 100
 77 31.5 20.5 61.5









#77
Bromobenzene
Concen: 47.164 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:156 Resp: 62702
Ion Ratio Lower Upper
156 100
77 181.9 92.8 278.3
158 93.6 48.8 146.3



#78
n-propylbenzene
Concen: 48.814 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 91 Resp: 325019
Ion Ratio Lower Upper
91 100
120 21.8 10.8 32.4









#79
2-Chlorotoluene
Concen: 46.185 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 198350
Ion Ratio Lower Upper
91 100
126 31.2 15.6 46.7



#80
1,3,5-Trimethylbenzene
Concen: 48.737 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 105 Resp: 233158
Ion Ratio Lower Upper
105 100
120 47.0 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 46.952 ug/l
RT: 11.012 min Scan# 1629
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 25533
Ion Ratio Lower Upper
75 100
53 124.6 100.5 150.7
89 47.1 37.0 55.6



#82
4-Chlorotoluene
Concen: 47.823 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 91 Resp: 227761
Ion Ratio Lower Upper
91 100
126 26.9 13.3 39.8









#83
 tert-Butylbenzene
 Concen: 47.471 ug/l
 RT: 11.713 min Scan# 1743
 Delta R.T. 0.000 min
 Lab File: VX046078.D
 Acq: 08 May 2025 10:05




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:119 Resp: 228755
 Ion Ratio Lower Upper
 119 100
 91 66.0 32.9 98.7
 134 22.5 11.4 34.1



#84
 1,2,4-Trimethylbenzene
 Concen: 48.637 ug/l
 RT: 11.750 min Scan# 1749
 Delta R.T. 0.000 min
 Lab File: VX046078.D
 Acq: 08 May 2025 10:05

Tgt Ion:105 Resp: 235628
 Ion Ratio Lower Upper
 105 100
 120 42.4 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 48.992 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX046078.D
 Acq: 08 May 2025 10:05




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:105 Resp: 289872
 Ion Ratio Lower Upper
 105 100
 134 19.2 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 49.423 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX046078.D
 Acq: 08 May 2025 10:05

Tgt Ion:119 Resp: 241375
 Ion Ratio Lower Upper
 119 100
 134 25.1 12.5 37.5
 91 27.6 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 47.766 ug/l

RT: 11.963 min Scan# 1784

Delta R.T. -0.006 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:146 Resp: 115894

Ion Ratio Lower Upper

146 100

111 44.6 22.1 66.3

148 63.1 32.1 96.5



#88

1,4-Dichlorobenzene

Concen: 46.133 ug/l

RT: 12.037 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05




Tgt Ion:146 Resp: 114313


Ion Ratio Lower Upper

146 100

111 43.2 21.3 63.9

148 64.6 31.9 95.5





#89

n-Butylbenzene

Concen: 50.699 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046078.D

Acq: 08 May 2025 10:05

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 91 Resp: 217191

Ion Ratio Lower Upper

91 100

92 54.1 26.9 80.7

134 23.9 11.8 35.3



#90

Hexachloroethane

Concen: 48.677 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046078.D




Acq: 08 May 2025 10:05


Tgt Ion: 117 Resp: 41883

Ion Ratio Lower Upper

117 100

201 63.1 31.6 94.7









#91
1,2-Dichlorobenzene
Concen: 46.645 ug/l
RT: 12.329 min Scan# 1845
Delta R.T. -0.006 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:146 Resp: 113569
Ion Ratio Lower Upper
146 100
111 45.8 22.4 67.3
148 63.6 31.6 94.8



#92
1,2-Dibromo-3-Chloropropane
Concen: 47.926 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion: 75 Resp: 21306
Ion Ratio Lower Upper
75 100
155 70.5 34.9 104.8
157 87.3 43.8 131.4









#93
 1,2,4-Trichlorobenzene
 Concen: 49.033 ug/l
 RT: 13.585 min Scan# 2050
 Delta R.T. 0.000 min
 Lab File: VX046078.D
 Acq: 08 May 2025 10:05




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:180 Resp: 68569
 Ion Ratio Lower Upper
 180 100
 182 95.5 47.9 143.8
 145 34.6 17.1 51.2



#94
 Hexachlorobutadiene
 Concen: 48.535 ug/l
 RT: 13.719 min Scan# 2072
 Delta R.T. -0.006 min
 Lab File: VX046078.D
 Acq: 08 May 2025 10:05

Tgt Ion:225 Resp: 29643
 Ion Ratio Lower Upper
 225 100
 223 63.5 30.8 92.4
 227 63.9 30.9 92.7









#95
Naphthalene
Concen: 46.620 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:128 Resp: 239114
Ion Ratio Lower Upper
128 100
127 13.2 10.4 15.6
129 11.1 8.6 13.0



#96
1,2,3-Trichlorobenzene
Concen: 47.508 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046078.D
Acq: 08 May 2025 10:05

Tgt Ion:180 Resp: 68552
Ion Ratio Lower Upper
180 100
182 96.0 47.8 143.3
145 36.3 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046081.D
 Acq On : 08 May 2025 11:20
 Operator : JC/MD
 Sample : VX0508WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBS01

Quant Time: May 09 01:55:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	97732	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	172839	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	147787	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	68652	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	90728	49.795	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.580%
35) Dibromofluoromethane	5.379	113	62462	50.185	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.380%
50) Toluene-d8	8.647	98	211573	49.114	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.220%
62) 4-Bromofluorobenzene	11.079	95	79658	48.207	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.420%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	29711	19.862	ug/l	96
3) Chloromethane	1.313	50	26941	18.572	ug/l	98
4) Vinyl Chloride	1.374	62	24955	18.484	ug/l	99
5) Bromomethane	1.599	94	11222	17.921	ug/l	100
6) Chloroethane	1.672	64	13881	19.260	ug/l	95
7) Trichlorofluoromethane	1.874	101	38496	19.294	ug/l	100
8) Diethyl Ether	2.130	74	12418	18.283	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.319	101	22890	18.538	ug/l	97
10) Methyl Iodide	2.447	142	26968	18.458	ug/l	99
11) Tert butyl alcohol	2.971	59	23936	93.588	ug/l	100
12) 1,1-Dichloroethene	2.312	96	21341	18.416	ug/l	95
13) Acrolein	2.233	56	27935	95.908	ug/l	98
14) Allyl chloride	2.654	41	41968	18.949	ug/l	98
15) Acrylonitrile	3.062	53	71056	97.160	ug/l	99
16) Acetone	2.380	43	67906	92.952	ug/l	98
17) Carbon Disulfide	2.508	76	47650	17.344	ug/l	100
18) Methyl Acetate	2.703	43	33048	19.494	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	78579	19.341	ug/l	98
20) Methylene Chloride	2.788	84	24521	17.515	ug/l	91
21) trans-1,2-Dichloroethene	3.087	96	21957	18.841	ug/l	96
22) Diisopropyl ether	3.763	45	82904	19.378	ug/l	97
23) Vinyl Acetate	3.721	43	357968	95.133	ug/l	99
24) 1,1-Dichloroethane	3.605	63	46330	19.443	ug/l	97
25) 2-Butanone	4.556	43	102390	96.539	ug/l	97
26) 2,2-Dichloropropane	4.465	77	33519	17.972	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	27420	19.544	ug/l	97
28) Bromochloromethane	4.891	49	23913	20.849	ug/l	99
29) Tetrahydrofuran	5.007	42	64605	97.209	ug/l	100
30) Chloroform	5.086	83	49183	19.803	ug/l	99
31) Cyclohexane	5.464	56	41483	19.104	ug/l	95
32) 1,1,1-Trichloroethane	5.379	97	41440	19.248	ug/l	99
36) 1,1-Dichloropropene	5.684	75	30344	18.145	ug/l	99
37) Ethyl Acetate	4.715	43	37823	18.307	ug/l	99
38) Carbon Tetrachloride	5.672	117	35426	18.854	ug/l	98
39) Methylcyclohexane	7.373	83	39142	18.181	ug/l	97
40) Benzene	6.031	78	94792	19.352	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046081.D
 Acq On : 08 May 2025 11:20
 Operator : JC/MD
 Sample : VX0508WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBS01

Quant Time: May 09 01:55:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	22072	20.422	ug/l	97
42) 1,2-Dichloroethane	6.086	62	40044	18.942	ug/l	100
43) Isopropyl Acetate	6.342	43	61195	19.415	ug/l	98
44) Trichloroethene	7.123	130	21873	18.553	ug/l	100
45) 1,2-Dichloropropane	7.427	63	23731	19.484	ug/l	100
46) Dibromomethane	7.580	93	18545	19.305	ug/l	97
47) Bromodichloromethane	7.818	83	36386	19.231	ug/l	96
48) Methyl methacrylate	7.696	41	31721	19.706	ug/l	99
49) 1,4-Dioxane	7.659	88	12727	416.399	ug/l	99
51) 4-Methyl-2-Pentanone	8.574	43	204540	97.762	ug/l	98
52) Toluene	8.714	92	57389	19.107	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	29985	17.830	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	34522	18.573	ug/l	96
55) 1,1,2-Trichloroethane	9.147	97	22879	19.319	ug/l	98
56) Ethyl methacrylate	9.116	69	36760	19.475	ug/l	98
57) 1,3-Dichloropropane	9.305	76	41307	19.421	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	91264	94.841	ug/l	98
59) 2-Hexanone	9.427	43	151942	98.160	ug/l	100
60) Dibromochloromethane	9.519	129	25083	19.285	ug/l	98
61) 1,2-Dibromoethane	9.610	107	23506	19.096	ug/l	96
64) Tetrachloroethene	9.269	164	19802	18.937	ug/l	96
65) Chlorobenzene	10.079	112	60865	18.816	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	20947	18.964	ug/l	98
67) Ethyl Benzene	10.189	91	111800	19.608	ug/l	100
68) m/p-Xylenes	10.299	106	81709	39.181	ug/l	98
69) o-Xylene	10.640	106	40954	20.144	ug/l	98
70) Styrene	10.652	104	66398	19.937	ug/l	99
71) Bromoform	10.799	173	15535	18.733	ug/l #	96
73) Isopropylbenzene	10.957	105	108018	20.210	ug/l	99
74) N-amyl acetate	10.841	43	51097	19.347	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	36452	19.462	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	39007	23.605	ug/l	87
77) Bromobenzene	11.195	156	24434	19.691	ug/l	97
78) n-propylbenzene	11.299	91	123628	19.893	ug/l	99
79) 2-Chlorotoluene	11.360	91	77625	19.365	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	90424	20.251	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	8871	17.477	ug/l	95
82) 4-Chlorotoluene	11.451	91	87403	19.662	ug/l	99
83) tert-Butylbenzene	11.713	119	90081	20.028	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	90858	20.093	ug/l	100
85) sec-Butylbenzene	11.890	105	109461	19.821	ug/l	99
86) p-Isopropyltoluene	12.006	119	89406	19.614	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	44061	19.456	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	44060	19.051	ug/l	99
89) n-Butylbenzene	12.329	91	75618	18.912	ug/l	99
90) Hexachloroethane	12.536	117	15534	19.343	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	45764	20.138	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.939	75	7740	18.654	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	24066	18.438	ug/l	98
94) Hexachlorobutadiene	13.725	225	10658	18.697	ug/l	97
95) Naphthalene	13.774	128	87722	18.324	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	25237	18.739	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046081.D
Acq On : 08 May 2025 11:20
Operator : JC/MD
Sample : VX0508WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Quant Time: May 09 01:55:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 09 01:55:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion:168 Resp: 97732
Ion Ratio Lower Upper
168 100
99 70.8 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 19.862 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 85 Resp: 29711
Ion Ratio Lower Upper
85 100
87 33.4 15.7 47.1









#3
Chloromethane
Concen: 18.572 ug/l
RT: 1.313 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 50 Resp: 26941
Ion Ratio Lower Upper
50 100
52 32.9 25.4 38.2



#4
Vinyl Chloride
Concen: 18.484 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 62 Resp: 24955
Ion Ratio Lower Upper
62 100
64 31.0 25.2 37.8









#5
Bromomethane
Concen: 17.921 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 94 Resp: 11222
Ion Ratio Lower Upper
94 100
96 90.7 72.8 109.2



#6
Chloroethane
Concen: 19.260 ug/l
RT: 1.672 min Scan# 96
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 64 Resp: 13881
Ion Ratio Lower Upper
64 100
66 33.1 24.3 36.5









#7
Trichlorofluoromethane
Concen: 19.294 ug/l
RT: 1.874 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion:101 Resp: 38496
Ion Ratio Lower Upper
101 100
103 64.0 51.0 76.4



#8
Diethyl Ether
Concen: 18.283 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 74 Resp: 12418
Ion Ratio Lower Upper
74 100
45 108.5 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.538 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion:	101	Resp:	22890
Ion	Ratio	Lower	Upper
101	100		
85	48.2	38.6	58.0
151	73.4	55.2	82.8



#10
Methyl Iodide
Concen: 18.458 ug/l
RT: 2.447 min Scan# 223
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20


Tgt Ion:	142	Resp:	26968
Ion	Ratio	Lower	Upper
142	100		
127	52.7	41.7	62.5
141	13.8	11.5	17.3








#11
Tert butyl alcohol
Concen: 93.588 ug/l
RT: 2.971 min Scan# 309
Delta R.T. -0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01







Tgt Ion: 59 Resp: 23936
Ion Ratio Lower Upper
59 100
57 10.7 8.6 12.8



#12
1,1-Dichloroethene
Concen: 18.416 ug/l
RT: 2.312 min Scan# 201
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 96 Resp: 21341
Ion Ratio Lower Upper
96 100
61 190.8 146.2 219.2
98 67.7 51.0 76.6





#13

Acrolein

Concen: 95.908 ug/l

RT: 2.233 min Scan# 188

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :





VX0508WBS01

Tgt Ion: 56 Resp: 27935

Ion Ratio Lower Upper

56 100

55 71.9 56.2 84.4



#14

Allyl chloride

Concen: 18.949 ug/l

RT: 2.654 min Scan# 257

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20




Tgt Ion: 41 Resp: 41968


Ion Ratio Lower Upper

41 100

39 74.5 60.6 90.8

76 29.7 24.9 37.3









#15
Acrylonitrile
Concen: 97.160 ug/l
RT: 3.062 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 53 Resp: 71056
Ion Ratio Lower Upper
53 100
52 82.8 65.3 97.9
51 37.9 29.8 44.8



#16
Acetone
Concen: 92.952 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 43 Resp: 67906
Ion Ratio Lower Upper
43 100
58 25.7 21.2 31.8









#17
Carbon Disulfide
Concen: 17.344 ug/l
RT: 2.508 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 76 Resp: 47650
Ion Ratio Lower Upper
76 100
78 8.6 7.0 10.4



#18
Methyl Acetate
Concen: 19.494 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 43 Resp: 33048
Ion Ratio Lower Upper
43 100
74 21.3 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 19.341 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 73 Resp: 78579
Ion Ratio Lower Upper
73 100
57 21.3 17.7 26.5



#20
Methylene Chloride
Concen: 17.515 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 84 Resp: 24521
Ion Ratio Lower Upper
84 100
49 126.7 113.9 170.9
51 38.1 33.5 50.3
86 65.1 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 18.841 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 96 Resp: 21957
Ion Ratio Lower Upper
96 100
61 154.1 119.5 179.3
98 66.6 50.0 75.0



#22
Diisopropyl ether
Concen: 19.378 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 45 Resp: 82904
Ion Ratio Lower Upper
45 100
43 80.3 66.6 100.0
87 24.8 19.8 29.6
59 11.5 8.6 12.8









#23
 Vinyl Acetate
 Concen: 95.133 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046081.D
 Acq: 08 May 2025 11:20




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBS01


Tgt Ion: 43 Resp: 357968
 Ion Ratio Lower Upper
 43 100
 86 9.7 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 19.443 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046081.D
 Acq: 08 May 2025 11:20

Tgt Ion: 63 Resp: 46330
 Ion Ratio Lower Upper
 63 100
 98 5.6 3.6 10.8
 100 4.4 2.1 6.3









#25
2-Butanone
Concen: 96.539 ug/l
RT: 4.556 min Scan# 569
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 43 Resp: 102390
Ion Ratio Lower Upper
43 100
72 21.6 18.4 27.6



#26
2,2-Dichloropropane
Concen: 17.972 ug/l
RT: 4.465 min Scan# 554
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 77 Resp: 33519
Ion Ratio Lower Upper
77 100
97 21.4 10.5 31.5









#27
 cis-1,2-Dichloroethene
 Concen: 19.544 ug/l
 RT: 4.483 min Scan# 51
 Delta R.T. -0.000 min
 Lab File: VX046081.D
 Acq: 08 May 2025 11:20




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBS01


Tgt Ion: 96 Resp: 27420
 Ion Ratio Lower Upper
 96 100
 61 157.7 0.0 322.8
 98 61.6 0.0 129.0



#28
 Bromochloromethane
 Concen: 20.849 ug/l
 RT: 4.891 min Scan# 624
 Delta R.T. -0.006 min
 Lab File: VX046081.D
 Acq: 08 May 2025 11:20

Tgt Ion: 49 Resp: 23913
 Ion Ratio Lower Upper
 49 100
 129 1.7 0.0 4.0
 130 69.4 56.2 84.2









#29
Tetrahydrofuran
Concen: 97.209 ug/l
RT: 5.007 min Scan# 64
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 42 Resp: 64605
Ion Ratio Lower Upper
42 100
72 41.8 33.7 50.5
71 39.2 31.4 47.2



#30
Chloroform
Concen: 19.803 ug/l
RT: 5.086 min Scan# 656
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 83 Resp: 49183
Ion Ratio Lower Upper
83 100
85 62.7 49.3 73.9









#31
Cyclohexane
Concen: 19.104 ug/l
RT: 5.464 min Scan# 718
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion:	56	Resp:	41483
Ion	Ratio	Lower	Upper
56	100		
69	31.5	24.4	36.6
84	78.6	66.9	100.3



#32
1,1,1-Trichloroethane
Concen: 19.248 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion:	97	Resp:	41440
Ion	Ratio	Lower	Upper
97	100		
99	64.0	51.8	77.6
61	51.2	40.1	60.1





#33

1,2-Dichloroethane-d4

Concen: 49.795 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :





VX0508WBS01

Tgt Ion: 65 Resp: 90728

Ion Ratio Lower Upper

65 100

67 50.3 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20




Tgt Ion: 114 Resp: 172839


Ion Ratio Lower Upper

114 100

63 24.7 0.0 49.2

88 17.8 0.0 33.6









#35
 Dibromofluoromethane
 Concen: 50.185 ug/l
 RT: 5.379 min Scan# 704
 Delta R.T. -0.000 min
 Lab File: VX046081.D
 Acq: 08 May 2025 11:20




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBS01


Tgt Ion:113 Resp: 62462
 Ion Ratio Lower Upper
 113 100
 111 102.1 83.1 124.7
 192 17.2 13.3 19.9



#36
 1,1-Dichloropropene
 Concen: 18.145 ug/l
 RT: 5.684 min Scan# 754
 Delta R.T. -0.000 min
 Lab File: VX046081.D
 Acq: 08 May 2025 11:20

Tgt Ion: 75 Resp: 30344
 Ion Ratio Lower Upper
 75 100
 110 32.6 16.3 48.9
 77 32.0 24.3 36.5









#37
Ethyl Acetate
Concen: 18.307 ug/l
RT: 4.715 min Scan# 595
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 43 Resp: 37823
Ion Ratio Lower Upper
43 100
61 12.8 10.3 15.5
70 10.5 7.9 11.9



#38
Carbon Tetrachloride
Concen: 18.854 ug/l
RT: 5.672 min Scan# 752
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 117 Resp: 35426
Ion Ratio Lower Upper
117 100
119 96.0 75.2 112.8
121 29.9 24.2 36.4









#39
Methylcyclohexane
Concen: 18.181 ug/l
RT: 7.373 min Scan# 1031
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 83 Resp: 39142
Ion Ratio Lower Upper
83 100
55 84.7 64.7 97.1
98 47.8 37.4 56.2



#40
Benzene
Concen: 19.352 ug/l
RT: 6.031 min Scan# 811
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 78 Resp: 94792
Ion Ratio Lower Upper
78 100
77 23.9 19.0 28.4









#41
Methacrylonitrile
Concen: 20.422 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 41	Resp: 22072
Ion Ratio	Lower Upper
41 100	
39 56.0	47.2 70.8
67 65.5	50.7 76.1
52 30.4	24.6 37.0



#42
1,2-Dichloroethane
Concen: 18.942 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 62	Resp: 40044
Ion Ratio	Lower Upper
62 100	
98 7.6	0.0 15.2









#43
Isopropyl Acetate
Concen: 19.415 ug/l
RT: 6.342 min Scan# 861
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 43 Resp: 61195
Ion Ratio Lower Upper
43 100
61 17.5 14.3 21.5
87 10.9 9.5 14.3



#44
Trichloroethene
Concen: 18.553 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 130 Resp: 21873
Ion Ratio Lower Upper
130 100
95 102.0 0.0 204.2





#45

1,2-Dichloropropane

Concen: 19.484 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :





VX0508WBS01

Tgt Ion: 63 Resp: 23731

Ion Ratio Lower Upper

63 100

65 30.0 24.1 36.1



#46

Dibromomethane

Concen: 19.305 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20




Tgt Ion: 93 Resp: 18545


Ion Ratio Lower Upper

93 100

95 80.6 65.6 98.4

174 81.5 68.2 102.2









#47
Bromodichloromethane
Concen: 19.231 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion:	83	Resp:	36386
Ion	Ratio	Lower	Upper
83	100		
85	66.4	50.5	75.7
127	8.3	6.5	9.7



#48
Methyl methacrylate
Concen: 19.706 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion:	41	Resp:	31721
Ion	Ratio	Lower	Upper
41	100		
69	72.4	58.5	87.7
39	63.7	51.7	77.5









#49
1,4-Dioxane
Concen: 416.399 ug/l
RT: 7.659 min Scan# 1078
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 88 Resp: 12727
Ion Ratio Lower Upper
88 100
43 41.1 33.4 50.2
58 73.7 58.6 88.0



#50
Toluene-d8
Concen: 49.114 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 98 Resp: 211573
Ion Ratio Lower Upper
98 100
100 64.4 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 97.762 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 43 Resp: 204540
Ion Ratio Lower Upper
43 100
58 35.2 28.9 43.3



#52
Toluene
Concen: 19.107 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 92 Resp: 57389
Ion Ratio Lower Upper
92 100
91 172.1 136.6 205.0









#53
 t-1,3-Dichloropropene
 Concen: 17.830 ug/l
 RT: 8.976 min Scan# 1194
 Delta R.T. -0.000 min
 Lab File: VX046081.D
 Acq: 08 May 2025 11:20




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBS01


Tgt Ion: 75 Resp: 29985
 Ion Ratio Lower Upper
 75 100
 77 30.1 25.0 37.6



#54
 cis-1,3-Dichloropropene
 Concen: 18.573 ug/l
 RT: 8.366 min Scan# 1194
 Delta R.T. 0.006 min
 Lab File: VX046081.D
 Acq: 08 May 2025 11:20

Tgt Ion: 75 Resp: 34522
 Ion Ratio Lower Upper
 75 100
 77 33.4 25.4 38.0
 39 62.1 52.2 78.4









#55
1,1,2-Trichloroethane
Concen: 19.319 ug/l
RT: 9.147 min Scan# 1322
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion:	97	Resp:	22879
Ion	Ratio	Lower	Upper
97	100		
83	85.3	69.0	103.6
85	57.4	44.3	66.5
99	61.4	47.8	71.6



#56
Ethyl methacrylate
Concen: 19.475 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion:	69	Resp:	36760
Ion	Ratio	Lower	Upper
69	100		
41	78.1	60.8	91.2
39	48.9	39.0	58.6





#57

1,3-Dichloropropane

Concen: 19.421 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :





VX0508WBS01

Tgt Ion: 76 Resp: 41307

Ion Ratio Lower Upper

76 100

78 31.8 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 94.841 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046081.D




Acq: 08 May 2025 11:20


Tgt Ion: 63 Resp: 91264

Ion Ratio Lower Upper

63 100

106 26.1 21.5 32.3









#59
2-Hexanone
Concen: 98.160 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 43 Resp: 151942
Ion Ratio Lower Upper
43 100
58 49.5 24.9 74.6



#60
Dibromochloromethane
Concen: 19.285 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 129 Resp: 25083
Ion Ratio Lower Upper
129 100
127 77.2 39.3 117.8









#61
1,2-Dibromoethane
Concen: 19.096 ug/l
RT: 9.610 min Scan# 1397
Delta R.T. 0.006 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion:107 Resp: 23506
Ion Ratio Lower Upper
107 100
109 95.4 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 48.207 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 95 Resp: 79658
Ion Ratio Lower Upper
95 100
174 68.4 0.0 135.8
176 65.4 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion:117 Resp: 147787
Ion Ratio Lower Upper
117 100
82 64.6 50.6 76.0
119 32.2 25.8 38.6



#64
Tetrachloroethene
Concen: 18.937 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion:164 Resp: 19802
Ion Ratio Lower Upper
164 100
166 124.3 105.0 157.6
129 99.9 83.5 125.3
131 95.2 76.5 114.7









#65
Chlorobenzene
Concen: 18.816 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion:112 Resp: 60865
Ion Ratio Lower Upper
112 100
114 31.5 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 18.964 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion:131 Resp: 20947
Ion Ratio Lower Upper
131 100
133 95.7 47.3 141.9
119 65.7 31.6 95.0





#67

Ethyl Benzene

Concen: 19.608 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :





VX0508WBS01

Tgt Ion: 91 Resp: 111800

Ion Ratio Lower Upper

91 100

106 29.3 23.4 35.2



#68

m/p-Xylenes

Concen: 39.181 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046081.D




Acq: 08 May 2025 11:20


Tgt Ion: 106 Resp: 81709

Ion Ratio Lower Upper

106 100

91 216.4 171.2 256.8









#69
o-Xylene
Concen: 20.144 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion:106 Resp: 40954
Ion Ratio Lower Upper
106 100
91 228.6 112.7 338.1



#70
Styrene
Concen: 19.937 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion:104 Resp: 66398
Ion Ratio Lower Upper
104 100
78 57.6 45.7 68.5
103 55.5 43.7 65.5





#71

Bromoform

Concen: 18.733 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01





Tgt Ion:173 Resp: 15535

Ion Ratio Lower Upper

173 100

175 49.4 23.4 70.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20




Tgt Ion:152 Resp: 68652


Ion Ratio Lower Upper

152 100

115 75.6 46.9 140.7

150 162.3 0.0 351.0









#73
Isopropylbenzene
Concen: 20.210 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion:105 Resp: 108018
Ion Ratio Lower Upper
105 100
120 25.2 12.8 38.4



#74
N-ethyl acetate
Concen: 19.347 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 43 Resp: 51097
Ion Ratio Lower Upper
43 100
70 38.1 30.9 46.3
55 23.4 18.7 28.1
61 22.0 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 19.462 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 83 Resp: 36452
Ion Ratio Lower Upper
83 100
131 10.0 5.0 14.9
85 63.8 31.9 95.7



#76
1,2,3-Trichloropropane
Concen: 23.605 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 75 Resp: 39007
Ion Ratio Lower Upper
75 100
77 33.1 20.5 61.5









#77
Bromobenzene
Concen: 19.691 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion:156 Resp: 24434
Ion Ratio Lower Upper
156 100
77 181.1 92.8 278.3
158 94.6 48.8 146.3



#78
n-propylbenzene
Concen: 19.893 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 91 Resp: 123628
Ion Ratio Lower Upper
91 100
120 22.2 10.8 32.4









#79
2-Chlorotoluene
Concen: 19.365 ug/l
RT: 11.360 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 91 Resp: 77625
Ion Ratio Lower Upper
91 100
126 30.9 15.6 46.7



#80
1,3,5-Trimethylbenzene
Concen: 20.251 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 105 Resp: 90424
Ion Ratio Lower Upper
105 100
120 45.7 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 17.477 ug/l
RT: 11.018 min Scan# 1100
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 75 Resp: 8871
Ion Ratio Lower Upper
75 100
53 132.4 100.5 150.7
89 44.8 37.0 55.6



#82
4-Chlorotoluene
Concen: 19.662 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 91 Resp: 87403
Ion Ratio Lower Upper
91 100
126 26.8 13.3 39.8





#83

tert-Butylbenzene

Concen: 20.028 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01





Tgt Ion:119 Resp: 90081

Ion Ratio Lower Upper

119 100

91 65.4 32.9 98.7

134 23.2 11.4 34.1



#84

1,2,4-Trimethylbenzene

Concen: 20.093 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX046081.D




Acq: 08 May 2025 11:20


Tgt Ion:105 Resp: 90858

Ion Ratio Lower Upper

105 100

120 42.4 21.2 63.6









#85
sec-Butylbenzene
Concen: 19.821 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion:105 Resp: 109461
Ion Ratio Lower Upper
105 100
134 19.1 9.7 29.1



#86
p-Isopropyltoluene
Concen: 19.614 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion:119 Resp: 89406
Ion Ratio Lower Upper
119 100
134 25.3 12.5 37.5
91 28.2 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 19.456 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01





Tgt Ion:146 Resp: 44061

Ion Ratio Lower Upper

146 100

111 44.6 22.1 66.3

148 63.2 32.1 96.5



#88

1,4-Dichlorobenzene

Concen: 19.051 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20




Tgt Ion:146 Resp: 44060


Ion Ratio Lower Upper

146 100

111 44.3 21.3 63.9

148 63.5 31.9 95.5









#89
n-Butylbenzene
Concen: 18.912 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion: 91	Resp:	75618
Ion	Ratio	Lower Upper
91	100	
92	54.3	26.9 80.7
134	23.7	11.8 35.3



#90
Hexachloroethane
Concen: 19.343 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion: 117	Resp:	15534
Ion	Ratio	Lower Upper
117	100	
201	63.1	31.6 94.7





#91

1,2-Dichlorobenzene

Concen: 20.138 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBS01





Tgt Ion:146 Resp: 45764

Ion Ratio Lower Upper

146 100

111 43.5 22.4 67.3

148 62.4 31.6 94.8



#92

1,2-Dibromo-3-Chloropropane

Concen: 18.654 ug/l

RT: 12.939 min Scan# 1944

Delta R.T. -0.000 min

Lab File: VX046081.D

Acq: 08 May 2025 11:20




Tgt Ion: 75 Resp: 7740


Ion Ratio Lower Upper

75 100

155 74.7 34.9 104.8

157 88.9 43.8 131.4









#93
1,2,4-Trichlorobenzene
Concen: 18.438 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01


Tgt Ion:180 Resp: 24066
Ion Ratio Lower Upper
180 100
182 93.8 47.9 143.8
145 34.0 17.1 51.2



#94
Hexachlorobutadiene
Concen: 18.697 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion:225 Resp: 10658
Ion Ratio Lower Upper
225 100
223 60.0 30.8 92.4
227 64.0 30.9 92.7









#95
Naphthalene
Concen: 18.324 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBS01

Tgt Ion:128	Resp:	87722
Ion	Ratio	Lower Upper
128	100	
127	12.7	10.4 15.6
129	10.9	8.6 13.0



#96
1,2,3-Trichlorobenzene
Concen: 18.739 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046081.D
Acq: 08 May 2025 11:20

Tgt Ion:180	Resp:	25237
Ion	Ratio	Lower Upper
180	100	
182	94.2	47.8 143.3
145	35.7	18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046082.D
 Acq On : 08 May 2025 11:45
 Operator : JC/MD
 Sample : VX0508WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBSD01

Quant Time: May 09 01:55:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	89125	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	157859	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	138205	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	63610	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	84500	50.855	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.720%	
35) Dibromofluoromethane	5.379	113	56900	50.055	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.100%	
50) Toluene-d8	8.647	98	196138	49.851	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.700%	
62) 4-Bromofluorobenzene	11.079	95	74503	49.366	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 98.740%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	27046	19.827	ug/l	97
3) Chloromethane	1.306	50	25411	19.209	ug/l	100
4) Vinyl Chloride	1.374	62	22115	17.963	ug/l	94
5) Bromomethane	1.593	94	10264	17.974	ug/l	98
6) Chloroethane	1.672	64	12495	19.011	ug/l	99
7) Trichlorofluoromethane	1.873	101	35016	19.244	ug/l	99
8) Diethyl Ether	2.130	74	11766	18.996	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.325	101	21725	19.293	ug/l	98
10) Methyl Iodide	2.447	142	25437	19.091	ug/l	98
11) Tert butyl alcohol	2.971	59	24333	104.328	ug/l	99
12) 1,1-Dichloroethene	2.312	96	19388	18.346	ug/l	98
13) Acrolein	2.233	56	28048	105.595	ug/l	97
14) Allyl chloride	2.660	41	39668	19.640	ug/l	98
15) Acrylonitrile	3.062	53	70056	105.044	ug/l	99
16) Acetone	2.379	43	65406	98.176	ug/l	99
17) Carbon Disulfide	2.501	76	43556	17.384	ug/l	99
18) Methyl Acetate	2.703	43	32347	20.923	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	74256	20.042	ug/l	98
20) Methylene Chloride	2.782	84	23482	18.393	ug/l	93
21) trans-1,2-Dichloroethene	3.087	96	20339	19.138	ug/l	94
22) Diisopropyl ether	3.757	45	78706	20.174	ug/l	94
23) Vinyl Acetate	3.721	43	350984	102.285	ug/l	99
24) 1,1-Dichloroethane	3.605	63	42920	19.752	ug/l	98
25) 2-Butanone	4.556	43	101125	104.554	ug/l	100
26) 2,2-Dichloropropane	4.471	77	31214	18.352	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	25319	19.790	ug/l	98
28) Bromochloromethane	4.891	49	22526	21.536	ug/l	99
29) Tetrahydrofuran	5.007	42	64907	107.095	ug/l	99
30) Chloroform	5.080	83	45164	19.940	ug/l	97
31) Cyclohexane	5.458	56	38562	19.474	ug/l	97
32) 1,1,1-Trichloroethane	5.373	97	38497	19.608	ug/l	99
36) 1,1-Dichloropropene	5.690	75	28723	18.806	ug/l	98
37) Ethyl Acetate	4.714	43	37116	19.669	ug/l	99
38) Carbon Tetrachloride	5.672	117	32427	18.896	ug/l	98
39) Methylcyclohexane	7.372	83	36280	18.451	ug/l	99
40) Benzene	6.031	78	86138	19.254	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046082.D
 Acq On : 08 May 2025 11:45
 Operator : JC/MD
 Sample : VX0508WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBSD01

Quant Time: May 09 01:55:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	21603	21.885	ug/l	99
42) 1,2-Dichloroethane	6.080	62	38922	20.158	ug/l	100
43) Isopropyl Acetate	6.342	43	59168	20.553	ug/l	99
44) Trichloroethene	7.123	130	20766	19.286	ug/l	98
45) 1,2-Dichloropropane	7.427	63	22527	20.250	ug/l	99
46) Dibromomethane	7.580	93	17952	20.461	ug/l	96
47) Bromodichloromethane	7.818	83	34186	19.783	ug/l	99
48) Methyl methacrylate	7.689	41	30850	20.983	ug/l	99
49) 1,4-Dioxane	7.659	88	12525	448.677	ug/l	98
51) 4-Methyl-2-Pentanone	8.567	43	201237	105.311	ug/l	98
52) Toluene	8.714	92	53498	19.502	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	29261	19.050	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	33824	19.924	ug/l	97
55) 1,1,2-Trichloroethane	9.147	97	22314	20.629	ug/l	99
56) Ethyl methacrylate	9.116	69	35662	20.686	ug/l	99
57) 1,3-Dichloropropane	9.305	76	39386	20.275	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	87699	99.784	ug/l	99
59) 2-Hexanone	9.427	43	149616	105.830	ug/l	100
60) Dibromochloromethane	9.518	129	24175	20.351	ug/l	94
61) 1,2-Dibromoethane	9.610	107	22397	19.922	ug/l	97
64) Tetrachloroethene	9.268	164	18567	18.987	ug/l	97
65) Chlorobenzene	10.079	112	56955	18.828	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	19827	19.195	ug/l	98
67) Ethyl Benzene	10.189	91	103863	19.478	ug/l	99
68) m/p-Xylenes	10.299	106	76787	39.374	ug/l	99
69) o-Xylene	10.640	106	37621	19.787	ug/l	98
70) Styrene	10.652	104	62355	20.021	ug/l	100
71) Bromoform	10.799	173	15034	19.386	ug/l #	96
73) Isopropylbenzene	10.957	105	99423	20.076	ug/l	99
74) N-amyl acetate	10.841	43	49759	20.334	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	35488	20.449	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	36219	23.655	ug/l	90
77) Bromobenzene	11.195	156	22147	19.263	ug/l	99
78) n-propylbenzene	11.299	91	115197	20.006	ug/l	100
79) 2-Chlorotoluene	11.360	91	73340	19.747	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	84507	20.426	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	8472	18.014	ug/l	94
82) 4-Chlorotoluene	11.451	91	81988	19.906	ug/l	99
83) tert-Butylbenzene	11.713	119	84407	20.254	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	85289	20.357	ug/l	98
85) sec-Butylbenzene	11.890	105	102105	19.955	ug/l	99
86) p-Isopropyltoluene	12.006	119	84675	20.048	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	41242	19.655	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	41693	19.456	ug/l	97
89) n-Butylbenzene	12.329	91	72472	19.562	ug/l	99
90) Hexachloroethane	12.536	117	14264	19.170	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	42845	20.348	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.939	75	8274	21.521	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	23756	19.643	ug/l	98
94) Hexachlorobutadiene	13.719	225	10435	19.756	ug/l	98
95) Naphthalene	13.774	128	89648	20.211	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	24789	19.865	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046082.D
Acq On : 08 May 2025 11:45
Operator : JC/MD
Sample : VX0508WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Quant Time: May 09 01:55:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 09 01:55:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion:168 Resp: 89125
Ion Ratio Lower Upper
168 100
99 72.8 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 19.827 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 85 Resp: 27046
Ion Ratio Lower Upper
85 100
87 32.9 15.7 47.1









#3
Chloromethane
Concen: 19.209 ug/l
RT: 1.306 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion: 50 Resp: 25411
Ion Ratio Lower Upper
50 100
52 31.9 25.4 38.2



#4
Vinyl Chloride
Concen: 17.963 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 62 Resp: 22115
Ion Ratio Lower Upper
62 100
64 34.6 25.2 37.8









#5
Bromomethane
Concen: 17.974 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion: 94 Resp: 10264
Ion Ratio Lower Upper
94 100
96 93.3 72.8 109.2



#6
Chloroethane
Concen: 19.011 ug/l
RT: 1.672 min Scan# 96
Delta R.T. 0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 64 Resp: 12495
Ion Ratio Lower Upper
64 100
66 31.0 24.3 36.5









#7
 Trichlorofluoromethane
 Concen: 19.244 ug/l
 RT: 1.873 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX046082.D
 Acq: 08 May 2025 11:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBSD01


Tgt Ion:101 Resp: 35016
 Ion Ratio Lower Upper
 101 100
 103 64.7 51.0 76.4



#8
 Diethyl Ether
 Concen: 18.996 ug/l
 RT: 2.130 min Scan# 171
 Delta R.T. -0.006 min
 Lab File: VX046082.D
 Acq: 08 May 2025 11:45

Tgt Ion: 74 Resp: 11766
 Ion Ratio Lower Upper
 74 100
 45 109.6 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.293 ug/l
RT: 2.325 min Scan# 203
Delta R.T. 0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion:101 Resp: 21725
Ion Ratio Lower Upper
101 100
85 46.6 38.6 58.0
151 69.7 55.2 82.8



#10
Methyl Iodide
Concen: 19.091 ug/l
RT: 2.447 min Scan# 223
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion:142 Resp: 25437
Ion Ratio Lower Upper
142 100
127 53.3 41.7 62.5
141 14.0 11.5 17.3









#11
Tert butyl alcohol
Concen: 104.328 ug/l
RT: 2.971 min Scan# 310
Delta R.T. -0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion: 59 Resp: 24333
Ion Ratio Lower Upper
59 100
57 11.2 8.6 12.8



#12
1,1-Dichloroethene
Concen: 18.346 ug/l
RT: 2.312 min Scan# 201
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 96 Resp: 19388
Ion Ratio Lower Upper
96 100
61 183.0 146.2 219.2
98 58.7 51.0 76.6









#13
Acrolein
Concen: 105.595 ug/l
RT: 2.233 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion: 56 Resp: 28048
Ion Ratio Lower Upper
56 100
55 67.9 56.2 84.4



#14
Allyl chloride
Concen: 19.640 ug/l
RT: 2.660 min Scan# 258
Delta R.T. 0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 41 Resp: 39668
Ion Ratio Lower Upper
41 100
39 74.0 60.6 90.8
76 29.5 24.9 37.3









#15
Acrylonitrile
Concen: 105.044 ug/l
RT: 3.062 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion: 53 Resp: 70056
Ion Ratio Lower Upper
53 100
52 81.2 65.3 97.9
51 38.0 29.8 44.8



#16
Acetone
Concen: 98.176 ug/l
RT: 2.379 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 43 Resp: 65406
Ion Ratio Lower Upper
43 100
58 27.0 21.2 31.8









#17
Carbon Disulfide
Concen: 17.384 ug/l
RT: 2.501 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion: 76 Resp: 43556
Ion Ratio Lower Upper
76 100
78 8.9 7.0 10.4



#18
Methyl Acetate
Concen: 20.923 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 43 Resp: 32347
Ion Ratio Lower Upper
43 100
74 21.8 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 20.042 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion: 73 Resp: 74256
Ion Ratio Lower Upper
73 100
57 23.1 17.7 26.5



#20
Methylene Chloride
Concen: 18.393 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 84 Resp: 23482
Ion Ratio Lower Upper
84 100
49 133.7 113.9 170.9
51 40.0 33.5 50.3
86 58.7 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 19.138 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion: 96 Resp: 20339
Ion Ratio Lower Upper
96 100
61 139.1 119.5 179.3
98 61.2 50.0 75.0



#22
Diisopropyl ether
Concen: 20.174 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 45 Resp: 78706
Ion Ratio Lower Upper
45 100
43 90.4 66.6 100.0
87 23.5 19.8 29.6
59 11.1 8.6 12.8









#23
 Vinyl Acetate
 Concen: 102.285 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046082.D
 Acq: 08 May 2025 11:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBSD01


Tgt Ion: 43 Resp: 350984
 Ion Ratio Lower Upper
 43 100
 86 9.8 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 19.752 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046082.D
 Acq: 08 May 2025 11:45

Tgt Ion: 63 Resp: 42920
 Ion Ratio Lower Upper
 63 100
 98 6.6 3.6 10.8
 100 3.7 2.1 6.3









#25
2-Butanone
Concen: 104.554 ug/l
RT: 4.556 min Scan# 569
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion: 43 Resp: 101125
Ion Ratio Lower Upper
43 100
72 23.2 18.4 27.6



#26
2,2-Dichloropropane
Concen: 18.352 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 77 Resp: 31214
Ion Ratio Lower Upper
77 100
97 21.3 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 19.790 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01





Tgt Ion: 96 Resp: 25319

Ion Ratio Lower Upper

96 100

61 157.4 0.0 322.8

98 64.2 0.0 129.0



#28

Bromochloromethane

Concen: 21.536 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45




Tgt Ion: 49 Resp: 22526


Ion Ratio Lower Upper

49 100

129 1.5 0.0 4.0

130 69.6 56.2 84.2





#29

Tetrahydrofuran

Concen: 107.095 ug/l

RT: 5.007 min Scan# 64

Delta R.T. 0.006 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01





Tgt Ion: 42 Resp: 64907

Ion Ratio Lower Upper

42 100

72 41.0 33.7 50.5

71 38.8 31.4 47.2



#30

Chloroform

Concen: 19.940 ug/l

RT: 5.080 min Scan# 655

Delta R.T. -0.006 min

Lab File: VX046082.D




Acq: 08 May 2025 11:45


Tgt Ion: 83 Resp: 45164

Ion Ratio Lower Upper

83 100


85 59.7 49.3 73.9








#31
Cyclohexane
Concen: 19.474 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01







Tgt Ion:	56	69	84
Resp:	38562	29.2	86.6
Lower		24.4	66.9
Upper		36.6	100.3



#32
1,1,1-Trichloroethane
Concen: 19.608 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion:	97	99	61
Resp:	38497	63.4	50.4
Lower		51.8	40.1
Upper		77.6	60.1





#33

1,2-Dichloroethane-d4

Concen: 50.855 ug/l

RT: 5.946 min Scan# 798

Delta R.T. -0.006 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :





VX0508WBSD01

Tgt Ion: 65 Resp: 84500

Ion Ratio Lower Upper

65 100

67 49.9 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45




Tgt Ion: 114 Resp: 157859


Ion Ratio Lower Upper

114 100

63 23.9 0.0 49.2


88 16.9 0.0 33.6








#35
Dibromofluoromethane
Concen: 50.055 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45


Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01





Tgt Ion:113 Resp: 56900
Ion Ratio Lower Upper
113 100
111 103.8 83.1 124.7
192 16.8 13.3 19.9




#36
1,1-Dichloropropene
Concen: 18.806 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45



Tgt Ion: 75 Resp: 28723
Ion Ratio Lower Upper
75 100
110 31.9 16.3 48.9
77 31.5 24.3 36.5





#37

Ethyl Acetate

Concen: 19.669 ug/l

RT: 4.714 min Scan# 595

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01





Tgt Ion: 43 Resp: 37116

Ion Ratio Lower Upper

43 100

61 13.4 10.3 15.5

70 9.9 7.9 11.9



#38

Carbon Tetrachloride

Concen: 18.896 ug/l

RT: 5.672 min Scan# 752

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45




Tgt Ion: 117 Resp: 32427


Ion Ratio Lower Upper

117 100

119 95.5 75.2 112.8

121 28.2 24.2 36.4





#39

Methylcyclohexane

Concen: 18.451 ug/l

RT: 7.372 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01





Tgt Ion: 83 Resp: 36280

Ion Ratio Lower Upper

83 100

55 81.7 64.7 97.1

98 47.9 37.4 56.2



#40

Benzene

Concen: 19.254 ug/l

RT: 6.031 min Scan# 811

Delta R.T. -0.000 min

Lab File: VX046082.D




Acq: 08 May 2025 11:45


Tgt Ion: 78 Resp: 86138

Ion Ratio Lower Upper

78 100

77 22.9 19.0 28.4









#41
Methacrylonitrile
Concen: 21.885 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion:	41	Resp:	21603
Ion	Ratio	Lower	Upper
41	100		
39	58.4	47.2	70.8
67	64.9	50.7	76.1
52	30.4	24.6	37.0



#42
1,2-Dichloroethane
Concen: 20.158 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion:	62	Resp:	38922
Ion	Ratio	Lower	Upper
62	100		
98	7.7	0.0	15.2









#43
Isopropyl Acetate
Concen: 20.553 ug/l
RT: 6.342 min Scan# 861
Delta R.T. 0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion: 43 Resp: 59168
Ion Ratio Lower Upper
43 100
61 17.6 14.3 21.5
87 11.8 9.5 14.3



#44
Trichloroethene
Concen: 19.286 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion:130 Resp: 20766
Ion Ratio Lower Upper
130 100
95 104.4 0.0 204.2





#45

1,2-Dichloropropane

Concen: 20.250 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :





VX0508WBSD01

Tgt Ion: 63 Resp: 22527

Ion Ratio Lower Upper

63 100

65 29.5 24.1 36.1



#46

Dibromomethane

Concen: 20.461 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45




Tgt Ion: 93 Resp: 17952


Ion Ratio Lower Upper

93 100

95 79.0 65.6 98.4

174 80.4 68.2 102.2









#47
Bromodichloromethane
Concen: 19.783 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion:	83	Resp:	34186
Ion	Ratio	Lower	Upper
83	100		
85	62.3	50.5	75.7
127	8.3	6.5	9.7



#48
Methyl methacrylate
Concen: 20.983 ug/l
RT: 7.689 min Scan# 1083
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion:	41	Resp:	30850
Ion	Ratio	Lower	Upper
41	100		
69	72.6	58.5	87.7
39	65.1	51.7	77.5









#49
1,4-Dioxane
Concen: 448.677 ug/l
RT: 7.659 min Scan# 110
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion: 88 Resp: 12525
Ion Ratio Lower Upper
88 100
43 42.4 33.4 50.2
58 75.7 58.6 88.0



#50
Toluene-d8
Concen: 49.851 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 98 Resp: 196138
Ion Ratio Lower Upper
98 100
100 64.5 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 105.311 ug/l
RT: 8.567 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion: 43 Resp: 201237
Ion Ratio Lower Upper
43 100
58 35.0 28.9 43.3



#52
Toluene
Concen: 19.502 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 92 Resp: 53498
Ion Ratio Lower Upper
92 100
91 171.6 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 19.050 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :





VX0508WBSD01

Tgt Ion: 75 Resp: 29261

Ion Ratio Lower Upper

75 100

77 31.9 25.0 37.6



#54

cis-1,3-Dichloropropene

Concen: 19.924 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45




Tgt Ion: 75 Resp: 33824


Ion Ratio Lower Upper

75 100

77 32.9 25.4 38.0

39 62.9 52.2 78.4









#55
1,1,2-Trichloroethane
Concen: 20.629 ug/l
RT: 9.147 min Scan# 1322
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion:	97	Resp:	22314
Ion	Ratio	Lower	Upper
97	100		
83	85.5	69.0	103.6
85	55.9	44.3	66.5
99	59.8	47.8	71.6



#56
Ethyl methacrylate
Concen: 20.686 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion:	69	Resp:	35662
Ion	Ratio	Lower	Upper
69	100		
41	77.4	60.8	91.2
39	48.8	39.0	58.6





#57

1,3-Dichloropropane

Concen: 20.275 ug/l

RT: 9.305 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :





VX0508WBSD01

Tgt Ion: 76 Resp: 39386

Ion Ratio Lower Upper

76 100

78 31.6 26.3 39.5



#58

2-Chloroethyl Vinyl ether

Concen: 99.784 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046082.D




Acq: 08 May 2025 11:45


Tgt Ion: 63 Resp: 87699

Ion Ratio Lower Upper

63 100

106 27.5 21.5 32.3





#59

2-Hexanone

Concen: 105.830 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :





VX0508WBSD01

Tgt Ion: 43 Resp: 149616

Ion Ratio Lower Upper

43 100

58 49.7 24.9 74.6



#60

Dibromochloromethane

Concen: 20.351 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046082.D




Acq: 08 May 2025 11:45


Tgt Ion: 129 Resp: 24175

Ion Ratio Lower Upper

129 100

127 73.6 39.3 117.8









#61
1,2-Dibromoethane
Concen: 19.922 ug/l
RT: 9.610 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion:107 Resp: 22397
Ion Ratio Lower Upper
107 100
109 94.7 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 49.366 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 95 Resp: 74503
Ion Ratio Lower Upper
95 100
174 67.0 0.0 135.8
176 64.0 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion:117 Resp: 138205
Ion Ratio Lower Upper
117 100
82 61.6 50.6 76.0
119 31.5 25.8 38.6



#64
Tetrachloroethene
Concen: 18.987 ug/l
RT: 9.268 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion:164 Resp: 18567
Ion Ratio Lower Upper
164 100
166 124.3 105.0 157.6
129 103.7 83.5 125.3
131 97.1 76.5 114.7









#65
Chlorobenzene
Concen: 18.828 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion:112 Resp: 56955
Ion Ratio Lower Upper
112 100
114 32.1 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 19.195 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion:131 Resp: 19827
Ion Ratio Lower Upper
131 100
133 95.8 47.3 141.9
119 65.5 31.6 95.0





#67

Ethyl Benzene

Concen: 19.478 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :





VX0508WBSD01

Tgt Ion: 91 Resp: 103863

Ion Ratio Lower Upper

91 100

106 28.9 23.4 35.2



#68

m/p-Xylenes

Concen: 39.374 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046082.D




Acq: 08 May 2025 11:45


Tgt Ion: 106 Resp: 76787

Ion Ratio Lower Upper

106 100

91 215.6 171.2 256.8









#69
o-Xylene
Concen: 19.787 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion:106 Resp: 37621
Ion Ratio Lower Upper
106 100
91 228.7 112.7 338.1



#70
Styrene
Concen: 20.021 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion:104 Resp: 62355
Ion Ratio Lower Upper
104 100
78 57.8 45.7 68.5
103 54.6 43.7 65.5









#71
Bromoform
Concen: 19.386 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion:173 Resp: 15034
Ion Ratio Lower Upper
173 100
175 49.3 23.4 70.0
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion:152 Resp: 63610
Ion Ratio Lower Upper
152 100
115 77.1 46.9 140.7
150 164.6 0.0 351.0









#73
Isopropylbenzene
Concen: 20.076 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion:105 Resp: 99423
Ion Ratio Lower Upper
105 100
120 26.0 12.8 38.4



#74
N-ethyl acetate
Concen: 20.334 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 43 Resp: 49759
Ion Ratio Lower Upper
43 100
70 38.0 30.9 46.3
55 23.7 18.7 28.1
61 21.4 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 20.449 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion: 83 Resp: 35488
Ion Ratio Lower Upper
83 100
131 9.5 5.0 14.9
85 65.6 31.9 95.7



#76
1,2,3-Trichloropropane
Concen: 23.655 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 75 Resp: 36219
Ion Ratio Lower Upper
75 100
77 35.0 20.5 61.5









#77
Bromobenzene
Concen: 19.263 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion:156 Resp: 22147
Ion Ratio Lower Upper
156 100
77 186.2 92.8 278.3
158 100.9 48.8 146.3



#78
n-propylbenzene
Concen: 20.006 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 91 Resp: 115197
Ion Ratio Lower Upper
91 100
120 21.5 10.8 32.4









#79
2-Chlorotoluene
Concen: 19.747 ug/l
RT: 11.360 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion: 91 Resp: 73340
Ion Ratio Lower Upper
91 100
126 30.7 15.6 46.7



#80
1,3,5-Trimethylbenzene
Concen: 20.426 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 105 Resp: 84507
Ion Ratio Lower Upper
105 100
120 46.1 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 18.014 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion: 75 Resp: 8472
Ion Ratio Lower Upper
75 100
53 134.7 100.5 150.7
89 47.5 37.0 55.6



#82
4-Chlorotoluene
Concen: 19.906 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 91 Resp: 81988
Ion Ratio Lower Upper
91 100
126 26.8 13.3 39.8









#83
 tert-Butylbenzene
 Concen: 20.254 ug/l
 RT: 11.713 min Scan# 1743
 Delta R.T. -0.000 min
 Lab File: VX046082.D
 Acq: 08 May 2025 11:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBSD01


Tgt Ion:119 Resp: 84407
 Ion Ratio Lower Upper
 119 100
 91 66.9 32.9 98.7
 134 22.5 11.4 34.1



#84
 1,2,4-Trimethylbenzene
 Concen: 20.357 ug/l
 RT: 11.750 min Scan# 1749
 Delta R.T. -0.000 min
 Lab File: VX046082.D
 Acq: 08 May 2025 11:45

Tgt Ion:105 Resp: 85289
 Ion Ratio Lower Upper
 105 100
 120 41.3 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 19.955 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046082.D
 Acq: 08 May 2025 11:45




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0508WBSD01


Tgt Ion:105 Resp: 102105
 Ion Ratio Lower Upper
 105 100
 134 19.1 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 20.048 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046082.D
 Acq: 08 May 2025 11:45

Tgt Ion:119 Resp: 84675
 Ion Ratio Lower Upper
 119 100
 134 25.4 12.5 37.5
 91 27.7 13.8 41.4









#87
1,3-Dichlorobenzene
Concen: 19.655 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion:146 Resp: 41242
Ion Ratio Lower Upper
146 100
111 44.3 22.1 66.3
148 63.9 32.1 96.5



#88
1,4-Dichlorobenzene
Concen: 19.456 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion:146 Resp: 41693
Ion Ratio Lower Upper
146 100
111 44.4 21.3 63.9
148 66.1 31.9 95.5





#89

n-Butylbenzene

Concen: 19.562 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046082.D

Acq: 08 May 2025 11:45

Instrument :

MSVOA_X

ClientSampleId :

VX0508WBSD01





Tgt Ion: 91 Resp: 72472

Ion Ratio Lower Upper

91 100

92 53.0 26.9 80.7

134 24.6 11.8 35.3



#90

Hexachloroethane

Concen: 19.170 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046082.D




Acq: 08 May 2025 11:45


Tgt Ion: 117 Resp: 14264

Ion Ratio Lower Upper

117 100

201 62.0 31.6 94.7









#91
1,2-Dichlorobenzene
Concen: 20.348 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion:146 Resp: 42845
Ion Ratio Lower Upper
146 100
111 46.7 22.4 67.3
148 63.8 31.6 94.8



#92
1,2-Dibromo-3-Chloropropane
Concen: 21.521 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion: 75 Resp: 8274
Ion Ratio Lower Upper
75 100
155 68.1 34.9 104.8
157 85.6 43.8 131.4









#93
1,2,4-Trichlorobenzene
Concen: 19.643 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01


Tgt Ion:	180	Resp:	23756
Ion Ratio	Lower	Upper	
180	100		
182	93.9	47.9	143.8
145	34.9	17.1	51.2



#94
Hexachlorobutadiene
Concen: 19.756 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion:	225	Resp:	10435
Ion Ratio	Lower	Upper	
225	100		
223	63.7	30.8	92.4
227	62.9	30.9	92.7









#95
Naphthalene
Concen: 20.211 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45




Instrument :
MSVOA_X
ClientSampleId :
VX0508WBSD01

Tgt Ion:	128	Resp:	89648
Ion	Ratio	Lower	Upper
128	100		
127	13.2	10.4	15.6
129	10.8	8.6	13.0



#96
1,2,3-Trichlorobenzene
Concen: 19.865 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046082.D
Acq: 08 May 2025 11:45

Tgt Ion:	180	Resp:	24789
Ion	Ratio	Lower	Upper
180	100		
182	94.3	47.8	143.3
145	36.0	18.1	54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046100.D
 Acq On : 08 May 2025 18:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Quant Time: May 09 01:59:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	90039	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	158698	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	137076	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66563	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	81456	48.526	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.060%
35) Dibromofluoromethane	5.379	113	55614	48.665	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.320%
50) Toluene-d8	8.647	98	192501	48.668	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.340%
62) 4-Bromofluorobenzene	11.079	95	73670	48.556	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.120%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	74341	53.944	ug/l	100
3) Chloromethane	1.307	50	68560	51.301	ug/l	96
4) Vinyl Chloride	1.374	62	62758	50.457	ug/l	98
5) Bromomethane	1.593	94	27111	46.995	ug/l	98
6) Chloroethane	1.666	64	33953	51.135	ug/l	94
7) Trichlorofluoromethane	1.873	101	94604	51.465	ug/l	100
8) Diethyl Ether	2.130	74	31351	50.101	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.319	101	57674	50.699	ug/l	99
10) Methyl Iodide	2.440	142	72888	54.149	ug/l	99
11) Tert butyl alcohol	2.977	59	63473	269.379	ug/l	99
12) 1,1-Dichloroethene	2.306	96	53022	49.663	ug/l	98
13) Acrolein	2.239	56	82589	307.775	ug/l	100
14) Allyl chloride	2.654	41	105974	51.936	ug/l	98
15) Acrylonitrile	3.062	53	179590	266.549	ug/l	99
16) Acetone	2.386	43	175020	260.042	ug/l	99
17) Carbon Disulfide	2.501	76	135805	53.653	ug/l	99
18) Methyl Acetate	2.703	43	86369	55.300	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	196589	52.521	ug/l	99
20) Methylene Chloride	2.782	84	61933	48.019	ug/l	94
21) trans-1,2-Dichloroethene	3.087	96	53687	50.003	ug/l	94
22) Diisopropyl ether	3.763	45	208277	52.843	ug/l	92
23) Vinyl Acetate	3.721	43	931721	268.769	ug/l	100
24) 1,1-Dichloroethane	3.605	63	114758	52.275	ug/l	99
25) 2-Butanone	4.556	43	261893	268.026	ug/l	99
26) 2,2-Dichloropropane	4.471	77	69517	40.458	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	66703	51.607	ug/l	96
28) Bromochloromethane	4.891	49	52572	49.751	ug/l	98
29) Tetrahydrofuran	5.001	42	165543	270.368	ug/l	99
30) Chloroform	5.086	83	118685	51.869	ug/l	99
31) Cyclohexane	5.458	56	100717	50.347	ug/l	99
32) 1,1,1-Trichloroethane	5.379	97	103512	52.187	ug/l	98
36) 1,1-Dichloropropene	5.684	75	75619	49.249	ug/l	98
37) Ethyl Acetate	4.714	43	97419	51.353	ug/l	100
38) Carbon Tetrachloride	5.672	117	88790	51.466	ug/l	97
39) Methylcyclohexane	7.379	83	96361	48.747	ug/l	98
40) Benzene	6.031	78	230602	51.273	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
 Data File : VX046100.D
 Acq On : 08 May 2025 18:46
 Operator : JC/MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDC0050EC

Quant Time: May 09 01:59:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	55534	55.961	ug/l	98
42) 1,2-Dichloroethane	6.080	62	99228	51.120	ug/l	100
43) Isopropyl Acetate	6.342	43	156399	54.041	ug/l	99
44) Trichloroethene	7.123	130	54530	50.376	ug/l	98
45) 1,2-Dichloropropane	7.427	63	59216	52.950	ug/l	99
46) Dibromomethane	7.574	93	45589	51.686	ug/l	98
47) Bromodichloromethane	7.818	83	92853	53.449	ug/l	98
48) Methyl methacrylate	7.689	41	82409	55.756	ug/l	99
49) 1,4-Dioxane	7.659	88	31325	1116.209	ug/l	98
51) 4-Methyl-2-Pentanone	8.573	43	522471	271.972	ug/l	99
52) Toluene	8.714	92	140581	50.976	ug/l	98
53) t-1,3-Dichloropropene	8.976	75	80387	52.059	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	88964	52.127	ug/l	97
55) 1,1,2-Trichloroethane	9.153	97	55968	51.469	ug/l	98
56) Ethyl methacrylate	9.116	69	96903	55.913	ug/l	99
57) 1,3-Dichloropropane	9.305	76	98930	50.658	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	233750	264.555	ug/l	100
59) 2-Hexanone	9.427	43	393840	277.107	ug/l	100
60) Dibromochloromethane	9.518	129	64916	54.358	ug/l	98
61) 1,2-Dibromoethane	9.610	107	59495	52.641	ug/l	99
64) Tetrachloroethene	9.268	164	48558	50.066	ug/l	98
65) Chlorobenzene	10.073	112	149645	49.877	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	52783	51.521	ug/l	99
67) Ethyl Benzene	10.189	91	276173	52.220	ug/l	100
68) m/p-Xylenes	10.299	106	200363	103.585	ug/l	98
69) o-Xylene	10.640	106	99641	52.839	ug/l	98
70) Styrene	10.652	104	167552	54.240	ug/l	99
71) Bromoform	10.799	173	42429	55.162	ug/l #	97
73) Isopropylbenzene	10.957	105	264733	51.086	ug/l	100
74) N-amyl acetate	10.841	43	138217	53.976	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	90125	49.628	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	99283	61.967	ug/l	85
77) Bromobenzene	11.195	156	59559	49.505	ug/l	99
78) n-propylbenzene	11.305	91	309136	51.305	ug/l	100
79) 2-Chlorotoluene	11.360	91	190049	48.900	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	220577	50.950	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	23359	47.466	ug/l	99
82) 4-Chlorotoluene	11.451	91	219982	51.040	ug/l	100
83) tert-Butylbenzene	11.713	119	220925	50.661	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	223626	51.007	ug/l	99
85) sec-Butylbenzene	11.890	105	273948	51.163	ug/l	100
86) p-Isopropyltoluene	12.006	119	227600	51.497	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	109705	49.964	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	108913	48.570	ug/l	99
89) n-Butylbenzene	12.329	91	198814	51.283	ug/l	100
90) Hexachloroethane	12.536	117	40761	52.349	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	110746	50.262	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.939	75	22030	54.760	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	63121	49.878	ug/l	98
94) Hexachlorobutadiene	13.725	225	26797	48.483	ug/l	98
95) Naphthalene	13.774	128	236240	50.897	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	65100	49.854	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050825\
Data File : VX046100.D
Acq On : 08 May 2025 18:46
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Quant Time: May 09 01:59:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: May 09 01:59:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC







Tgt Ion:168 Resp: 90039
Ion Ratio Lower Upper
168 100
99 73.3 54.9 82.3



#2
Dichlorodifluoromethane
Concen: 53.944 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 85 Resp: 74341
Ion Ratio Lower Upper
85 100
87 31.6 15.7 47.1









#3
Chloromethane
Concen: 51.301 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 50 Resp: 68560
Ion Ratio Lower Upper
50 100
52 34.1 25.4 38.2



#4
Vinyl Chloride
Concen: 50.457 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 62 Resp: 62758
Ion Ratio Lower Upper
62 100
64 32.9 25.2 37.8









#5
Bromomethane
Concen: 46.995 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 94 Resp: 27111
Ion Ratio Lower Upper
94 100
96 92.6 72.8 109.2



#6
Chloroethane
Concen: 51.135 ug/l
RT: 1.666 min Scan# 95
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 64 Resp: 33953
Ion Ratio Lower Upper
64 100
66 33.6 24.3 36.5









#7
 Trichlorofluoromethane
 Concen: 51.465 ug/l
 RT: 1.873 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX046100.D
 Acq: 08 May 2025 18:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:101 Resp: 94604
 Ion Ratio Lower Upper
 101 100
 103 63.4 51.0 76.4



#8
 Diethyl Ether
 Concen: 50.101 ug/l
 RT: 2.130 min Scan# 171
 Delta R.T. -0.006 min
 Lab File: VX046100.D
 Acq: 08 May 2025 18:46

Tgt Ion: 74 Resp: 31351
 Ion Ratio Lower Upper
 74 100
 45 105.3 54.9 164.8









#9
 1,1,2-Trichlorotrifluoroethane
 Concen: 50.699 ug/l
 RT: 2.319 min Scan# 202
 Delta R.T. -0.000 min
 Lab File: VX046100.D
 Acq: 08 May 2025 18:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:101 Resp: 57674
 Ion Ratio Lower Upper
 101 100
 85 46.9 38.6 58.0
 151 68.6 55.2 82.8



#10
 Methyl Iodide
 Concen: 54.149 ug/l
 RT: 2.440 min Scan# 222
 Delta R.T. -0.006 min
 Lab File: VX046100.D
 Acq: 08 May 2025 18:46

Tgt Ion:142 Resp: 72888
 Ion Ratio Lower Upper
 142 100
 127 53.2 41.7 62.5
 141 14.6 11.5 17.3









#11
Tert butyl alcohol
Concen: 269.379 ug/l
RT: 2.977 min Scan# 310
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 59 Resp: 63473
Ion Ratio Lower Upper
59 100
57 11.0 8.6 12.8



#12
1,1-Dichloroethene
Concen: 49.663 ug/l
RT: 2.306 min Scan# 200
Delta R.T. -0.006 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 96 Resp: 53022
Ion Ratio Lower Upper
96 100
61 187.1 146.2 219.2
98 63.8 51.0 76.6









#13
Acrolein
Concen: 307.775 ug/l
RT: 2.239 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 82589
Ion Ratio Lower Upper
56 100
55 70.6 56.2 84.4



#14
Allyl chloride
Concen: 51.936 ug/l
RT: 2.654 min Scan# 257
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 41 Resp: 105974
Ion Ratio Lower Upper
41 100
39 73.2 60.6 90.8
76 30.4 24.9 37.3





#15

Acrylonitrile

Concen: 266.549 ug/l

RT: 3.062 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 53 Resp: 179590

Ion Ratio Lower Upper

53 100

52 82.8 65.3 97.9

51 37.3 29.8 44.8



#16

Acetone

Concen: 260.042 ug/l

RT: 2.386 min Scan# 213

Delta R.T. -0.000 min

Lab File: VX046100.D




Acq: 08 May 2025 18:46


Tgt Ion: 43 Resp: 175020

Ion Ratio Lower Upper

43 100

58 25.9 21.2 31.8





#17

Carbon Disulfide

Concen: 53.653 ug/l

RT: 2.501 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 76 Resp: 135805

Ion Ratio Lower Upper

76 100

78 9.1 7.0 10.4



#18

Methyl Acetate

Concen: 55.300 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046100.D




Acq: 08 May 2025 18:46


Tgt Ion: 43 Resp: 86369

Ion Ratio Lower Upper

43 100

74 22.1 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 52.521 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 73 Resp: 196589
Ion Ratio Lower Upper
73 100
57 22.3 17.7 26.5



#20
Methylene Chloride
Concen: 48.019 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 84 Resp: 61933
Ion Ratio Lower Upper
84 100
49 135.8 113.9 170.9
51 41.4 33.5 50.3
86 60.3 53.8 80.8









#21
trans-1,2-Dichloroethene
Concen: 50.003 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 96 Resp: 53687
Ion Ratio Lower Upper
96 100
61 159.1 119.5 179.3
98 63.5 50.0 75.0



#22
Diisopropyl ether
Concen: 52.843 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 45 Resp: 208277
Ion Ratio Lower Upper
45 100
43 74.3 66.6 100.0
87 26.3 19.8 29.6
59 11.0 8.6 12.8









#23
 Vinyl Acetate
 Concen: 268.769 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046100.D
 Acq: 08 May 2025 18:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 43 Resp: 931721
 Ion Ratio Lower Upper
 43 100
 86 9.4 7.5 11.3



#24
 1,1-Dichloroethane
 Concen: 52.275 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX046100.D
 Acq: 08 May 2025 18:46

Tgt Ion: 63 Resp: 114758
 Ion Ratio Lower Upper
 63 100
 98 6.6 3.6 10.8
 100 4.4 2.1 6.3









#25
2-Butanone
Concen: 268.026 ug/l
RT: 4.556 min Scan# 569
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 261893
Ion Ratio Lower Upper
43 100
72 22.6 18.4 27.6



#26
2,2-Dichloropropane
Concen: 40.458 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.006 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 77 Resp: 69517
Ion Ratio Lower Upper
77 100
97 21.3 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 51.607 ug/l

RT: 4.483 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 96 Resp: 66703

Ion Ratio Lower Upper

96 100

61 155.4 0.0 322.8

98 63.5 0.0 129.0



#28

Bromochloromethane

Concen: 49.751 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46




Tgt Ion: 49 Resp: 52572


Ion Ratio Lower Upper

49 100

129 1.9 0.0 4.0

130 68.3 56.2 84.2





#29

Tetrahydrofuran

Concen: 270.368 ug/l

RT: 5.001 min Scan# 642

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 42 Resp: 165543

Ion Ratio Lower Upper

42 100

72 41.7 33.7 50.5

71 38.7 31.4 47.2



#30

Chloroform

Concen: 51.869 ug/l

RT: 5.086 min Scan# 656

Delta R.T. -0.000 min

Lab File: VX046100.D




Acq: 08 May 2025 18:46


Tgt Ion: 83 Resp: 118685

Ion Ratio Lower Upper

83 100

85 62.7 49.3 73.9









#31
Cyclohexane
Concen: 50.347 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 100717
Ion Ratio Lower Upper
56 100
69 30.2 24.4 36.6
84 82.4 66.9 100.3



#32
1,1,1-Trichloroethane
Concen: 52.187 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.006 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 97 Resp: 103512
Ion Ratio Lower Upper
97 100
99 63.7 51.8 77.6
61 48.7 40.1 60.1









#33
 1,2-Dichloroethane-d4
 Concen: 48.526 ug/l
 RT: 5.952 min Scan# 798
 Delta R.T. -0.000 min
 Lab File: VX046100.D
 Acq: 08 May 2025 18:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 65 Resp: 81456
 Ion Ratio Lower Upper
 65 100
 67 50.1 0.0 99.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.757 min Scan# 930
 Delta R.T. -0.000 min
 Lab File: VX046100.D
 Acq: 08 May 2025 18:46

Tgt Ion: 114 Resp: 158698
 Ion Ratio Lower Upper
 114 100
 63 24.0 0.0 49.2
 88 17.3 0.0 33.6





#35

Dibromofluoromethane

Concen: 48.665 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:113 Resp: 55614

Ion Ratio Lower Upper

113 100

111 104.4 83.1 124.7

192 17.1 13.3 19.9



#36

1,1-Dichloropropene

Concen: 49.249 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46




Tgt Ion: 75 Resp: 75619


Ion Ratio Lower Upper

75 100

110 33.0 16.3 48.9

77 31.9 24.3 36.5





#37

Ethyl Acetate

Concen: 51.353 ug/l

RT: 4.714 min Scan# 595

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 43 Resp: 97419

Ion Ratio Lower Upper

43 100

61 12.7 10.3 15.5

70 9.9 7.9 11.9



#38

Carbon Tetrachloride

Concen: 51.466 ug/l

RT: 5.672 min Scan# 752

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46




Tgt Ion: 117 Resp: 88790


Ion Ratio Lower Upper

117 100

119 96.0 75.2 112.8

121 32.4 24.2 36.4









#39
Methylcyclohexane
Concen: 48.747 ug/l
RT: 7.379 min Scan# 1031
Delta R.T. 0.006 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 96361
Ion Ratio Lower Upper
83 100
55 83.9 64.7 97.1
98 46.7 37.4 56.2



#40
Benzene
Concen: 51.273 ug/l
RT: 6.031 min Scan# 811
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 78 Resp: 230602
Ion Ratio Lower Upper
78 100
77 23.6 19.0 28.4









#41
Methacrylonitrile
Concen: 55.961 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	41	Resp:	55534
Ion	Ratio	Lower	Upper
41	100		
39	57.0	47.2	70.8
67	64.6	50.7	76.1
52	30.3	24.6	37.0



#42
1,2-Dichloroethane
Concen: 51.120 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion:	62	Resp:	99228
Ion	Ratio	Lower	Upper
62	100		
98	7.7	0.0	15.2





#43

Isopropyl Acetate

Concen: 54.041 ug/l

RT: 6.342 min Scan# 861

Delta R.T. 0.006 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 43 Resp: 156399

Ion Ratio Lower Upper

43 100

61 17.5 14.3 21.5

87 11.7 9.5 14.3



#44

Trichloroethene

Concen: 50.376 ug/l

RT: 7.123 min Scan# 990

Delta R.T. -0.000 min

Lab File: VX046100.D




Acq: 08 May 2025 18:46


Tgt Ion:130 Resp: 54530

Ion Ratio Lower Upper

130 100

95 104.0 0.0 204.2





#45

1,2-Dichloropropane

Concen: 52.950 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 63 Resp: 59216

Ion Ratio Lower Upper

63 100

65 30.5 24.1 36.1



#46

Dibromomethane

Concen: 51.686 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46




Tgt Ion: 93 Resp: 45589


Ion Ratio Lower Upper

93 100

95 80.8 65.6 98.4

174 82.6 68.2 102.2









#47
Bromodichloromethane
Concen: 53.449 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 92853
Ion Ratio Lower Upper
83 100
85 61.5 50.5 75.7
127 8.5 6.5 9.7



#48
Methyl methacrylate
Concen: 55.756 ug/l
RT: 7.689 min Scan# 1083
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 41 Resp: 82409
Ion Ratio Lower Upper
41 100
69 71.7 58.5 87.7
39 64.0 51.7 77.5









#49
1,4-Dioxane
Concen: 1116.209 ug/l
RT: 7.659 min Scan# 1078
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	88	Resp:	31325
Ion	Ratio	Lower	Upper
88	100		
43	41.3	33.4	50.2
58	75.3	58.6	88.0



#50
Toluene-d8
Concen: 48.668 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion:	98	Resp:	192501
Ion	Ratio	Lower	Upper
98	100		
100	66.2	53.5	80.3









#51
4-Methyl-2-Pentanone
Concen: 271.972 ug/l
RT: 8.573 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 522471
Ion Ratio Lower Upper
43 100
58 35.6 28.9 43.3



#52
Toluene
Concen: 50.976 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 92 Resp: 140581
Ion Ratio Lower Upper
92 100
91 173.5 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 52.059 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 75 Resp: 80387

Ion Ratio Lower Upper

75 100

77 31.5 25.0 37.6



#54

cis-1,3-Dichloropropene

Concen: 52.127 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46




Tgt Ion: 75 Resp: 88964


Ion Ratio Lower Upper

75 100

77 31.4 25.4 38.0

39 62.1 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 51.469 ug/l

RT: 9.153 min Scan# 1

Delta R.T. 0.006 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 97 Resp: 55968





Ion Ratio Lower Upper

97 100

83 85.4 69.0 103.6

85 54.2 44.3 66.5

99 63.3 47.8 71.6



#56

Ethyl methacrylate

Concen: 55.913 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46




Tgt Ion: 69 Resp: 96903


Ion Ratio Lower Upper

69 100

41 75.9 60.8 91.2

39 47.7 39.0 58.6









#57
1,3-Dichloropropane
Concen: 50.658 ug/l
RT: 9.305 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 98930
Ion Ratio Lower Upper
76 100
78 32.4 26.3 39.5



#58
2-Chloroethyl Vinyl ether
Concen: 264.555 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 63 Resp: 233750
Ion Ratio Lower Upper
63 100
106 26.7 21.5 32.3









#59
2-Hexanone
Concen: 277.107 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 393840
Ion Ratio Lower Upper
43 100
58 49.8 24.9 74.6



#60
Dibromochloromethane
Concen: 54.358 ug/l
RT: 9.518 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 129 Resp: 64916
Ion Ratio Lower Upper
129 100
127 77.2 39.3 117.8









#61
1,2-Dibromoethane
Concen: 52.641 ug/l
RT: 9.610 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:107 Resp: 59495
Ion Ratio Lower Upper
107 100
109 92.6 73.3 109.9



#62
4-Bromofluorobenzene
Concen: 48.556 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 95 Resp: 73670
Ion Ratio Lower Upper
95 100
174 67.8 0.0 135.8
176 66.3 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:117 Resp: 137076
Ion Ratio Lower Upper
117 100
82 65.0 50.6 76.0
119 32.8 25.8 38.6



#64
Tetrachloroethene
Concen: 50.066 ug/l
RT: 9.268 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion:164 Resp: 48558
Ion Ratio Lower Upper
164 100
166 125.5 105.0 157.6
129 103.8 83.5 125.3
131 95.3 76.5 114.7









#65
Chlorobenzene
Concen: 49.877 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:112 Resp: 149645
Ion Ratio Lower Upper
112 100
114 31.6 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 51.521 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion:131 Resp: 52783
Ion Ratio Lower Upper
131 100
133 95.2 47.3 141.9
119 64.8 31.6 95.0









#67
Ethyl Benzene
Concen: 52.220 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 276173
Ion Ratio Lower Upper
91 100
106 29.5 23.4 35.2



#68
m/p-Xylenes
Concen: 103.585 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 106 Resp: 200363
Ion Ratio Lower Upper
106 100
91 217.0 171.2 256.8





#69

o-Xylene

Concen: 52.839 ug/l

RT: 10.640 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tgt Ion:106 Resp: 99641

Ion Ratio Lower Upper

106 100

91 228.9 112.7 338.1



#70

Styrene

Concen: 54.240 ug/l

RT: 10.652 min Scan# 1569

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46




Tgt Ion:104 Resp: 167552


Ion Ratio Lower Upper

104 100

78 57.9 45.7 68.5

103 54.8 43.7 65.5









#71
Bromoform
Concen: 55.162 ug/l
RT: 10.799 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:173 Resp: 42429
Ion Ratio Lower Upper
173 100
175 48.5 23.4 70.0
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion:152 Resp: 66563
Ion Ratio Lower Upper
152 100
115 91.2 46.9 140.7
150 171.5 0.0 351.0









#73
Isopropylbenzene
Concen: 51.086 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:105 Resp: 264733
Ion Ratio Lower Upper
105 100
120 25.4 12.8 38.4



#74
N-ethyl acetate
Concen: 53.976 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 43 Resp: 138217
Ion Ratio Lower Upper
43 100
70 38.4 30.9 46.3
55 23.1 18.7 28.1
61 22.0 17.1 25.7









#75
1,1,2-Tetrachloroethane
Concen: 49.628 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	83	Resp:	90125
Ion	Ratio	Lower	Upper
83	100		
131	9.7	5.0	14.9
85	63.5	31.9	95.7



#76
1,2,3-Trichloropropane
Concen: 61.967 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion:	75	Resp:	99283
Ion	Ratio	Lower	Upper
75	100		
77	31.9	20.5	61.5





#77

Bromobenzene

Concen: 49.505 ug/l

RT: 11.195 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:156 Resp: 59559

Ion Ratio Lower Upper

156 100

77 187.5 92.8 278.3

158 96.2 48.8 146.3



#78

n-propylbenzene

Concen: 51.305 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. 0.006 min

Lab File: VX046100.D




Acq: 08 May 2025 18:46


Tgt Ion: 91 Resp: 309136

Ion Ratio Lower Upper

91 100

120 21.7 10.8 32.4









#79
2-Chlorotoluene
Concen: 48.900 ug/l
RT: 11.360 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 190049
Ion Ratio Lower Upper
91 100
126 31.6 15.6 46.7



#80
1,3,5-Trimethylbenzene
Concen: 50.950 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 105 Resp: 220577
Ion Ratio Lower Upper
105 100
120 46.7 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 47.466 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 23359
Ion Ratio Lower Upper
75 100
53 127.0 100.5 150.7
89 46.3 37.0 55.6



#82
4-Chlorotoluene
Concen: 51.040 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 91 Resp: 219982
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83

tert-Butylbenzene

Concen: 50.661 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:119 Resp: 220925

Ion Ratio Lower Upper

119 100

91 66.0 32.9 98.7

134 22.7 11.4 34.1



#84

1,2,4-Trimethylbenzene

Concen: 51.007 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX046100.D




Acq: 08 May 2025 18:46


Tgt Ion:105 Resp: 223626

Ion Ratio Lower Upper

105 100

120 43.3 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 51.163 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046100.D
 Acq: 08 May 2025 18:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:105 Resp: 273948
 Ion Ratio Lower Upper
 105 100
 134 19.3 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 51.497 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046100.D
 Acq: 08 May 2025 18:46

Tgt Ion:119 Resp: 227600
 Ion Ratio Lower Upper
 119 100
 134 24.7 12.5 37.5
 91 27.4 13.8 41.4









#87
1,3-Dichlorobenzene
Concen: 49.964 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:146 Resp: 109705
Ion Ratio Lower Upper
146 100
111 44.4 22.1 66.3
148 63.4 32.1 96.5



#88
1,4-Dichlorobenzene
Concen: 48.570 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion:146 Resp: 108913
Ion Ratio Lower Upper
146 100
111 43.4 21.3 63.9
148 64.3 31.9 95.5









#89
n-Butylbenzene
Concen: 51.283 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 198814
Ion Ratio Lower Upper
91 100
92 53.6 26.9 80.7
134 23.9 11.8 35.3



#90
Hexachloroethane
Concen: 52.349 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 117 Resp: 40761
Ion Ratio Lower Upper
117 100
201 61.9 31.6 94.7









#91
1,2-Dichlorobenzene
Concen: 50.262 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tgt Ion:146 Resp: 110746
Ion Ratio Lower Upper
146 100
111 46.4 22.4 67.3
148 64.3 31.6 94.8



#92
1,2-Dibromo-3-Chloropropane
Concen: 54.760 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046100.D
Acq: 08 May 2025 18:46

Tgt Ion: 75 Resp: 22030
Ion Ratio Lower Upper
75 100
155 68.8 34.9 104.8
157 88.3 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 49.878 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:180 Resp: 63121

Ion Ratio Lower Upper

180 100

182 93.5 47.9 143.8

145 34.8 17.1 51.2



#94

Hexachlorobutadiene

Concen: 48.483 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46




Tgt Ion:225 Resp: 26797


Ion Ratio Lower Upper

225 100

223 64.6 30.8 92.4

227 62.4 30.9 92.7





#95

Naphthalene

Concen: 50.897 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:128 Resp: 236240

Ion Ratio Lower Upper

128 100

127 12.6 10.4 15.6

129 10.8 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 49.854 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. -0.000 min

Lab File: VX046100.D

Acq: 08 May 2025 18:46

Tgt Ion:180 Resp: 65100

Ion Ratio Lower Upper

180 100

182 95.0 47.8 143.3

145 36.8 18.1 54.3

