

Surrogate Summary

SDG No.: Q1985

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1985-01	RW8-BW-20250507	1,2-Dichloroethane-d4	50	53.6	107	81	118
		Dibromofluoromethane	50	51.1	102	80	119
		Toluene-d8	50	49.5	99	89	112
		4-Bromofluorobenzene	50	49.3	99	85	114
Q1985-02	TB-20250507	1,2-Dichloroethane-d4	50	51.4	103	81	118
		Dibromofluoromethane	50	49.4	99	80	119
		Toluene-d8	50	49.2	98	89	112
		4-Bromofluorobenzene	50	49.4	99	85	114
VX0508WBL01	VX0508WBL01	1,2-Dichloroethane-d4	50	52.5	105	81	118
		Dibromofluoromethane	50	49.3	99	80	119
		Toluene-d8	50	49.8	100	89	112
		4-Bromofluorobenzene	50	48.3	97	85	114
VX0508WBS01	VX0508WBS01	1,2-Dichloroethane-d4	50	49.8	100	81	118
		Dibromofluoromethane	50	50.2	100	80	119
		Toluene-d8	50	49.1	98	89	112
		4-Bromofluorobenzene	50	48.2	96	85	114
VX0508WBSD0	VX0508WBSD01	1,2-Dichloroethane-d4	50	50.9	102	81	118
		Dibromofluoromethane	50	50.0	100	80	119
		Toluene-d8	50	49.9	100	89	112
		4-Bromofluorobenzene	50	49.4	99	85	114