

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1985

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VX046082.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0508WBSD01	Chloromethane	20	19.2	ug/L	96	3		50	139	20
	Vinyl chloride	20	18.0	ug/L	90	3		58	137	20
	Bromomethane	20	18.0	ug/L	90	0		53	141	20
	Chloroethane	20	19.0	ug/L	95	2		60	138	20
	Trichlorofluoromethane	20	19.2	ug/L	96	1		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	19.3	ug/L	97	4		70	136	20
	1,1-Dichloroethene	20	18.3	ug/L	92	0		71	131	20
	Acetone	100	98.2	ug/L	98	5		39	160	20
	Carbon disulfide	20	17.4	ug/L	87	1		64	133	20
	Methyl tert-butyl Ether	20	20.0	ug/L	100	3		71	124	20
	Methylene Chloride	20	18.4	ug/L	92	4		74	124	20
	trans-1,2-Dichloroethene	20	19.1	ug/L	96	2		75	124	20
	1,1-Dichloroethane	20	19.8	ug/L	99	2		77	125	20
	2-Butanone	100	100	ug/L	100	3		56	143	20
	Carbon Tetrachloride	20	18.9	ug/L	95	0		72	136	20
	cis-1,2-Dichloroethene	20	19.8	ug/L	99	1		78	123	20
	Chloroform	20	19.9	ug/L	100	1		79	124	20
	1,1,1-Trichloroethane	20	19.6	ug/L	98	2		74	131	20
	Methylcyclohexane	20	18.5	ug/L	93	2		72	132	20
	Benzene	20	19.3	ug/L	97	0		79	120	20
	1,2-Dichloroethane	20	20.2	ug/L	101	6		73	128	20
	Trichloroethene	20	19.3	ug/L	97	4		79	123	20
	1,2-Dichloropropane	20	20.3	ug/L	102	4		78	122	20
	Bromodichloromethane	20	19.8	ug/L	99	3		79	125	20
	4-Methyl-2-Pentanone	100	110	ug/L	110	12		67	130	20
	Toluene	20	19.5	ug/L	98	2		80	121	20
	t-1,3-Dichloropropene	20	19.1	ug/L	96	8		73	127	20
	cis-1,3-Dichloropropene	20	19.9	ug/L	100	7		75	124	20
	1,1,2-Trichloroethane	20	20.6	ug/L	103	6		80	119	20
	2-Hexanone	100	110	ug/L	110	12		57	139	20
	Dibromochloromethane	20	20.4	ug/L	102	5		74	126	20
	Tetrachloroethene	20	19.0	ug/L	95	0		74	129	20
	Chlorobenzene	20	18.8	ug/L	94	0		82	118	20
	Ethyl Benzene	20	19.5	ug/L	98	0		79	121	20
	m/p-Xylenes	40	39.4	ug/L	99	1		80	121	20
	o-Xylene	20	19.8	ug/L	99	2		78	122	20
	Styrene	20	20.0	ug/L	100	0		78	123	20
	Bromoform	20	19.4	ug/L	97	3		66	130	20
	Isopropylbenzene	20	20.1	ug/L	101	0		72	131	20
	1,1,2,2-Tetrachloroethane	20	20.4	ug/L	102	4		71	121	20
	1,3-Dichlorobenzene	20	19.7	ug/L	99	1		80	119	20
	1,4-Dichlorobenzene	20	19.5	ug/L	98	2		79	118	20
	1,2-Dichlorobenzene	20	20.3	ug/L	102	1		80	119	20