

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q1985 SAS No.: Q1985 SDG NO.: Q1985
 Lab File ID: VX046078.D Date Analyzed: 05/08/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:05
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	105825	5.54	181991	6.75	158314	10.05
UPPER LIMIT	211650	6.044	363982	7.251	316628	10.549
LOWER LIMIT	52912.5	5.044	90995.5	6.251	79157	9.549
EPA SAMPLE NO.						
RW8-BW-20250507	65710	5.54	131109	6.76	121918	10.05
TB-20250507	64053	5.55	126115	6.76	114932	10.05
VX0508WBL01	66087	5.54	132156	6.76	121254	10.05
VX0508WBS01	97732	5.54	172839	6.76	147787	10.05
VX0508WBSD01	89125	5.54	157859	6.76	138205	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.