

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q1985 SAS No.: Q1985 SDG NO.: Q1985
 Lab File ID: VX046078.D Date Analyzed: 05/08/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:05
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	73554	12.018				
UPPER LIMIT	147108	12.518				
LOWER LIMIT	36777	11.518				
EPA SAMPLE NO.						
RW8-BW-20250507	50270	12.02				
TB-20250507	49527	12.02				
VX0508WBL01	49493	12.02				
VX0508WBS01	68652	12.02				
VX0508WBSD01	63610	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.