

rteres

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC05132
Data File : FC068847.D
Signal(s) : FID1A.ch
Acq On : 13 May 2025 13:26
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 05/14/2025
Supervised By :mohammad ahmed 05/15/2025

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 050625.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.441	3.377	3.479	BB	200564	1664991	81.55%	4.363%
2	4.516	4.459	4.562	BB	191509	1708898	83.70%	4.478%
3	6.113	6.055	6.172	BB	203075	1860581	91.13%	4.875%
4	6.541	6.480	6.582	BB	174313	1702226	83.38%	4.460%
5	7.171	7.130	7.204	BB	195773	1832129	89.74%	4.801%
6	8.337	8.280	8.379	BB	161115	1671463	81.87%	4.380%
7	9.940	9.875	9.980	BB	152864	1670747	81.84%	4.378%
8	11.383	11.340	11.434	BB	148964	1655681	81.10%	4.338%
9	11.707	11.647	11.749	BB	188075	2041576	100.00%	5.349%
10	12.694	12.635	12.734	BB	134237	1619601	79.33%	4.244%
11	13.140	13.079	13.175	BB	126986	1510379	73.98%	3.957%
12	13.306	13.244	13.344	BB	134272	1603112	78.52%	4.200%
13	13.892	13.860	13.945	VB	130807	1599262	78.33%	4.190%
14	14.996	14.907	15.042	BB	130081	1601636	78.45%	4.197%
15	16.019	15.959	16.069	BB	126952	1601013	78.42%	4.195%
16	16.970	16.907	17.005	BB	129781	1692923	82.92%	4.436%
17	17.858	17.812	17.899	BB	142847	1922922	94.19%	5.038%
18	18.691	18.619	18.747	BB	148643	2025219	99.20%	5.306%
19	19.476	19.414	19.552	BB	132751	1910773	93.59%	5.007%
20	20.216	20.150	20.254	BV	131438	1807445	88.53%	4.736%
21	20.987	20.862	21.062	BB	95507	1739556	85.21%	4.558%
22	21.964	21.884	22.059	PB	68126	1723068	84.40%	4.515%
Sum of corrected areas:						38165202		

Aliphatic EPH 050625.M Wed May 14 03:36:07 2025