

Continuing Calibration Report for SequenceID : FC051325AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068847.D

Aliphatic C9-C12	5076115.000	60.000	3.341	6.641	84601.917	99133.167	14.658
Aliphatic C12-C16	3342210.000	40.000	6.642	10.040	83555.250	90984.404	8.165
Aliphatic C16-C21	4878394.000	60.000	10.041	13.406	81306.567	85760.459	5.193
Aliphatic C21-C28	6494834.000	80.000	13.407	17.068	81185.425	83091.075	2.293
Aliphatic C28-C40	11119253.000	120.000	17.069	22.062	92660.442	84622.692	-9.498
Aliphatic EPH	30910806.000	360.000	3.341	22.062	85863.350	87597.230	1.979