

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY042225

Review By	Semsettin Yesilyurt	Review On	4/23/2025 7:35:07 AM		
Supervise By	Mahesh Dadoda	Supervise On	4/23/2025 1:28:52 PM		
SubDirectory	VY042225	HP Acquire Method	MSVOA_Y	HP Processing Method	82y042225s.m

STD. NAME	STD REF.#
Tune/Reschk	VP133718
Initial Calibration Stds	VP133719,VP133720,VP133721,VP133722,VP133723,VP133724
CCC	
Internal Standard/PEM	VP131783
ICV/I.BLK	VP133725
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021952.D	22 Apr 2025 11:33		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY021953.D	22 Apr 2025 13:39		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY021954.D	22 Apr 2025 14:44	Com.#16 is on Linear Regression	SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY021955.D	22 Apr 2025 15:07		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY021956.D	22 Apr 2025 15:29		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY021957.D	22 Apr 2025 15:52		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY021958.D	22 Apr 2025 16:15		SY/MD	Ok,M
8	IBLK	IBLK	VY021959.D	22 Apr 2025 16:38		SY/MD	Ok
9	VSTDICV050	ICVVY042225	VY021960.D	22 Apr 2025 17:01		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY051225

Review By	Maresh Dadoda	Review On	5/13/2025 2:15:05 PM
Supervise By	Semsettin Yesilyurt	Supervise On	5/13/2025 2:16:30 PM
SubDirectory	VY051225	HP Acquire Method	MSVOA_Y
		HP Processing Method	82y042225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133882
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133883,VP133884 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022207.D	12 May 2025 07:56		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022208.D	12 May 2025 08:26		SY/MD	Ok,M
3	VY0512SBL01	VY0512SBL01	VY022209.D	12 May 2025 08:57		SY/MD	Ok
4	VY0512SBS01	VY0512SBS01	VY022210.D	12 May 2025 09:28		SY/MD	Ok,M
5	VY0512SBSD01	VY0512SBSD01	VY022211.D	12 May 2025 09:51		SY/MD	Ok,M
6	Q1978-01	72-12016	VY022212.D	12 May 2025 10:31	vial-B Not purged	SY/MD	Not Ok
7	Q1976-01RE	VNJ-207RE	VY022213.D	12 May 2025 10:54	vial-B ISTD Fail	SY/MD	Confirms
8	Q1983-02	OR-636-VOC-01	VY022214.D	12 May 2025 11:16	vial-A Internal standard fail	SY/MD	ReRun
9	Q1983-08	OR-636-VOC-02	VY022215.D	12 May 2025 11:39	vial-A Not purge	SY/MD	Not Ok
10	Q1983-14	OR-636-VOC-03	VY022216.D	12 May 2025 12:02	vial-A Internal standard fail	SY/MD	ReRun
11	Q1983-20	OR-636-VOC-04	VY022217.D	12 May 2025 12:24	vial-A Internal standard fail	SY/MD	ReRun
12	Q1983-26	OR-636-VOC-05	VY022218.D	12 May 2025 12:47	vial-A Internal standard fail	SY/MD	ReRun
13	Q1983-32	OR-636-VOC-06	VY022219.D	12 May 2025 13:10	vial-A Internal standard fail	SY/MD	ReRun
14	Q1983-38	OR-636-VOC-07	VY022220.D	12 May 2025 13:32	vial-A Internal standard fail	SY/MD	ReRun
15	Q1983-44	OR-636-VOC-08	VY022221.D	12 May 2025 13:55	vial-A	SY/MD	Ok
16	Q1983-50	OR-636-VOC-09	VY022222.D	12 May 2025 14:17	vial-A	SY/MD	Ok
17	Q1986-01	COMP-8	VY022223.D	12 May 2025 14:40	vial-A Internal standard fail	SY/MD	ReRun
18	Q1986-03	COMP-9	VY022224.D	12 May 2025 15:03	vial-A Internal standard fail	SY/MD	ReRun

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY051225

Review By	Mahesh Dadoda	Review On	5/13/2025 2:15:05 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/13/2025 2:16:30 PM		
SubDirectory	VY051225	HP Acquire Method	MSVOA_Y	HP Processing Method	82y042225s.m
STD. NAME	STD REF.#				
Tune/Reschk Initial Calibration Stds	VP133882				
CCC	VP133883,VP133884				
Internal Standard/PEM	VP131783				
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q1986-05	COMP-10	VY022225.D	12 May 2025 15:25	vial-A Internal standard fail	SY/MD	ReRun
20	Q1986-07	COMP-11	VY022226.D	12 May 2025 15:48	vial-A not req	SY/MD	Not Ok
21	Q1986-09	VNJ-218	VY022227.D	12 May 2025 16:11	vial-A Internal standard fail	SY/MD	ReRun
22	Q2002-01	EO-02-05092025	VY022228.D	12 May 2025 16:33	vial-A Internal standard fail;Surrogate fail	SY/MD	ReRun
23	Q1980-06	3836	VY022229.D	12 May 2025 16:56	vial-A Internal standard fail;Surrogate fail	SY/MD	ReRun
24	Q1987-01	GC1	VY022230.D	12 May 2025 17:19	vial-A	SY/MD	Ok
25	VSTDCCC050	VSTDCCC050EC	VY022231.D	12 May 2025 18:04		SY/MD	Ok,M

M : Manual Integration