

Surrogate Summary

SDG No.: Q1987

Client: G Environmental

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1987-01	GC1	1,2-Dichloroethane-d4	50	65.5	131 *	70 (63)	130 (155)
		Dibromofluoromethane	50	58.3	117	70 (70)	130 (134)
		Toluene-d8	50	52.1	104	70 (74)	130 (123)
		4-Bromofluorobenzene	50	42.2	84	70 (38)	130 (136)
VY0512SBL01	VY0512SBL01	1,2-Dichloroethane-d4	50	61.3	123	70 (63)	130 (155)
		Dibromofluoromethane	50	54.9	110	70 (70)	130 (134)
		Toluene-d8	50	49.5	99	70 (74)	130 (123)
		4-Bromofluorobenzene	50	42.4	85	70 (38)	130 (136)
VY0512SBS01	VY0512SBS01	1,2-Dichloroethane-d4	50	48.1	96	70 (63)	130 (155)
		Dibromofluoromethane	50	49.5	99	70 (70)	130 (134)
		Toluene-d8	50	50.0	100	70 (74)	130 (123)
		4-Bromofluorobenzene	50	49.6	99	70 (38)	130 (136)
VY0512SBSD01	VY0512SBSD01	1,2-Dichloroethane-d4	50	54.6	109	70 (63)	130 (155)
		Dibromofluoromethane	50	53.4	107	70 (70)	130 (134)
		Toluene-d8	50	53.7	107	70 (74)	130 (123)
		4-Bromofluorobenzene	50	52.9	106	70 (38)	130 (136)

