

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1987 SAS No.: Q1987 SDG NO.: Q1987
 Lab File ID: VY022208.D Date Analyzed: 05/12/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:26
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	176689	13.346				
UPPER LIMIT	353378	13.846				
LOWER LIMIT	88344.5	12.846				
EPA SAMPLE NO.						
GC1	98718	13.35				
VY0512SBL01	128920	13.35				
VY0512SBS01	169523	13.35				
VY0512SBSD01	162833	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.