

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1993 SAS No.: Q1993 SDG NO.: Q1993

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/13/2025

Lab File ID: BF142336.D Time Analyzed: 10:49

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	199243	6.904	783419	8.19	422251	9.95
UPPER LIMIT	398486	7.404	1566840	8.687	844502	10.445
LOWER LIMIT	99621.5	6.404	391710	7.687	211126	9.445
EPA SAMPLE NO.						
01 MW-3-20250508	220712	6.90	842702	8.19	448798	9.94

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.