

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1993 SAS No.: Q1993 SDG NO.: Q1993

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/16/2025

Lab File ID: BP024651.D Time Analyzed: 10:49

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	118329	7.657	478267	10.43	294926	14.29
UPPER LIMIT	236658	8.157	956534	10.928	589852	14.792
LOWER LIMIT	59164.5	7.157	239134	9.928	147463	13.792
EPA SAMPLE NO.						
01 MW-3-20250508-A	139534	7.66	567590	10.43	368139	14.30
02 MW-2-20250508	139709	7.66	573216	10.43	378521	14.29
03 MW-1-20250508	149432	7.66	637785	10.43	420480	14.29
04 MW-4-20250508	125391	7.66	504701	10.43	321435	14.30
05 EB-20250508	131811	7.66	533983	10.43	349924	14.29

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.