

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q1993 SAS No.: Q1993 SDG No.: Q1993

Instrument ID: MSVOA_X Calibration Date/Time: 05/13/2025 10:00

Lab File ID: VX046153.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.742	0.751	0.1	1.21	20
Vinyl Chloride	0.691	0.696		0.72	20
Bromomethane	0.320	0.304		-5	20
Chloroethane	0.369	0.375		1.63	20
1,1-Dichloroethene	0.593	0.578		-2.53	20
Acetone	0.374	0.410		9.63	20
Carbon Disulfide	1.406	1.375		-2.2	20
Methyl tert-butyl Ether	2.079	2.132		2.55	20
Methylene Chloride	0.716	0.668		-6.7	20
trans-1,2-Dichloroethene	0.596	0.589		-1.17	20
1,1-Dichloroethane	1.219	1.231	0.1	0.98	20
2-Butanone	0.543	0.561		3.32	20
Carbon Tetrachloride	0.544	0.574		5.51	20
cis-1,2-Dichloroethene	0.718	0.715		-0.42	20
Chloroform	1.271	1.300		2.28	20
1,1,1-Trichloroethane	1.101	1.117		1.45	20
Benzene	1.417	1.440		1.62	20
1,2-Dichloroethane	0.612	0.625		2.12	20
Trichloroethene	0.341	0.344		0.88	20
1,2-Dichloropropane	0.352	0.370		5.11	20
Bromodichloromethane	0.547	0.588		7.49	20
4-Methyl-2-Pentanone	0.605	0.632		4.46	20
Toluene	0.869	0.884		1.73	20
t-1,3-Dichloropropene	0.487	0.532		9.24	20
cis-1,3-Dichloropropene	0.538	0.581		7.99	20
1,1,2-Trichloroethane	0.343	0.351		2.33	20
2-Hexanone	0.448	0.467		4.24	20
Dibromochloromethane	0.376	0.418		11.17	20
Tetrachloroethene	0.354	0.368		3.95	20
Chlorobenzene	1.094	1.091	0.3	-0.27	20
Ethyl Benzene	1.929	1.981		2.7	20
m/p-Xylenes	0.706	0.736		4.25	20
o-Xylene	0.688	0.715		3.92	20
Styrene	1.127	1.209		7.28	20
Bromoform	0.281	0.307	0.1	9.25	20
1,1,2,2-Tetrachloroethane	1.364	1.283	0.3	-5.94	20
1,2-Dichloroethane-d4	0.932	0.899		-3.54	20
Dibromofluoromethane	0.360	0.371		3.06	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q1993 SAS No.: Q1993 SDG No.: Q1993

Instrument ID: MSVOA_X Calibration Date/Time: 05/13/2025 10:00

Lab File ID: VX046153.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Toluene-d8	1.246	1.236		-0.8	20
4-Bromofluorobenzene	0.478	0.484		1.25	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.