

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046047.D
 Acq On : 05 May 2025 16:27
 Operator : JC/MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC001

Quant Time: May 06 06:13:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	92040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	167022	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	140108	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59123	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount	50.000	Range	74 - 125	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount	50.000	Range	86 - 113	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount	50.000	Range	77 - 121	Recovery	=	0.000%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	1211	0.596	ug/l	98
3) Chloromethane	1.307	50	1278	0.639	ug/l	97
4) Vinyl Chloride	1.374	62	1239	0.707	ug/l #	82
6) Chloroethane	1.678	64	859	0.951	ug/l	91
7) Trichlorofluoromethane	1.880	101	1959	0.734	ug/l	95
8) Diethyl Ether	2.130	74	742	0.845	ug/l	87
9) 1,1,2-Trichlorotrifluo...	2.319	101	1166	0.731	ug/l	96
12) 1,1-Dichloroethene	2.313	96	1093	0.704	ug/l #	87
14) Allyl chloride	2.654	41	1936	0.660	ug/l	89
15) Acrylonitrile	3.075	53	3343	3.424	ug/l	98
16) Acetone	2.380	43	3500	3.736	ug/l	97
17) Carbon Disulfide	2.508	76	2619	0.728	ug/l	98
18) Methyl Acetate	2.709	43	1852	0.828	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	3588	0.662	ug/l	96
20) Methylene Chloride	2.782	84	1571	0.828	ug/l #	90
21) trans-1,2-Dichloroethene	3.081	96	1111	0.701	ug/l #	71
22) Diisopropyl ether	3.757	45	3857	0.703	ug/l #	26
23) Vinyl Acetate	3.727	43	15283	3.154	ug/l	96
24) 1,1-Dichloroethane	3.605	63	2054	0.637	ug/l	97
25) 2-Butanone	4.593	43	4558	3.371	ug/l	92
26) 2,2-Dichloropropane	4.458	77	1777	0.728	ug/l	84
27) cis-1,2-Dichloroethene	4.483	96	1324	0.693	ug/l	91
28) Bromochloromethane	4.910	49	1061m	0.619	ug/l	
29) Tetrahydrofuran	5.019	42	2931	3.338	ug/l	99
30) Chloroform	5.093	83	2328	0.695	ug/l	98
32) 1,1,1-Trichloroethane	5.379	97	1869	0.647	ug/l #	51
36) 1,1-Dichloropropene	5.684	75	1648	0.758	ug/l #	83
37) Ethyl Acetate	4.733	43	1956m	0.717	ug/l	
38) Carbon Tetrachloride	5.678	117	1807	0.724	ug/l #	78
39) Methylcyclohexane	7.373	83	2096	0.770	ug/l	97
40) Benzene	6.044	78	4502	0.668	ug/l #	89
41) Methacrylonitrile	4.946	41	778	0.513	ug/l #	56
42) 1,2-Dichloroethane	6.092	62	1934	0.694	ug/l	81
43) Isopropyl Acetate	6.354	43	2553	0.612	ug/l #	81
44) Trichloroethene	7.135	130	1083	0.680	ug/l	85
45) 1,2-Dichloropropane	7.440	63	1059	0.633	ug/l	89

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046047.D
 Acq On : 05 May 2025 16:27
 Operator : JC/MD
 Sample : VSTDIC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001

Quant Time: May 06 06:13:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	7.586	93	880	0.667	ug/l	97
47) Bromodichloromethane	7.818	83	1620	0.633	ug/l #	91
48) Methyl methacrylate	7.714	41	1235m	0.577	ug/l	
49) 1,4-Dioxane	7.677	88	498	11.703	ug/l #	66
51) 4-Methyl-2-Pentanone	8.580	43	9378	3.430	ug/l	97
52) Toluene	8.720	92	2684	0.672	ug/l	96
53) t-1,3-Dichloropropene	8.988	75	1238	0.576	ug/l #	79
54) cis-1,3-Dichloropropene	8.379	75	1414	0.564	ug/l #	87
55) 1,1,2-Trichloroethane	9.153	97	1029	0.637	ug/l	92
56) Ethyl methacrylate	9.128	69	1260	0.504	ug/l #	76
57) 1,3-Dichloropropane	9.311	76	2036	0.711	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.244	63	3846	3.420	ug/l	94
59) 2-Hexanone	9.439	43	6424	3.096	ug/l	88
60) Dibromochloromethane	9.525	129	1023	0.582	ug/l	93
61) 1,2-Dibromoethane	9.610	107	1077	0.650	ug/l	98
64) Tetrachloroethene	9.275	164	972	0.679	ug/l	95
65) Chlorobenzene	10.079	112	3170	0.735	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	1035	0.720	ug/l #	67
67) Ethyl Benzene	10.195	91	5052	0.687	ug/l	94
68) m/p-Xylenes	10.299	106	3630	1.367	ug/l	97
69) o-Xylene	10.640	106	1799	0.672	ug/l	94
70) Styrene	10.659	104	2664	0.622	ug/l	98
71) Bromoform	10.799	173	657	0.608	ug/l #	100
73) Isopropylbenzene	10.963	105	4480	0.695	ug/l	96
74) N-amyl acetate	10.854	43	2028	0.631	ug/l #	91
75) 1,1,2,2-Tetrachloroethane	11.213	83	1835	0.800	ug/l	94
76) 1,2,3-Trichloropropane	11.244	75	1661m	0.674	ug/l	
77) Bromobenzene	11.201	156	1095	0.734	ug/l	96
78) n-propylbenzene	11.305	91	5052	0.699	ug/l	97
79) 2-Chlorotoluene	11.360	91	3765	0.773	ug/l	94
80) 1,3,5-Trimethylbenzene	11.451	105	3590	0.671	ug/l	96
82) 4-Chlorotoluene	11.457	91	3815	0.713	ug/l	97
83) tert-Butylbenzene	11.713	119	3951	0.751	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	3725	0.697	ug/l	97
85) sec-Butylbenzene	11.890	105	4385	0.672	ug/l	100
86) p-Isopropyltoluene	12.006	119	3577	0.679	ug/l	94
87) 1,3-Dichlorobenzene	11.969	146	1914	0.689	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	2149m	0.789	ug/l	
89) n-Butylbenzene	12.335	91	2889	0.631	ug/l	92
90) Hexachloroethane	12.536	117	604	0.616	ug/l	99
91) 1,2-Dichlorobenzene	12.341	146	2022	0.747	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	12.945	75	306	0.586	ug/l	83
93) 1,2,4-Trichlorobenzene	13.591	180	1019	0.688	ug/l	94
94) Hexachlorobutadiene	13.725	225	465	0.671	ug/l	89
95) Naphthalene	13.780	128	4138	0.767	ug/l	97
96) 1,2,3-Trichlorobenzene	13.963	180	1113	0.716	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations

Quant Time: May 06 06:13:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

