

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
 Lab Code: CHEM Case No.: Q1993 SAS No.: Q1993 SDG NO.: Q1993
 Lab File ID: VX046153.D Date Analyzed: 05/13/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:00
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	98502	5.54	167387	6.75	146749	10.05
UPPER LIMIT	197004	6.043	334774	7.25	293498	10.549
LOWER LIMIT	49251	5.043	83693.5	6.25	73374.5	9.549
EPA SAMPLE NO.						
EB-20250508	65629	5.54	132457	6.76	126910	10.05
TB	66175	5.54	131830	6.76	126040	10.06
VX0513WBL01	67277	5.55	133345	6.76	124650	10.05
VX0513WBS01	89387	5.54	160585	6.76	139963	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.