

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
 Lab Code: CHEM Case No.: Q1993 SAS No.: Q1993 SDG NO.: Q1993
 Lab File ID: VX046102.D Date Analyzed: 05/09/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:05
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	64593	12.018				
UPPER LIMIT	129186	12.518				
LOWER LIMIT	32296.5	11.518				
EPA SAMPLE NO.						
MW-3-20250508	49722	12.02				
MW-3-20250508MS	66944	12.02				
MW-3-20250508MSD	65108	12.02				
MW-3-20250508-A	51999	12.02				
MW-2-20250508	51775	12.02				
MW-1-20250508	50295	12.02				
MW-4-20250508	51028	12.02				
VX0509WBL01	52925	12.02				
VX0509WBS01	64755	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.