

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1994 SAS No.: Q1994 SDG NO.: Q1994

EPA Sample No.: SSTD020369 Date Analyzed: 05/20/2025

Lab File ID: BN037067.D Time Analyzed: 10:16

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	426404	17.01	426459	21.239	403677	24.122
UPPER LIMIT	852808	17.51	852918	21.739	807354	24.622
LOWER LIMIT	213202	16.51	213230	20.739	201839	23.622
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	367094	17.01	322665	21.24	294318	24.12

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.