

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q1994 SAS No.: Q1994 SDG NO.: Q1994
 Lab File ID: VX046102.D Date Analyzed: 05/09/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:05
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	90819	5.54	154646	6.75	135275	10.05
UPPER LIMIT	181638	6.044	309292	7.251	270550	10.549
LOWER LIMIT	45409.5	5.044	77323	6.251	67637.5	9.549
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	65371	5.54	129831	6.76	120192	10.05
STORAGE-BLANK-WATER-REF-5	70636	5.54	138613	6.76	127053	10.06
STORAGE-BLANK-WATER-REF-4	66237	5.55	131755	6.76	123029	10.05
STORAGE-BLANK-SAMPLE-MANAGEMENT	68764	5.54	136955	6.76	128312	10.05

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.