

# **DATA PACKAGE**

**GENERAL CHEMISTRY** 

PROJECT NAME: REAC - 83150.1.029.08ME.A.LPV818.0800

**TECHLAW CONSULTANTS, INC** 

14840 Conference Center Dr.

**Suite #200** 

**Chantily, VA - 20151** 

Phone No: 703-818-3205

**ORDER ID: Q1995** 

**ATTENTION: Jim Burden** 







# Table Of Contents for Q1995

1) Signature Page	3
2) Case Narrative	4
2.1) Genchem- Case Narrative	4
3) Qualifier Page	5
4) QA Checklist	6
5) Genchem Data	7
6) Shipping Document	26
6.1) CHAIN OF CUSTODY	27
6.2) Lab Certificate	28

Q1995 **2 of 28** 



# **Cover Page**

**Order ID:** Q1995

**Project ID:** REAC - 83150.1.029.08ME.A.LPV818.0800

Client: TechLaw Consultants, Inc

Lab Sample Number	Client Sample Number
Q1995-01	ERT-273
Q1995-02	ERT-274
Q1995-03	ERT-275
Q1995-04	ERT-276
Q1995-05	ERT-277
Q1995-06	ERT-278
Q1995-07	ERT-279

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

By Nimisha Pandya, QA/QC Supervisor at 11:01 am, May 20, 2025

NYDOH CERTIFICATION NO - 11376 NJDEP CERTIFICATION NO - 20012

5/20/2025

Date:

Q1995 3 of 28



#### **CASE NARRATIVE**

**TechLaw Consultants, Inc** 

Project Name: REAC - 83150.1.029.08ME.A.LPV818.0800

Project # N/A **Order ID # O1995 Test Name: TOC** 

#### A. Number of Samples and Date of Receipt:

7 Solid samples were received on 05/09/2025.

#### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: TOC. This data package contains results for TOC.

#### C. Analytical Techniques:

The analysis of TOC was based on method 9060A.

#### D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike analysis met criteria for all samples.

The Matrix Spike Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

#### E. Additional Comments:

Lab has used least representable sample weight for the sample Q1995-06 for TOC analysis. Therefore Lab has reported the TOC result with "OR" qualifier.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature By Nimisha Pandya, QA/QC Supervisor at 11:02 am, May 20, 2025



#### DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following "Results Qualifiers" are used:

- J Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
- U Indicates the analyte was analyzed for, but not detected.
- ND Indicates the analyte was analyzed for, but not detected
- E Indicates the reported value is estimated because of the presence of interference
- M Indicates Duplicate injection precision not met.
- N Indicates the spiked sample recovery is not within control limits.
- S Indicates the reported value was determined by the Method of Standard Addition (MSA).
- \* Indicates that the duplicate analysis is not within control limits.
- + Indicates the correlation coefficient for the MSA is less than 0.995.
- D Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
- M Method qualifiers
  - **"P"** for ICP instrument
  - "PM" for ICP when Microwave Digestion is used
  - "CV" for Manual Cold Vapor AA
  - "AV" for automated Cold Vapor AA
  - "CA" for MIDI-Distillation Spectrophotometric "AS" for Semi –Automated Spectrophotometric
  - "C" for Manual Spectrophotometric
  - **"T"** for Titrimetric
  - "NR" for analyte not required to be analyzed
- OR Indicates the analyte's concentration exceeds the calibrated range of the
  - instrument for that specific analysis.
- Q Indicates the LCS did not meet the control limits requirements
- H Sample Analysis Out Of Hold Time



#### APPENDIX A

#### **QA REVIEW GENERAL DOCUMENTATION**

Project #: Q1995

	Completed
For thorough review, the report must have the following:	
GENERAL:	
Are all original paperwork present (chain of custody, record of communication, airbill, sample management lab chronicle, login page)	<u> </u>
Check chain-of-custody for proper relinquish/return of samples	<u> </u>
Is the chain of custody signed and complete	<u> </u>
Check internal chain-of-custody for proper relinquish/return of samples /sample extracts	<del>'</del> <del>'</del> <del>'</del> <del>'</del> <del>'</del> <del>'</del>
Collect information for each project id from server. Were all requirements followed	<u> </u>
COVER PAGE:	
Do numbers of samples correspond to the number of samples in the Chain of Custody on login page	<u> </u>
Do lab numbers and client Ids on cover page agree with the Chain of Custody	<u> </u>
CHAIN OF CUSTODY:	
Do requested analyses on Chain of Custody agree with form I results	<u> </u>
Do requested analyses on Chain of Custody agree with the log-in page	<u> </u>
Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody	<u> </u>
Were the samples received within hold time	<u> </u>
Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle	<u> </u>
ANALYTICAL:	
Was method requirement followed?	<u> </u>
Was client requirement followed?	<u> </u>
Does the case narrative summarize all QC failure?	<del>'</del> <del>'</del> <del>'</del> <del>'</del> <del>'</del> <del>'</del>
All runlogs and manual integration are reviewed for requirements	<u> </u>
All manual calculations and /or hand notations verified	<u> </u>

QA Review Signature: SOHIL JODHANI Date: 05/20/2025

Q1995 **6 of 28** 



#### **LAB CHRONICLE**

OrderID: Q1995

Client: TechLaw Consultants, Inc

Contact: Jim Burden

**OrderDate:** 5/9/2025 10:41:40 AM

**Project:** REAC - 83150.1.029.08ME.A.LPV818.0800

Location: L51

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1995-01	ERT-273	SOIL			05/06/25 11:15			05/09/25
			TOC	9060A			05/16/25 09:30	
Q1995-02	ERT-274	SOIL			05/07/25 08:05			05/09/25
			TOC	9060A			05/16/25 09:41	
Q1995-03	ERT-275	SOIL			05/07/25 08:15			05/09/25
			TOC	9060A			05/16/25 09:50	
Q1995-04	ERT-276	SOIL			05/07/25 09:50			05/09/25
			TOC	9060A			05/16/25 10:16	
Q1995-05	ERT-277	SOIL			05/07/25 10:05			05/09/25
			TOC	9060A			05/16/25 10:27	
Q1995-06	ERT-278	SOIL			05/07/25 10:07			05/09/25
			тос	9060A			05/16/25 11:07	
Q1995-07	ERT-279	SOIL			05/07/25 10:10			05/09/25
			TOC	9060A			05/16/25 11:49	

Q1995 **7 of 28** 



# 5

Α

C

SAMPLE DATA

SOIL



284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900, Fax: 908 789 8922

Matrix:

#### **Report of Analysis**

Client: TechLaw Consultants, Inc Date Collected: 05/06/25 11:15

Project: REAC - 83150.1.029.08ME.A.LPV818.0800 Date Received: 05/09/25

Client Sample ID: SDG No.: Q1995 ERT-273 Lab Sample ID: Q1995-01

94.9 % Solid:

Parameter	Conc. Qua.	DF MD	LOQ / CRQL	Units(Dry Weight) Prep Date	Date Ana.	Ana Met.
TOC	1600	1 28.9	250	mg/Kg	05/16/25 09:30	9060A

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

9 of 28



Fax: 908 789 8922

#### **Report of Analysis**

Client: TechLaw Consultants, Inc Date Collected: 05/07/25 08:05

Project: REAC - 83150.1.029.08ME.A.LPV818.0800 Date Received: 05/09/25

Client Sample ID: SDG No.: Q1995 ERT-274 Lab Sample ID: Q1995-02 Matrix: SOIL

> % Solid: 72.4

Parameter	Conc. Qua.	DF MDL	LOQ / CRQL	Units(Dry Weight) Prep Date	Date Ana.	Ana Met.
TOC	3600	1 28.9	250	mg/Kg	05/16/25 09:41	9060A

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

10 of 28

Q1995



Fax: 908 789 8922

#### **Report of Analysis**

Client: TechLaw Consultants, Inc Date Collected: 05/07/25 08:15

Project: REAC - 83150.1.029.08ME.A.LPV818.0800 Date Received: 05/09/25

Client Sample ID: SDG No.: Q1995 ERT-275 Lab Sample ID: Q1995-03 Matrix: SOIL

97.9 % Solid:

Parameter	Conc. Qua.	DF MDL	LOQ / CRQL	Units(Dry Weight) Prep Date	Date Ana. Ana	Met.
TOC	2950	1 28.9	250	mg/Kg	05/16/25 09:50 900	60A

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Q1995

Matrix:

SOIL



Lab Sample ID:

Fax: 908 789 8922

Q1995-04

#### **Report of Analysis**

Client: TechLaw Consultants, Inc Date Collected: 05/07/25 09:50

Project: REAC - 83150.1.029.08ME.A.LPV818.0800 Date Received: 05/09/25

Client Sample ID: ERT-276 SDG No.: Q1995

% Solid: 82.5

Parameter	Conc. Qua.	DF MDL	LOQ / CRQL Units(Dry V	Veight) Prep Date Dat	te Ana. Ana Met.
TOC	1690	1 28.9	250 mg/k	g 05/	16/25 10:16 9060A

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Q1995

12 of 28



#### **Report of Analysis**

Client: TechLaw Consultants, Inc Date Collected: 05/07/25 10:05

Project: REAC - 83150.1.029.08ME.A.LPV818.0800 Date Received: 05/09/25

Client Sample ID: SDG No.: Q1995 ERT-277 Lab Sample ID: Q1995-05 Matrix: SOIL

> % Solid: 88

Parameter	Conc. Qua.	DF MDL	LOQ / CRQL	Units(Dry Weight) Prep Date	Date Ana. Ana Met.
TOC	6500	1 28.9	250	mg/Kg	05/16/25 10:27 9060A

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Q1995



#### **Report of Analysis**

Client: TechLaw Consultants, Inc Date Collected: 05/07/25 10:07

Project: REAC - 83150.1.029.08ME.A.LPV818.0800 Date Received: 05/09/25

Client Sample ID: SDG No.: Q1995 ERT-278 Lab Sample ID: Q1995-06 Matrix: SOIL

> % Solid: 77.9

Parameter	Conc. Qua. DF MDL	LOQ / CRQL Units(Dry Weight) Prep Date	Date Ana. Ana Met.
TOC	11400 OR 1 28.9	250 mg/Kg	05/16/25 11:07 9060A

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

14 of 28



#### **Report of Analysis**

Client: TechLaw Consultants, Inc Date Collected: 05/07/25 10:10

Project: REAC - 83150.1.029.08ME.A.LPV818.0800 Date Received: 05/09/25

Client Sample ID: ERT-279 SDG No.: Q1995

Lab Sample ID: Q1995-07 Matrix: SOIL

% Solid: 85.7

Parameter	Conc. Qua.	DF MDL	LOQ / CRQL	Units(Dry Weight) Prep Date	Date Ana. Ana Met.	
TOC	4740	1 28.9	250	mg/Kg	05/16/25 11:49 9060A	

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Q1995



# QC RESULT SUMMARY

5

Α



D



### **Initial and Continuing Calibration Verification**

Client: TechLaw Consultants, Inc SDG No.: Q1995

**Project:** REAC - 83150.1.029.08ME.A.LPV818.0800 **RunNo.:** LB135795

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1	mg/L	1030	1000	103	90-110	03/14/2025
Sample ID:	CCV1	mg/L	992	1000	99	90-110	05/16/2025
Sample ID:	CCV2	mg/L	1030	1000	103	90-110	05/16/2025
Sample ID:	CCV3	mg/L	1030	1000	103	90-110	05/16/2025

Q1995 **17 of 28** 



В

С



**Initial and Continuing Calibration Verification** 

Client: TechLaw Consultants, Inc SDG No.: Q1995

**Project:** REAC - 83150.1.029.08ME.A.LPV818.0800 **RunNo.:** LB135795

Q1995 **18 of 28** 



### **Initial and Continuing Calibration Blank Summary**

Client: TechLaw Consultants, Inc SDG No.: Q1995

**Project:** REAC - 83150.1.029.08ME.A.LPV818.0800 **RunNo.:** LB135795

Analyte		Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID:	ICB1	mg/L	< 125.0000	125.0000	Ū	22.3	250	03/14/2025
Sample ID:	CCB1	mg/L	< 125.0000	125.0000	U	22.3	250	05/16/2025
Sample ID:	CCB2	mg/L	< 125.0000	125.0000	U	22.3	250	05/16/2025
Sample ID:	CCB3	mg/L	< 125.0000	125.0000	U	22.3	250	05/16/2025

Q1995 **19 of 28** 



В

С



#### **Initial and Continuing Calibration Blank Summary**

Client: TechLaw Consultants, Inc SDG No.: Q1995

Project: REAC - 83150.1.029.08ME.A.LPV818.0800 RunNo.: LB135795

Q1995 **20 of 28** 



# **Preparation Blank Summary**

Client: TechLaw Consultants, Inc SDG No.: Q1995

**Project:** REAC - 83150.1.029.08ME.A.LPV818.0800

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID:	LB135795BLS mg/Kg	< 125.0000	125.0000	U	28.9	250	05/16/2025

Q1995 **21 of 28** 



# **Matrix Spike Summary**

Client: TechLaw Consultants, Inc SDG No.: Q1995

**Project:** REAC - 83150.1.029.08ME.A.LPV818.0800 **Sample ID:** Q1995-05

Client ID: ERT-277MS Percent Solids for Spike Sample: 88

		Acceptance	Spiked	Conc.	Sample	Conc.	Spike	Dilution	<b>%</b>		Analysis
Analyte	Units	Limit %R	Result	Qualifier	Result	Qualifier	Added	Factor	Rec	Qual	Date
TOC	mg/Kg	75-125	7460		6500		1000	1	96		05/16/2025

5

А

С



Q1995 **22 of 28** 



# **Matrix Spike Summary**

Client: TechLaw Consultants, Inc SDG No.: Q1995

**Project:** REAC - 83150.1.029.08ME.A.LPV818.0800 **Sample ID:** Q1995-05

Client ID: ERT-277MSD Percent Solids for Spike Sample: 88

		Acceptance	Spiked	Conc.	Sample	Conc.	Spike	Dilution	%		Analysis
Analyte	Units	Limit %R	Result	Qualifier	Result	Qualifier	Added	Factor	Rec	Qual	Date
TOC	mg/Kg	75-125	7560		6500		1000	1	106		05/16/2025

A

\_

Q1995 **23 of 28** 



# **Duplicate Sample Summary**

Client: TechLaw Consultants, Inc SDG No.: Q1995

**Project:** REAC - 83150.1.029.08ME.A.LPV818.0800 **Sample ID:** Q1995-05

Client ID: ERT-277MSD Percent Solids for Spike Sample: 88

		Acceptance	Sample	Conc.	Duplicate	Conc.	Dilution	RPD/		Analysis	
Analyte	Units	Limit	Result	Qualifier	Result	Qualifier	Factor	AD	Qual	Date	
тос	mg/Kg	+/-20	7460		7560		1	1		05/16/2025	_

Α

С

\_

Q1995 **24 of 28** 



### **Laboratory Control Sample Summary**

Client: TechLaw Consultants, Inc SDG No.: Q1995

**Project:** REAC - 83150.1.029.08ME.A.LPV818.0800 **Run No.:** LB135795

Analyte		Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135795BSS								
TOC		mg/Kg	1000	1010		101	1	90-110	05/16/2025

Q1995 **25 of 28** 



# SHIPPING DOCUMENTS

Q1995 **26 of 28** 

#### USEPA / REAC

DateShipped: 5/8/2025 CarrierName: FedEx

Special Instructions:

Project Code: TechLaw REAC

#### **CHAIN OF CUSTODY RECORD**

Site #: MTD986073252 Contact Name: Jean Bolduc Contact Phone: 949-321-4280 01995

No: 8-050825-092440-0016

SAMPLES TRANSFERRED FROM

**CHAIN OF CUSTODY #** 

Cooler #: 1 Lab: CHEMTECH Lab Phone: 908-728-3144

Lab #	Sample #	Location	Sub Location	Analyses	Matrix	Sample Date	Sample Time	Numb Cont	Container	Preservative	Lab QC
	ERT-273	DL-08B	9' - 10'	TOC	Soil	5/6/2025	11:15	1	16 oz Glass Jar	4 C	N
	ERT-273	DL-08B	9' - 10'	% Moisture	Soil	5/6/2025	11:15	1	2 oz Glass Jar	4 C	N
	ERT-274	DL-09	19.5' - 20.5'	TOC	Soil	5/7/2025	08:05	1	16 oz Glass Jar	4 C	N
	ERT-274	DL-09	19.5' - 20.5'	% Moisture	Soil	5/7/2025	08:05	1	2 oz Glass Jar	4 C	N
	ERT-275	DL-09	24' - 25'	TOC	Soil	5/7/2025	08:15	1	16 oz Glass Jar	4 C	N
	ERT-275	DL-09	24' - 25'	% Moisture	Soil	5/7/2025	08:15	1	2 oz Glass Jar	4 C	N
	ERT-276	DL-14	14' - 15'	TOC	Soil	5/7/2025	09:50	1	16 oz Glass Jar	4 C	N
	ERT-276	DL-14	14' - 15'	% Moisture	Soil	5/7/2025	09:50	.1	2 oz Glass Jar	4 C	N
	ERT-277	DL-14	19' - 20'	TOC	Soil	5/7/2025	10:05	1	16 oz Glass Jar	4 C	N
	ERT-277	DL-14	19' - 20'	% Moisture	Soil	5/7/2025	10:05	1	2 oz Glass Jar	4 C	N
	ERT-278	DL-14	24' - 25'	TOC	Soil	5/7/2025	10:07	1	16 oz Glass Jar	4 C	N
	ERT-278	DL-14	24' - 25'	% Moisture	Soil	5/7/2025	10:07	1	2 oz Glass Jar	4 C	N
	ERT-279	DL-14	20' - 22'	TOC	Soil	5/7/2025	10:10	1	16 oz Glass Jar	4 C	N
	ERT-279	DL-14	20' - 22'	% Moisture	Soil	5/7/2025	10:10	1	2 oz Glass Jar	4 C	N
											-
											17

Items/Reason	Relinquished by (Signature and Organization)	Date/Time	Received by (Signature and Organization)	Date/Time	Sample Condition Upon Receipt
122 / ANNEYSIS	lett EPA/EST	5/8/2025	O.	10:18	IR.Gn#1 4.7.C
					Custody Seal Intact
					NO Temp Blank
					PADJUST FACTOR + 11



# Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488