

Initial and Continuing Calibration Verification

Client: TechLaw Consultants, Inc

SDG No.: Q1995

Project: REAC - 83150.1.029.08ME.A.LPV818.0800

RunNo.: LB135795

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 TOC	mg/L	1030	1000	103	90-110	03/14/2025
Sample ID: CCV1 TOC	mg/L	992	1000	99	90-110	05/16/2025
Sample ID: CCV2 TOC	mg/L	1030	1000	103	90-110	05/16/2025
Sample ID: CCV3 TOC	mg/L	1030	1000	103	90-110	05/16/2025

Initial and Continuing Calibration Verification

Client: TechLaw Consultants, Inc

SDG No.: Q1995

Project: REAC - 83150.1.029.08ME.A.LPV818.0800

RunNo.: LB135795

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------