

Laboratory Control Sample Summary

Client:	TechLaw Consultants, Inc	SDG No.:	Q1995
Project:	REAC - 83150.1.029.08ME.A.LPV818.0800	Run No.:	LB135795

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135795BSS							
TOC	mg/Kg	1000	1010		101	1	90-110	05/16/2025