

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ALLI03
Lab Code: CHEM Case No.: Q2000 SAS No.: Q2000 SDG No.: Q2000
Instrument ID: BNA_F Calibration Date(s): 05/05/2025 05/05/2025
Calibration Time(s): 13:54 17:15

LAB FILE ID:		RRF2.5 = BF142295.D		RRF005 = BF142296.D		RRF010 = BF142297.D		
		RRF020 = BF142298.D		RRF040 = BF142299.D		RRF050 = BF142300.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.212	1.177	1.211	1.126	1.224	1.172	4.4
Phenol-d6		1.510	1.458	1.489	1.375	1.500	1.441	4.6
Nitrobenzene-d5		0.289	0.304	0.336	0.327	0.360	0.328	7.5
2,4-Dimethylphenol		0.311	0.314	0.321	0.298	0.329	0.312	3.5
2-Fluorobiphenyl		1.737	1.605	1.532	1.272	1.337	1.403	16.3
2,4,6-Tribromophenol		0.161	0.176	0.198	0.191	0.218	0.193	10.2
Terphenyl-d14		1.493	1.424	1.460	1.255	1.390	1.351	9.6