

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ALLI03
Lab Code: CHEM Case No.: Q2000 SAS No.: Q2000 SDG No.: Q2000
Instrument ID: BNA_P Calibration Date(s): 05/13/2025 05/13/2025
Calibration Time(s): 10:41 15:26

LAB FILE ID:		RRF2.5 = BP024614.D		RRF005 = BP024615.D		RRF010 = BP024616.D		
		RRF020 = BP024617.D		RRF040 = BP024618.D		RRF050 = BP024619.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.134	1.096	1.162	1.105	1.272	1.169	5.5
Phenol-d6		1.301	1.344	1.486	1.472	1.671	1.492	9.1
Nitrobenzene-d5		0.381	0.376	0.401	0.382	0.429	0.396	4.7
2,4-Dimethylphenol		0.273	0.279	0.300	0.298	0.329	0.301	6.6
2-Fluorobiphenyl		1.601	1.498	1.553	1.432	1.602	1.527	4.1
2,4,6-Tribromophenol		0.324	0.321	0.335	0.323	0.369	0.339	5.5
Terphenyl-d14		1.192	1.113	1.165	1.125	1.226	1.166	3.3