

Surrogate Summary

SW-846

SDG No.: Q2000

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167951BL	PB167951BL	2-Fluorophenol	150	123	82		10	139
		Phenol-d6	150	119	79		10	134
		2,4,6-Tribromophenol	150	137	92		44	137
PB167951BS	PB167951BS	2-Fluorophenol	150	121	80		10	139
		Phenol-d6	150	122	81		10	134
		2,4,6-Tribromophenol	150	144	96		44	137
Q1993-02MS	MW-3-20250508MS	2-Fluorophenol	150	63.4	42		10	139
		Phenol-d6	150	41.2	27		10	134
		2,4,6-Tribromophenol	150	139	93		44	137
Q1993-03MSD	MW-3-20250508MSD	2-Fluorophenol	150	67.3	45		10	139
		Phenol-d6	150	44.1	29		10	134
		2,4,6-Tribromophenol	150	149	99		44	137
Q2000-02	38073-040323	2-Fluorophenol	150	159	106		10	139
		Phenol-d6	150	165	110		10	134
		2,4,6-Tribromophenol	150	160	107		44	137
Q2000-02DL	38073-040323DL	2-Fluorophenol	150	152	101		10	139
		Phenol-d6	150	152	101		10	134
		2,4,6-Tribromophenol	150	153	102		44	137