

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2000

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E DataFile: BF142338.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB167951BS	2,4-Dimethylphenol	50	45.2	ug/L	90				42	142	