

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark		Date Collected:	05/08/25
Project:	NJ Waste Water PT		Date Received:	05/09/25
Client Sample ID:	38072-062223DL		SDG No.:	Q2000
Lab Sample ID:	Q2000-01DL		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOCMS Group5
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037012.D	50	05/13/25 08:49	05/14/25 13:29	PB167952

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
208-96-8	Acenaphthylene	64.2	D	1.90	5.00	ug/L
SURROGATES						
7297-45-2	2-Methylnaphthalene-d10	0.45		20 - 139	113%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.45		30 - 150	113%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.50		27 - 154	125%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.40		25 - 149	100%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.50		54 - 175	125%	SPK: 0.4
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	2750		7.618		
1146-65-2	Naphthalene-d8	7480		10.394		
15067-26-2	Acenaphthene-d10	4290		14.267		
1517-22-2	Phenanthrene-d10	8580		17.009		
1719-03-5	Chrysene-d12	8430		21.215		
1520-96-3	Perylene-d12	7350		23.418		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products