



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

### **ENTACT**

**Project Name: 540 Degraw St, Brooklyn, NY - E9309**

**Project # N/A**

**Order ID # Q2001**

**Test Name: ASTM Ammonia,ASTM COD,ASTM Oil and Grease,ASTM TS,Corrosivity,Ignitability,Oil and Grease,Paint Filter,pH,Reactive Cyanide,Reactive Sulfide,TS,TVS**

### **A. Number of Samples and Date of Receipt:**

16 Solid samples were received on 05/09/2025.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: ASTM Ammonia, ASTM COD, ASTM Leach Extraction, ASTM Oil and Grease, ASTM TS, Corrosivity, Ignitability, Oil and Grease, Paint Filter, PCB, pH, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, TCLP Metals Group2, TS and TVS. This data package contains results for ASTM Ammonia,ASTM COD,ASTM Oil and Grease,ASTM TS,Corrosivity,Ignitability,Oil and Grease,Paint Filter,pH,Reactive Cyanide,Reactive Sulfide,TS,TVS.

### **C. Analytical Techniques:**

The analysis of Ignitability was based on method 1030, The analysis of TVS was based on method 160.4, The analysis of ASTM Oil and Grease was based on method 1664A, The analysis of Reactive Cyanide was based on method 9012B, The analysis of Reactive Sulfide was based on method 9034, The analysis of Corrosivity,pH was based on method 9045D, The analysis of Oil and Grease was based on method 9071B, The analysis of Paint Filter was based on method 9095B, The analysis of ASTM TS,TS was based on method SM2540 B, The analysis of ASTM Ammonia was based on method SM4500-NH3 and The analysis of ASTM COD was based on method SM5220 D.

### **D. QA/ QC Samples:**

The Holding Times were met for all samples except for WC-A1-05-C of pH, for WC-A1-05-C of Corrosivity, for WC-A1-06-C of pH, for WC-A1-06-C of Corrosivity, for WC-A1-07-C of pH, for WC-A1-07-C of Corrosivity, for WC-A4-03-C of pH and for WC-A4-03-C of Corrosivity. As these samples are received out of hold.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike (WC-A4-03-CMS) analysis met criteria for all samples except for Oil and Grease due to matrix interference.



The Matrix Spike Duplicate (WC-A4-03-CMSD) analysis met criteria for all samples except for Oil and Grease due to matrix interference.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

**E. Additional Comments:**

As per method 1664A, MS/MSD is required to be performed with the sample analysis. However, Lab did not receive sufficient volume to perform the MS/MSD therefore MS/MSD were not performed for this project.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_