

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2001

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135741

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	6.99	7	100	90-110	05/12/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	05/12/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	05/12/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2001

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135742

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	6.99	7	100	90-110	05/12/2025
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	05/12/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	05/12/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2001

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135764

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.094	0.099	95	85-115	05/14/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	05/14/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	05/14/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	05/14/2025
Sample ID: CCV4 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	05/14/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2001

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135772

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV ASTM COD	mg/L	48.972	50	98	95-105	05/06/2025
Sample ID: CCV1 ASTM COD	mg/L	51.007	50	102	95-105	05/14/2025
Sample ID: CCV2 ASTM COD	mg/L	51.007	50	102	95-105	05/14/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2001

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135775

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 ASTM Ammonia	mg/L	0.98	1	98	90-110	05/14/2025
Sample ID: CCV1 ASTM Ammonia	mg/L	0.93	1	93	90-110	05/14/2025
Sample ID: CCV2 ASTM Ammonia	mg/L	0.94	1	94	90-110	05/14/2025
Sample ID: CCV3 ASTM Ammonia	mg/L	1	1	100	90-110	05/14/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2001

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135775

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------