

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q2001
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q1980-04
Client ID:	3848DUP	Percent Solids for Spike Sample:	100

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ignitability	oC	+/-20	NO		NO		1	0		05/12/2025

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q2001
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2001-02
Client ID:	WC-A4-03-CDUP	Percent Solids for Spike Sample:	79.1

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Paint Filter	ml/100gm	+/-20	1.00	U	1.00	U	1	0		05/14/2025
Oil and Grease	mg/Kg	+/-20	732		744		1	1.66		05/14/2025

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q2001
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2001-02
Client ID:	WC-A4-03-CMSD	Percent Solids for Spike Sample:	79.1

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/Kg	+/-20	940		946		1	0.57		05/14/2025

Duplicate Sample Summary

Client: ENTACT	SDG No.: Q2001
Project: 540 Degraw St, Brooklyn, NY - E9309	Sample ID: Q2001-03
Client ID: WC-A4-03-CDUP	Percent Solids for Spike Sample: 100

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Reactive Sulfide	mg/Kg	+/-20	3.16	J	3.16	J	1	0		05/12/2025
Reactive Cyanide	mg/Kg	+/-20	0.0084	U	0.0084	U	1	0		05/14/2025

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: WC-A4-03-CDUP	SDG No.: Q2001 Sample ID: Q2001-04 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
ASTM COD	mg/L	+/-20	64.2		63.2		1	1.57		05/14/2025
ASTM Ammonia	mg/L	+/-20	0.62		0.66		1	6		05/14/2025

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: WC-A4-03-CMSD	SDG No.: Q2001 Sample ID: Q2001-04 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
ASTM COD	mg/L	+/-20	109		108		1	0.92		05/14/2025
ASTM Ammonia	mg/L	+/-20	1.70		1.60		1	6		05/14/2025

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: WC-A1-07-CDUP	SDG No.: Q2001 Sample ID: Q2001-14 Percent Solids for Spike Sample: 100
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TS	%	+/-5	81.8		82.2		1	0.49		05/12/2025
TVS	%	+/-5	3.40	J	3.30	J	1	2.99		05/12/2025
pH	pH	+/-20	12.3		12.3		1	0.08		05/12/2025

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q2001
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2001-15
Client ID:	WC-A1-07-CDUP	Percent Solids for Spike Sample:	100

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Corrosivity	pH	+/-20	12.3		12.3		1	0.08		05/12/2025

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q2001
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2001-16
Client ID:	WC-A1-07-CDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
ASTM TS	mg/L	+/-5	1490		1510		1	1.4		05/13/2025
ASTM Oil and Grease	mg/L	+/-18	0.30	J	0.30	J	1	0		05/14/2025

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	Q2001
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2001-16
Client ID:	WC-A1-07-CMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
ASTM Oil and Grease	mg/L	+/-18	20.4		20.4		1	0		05/14/2025