

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2001

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135759

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135759BS							
ASTM Oil and Grease	mg/L	20.0	16.9		84	1	78-114	05/14/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2001

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135760

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135760BS							
Oil and Grease	mg/Kg	100	94.9		95	1	80-120	05/14/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2001

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135772

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135772BS							
ASTM COD	mg/L	50	48.0		96	1	90-110	05/14/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2001

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135775

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167986BS							
ASTM Ammonia	mg/L	1	0.98		98	1	90-110	05/14/2025