

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** Q2001
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2001 **SAS No.:** Q2001
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.18	0.2	89	70 - 130	CV	05/14/2025	11:54	LB135763
CRI01	Arsenic	19.9	20.0	99	65 - 135	P	05/14/2025	13:43	LB135778
	Barium	85.9	100	86	65 - 135	P	05/14/2025	13:43	LB135778
	Cadmium	6.01	6.0	100	65 - 135	P	05/14/2025	13:43	LB135778
	Chromium	9.05	10.0	90	65 - 135	P	05/14/2025	13:43	LB135778
	Copper	22.3	20.0	112	65 - 135	P	05/14/2025	13:43	LB135778
	Lead	11.1	12.0	92	65 - 135	P	05/14/2025	13:43	LB135778
	Nickel	38.8	40.0	97	65 - 135	P	05/14/2025	13:43	LB135778
	Selenium	22.6	20.0	113	65 - 135	P	05/14/2025	13:43	LB135778
	Silver	9.49	10.0	95	65 - 135	P	05/14/2025	13:43	LB135778
	Zinc	42.2	40.0	106	65 - 135	P	05/14/2025	13:43	LB135778
CRI01	Arsenic	21.4	20.0	107	65 - 135	P	05/15/2025	13:51	LB135794
	Barium	94.3	100	94	65 - 135	P	05/15/2025	13:51	LB135794
	Cadmium	6.05	6.0	101	65 - 135	P	05/15/2025	13:51	LB135794
	Chromium	9.78	10.0	98	65 - 135	P	05/15/2025	13:51	LB135794
	Copper	22.2	20.0	111	65 - 135	P	05/15/2025	13:51	LB135794
	Lead	12.3	12.0	102	65 - 135	P	05/15/2025	13:51	LB135794
	Nickel	39.1	40.0	98	65 - 135	P	05/15/2025	13:51	LB135794
	Selenium	22.4	20.0	112	65 - 135	P	05/15/2025	13:51	LB135794
	Silver	10.1	10.0	101	65 - 135	P	05/15/2025	13:51	LB135794
	Zinc	42.5	40.0	106	65 - 135	P	05/15/2025	13:51	LB135794
CRI01	Arsenic	19.3	20.0	96	65 - 135	P	05/16/2025	13:09	LB135802
	Barium	93.2	100	93	65 - 135	P	05/16/2025	13:09	LB135802
	Cadmium	6.13	6.0	102	65 - 135	P	05/16/2025	13:09	LB135802
	Chromium	9.83	10.0	98	65 - 135	P	05/16/2025	13:09	LB135802
	Copper	22.1	20.0	110	65 - 135	P	05/16/2025	13:09	LB135802
	Lead	13.3	12.0	111	65 - 135	P	05/16/2025	13:09	LB135802
	Nickel	39.6	40.0	99	65 - 135	P	05/16/2025	13:09	LB135802
	Selenium	18.9	20.0	95	65 - 135	P	05/16/2025	13:09	LB135802
	Silver	10.0	10.0	100	65 - 135	P	05/16/2025	13:09	LB135802
	Zinc	42.3	40.0	106	65 - 135	P	05/16/2025	13:09	LB135802