



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT				SDG No.:	Q2001				
Contract:	ENTA05	Lab Code:	CHEM		Case No.:	Q2001	SAS No.:	Q2001		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB08	Mercury	0.20	+/-0.20	U	0.20	CV	05/14/2025	11:44	LB135763	



- 3a -

Client:	ENTACT				SDG No.:	Q2001			
Contract:	ENTA05	Lab Code:	CHEM		Case No.:	Q2001	SAS No.:	Q2001	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB22	Mercury	0.20	+/-0.20	U	0.20	CV	05/14/2025	11:51	LB135763
CCB23	Mercury	0.20	+/-0.20	U	0.20	CV	05/14/2025	12:26	LB135763
CCB24	Mercury	0.20	+/-0.20	U	0.20	CV	05/14/2025	12:54	LB135763
CCB25	Mercury	0.20	+/-0.20	U	0.20	CV	05/14/2025	13:30	LB135763
CCB26	Mercury	0.20	+/-0.20	U	0.20	CV	05/14/2025	13:57	LB135763
CCB27	Mercury	0.20	+/-0.20	U	0.20	CV	05/14/2025	14:25	LB135763
CCB28	Mercury	0.20	+/-0.20	U	0.20	CV	05/14/2025	14:41	LB135763
CCB29	Mercury	0.20	+/-0.20	U	0.20	CV	05/14/2025	14:59	LB135763

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q2001</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2001</u>	SAS No.: <u>Q2001</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/14/2025	13:38	LB135778
	Barium	100	+/-100	U	100	P	05/14/2025	13:38	LB135778
	Cadmium	6.00	+/-6.00	U	6.00	P	05/14/2025	13:38	LB135778
	Chromium	10.0	+/-10.0	U	10.0	P	05/14/2025	13:38	LB135778
	Copper	20.0	+/-20.0	U	20.0	P	05/14/2025	13:38	LB135778
	Lead	12.0	+/-12.0	U	12.0	P	05/14/2025	13:38	LB135778
	Nickel	40.0	+/-40.0	U	40.0	P	05/14/2025	13:38	LB135778
	Selenium	20.0	+/-20.0	U	20.0	P	05/14/2025	13:38	LB135778
	Silver	10.0	+/-10.0	U	10.0	P	05/14/2025	13:38	LB135778
	Zinc	40.0	+/-40.0	U	40.0	P	05/14/2025	13:38	LB135778

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT		SDG No.: Q2001							
Contract: ENTA05	Lab Code: CHEM	Case No.: Q2001	SAS No.: Q2001						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/14/2025	14:56	LB135778
	Barium	100	+/-100	U	100	P	05/14/2025	14:56	LB135778
	Cadmium	6.00	+/-6.00	U	6.00	P	05/14/2025	14:56	LB135778
	Chromium	10.0	+/-10.0	U	10.0	P	05/14/2025	14:56	LB135778
	Copper	20.0	+/-20.0	U	20.0	P	05/14/2025	14:56	LB135778
	Lead	12.0	+/-12.0	U	12.0	P	05/14/2025	14:56	LB135778
	Nickel	40.0	+/-40.0	U	40.0	P	05/14/2025	14:56	LB135778
	Selenium	20.0	+/-20.0	U	20.0	P	05/14/2025	14:56	LB135778
	Silver	10.0	+/-10.0	U	10.0	P	05/14/2025	14:56	LB135778
	Zinc	40.0	+/-40.0	U	40.0	P	05/14/2025	14:56	LB135778
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	05/14/2025	15:57	LB135778
	Barium	100	+/-100	U	100	P	05/14/2025	15:57	LB135778
	Cadmium	6.00	+/-6.00	U	6.00	P	05/14/2025	15:57	LB135778
	Chromium	10.0	+/-10.0	U	10.0	P	05/14/2025	15:57	LB135778
	Copper	20.0	+/-20.0	U	20.0	P	05/14/2025	15:57	LB135778
	Lead	12.0	+/-12.0	U	12.0	P	05/14/2025	15:57	LB135778
	Nickel	40.0	+/-40.0	U	40.0	P	05/14/2025	15:57	LB135778
	Selenium	20.0	+/-20.0	U	20.0	P	05/14/2025	15:57	LB135778
	Silver	10.0	+/-10.0	U	10.0	P	05/14/2025	15:57	LB135778
	Zinc	40.0	+/-40.0	U	40.0	P	05/14/2025	15:57	LB135778
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	05/14/2025	16:48	LB135778
	Barium	100	+/-100	U	100	P	05/14/2025	16:48	LB135778
	Cadmium	6.00	+/-6.00	U	6.00	P	05/14/2025	16:48	LB135778
	Chromium	10.0	+/-10.0	U	10.0	P	05/14/2025	16:48	LB135778
	Copper	20.0	+/-20.0	U	20.0	P	05/14/2025	16:48	LB135778
	Lead	12.0	+/-12.0	U	12.0	P	05/14/2025	16:48	LB135778
	Nickel	40.0	+/-40.0	U	40.0	P	05/14/2025	16:48	LB135778
	Selenium	20.0	+/-20.0	U	20.0	P	05/14/2025	16:48	LB135778
	Silver	10.0	+/-10.0	U	10.0	P	05/14/2025	16:48	LB135778
	Zinc	40.0	+/-40.0	U	40.0	P	05/14/2025	16:48	LB135778
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	05/14/2025	18:12	LB135778
	Barium	100	+/-100	U	100	P	05/14/2025	18:12	LB135778
	Cadmium	6.00	+/-6.00	U	6.00	P	05/14/2025	18:12	LB135778
	Chromium	10.0	+/-10.0	U	10.0	P	05/14/2025	18:12	LB135778
	Copper	20.0	+/-20.0	U	20.0	P	05/14/2025	18:12	LB135778
	Lead	12.0	+/-12.0	U	12.0	P	05/14/2025	18:12	LB135778
	Nickel	40.0	+/-40.0	U	40.0	P	05/14/2025	18:12	LB135778
	Selenium	20.0	+/-20.0	U	20.0	P	05/14/2025	18:12	LB135778
	Silver	10.0	+/-10.0	U	10.0	P	05/14/2025	18:12	LB135778

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q2001</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2001</u>	SAS No.: <u>Q2001</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Zinc	40.0	+/-40.0	U	40.0	P	05/14/2025	18:12	LB135778
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	05/14/2025	19:25	LB135778
	Barium	100	+/-100	U	100	P	05/14/2025	19:25	LB135778
	Cadmium	6.00	+/-6.00	U	6.00	P	05/14/2025	19:25	LB135778
	Chromium	10.0	+/-10.0	U	10.0	P	05/14/2025	19:25	LB135778
	Copper	20.0	+/-20.0	U	20.0	P	05/14/2025	19:25	LB135778
	Lead	12.0	+/-12.0	U	12.0	P	05/14/2025	19:25	LB135778
	Nickel	40.0	+/-40.0	U	40.0	P	05/14/2025	19:25	LB135778
	Selenium	20.0	+/-20.0	U	20.0	P	05/14/2025	19:25	LB135778
	Silver	10.0	+/-10.0	U	10.0	P	05/14/2025	19:25	LB135778
	Zinc	40.0	+/-40.0	U	40.0	P	05/14/2025	19:25	LB135778

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q2001</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2001</u>	SAS No.: <u>Q2001</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/15/2025	13:47	LB135794
	Barium	100	+/-100	U	100	P	05/15/2025	13:47	LB135794
	Cadmium	6.00	+/-6.00	U	6.00	P	05/15/2025	13:47	LB135794
	Chromium	10.0	+/-10.0	U	10.0	P	05/15/2025	13:47	LB135794
	Copper	20.0	+/-20.0	U	20.0	P	05/15/2025	13:47	LB135794
	Lead	12.0	+/-12.0	U	12.0	P	05/15/2025	13:47	LB135794
	Nickel	40.0	+/-40.0	U	40.0	P	05/15/2025	13:47	LB135794
	Selenium	20.0	+/-20.0	U	20.0	P	05/15/2025	13:47	LB135794
	Silver	10.0	+/-10.0	U	10.0	P	05/15/2025	13:47	LB135794
	Zinc	40.0	+/-40.0	U	40.0	P	05/15/2025	13:47	LB135794

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT		SDG No.: Q2001							
Contract: ENTA05	Lab Code: CHEM	Case No.: Q2001	SAS No.: Q2001						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/15/2025	14:22	LB135794
	Barium	100	+/-100	U	100	P	05/15/2025	14:22	LB135794
	Cadmium	6.00	+/-6.00	U	6.00	P	05/15/2025	14:22	LB135794
	Chromium	10.0	+/-10.0	U	10.0	P	05/15/2025	14:22	LB135794
	Copper	20.0	+/-20.0	U	20.0	P	05/15/2025	14:22	LB135794
	Lead	12.0	+/-12.0	U	12.0	P	05/15/2025	14:22	LB135794
	Nickel	40.0	+/-40.0	U	40.0	P	05/15/2025	14:22	LB135794
	Selenium	20.0	+/-20.0	U	20.0	P	05/15/2025	14:22	LB135794
	Silver	10.0	+/-10.0	U	10.0	P	05/15/2025	14:22	LB135794
	Zinc	40.0	+/-40.0	U	40.0	P	05/15/2025	14:22	LB135794
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	05/15/2025	15:09	LB135794
	Barium	100	+/-100	U	100	P	05/15/2025	15:09	LB135794
	Cadmium	6.00	+/-6.00	U	6.00	P	05/15/2025	15:09	LB135794
	Chromium	10.0	+/-10.0	U	10.0	P	05/15/2025	15:09	LB135794
	Copper	20.0	+/-20.0	U	20.0	P	05/15/2025	15:09	LB135794
	Lead	12.0	+/-12.0	U	12.0	P	05/15/2025	15:09	LB135794
	Nickel	40.0	+/-40.0	U	40.0	P	05/15/2025	15:09	LB135794
	Selenium	20.0	+/-20.0	U	20.0	P	05/15/2025	15:09	LB135794
	Silver	10.0	+/-10.0	U	10.0	P	05/15/2025	15:09	LB135794
	Zinc	40.0	+/-40.0	U	40.0	P	05/15/2025	15:09	LB135794
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	05/15/2025	15:57	LB135794
	Barium	100	+/-100	U	100	P	05/15/2025	15:57	LB135794
	Cadmium	6.00	+/-6.00	U	6.00	P	05/15/2025	15:57	LB135794
	Chromium	10.0	+/-10.0	U	10.0	P	05/15/2025	15:57	LB135794
	Copper	20.0	+/-20.0	U	20.0	P	05/15/2025	15:57	LB135794
	Lead	12.0	+/-12.0	U	12.0	P	05/15/2025	15:57	LB135794
	Nickel	40.0	+/-40.0	U	40.0	P	05/15/2025	15:57	LB135794
	Selenium	20.0	+/-20.0	U	20.0	P	05/15/2025	15:57	LB135794
	Silver	10.0	+/-10.0	U	10.0	P	05/15/2025	15:57	LB135794
	Zinc	40.0	+/-40.0	U	40.0	P	05/15/2025	15:57	LB135794
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	05/15/2025	16:44	LB135794
	Barium	100	+/-100	U	100	P	05/15/2025	16:44	LB135794
	Cadmium	6.00	+/-6.00	U	6.00	P	05/15/2025	16:44	LB135794
	Chromium	10.0	+/-10.0	U	10.0	P	05/15/2025	16:44	LB135794
	Copper	20.0	+/-20.0	U	20.0	P	05/15/2025	16:44	LB135794
	Lead	12.0	+/-12.0	U	12.0	P	05/15/2025	16:44	LB135794
	Nickel	40.0	+/-40.0	U	40.0	P	05/15/2025	16:44	LB135794
	Selenium	20.0	+/-20.0	U	20.0	P	05/15/2025	16:44	LB135794
	Silver	10.0	+/-10.0	U	10.0	P	05/15/2025	16:44	LB135794

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q2001</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2001</u>	SAS No.: <u>Q2001</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Zinc	40.0	+/-40.0	U	40.0	P	05/15/2025	16:44	LB135794
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	05/15/2025	17:56	LB135794
	Barium	100	+/-100	U	100	P	05/15/2025	17:56	LB135794
	Cadmium	6.00	+/-6.00	U	6.00	P	05/15/2025	17:56	LB135794
	Chromium	10.0	+/-10.0	U	10.0	P	05/15/2025	17:56	LB135794
	Copper	20.0	+/-20.0	U	20.0	P	05/15/2025	17:56	LB135794
	Lead	12.0	+/-12.0	U	12.0	P	05/15/2025	17:56	LB135794
	Nickel	40.0	+/-40.0	U	40.0	P	05/15/2025	17:56	LB135794
	Selenium	20.0	+/-20.0	U	20.0	P	05/15/2025	17:56	LB135794
	Silver	10.0	+/-10.0	U	10.0	P	05/15/2025	17:56	LB135794
	Zinc	40.0	+/-40.0	U	40.0	P	05/15/2025	17:56	LB135794
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	05/15/2025	18:52	LB135794
	Barium	100	+/-100	U	100	P	05/15/2025	18:52	LB135794
	Cadmium	6.00	+/-6.00	U	6.00	P	05/15/2025	18:52	LB135794
	Chromium	10.0	+/-10.0	U	10.0	P	05/15/2025	18:52	LB135794
	Copper	20.0	+/-20.0	U	20.0	P	05/15/2025	18:52	LB135794
	Lead	12.0	+/-12.0	U	12.0	P	05/15/2025	18:52	LB135794
	Nickel	40.0	+/-40.0	U	40.0	P	05/15/2025	18:52	LB135794
	Selenium	20.0	+/-20.0	U	20.0	P	05/15/2025	18:52	LB135794
	Silver	10.0	+/-10.0	U	10.0	P	05/15/2025	18:52	LB135794
	Zinc	40.0	+/-40.0	U	40.0	P	05/15/2025	18:52	LB135794

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q2001</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2001</u>	SAS No.: <u>Q2001</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/16/2025	13:05	LB135802
	Barium	100	+/-100	U	100	P	05/16/2025	13:05	LB135802
	Cadmium	6.00	+/-6.00	U	6.00	P	05/16/2025	13:05	LB135802
	Chromium	10.0	+/-10.0	U	10.0	P	05/16/2025	13:05	LB135802
	Copper	20.0	+/-20.0	U	20.0	P	05/16/2025	13:05	LB135802
	Lead	12.0	+/-12.0	U	12.0	P	05/16/2025	13:05	LB135802
	Nickel	40.0	+/-40.0	U	40.0	P	05/16/2025	13:05	LB135802
	Selenium	20.0	+/-20.0	U	20.0	P	05/16/2025	13:05	LB135802
	Silver	10.0	+/-10.0	U	10.0	P	05/16/2025	13:05	LB135802
	Zinc	40.0	+/-40.0	U	40.0	P	05/16/2025	13:05	LB135802

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT		SDG No.:		Q2001				
Contract:	ENTA05	Lab Code:	CHEM	Case No.:	Q2001	SAS No.:	Q2001		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/16/2025	13:39	LB135802
	Barium	100	+/-100	U	100	P	05/16/2025	13:39	LB135802
	Cadmium	6.00	+/-6.00	U	6.00	P	05/16/2025	13:39	LB135802
	Chromium	10.0	+/-10.0	U	10.0	P	05/16/2025	13:39	LB135802
	Copper	20.0	+/-20.0	U	20.0	P	05/16/2025	13:39	LB135802
	Lead	12.0	+/-12.0	U	12.0	P	05/16/2025	13:39	LB135802
	Nickel	40.0	+/-40.0	U	40.0	P	05/16/2025	13:39	LB135802
	Selenium	20.0	+/-20.0	U	20.0	P	05/16/2025	13:39	LB135802
	Silver	10.0	+/-10.0	U	10.0	P	05/16/2025	13:39	LB135802
	Zinc	40.0	+/-40.0	U	40.0	P	05/16/2025	13:39	LB135802
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	05/16/2025	14:40	LB135802
	Barium	100	+/-100	U	100	P	05/16/2025	14:40	LB135802
	Cadmium	6.00	+/-6.00	U	6.00	P	05/16/2025	14:40	LB135802
	Chromium	10.0	+/-10.0	U	10.0	P	05/16/2025	14:40	LB135802
	Copper	20.0	+/-20.0	U	20.0	P	05/16/2025	14:40	LB135802
	Lead	12.0	+/-12.0	U	12.0	P	05/16/2025	14:40	LB135802
	Nickel	40.0	+/-40.0	U	40.0	P	05/16/2025	14:40	LB135802
	Selenium	20.0	+/-20.0	U	20.0	P	05/16/2025	14:40	LB135802
	Silver	10.0	+/-10.0	U	10.0	P	05/16/2025	14:40	LB135802
	Zinc	40.0	+/-40.0	U	40.0	P	05/16/2025	14:40	LB135802
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	05/16/2025	15:26	LB135802
	Barium	100	+/-100	U	100	P	05/16/2025	15:26	LB135802
	Cadmium	6.00	+/-6.00	U	6.00	P	05/16/2025	15:26	LB135802
	Chromium	10.0	+/-10.0	U	10.0	P	05/16/2025	15:26	LB135802
	Copper	20.0	+/-20.0	U	20.0	P	05/16/2025	15:26	LB135802
	Lead	12.0	+/-12.0	U	12.0	P	05/16/2025	15:26	LB135802
	Nickel	40.0	+/-40.0	U	40.0	P	05/16/2025	15:26	LB135802
	Selenium	20.0	+/-20.0	U	20.0	P	05/16/2025	15:26	LB135802
	Silver	10.0	+/-10.0	U	10.0	P	05/16/2025	15:26	LB135802
	Zinc	40.0	+/-40.0	U	40.0	P	05/16/2025	15:26	LB135802
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	05/16/2025	16:12	LB135802
	Barium	100	+/-100	U	100	P	05/16/2025	16:12	LB135802
	Cadmium	6.00	+/-6.00	U	6.00	P	05/16/2025	16:12	LB135802
	Chromium	10.0	+/-10.0	U	10.0	P	05/16/2025	16:12	LB135802
	Copper	20.0	+/-20.0	U	20.0	P	05/16/2025	16:12	LB135802
	Lead	12.0	+/-12.0	U	12.0	P	05/16/2025	16:12	LB135802
	Nickel	40.0	+/-40.0	U	40.0	P	05/16/2025	16:12	LB135802
	Selenium	20.0	+/-20.0	U	20.0	P	05/16/2025	16:12	LB135802
	Silver	10.0	+/-10.0	U	10.0	P	05/16/2025	16:12	LB135802

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q2001</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2001</u>	SAS No.: <u>Q2001</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Zinc	40.0	+/-40.0	U	40.0	P	05/16/2025	16:12	LB135802
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	05/16/2025	16:58	LB135802
	Barium	100	+/-100	U	100	P	05/16/2025	16:58	LB135802
	Cadmium	6.00	+/-6.00	U	6.00	P	05/16/2025	16:58	LB135802
	Chromium	10.0	+/-10.0	U	10.0	P	05/16/2025	16:58	LB135802
	Copper	20.0	+/-20.0	U	20.0	P	05/16/2025	16:58	LB135802
	Lead	12.0	+/-12.0	U	12.0	P	05/16/2025	16:58	LB135802
	Nickel	40.0	+/-40.0	U	40.0	P	05/16/2025	16:58	LB135802
	Selenium	20.0	+/-20.0	U	20.0	P	05/16/2025	16:58	LB135802
	Silver	10.0	+/-10.0	U	10.0	P	05/16/2025	16:58	LB135802
	Zinc	40.0	+/-40.0	U	40.0	P	05/16/2025	16:58	LB135802
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	05/16/2025	17:32	LB135802
	Barium	100	+/-100	U	100	P	05/16/2025	17:32	LB135802
	Cadmium	6.00	+/-6.00	U	6.00	P	05/16/2025	17:32	LB135802
	Chromium	10.0	+/-10.0	U	10.0	P	05/16/2025	17:32	LB135802
	Copper	20.0	+/-20.0	U	20.0	P	05/16/2025	17:32	LB135802
	Lead	12.0	+/-12.0	U	12.0	P	05/16/2025	17:32	LB135802
	Nickel	40.0	+/-40.0	U	40.0	P	05/16/2025	17:32	LB135802
	Selenium	20.0	+/-20.0	U	20.0	P	05/16/2025	17:32	LB135802
	Silver	10.0	+/-10.0	U	10.0	P	05/16/2025	17:32	LB135802
	Zinc	40.0	+/-40.0	U	40.0	P	05/16/2025	17:32	LB135802

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT **SDG No.:** Q2001
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2001 **SAS No.:** Q2001

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------