

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q2001
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2001 **SAS No.:** Q2001

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167979BS Mercury	ug/L	4.0	4.51		113	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT SDG No.: Q2001
Contract: ENTA05 Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167984BS							
Arsenic	ug/L	4000	3830		96	80 - 120	P
Barium	ug/L	1000	858		86	80 - 120	P
Cadmium	ug/L	1000	930		93	80 - 120	P
Chromium	ug/L	2000	1930		96	80 - 120	P
Copper	ug/L	1500	1460		97	80 - 120	P
Lead	ug/L	5000	4660		93	80 - 120	P
Nickel	ug/L	2500	2290		92	80 - 120	P
Selenium	ug/L	10000	9430		94	80 - 120	P
Silver	ug/L	380	360		95	80 - 120	P
Zinc	ug/L	1000	973		97	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q2001
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2001 **SAS No.:** Q2001

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168012BS							
Arsenic	ug/L	4000	3850		96	80 - 120	P
Barium	ug/L	1000	877		88	80 - 120	P
Cadmium	ug/L	1000	949		95	80 - 120	P
Chromium	ug/L	2000	1950		98	80 - 120	P
Copper	ug/L	1500	1470		98	80 - 120	P
Lead	ug/L	5000	4740		95	80 - 120	P
Nickel	ug/L	2500	2330		93	80 - 120	P
Selenium	ug/L	10000	9560		96	80 - 120	P
Silver	ug/L	380	359		94	80 - 120	P
Zinc	ug/L	1000	967		97	80 - 120	P