

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/02/2025 05/02/2025

Client Sample No.: CCAL02 Date Analyzed: 05/13/2025

Lab Sample No.: AR1660CCC500 Data File : PO111041.D Time Analyzed: 17:33

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.773	4.673	4.873	441.850	500.000	-11.6
Aroclor-1016-2	4.792	4.692	4.892	451.170	500.000	-9.8
Aroclor-1016-3	4.848	4.748	4.948	420.910	500.000	-15.8
Aroclor-1016-4	4.968	4.868	5.068	454.170	500.000	-9.2
Aroclor-1016-5	5.226	5.125	5.325	452.120	500.000	-9.6
Aroclor-1260-1	6.264	6.165	6.365	412.050	500.000	-17.6
Aroclor-1260-2	6.454	6.355	6.555	412.210	500.000	-17.6
Aroclor-1260-3	6.821	6.722	6.922	427.000	500.000	-14.6
Aroclor-1260-4	7.081	6.982	7.182	419.970	500.000	-16.0
Aroclor-1260-5	7.323	7.225	7.425	419.140	500.000	-16.2
Decachlorobiphenyl	8.721	8.625	8.825	44.490	50.000	-11.0
Tetrachloro-m-xylene	3.684	3.583	3.783	49.800	50.000	-0.4

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/02/2025 05/02/2025

Client Sample No.: CCAL02 Date Analyzed: 05/13/2025

Lab Sample No.: AR1660CCC500 Data File : PO111041.D Time Analyzed: 17:33

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.758	4.660	4.860	518.770	500.000	3.8
Aroclor-1016-2	4.776	4.679	4.879	521.960	500.000	4.4
Aroclor-1016-3	4.952	4.854	5.054	521.280	500.000	4.3
Aroclor-1016-4	4.994	4.896	5.096	520.570	500.000	4.1
Aroclor-1016-5	5.206	5.109	5.309	522.110	500.000	4.4
Aroclor-1260-1	6.237	6.139	6.339	495.860	500.000	-0.8
Aroclor-1260-2	6.425	6.327	6.527	495.370	500.000	-0.9
Aroclor-1260-3	6.577	6.480	6.680	494.830	500.000	-1.0
Aroclor-1260-4	7.048	6.950	7.150	476.630	500.000	-4.7
Aroclor-1260-5	7.288	7.192	7.392	470.130	500.000	-6.0
Decachlorobiphenyl	8.671	8.576	8.776	55.420	50.000	10.8
Tetrachloro-m-xylene	3.680	3.580	3.780	54.320	50.000	8.6