

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL04 Date Analyzed: 05/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP072018.D Time Analyzed: 16:35

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.656	5.571	5.771	518.150	500.000	3.6
Aroclor-1016-2	5.678	5.593	5.793	499.780	500.000	0.0
Aroclor-1016-3	5.742	5.654	5.854	495.300	500.000	-0.9
Aroclor-1016-4	5.839	5.752	5.952	494.370	500.000	-1.1
Aroclor-1016-5	6.131	6.045	6.245	481.030	500.000	-3.8
Aroclor-1260-1	7.249	7.165	7.365	482.940	500.000	-3.4
Aroclor-1260-2	7.503	7.418	7.618	488.250	500.000	-2.4
Aroclor-1260-3	7.861	7.777	7.977	513.420	500.000	2.7
Aroclor-1260-4	8.085	8.002	8.202	494.760	500.000	-1.0
Aroclor-1260-5	8.404	8.322	8.522	521.650	500.000	4.3
Decachlorobiphenyl	10.212	10.137	10.337	52.800	50.000	5.6
Tetrachloro-m-xylene	4.505	4.417	4.617	50.480	50.000	1.0

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL04 Date Analyzed: 05/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP072018.D Time Analyzed: 16:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.885	4.798	4.998	536.020	500.000	7.2
Aroclor-1016-2	4.904	4.816	5.016	582.240	500.000	16.4
Aroclor-1016-3	5.081	4.993	5.193	526.540	500.000	5.3
Aroclor-1016-4	5.122	5.036	5.236	533.710	500.000	6.7
Aroclor-1016-5	5.337	5.250	5.450	561.140	500.000	12.2
Aroclor-1260-1	6.370	6.285	6.485	573.210	500.000	14.6
Aroclor-1260-2	6.558	6.473	6.673	564.270	500.000	12.9
Aroclor-1260-3	6.711	6.627	6.827	559.120	500.000	11.8
Aroclor-1260-4	7.181	7.098	7.298	568.790	500.000	13.8
Aroclor-1260-5	7.422	7.339	7.539	570.140	500.000	14.0
Decachlorobiphenyl	8.825	8.749	8.949	58.500	50.000	17.0
Tetrachloro-m-xylene	3.803	3.712	3.912	56.510	50.000	13.0