

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL05 Date Analyzed: 05/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP072027.D Time Analyzed: 20:39

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.661	5.571	5.771	484.730	500.000	-3.1
Aroclor-1016-2	5.683	5.593	5.793	492.850	500.000	-1.4
Aroclor-1016-3	5.745	5.654	5.854	491.010	500.000	-1.8
Aroclor-1016-4	5.842	5.752	5.952	485.690	500.000	-2.9
Aroclor-1016-5	6.135	6.045	6.245	476.500	500.000	-4.7
Aroclor-1260-1	7.253	7.165	7.365	473.840	500.000	-5.2
Aroclor-1260-2	7.506	7.418	7.618	475.770	500.000	-4.8
Aroclor-1260-3	7.865	7.777	7.977	496.620	500.000	-0.7
Aroclor-1260-4	8.089	8.002	8.202	480.540	500.000	-3.9
Aroclor-1260-5	8.407	8.322	8.522	514.770	500.000	3.0
Decachlorobiphenyl	10.216	10.137	10.337	52.300	50.000	4.6
Tetrachloro-m-xylene	4.508	4.417	4.617	49.540	50.000	-0.9

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL05 Date Analyzed: 05/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP072027.D Time Analyzed: 20:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.885	4.798	4.998	541.360	500.000	8.3
Aroclor-1016-2	4.903	4.816	5.016	541.580	500.000	8.3
Aroclor-1016-3	5.080	4.993	5.193	527.690	500.000	5.5
Aroclor-1016-4	5.121	5.036	5.236	532.650	500.000	6.5
Aroclor-1016-5	5.336	5.250	5.450	558.230	500.000	11.6
Aroclor-1260-1	6.368	6.285	6.485	570.060	500.000	14.0
Aroclor-1260-2	6.557	6.473	6.673	574.270	500.000	14.9
Aroclor-1260-3	6.709	6.627	6.827	568.040	500.000	13.6
Aroclor-1260-4	7.180	7.098	7.298	576.670	500.000	15.3
Aroclor-1260-5	7.422	7.339	7.539	593.320	500.000	18.7
Decachlorobiphenyl	8.825	8.749	8.949	62.250	50.000	24.5
Tetrachloro-m-xylene	3.802	3.712	3.912	53.550	50.000	7.1