

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL01 Date Analyzed: 05/14/2025

Lab Sample No.: HSTDCCC750 Data File : PS030190.D Time Analyzed: 17:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	8.851	8.750	8.950	696.850	712.500	-2.2
2,4-D	8.010	7.910	8.110	679.920	705.000	-3.6
2,4-DCAA	6.933	6.834	7.034	724.480	750.000	-3.4

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL01 Date Analyzed: 05/14/2025

Lab Sample No.: HSTDCCC750 Data File : PS030190.D Time Analyzed: 17:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.521	9.423	9.623	706.930	712.500	-0.8
2,4-D	8.643	8.544	8.744	697.370	705.000	-1.1
2,4-DCAA	7.453	7.354	7.554	734.300	750.000	-2.1