

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL03 Date Analyzed: 05/15/2025

Lab Sample No.: HSTDCCC750 Data File : PS030214.D Time Analyzed: 05:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	8.850	8.750	8.950	709.200	712.500	-0.5
2,4-D	8.009	7.910	8.110	691.270	705.000	-1.9
2,4-DCAA	6.933	6.834	7.034	737.200	750.000	-1.7

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL03 Date Analyzed: 05/15/2025

Lab Sample No.: HSTDCCC750 Data File : PS030214.D Time Analyzed: 05:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.521	9.423	9.623	719.650	712.500	1.0
2,4-D	8.643	8.544	8.744	710.310	705.000	0.8
2,4-DCAA	7.452	7.354	7.554	745.760	750.000	-0.6