

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL04 Date Analyzed: 05/15/2025

Lab Sample No.: HSTDCCC750 Data File : PS030224.D Time Analyzed: 10:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	8.851	8.750	8.950	671.550	712.500	-5.7
2,4-D	8.010	7.910	8.110	666.560	705.000	-5.5
2,4-DCAA	6.934	6.834	7.034	694.260	750.000	-7.4

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL04 Date Analyzed: 05/15/2025

Lab Sample No.: HSTDCCC750 Data File : PS030224.D Time Analyzed: 10:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.519	9.423	9.623	674.040	712.500	-5.4
2,4-D	8.641	8.544	8.744	668.710	705.000	-5.1
2,4-DCAA	7.451	7.354	7.554	698.470	750.000	-6.9