

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL05 Date Analyzed: 05/15/2025

Lab Sample No.: HSTDCCC750 Data File : PS030236.D Time Analyzed: 16:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	8.849	8.750	8.950	707.610	712.500	-0.7
2,4-D	8.008	7.910	8.110	690.730	705.000	-2.0
2,4-DCAA	6.932	6.834	7.034	734.390	750.000	-2.1

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL05 Date Analyzed: 05/15/2025

Lab Sample No.: HSTDCCC750 Data File : PS030236.D Time Analyzed: 16:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.519	9.423	9.623	722.990	712.500	1.5
2,4-D	8.641	8.544	8.744	710.880	705.000	0.8
2,4-DCAA	7.451	7.354	7.554	743.230	750.000	-0.9