



Report of Analysis

Client: ENTACT

540 Degraw St, Brooklyn, NY - E9309

Client Sample ID: WC-A4-03-CMSD

Lab Sample ID: Q2001-03MSD

Analytical Method: 8151A

Sample Wt/Vol: 100 Units: mL

Soil Aliquot Vol: uL

Extraction Type:

PS030204.D

Project:

GPC Factor: 1.0 PH:

Prep Method: SW3510C

File ID/Qc Batch: Dilution:

Prep Date Date Analyzed

tte Analyzed Prep Batch ID

05/06/25

05/09/25

Q2001

TCLP

10000

TCLP Herbicide

Decanted:

иL

Date Collected:

Date Received:

SDG No.:

Matrix:

% Solid:

Final Vol:

Injection Volume:

Test:

| CAS Number | Parameter | Conc. | Qualifier MDL | LOQ / CRQL | Units |
|------------|-------------------|-------|---------------------|------------|----------|
| TARGETS | | | | | |
| 94-75-7 | 2,4-D | 43.8 | 9.20 | 20.0 | ug/L |
| 93-72-1 | 2,4,5-TP (Silvex) | 40.6 | 7.80 | 20.0 | ug/L |
| SURROGATES | | | | | |
| 19719-28-9 | 2,4-DCAA | 390 | 70 (39) - 130 (175) | 78% | SPK: 500 |

Comments:

U = Not Detected

LOO = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit