

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

Continuing Calib Date: 05/14/2025 Initial Calibration Date(s): 05/13/2025 05/13/2025

Continuing Calib Time: 16:26 Initial Calibration Time(s): 12:24 13:18

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.11	9.12	9.02	9.22	0.01
Tetrachloro-m-xylene	3.59	3.59	3.49	3.69	0.00
gamma-BHC (Lindane)	4.37	4.38	4.28	4.48	0.01
Heptachlor	4.97	4.97	4.87	5.07	0.00
Heptachlor epoxide	5.73	5.74	5.64	5.84	0.01
Endrin	6.62	6.62	6.52	6.72	0.00
Methoxychlor	7.53	7.53	7.43	7.63	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

Continuing Calib Date: 05/14/2025 Initial Calibration Date(s): 05/13/2025 05/13/2025

Continuing Calib Time: 16:26 Initial Calibration Time(s): 12:24 13:18

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.08	8.09	7.99	8.19	0.01
Tetrachloro-m-xylene	2.91	2.91	2.81	3.01	0.01
gamma-BHC (Lindane)	3.75	3.75	3.65	3.85	0.00
Heptachlor	4.10	4.11	4.01	4.21	0.01
Heptachlor epoxide	4.89	4.90	4.80	5.00	0.01
Endrin	5.80	5.81	5.71	5.91	0.01
Methoxychlor	6.77	6.77	6.67	6.87	0.00