

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/13/2025 05/13/2025

Client Sample No.: CCAL01 Date Analyzed: 05/14/2025

Lab Sample No.: PSTDCCC050 Data File : PL095656.D Time Analyzed: 12:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.114	9.015	9.215	50.430	50.000	0.9
Endrin	6.617	6.519	6.719	49.390	50.000	-1.2
gamma-BHC (Lindane)	4.374	4.275	4.475	49.420	50.000	-1.2
Heptachlor	4.971	4.872	5.072	49.350	50.000	-1.3
Heptachlor epoxide	5.734	5.635	5.835	48.550	50.000	-2.9
Methoxychlor	7.533	7.434	7.634	50.710	50.000	1.4
Tetrachloro-m-xylene	3.592	3.493	3.693	47.830	50.000	-4.3

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/13/2025 05/13/2025

Client Sample No.: CCAL01 Date Analyzed: 05/14/2025

Lab Sample No.: PSTDCCC050 Data File : PL095656.D Time Analyzed: 12:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.085	7.987	8.187	47.300	50.000	-5.4
Endrin	5.807	5.708	5.908	45.250	50.000	-9.5
gamma-BHC (Lindane)	3.753	3.653	3.853	46.010	50.000	-8.0
Heptachlor	4.106	4.007	4.207	45.440	50.000	-9.1
Heptachlor epoxide	4.894	4.795	4.995	46.360	50.000	-7.3
Methoxychlor	6.772	6.673	6.873	46.130	50.000	-7.7
Tetrachloro-m-xylene	2.910	2.810	3.010	45.460	50.000	-9.1