

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/13/2025 05/13/2025

Client Sample No.: CCAL02 Date Analyzed: 05/14/2025

Lab Sample No.: PSTDCCC050 Data File : PL095667.D Time Analyzed: 16:26

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.112	9.015	9.215	49.600	50.000	-0.8
Endrin	6.617	6.519	6.719	48.320	50.000	-3.4
gamma-BHC (Lindane)	4.373	4.275	4.475	50.670	50.000	1.3
Heptachlor	4.970	4.872	5.072	49.670	50.000	-0.7
Heptachlor epoxide	5.733	5.635	5.835	49.410	50.000	-1.2
Methoxychlor	7.532	7.434	7.634	48.030	50.000	-3.9
Tetrachloro-m-xylene	3.592	3.493	3.693	49.560	50.000	-0.9

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/13/2025 05/13/2025

Client Sample No.: CCAL02 Date Analyzed: 05/14/2025

Lab Sample No.: PSTDCCC050 Data File : PL095667.D Time Analyzed: 16:26

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.080	7.987	8.187	46.030	50.000	-7.9
Endrin	5.801	5.708	5.908	43.950	50.000	-12.1
gamma-BHC (Lindane)	3.748	3.653	3.853	46.390	50.000	-7.2
Heptachlor	4.101	4.007	4.207	44.550	50.000	-10.9
Heptachlor epoxide	4.889	4.795	4.995	46.010	50.000	-8.0
Methoxychlor	6.767	6.673	6.873	42.910	50.000	-14.2
Tetrachloro-m-xylene	2.905	2.810	3.010	45.950	50.000	-8.1