

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ENTA05
Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG No.: Q2001
Instrument ID: BNA_F Calibration Date(s): 05/05/2025 05/05/2025
Calibration Time(s): 13:54 17:15

LAB FILE ID:		RRF2.5 = BF142295.D			RRF005 = BF142296.D		RRF010 = BF142297.D	
		RRF020 = BF142298.D			RRF040 = BF142299.D		RRF050 = BF142300.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Pyridine		1.159	1.233	1.210	1.176	1.311	1.239	5.0
2-Fluorophenol		1.212	1.177	1.211	1.126	1.224	1.172	4.4
Phenol-d6		1.510	1.458	1.489	1.375	1.500	1.441	4.6
1,4-Dichlorobenzene		1.562	1.499	1.486	1.364	1.476	1.438	6.5
2-Methylphenol		1.033	1.008	1.022	0.972	1.070	1.014	3.3
3+4-Methylphenols		1.354	1.281	1.322	1.218	1.333	1.265	6.5
Nitrobenzene-d5		0.289	0.304	0.336	0.327	0.360	0.328	7.5
Hexachloroethane		0.473	0.464	0.488	0.463	0.511	0.478	3.8
Nitrobenzene		0.281	0.293	0.316	0.306	0.337	0.309	6.0
Hexachlorobutadiene		0.188	0.182	0.181	0.168	0.185	0.177	5.3
2,4,6-Trichlorophenol		0.306	0.320	0.360	0.351	0.393	0.352	8.5
2-Fluorobiphenyl		1.737	1.605	1.532	1.272	1.337	1.403	16.3
2,4,5-Trichlorophenol		0.339	0.353	0.390	0.363	0.413	0.375	7.0
2,4-Dinitrotoluene		0.222	0.250	0.303	0.318	0.368	0.308	17.6
2,4,6-Tribromophenol		0.161	0.176	0.198	0.191	0.218	0.193	10.2
Hexachlorobenzene		0.264	0.246	0.253	0.237	0.258	0.249	3.8
Pentachlorophenol		0.094	0.109	0.130	0.139	0.153	0.131	16.7
Terphenyl-d14		1.493	1.424	1.460	1.255	1.390	1.351	9.6

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1