

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF050525

Review By	Rahul	Review On	5/6/2025 9:56:17 AM
Supervise By	Jagrut	Supervise On	5/6/2025 10:53:56 AM
SubDirectory	BF050525	HP Acquire Method	BNA_F
		HP Processing Method	bf050525

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12662,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142294.D	05 May 2025 13:24		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF142295.D	05 May 2025 13:54		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF142296.D	05 May 2025 14:23	Compound#32-54-65-77-85 removed from 5 ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BF142297.D	05 May 2025 14:52		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BF142298.D	05 May 2025 15:21		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BF142299.D	05 May 2025 15:50	Compound#26-54-65-80-84-85 kept on LR & #32 kept on QR	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BF142300.D	05 May 2025 16:18		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BF142301.D	05 May 2025 16:47		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BF142302.D	05 May 2025 17:15	Compound#9 removed from 80 ppm	RC/JU	Ok,M
10	SSTDICV040	ICVBF050525	BF142303.D	05 May 2025 18:13		RC/JU	Ok,M
11	PB167836BL	PB167836BL	BF142304.D	05 May 2025 18:41	Analyzed for contamination check.	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF051425

Review By	Rahul	Review On	5/15/2025 9:21:42 AM
Supervise By	Jagrut	Supervise On	5/15/2025 3:45:47 PM
SubDirectory	BF051425	HP Acquire Method	BNA_F
		HP Processing Method	bf050525

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12664,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142359.D	14 May 2025 11:22		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142360.D	14 May 2025 12:19		RC/JU	Ok,M
3	PB167929BL	PB167929BL	BF142361.D	14 May 2025 12:51		RC/JU	Ok
4	PB167929BS	PB167929BS	BF142362.D	14 May 2025 13:20		RC/JU	Ok,M
5	Q2004-01	VNJ-220	BF142363.D	14 May 2025 13:54		RC/JU	Ok
6	Q1982-01	TP-1	BF142364.D	14 May 2025 14:23		RC/JU	Ok
7	Q1982-03	TP-3	BF142365.D	14 May 2025 14:51		RC/JU	Ok
8	Q1982-03MS	TP-3MS	BF142366.D	14 May 2025 15:20		RC/JU	Ok,M
9	Q1982-03MSD	TP-3MSD	BF142367.D	14 May 2025 15:48		RC/JU	Ok,M
10	Q2014-03	MOO-25-0145	BF142368.D	14 May 2025 16:17		RC/JU	Ok
11	Q2004-03	2811	BF142369.D	14 May 2025 16:46		RC/JU	Ok
12	Q2017-04	MH-I	BF142370.D	14 May 2025 17:15		RC/JU	Ok
13	Q2001-03	WC-A4-03-C	BF142371.D	14 May 2025 17:44		RC/JU	Ok
14	Q2001-03MS	WC-A4-03-CMS	BF142372.D	14 May 2025 18:12		RC/JU	Ok,M
15	Q2001-03MSD	WC-A4-03-CMSD	BF142373.D	14 May 2025 18:41		RC/JU	Ok,M
16	Q2001-07	WC-A1-05-C	BF142374.D	14 May 2025 19:10		RC/JU	Ok
17	Q2001-11	WC-A1-06-C	BF142375.D	14 May 2025 19:39		RC/JU	Ok
18	Q2001-15	WC-A1-07-C	BF142376.D	14 May 2025 20:08		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF051425

Review By	Rahul	Review On	5/15/2025 9:21:42 AM		
Supervise By	Jagrut	Supervise On	5/15/2025 3:45:47 PM		
SubDirectory	BF051425	HP Acquire Method	BNA_F	HP Processing Method	bf050525
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791				
CCC	SP6787				
Internal Standard/PEM	S12664,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2017-08	MH-J	BF142377.D	14 May 2025 20:37		RC/JU	Ok
20	Q1982-02	TP-2B	BF142378.D	14 May 2025 21:06		RC/JU	Ok
21	Q2004-09	SOIL-STOCK	BF142379.D	14 May 2025 21:35		RC/JU	Ok
22	Q2017-01	MH-I	BF142380.D	14 May 2025 22:03		RC/JU	Ok
23	Q2017-05	MH-J	BF142381.D	14 May 2025 22:32		RC/JU	Ok
24	Q2005-01	252806	BF142382.D	14 May 2025 23:00	Internal Standard Fail	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF051925

Review By		Review On			
Supervise By		Supervise On			
SubDirectory	BF051925	HP Acquire Method	BNA_F	HP Processing Method	bf050525
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791				
CCC	SP6787				
Internal Standard/PEM	S12665,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142440.D	19 May 2025 09:40		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142441.D	19 May 2025 10:09		RC/JU	Ok
3	PB168017BL	PB168017BL	BF142442.D	19 May 2025 10:38		RC/JU	Ok
4	PB168017BS	PB168017BS	BF142443.D	19 May 2025 11:07		RC/JU	Ok,NR
5	PB167987BL	PB167987BL	BF142444.D	19 May 2025 11:36		RC/JU	Ok
6	PB167987BS	PB167987BS	BF142445.D	19 May 2025 12:05		RC/JU	Ok,NR
7	Q2032-01	TP-11	BF142446.D	19 May 2025 12:39		RC/JU	Ok
8	Q2032-02	TP-29	BF142447.D	19 May 2025 13:08		RC/JU	Ok
9	Q2032-03	TP-29-99	BF142448.D	19 May 2025 13:37		RC/JU	Ok
10	Q2048-01	L2-WC-1	BF142449.D	19 May 2025 14:06		RC/JU	Ok
11	Q2032-07	TP-37	BF142450.D	19 May 2025 14:35		RC/JU	Ok
12	Q2032-08	TP-32	BF142451.D	19 May 2025 15:04		RC/JU	Ok
13	Q2057-01	MH-L	BF142452.D	19 May 2025 15:33		RC/JU	Ok
14	Q1984-03	OU4-PCS-TC-34-05072	BF142453.D	19 May 2025 16:02		RC/JU	Ok
15	Q2047-01	SU-04-051425	BF142454.D	19 May 2025 16:31		RC/JU	Ok,NR
16	Q2062-01	L1-WC-1	BF142455.D	19 May 2025 17:00		RC/JU	Ok
17	Q2062-05	L1-WC-2	BF142456.D	19 May 2025 17:29		RC/JU	Ok,NR
18	Q2062-09	L1-WC-3	BF142457.D	19 May 2025 17:58		RC/JU	Ok,NR

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF051925

Review By	Review On
Supervise By	Supervise On
SubDirectory	BF051925
HP Acquire Method	BNA_F
HP Processing Method	bf050525
STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12665,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	Q2062-13	L1-WC-4	BF142458.D	19 May 2025 18:27		RC/JU	Ok,NR
20	Q2062-17	L1-WC-5	BF142459.D	19 May 2025 18:56		RC/JU	Ok
21	Q2062-21	L1-WC-6	BF142460.D	19 May 2025 19:25	Internal Standard Fail	RC/JU	ReRun
22	Q1984-09	OU4-TS-25-050725	BF142461.D	19 May 2025 19:54		RC/JU	Ok
23	Q1984-13	OU4-TS-27-050725	BF142462.D	19 May 2025 20:23		RC/JU	Ok
24	Q2071-17	L2-WC-6	BF142463.D	19 May 2025 20:52	Internal Standard Fail	RC/JU	ReRun
25	SSTDCCC040	SSTDCCC040EC	BF142464.D	19 May 2025 21:21		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP051325

Review By	Rahul	Review On	5/14/2025 2:20:20 PM		
Supervise By	Jagrut	Supervise On	5/14/2025 5:16:42 PM		
SubDirectory	BP051325	HP Acquire Method	BNA_P	HP Processing Method	bp051325

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12664,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024613.D	13 May 2025 10:00		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP024614.D	13 May 2025 10:41		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP024615.D	13 May 2025 11:22	Compound#32-54-56-65-77 removed from 5 ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP024616.D	13 May 2025 12:03		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BP024617.D	13 May 2025 12:43		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP024618.D	13 May 2025 13:24	Compound#32-54-56 Kept on LR	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BP024619.D	13 May 2025 14:05		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BP024620.D	13 May 2025 14:45		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BP024621.D	13 May 2025 15:26		RC/JU	Ok,M
10	SSTDICV040	ICVBP051325	BP024622.D	13 May 2025 17:01		RC/JU	Ok,M
11	PB167917TB	PB167917TB	BP024623.D	13 May 2025 18:23		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP051925

Review By	Review On
Supervise By	Supervise On
SubDirectory	BP051925
HP Acquire Method	BNA_P
HP Processing Method	BP051325
STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12665,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024685.D	19 May 2025 11:26		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP024686.D	19 May 2025 12:48		RC/JU	Ok
3	PB167963TB	PB167963TB	BP024687.D	19 May 2025 13:28		RC/JU	Ok
4	PB168041BL	PB168041BL	BP024688.D	19 May 2025 14:09		RC/JU	Ok
5	PB168041BS	PB168041BS	BP024689.D	19 May 2025 14:50	Recovery is low for compound #35	RC/JU	Not Ok
6	Q2032-04	TP-24	BP024690.D	19 May 2025 15:35		RC/JU	Ok
7	Q2032-05MS	TP-24MS	BP024691.D	19 May 2025 16:16		RC/JU	Ok
8	Q2032-06MSD	TP-24MSD	BP024692.D	19 May 2025 16:57		RC/JU	Ok
9	Q2032-11	FB-05132025	BP024693.D	19 May 2025 17:38	Internal Standard Fail	RC/JU	ReRun
10	Q2033-01	TW-WTS-08	BP024694.D	19 May 2025 18:19		RC/JU	Ok,NR
11	Q2068-03	LAW-25-0074	BP024695.D	19 May 2025 18:59	Surrogate Fail	RC/JU	ReRun
12	Q2067-01	303-PPR-FRAC	BP024696.D	19 May 2025 19:40	Need 2X Dilution	RC/JU	Dilution
13	Q2070-01	CONCRETE-PAD-0516	BP024697.D	19 May 2025 20:21		RC/JU	Ok
14	Q2071-09	L1-WC-9	BP024698.D	19 May 2025 21:02		RC/JU	Ok
15	Q2071-01	L1-WC-7	BP024699.D	19 May 2025 21:43		RC/JU	Ok
16	Q2071-05	L1-WC-8	BP024700.D	19 May 2025 22:24		RC/JU	Ok
17	Q2065-01	VACTRUCK-728068	BP024701.D	19 May 2025 23:04	Internal Standard Fail	RC/JU	ReRun

M : Manual Integration