

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2001

Client: ENTACT

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q2001-03MS		Client Sample ID: WC-A4-03-CMS		DataFile: BF142372.D							
Pyridine	500	0	330	ug/L	66				20 (10)	160 (109)	
1,4-Dichlorobenzene	500	0	350	ug/L	70				70 (55)	130 (125)	
2-Methylphenol	500	0	430	ug/L	86				70 (60)	130 (131)	
3+4-Methylphenols	500	0	440	ug/L	88				20 (54)	160 (136)	
Hexachloroethane	500	0	390	ug/L	78				20 (19)	160 (146)	
Nitrobenzene	500	0	450	ug/L	90				70 (62)	130 (112)	
Hexachlorobutadiene	500	0	410	ug/L	82				70 (52)	130 (125)	
2,4,6-Trichlorophenol	500	0	490	ug/L	98				70 (78)	130 (112)	
2,4,5-Trichlorophenol	500	0	470	ug/L	94				70 (71)	130 (111)	
2,4-Dinitrotoluene	500	0	620	ug/L	124				70 (74)	130 (137)	
Hexachlorobenzene	500	0	440	ug/L	88				70 (72)	130 (115)	
Pentachlorophenol	1000	0	1100	ug/L	110				20 (52)	160 (162)	