

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG NO.: Q2001

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/19/2025

Lab File ID: BP024686.D Time Analyzed: 12:48

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	148955	7.658	651543	10.43	426043	14.29
UPPER LIMIT	297910	8.158	1303090	10.928	852086	14.792
LOWER LIMIT	74477.5	7.158	325772	9.928	213022	13.792
EPA SAMPLE NO.						
01 PB167963TB	181587	7.66	675085	10.43	394044	14.29

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.