

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG No.: Q2001

Instrument ID: MSVOA_X Calibration Date/Time: 05/13/2025 10:00

Lab File ID: VX046153.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.691	0.696		0.72	20
1,1-Dichloroethene	0.593	0.578		-2.53	20
2-Butanone	0.543	0.561		3.32	20
Carbon Tetrachloride	0.544	0.574		5.51	20
Chloroform	1.271	1.300		2.28	20
Benzene	1.417	1.440		1.62	20
1,2-Dichloroethane	0.612	0.625		2.12	20
Trichloroethene	0.341	0.344		0.88	20
Tetrachloroethene	0.354	0.368		3.95	20
Chlorobenzene	1.094	1.091	0.3	-0.27	20
1,2-Dichloroethane-d4	0.932	0.899		-3.54	20
Dibromofluoromethane	0.360	0.371		3.06	20
Toluene-d8	1.246	1.236		-0.8	20
4-Bromofluorobenzene	0.478	0.484		1.25	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.