

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q2001 SAS No.: Q2001 SDG No.: Q2001

Instrument ID: MSVOA_X Calibration Date/Time: 05/14/2025 09:59

Lab File ID: VX046181.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.691	0.687		-0.58	20
1,1-Dichloroethene	0.593	0.584		-1.52	20
2-Butanone	0.543	0.560		3.13	20
Carbon Tetrachloride	0.544	0.568		4.41	20
Chloroform	1.271	1.302		2.44	20
Benzene	1.417	1.472		3.88	20
1,2-Dichloroethane	0.612	0.643		5.07	20
Trichloroethene	0.341	0.354		3.81	20
Tetrachloroethene	0.354	0.354		0	20
Chlorobenzene	1.094	1.097	0.3	0.27	20
1,2-Dichloroethane-d4	0.932	0.914		-1.93	20
Dibromofluoromethane	0.360	0.380		5.56	20
Toluene-d8	1.246	1.293		3.77	20
4-Bromofluorobenzene	0.478	0.507		6.07	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.